

Stainless Steel Temperature Probe

(Order Code TMP-BTA)



The Stainless Steel Temperature Probe is a rugged, general-purpose laboratory temperature sensor. It is designed to be used as you would use a thermometer for experiments in chemistry, physics, biology, Earth science, and environmental science.

Note: Do not completely submerge the sensor. The handle is not waterproof. This probe is the same as the probe shipped with CBL 2. Typical uses include the following:

- heat of fusion experiments
- monitoring endothermic and exothermic reactions
- weather studies
- specific heat experiments
- insulation studies

Collecting Data with the Stainless Steel Temperature Probe

This sensor can be used with the following interfaces to collect data:

- Vernier LabQuest™ as a standalone device or with a computer
- Vernier LabPro® with a computer, TI graphing calculator, or Palm® handheld
- Vernier Go!® Link
- Vernier EasyLink®
- Vernier SensorDAQ™
- CBL 2™

Here is the general procedure to follow when using the Stainless Steel Temperature Probe:

1. Connect the Stainless Steel Temperature Probe to the interface.
2. Start the data-collection software.
3. The software will identify the Stainless Steel Temperature Probe and load a default data-collection setup. You are now ready to collect data.

Data-Collection Software

This sensor can be used with an interface and the following data-collection software.

- **Logger Pro** This computer program is used with LabQuest, LabPro, or Go!Link.
- **Logger Lite** This computer program is used with LabQuest, LabPro, or Go!Link.
- **LabQuest App** This program is used when LabQuest is used as a stand-alone device.
- **EasyData App** This calculator application for the TI-83 Plus and TI-84 Plus can be used with CBL 2, LabPro, and Vernier EasyLink. We recommend version 2.0 or newer, which can be downloaded from the Vernier web site, www.vernier.com/easy/easydata.html, and then transferred to the calculator. See the Vernier web site, www.vernier.com/calc/software/index.html for more information on the App and Program Transfer Guidebook.
- **DataMate program** Use DataMate with LabPro or CBL 2 and TI-73, TI-83,

TI-84, TI-86, TI-89, and Voyage 200 calculators. See the LabPro and CBL 2 Guidebooks for instructions on transferring DataMate to the calculator.

- **Data Pro** This program is used with LabPro and a Palm handheld.
- **LabVIEW** National Instruments LabVIEW™ software is a graphical programming language sold by National Instruments. It is used with SensorDAQ and can be used with a number of other Vernier interfaces. See www.vernier.com/labview for more information.

NOTE: This product is to be used for educational purposes only. It is not appropriate for industrial, medical, research, or commercial applications.

Specifications

- Temperature range: -40 to 135°C (-40 to 275°F)
- Maximum temperature that the sensor can tolerate without damage: 150 C
- 13-bit resolution (LabQuest):
 - 0.09°C (-40 to 0°C)
 - 0.02°C (0 to 40°C)
 - 0.05°C (40 to 100°C)
 - 0.13°C (100 to 135°C)
- 12-bit resolution (LabQuest, LabPro):
 - 0.17°C (-40 to 0°C)
 - 0.03°C (0 to 40°C)
 - 0.1°C (40 to 100°C)
 - 0.25°C (100 to 135°C)
- 10-bit resolution (CBL 2):
 - 0.68°C (-40 to 0°C)
 - 0.12°C (0 to 40°C)
 - 0.4°C (40 to 100°C)
 - 1.0°C (100 to 135°C)
- Temperature sensor: 20 k NTC Thermistor
- Accuracy: ±0.2°C at 0°C, ±0.5°C at 100°C
- Response time (time for 90% change in reading):
 - 10 seconds (in water, with stirring)
 - 400 seconds (in still air)
 - 90 seconds (in moving air)
- Probe dimensions:
 - Probe length (handle plus body): 15.5 cm
 - Stainless steel body: length 10.5 cm, diameter 4.0 mm
 - Probe handle: length 5.0 cm, diameter 1.25 cm

This sensor is equipped with circuitry that supports auto-ID. When used with LabQuest, LabPro, Go! Link, SensorDAQ, EasyLink, or CBL 2, the data-collection software identifies the sensor and uses pre-defined parameters to configure an experiment appropriate to the recognized sensor.

How the Stainless Steel Temperature Probe Works

This probe uses the 20 k Ω NTC Thermistor. The thermistor is a variable resistor whose resistance decreases nonlinearly with increasing temperature. The best-fit approximation to this nonlinear characteristic is the Steinhart-Hart equation. At 25 $^{\circ}$ C, the resistance is approximately 4.3% per $^{\circ}$ C. The LabPro or CBL 2 interface measures the resistance value, R, at a particular temperature, and converts the resistance using the Steinhart-Hart equation:

$$T = [K_0 + K_1(\ln 1000R) + K_2(\ln 1000R)^3]^{-1} - 273.15$$

where T is temperature ($^{\circ}$ C), R is the measured resistance in k Ω , $K_0 = 1.02119 \times 10^{-3}$, $K_1 = 2.22468 \times 10^{-4}$, and $K_2 = 1.33342 \times 10^{-7}$. Our programs take care of this conversion for you, and provide readings in $^{\circ}$ C (or other units, if you load a different calibration).

Probe Chemical Tolerance

The Stainless Steel Temperature Probe body is constructed from grade 316 stainless steel.¹ This high-grade stainless steel provides a high level of corrosion resistance for use in the science classroom. Here are some general guidelines for usage:

1. The probe handle is constructed of molded plasticized Santoprene[®]. While this material is very chemical resistant, we recommend that you avoid submerging the probe beyond the stainless steel portion.
2. Always wash the probe thoroughly after use.
3. The probe can be left continuously in water at temperatures within the range of -40° to 150° C. Continuous usage in saltwater will cause only minor discoloration of the probe, with no negative effect on performance.
4. You can leave the probe continuously in most organic compounds, such as methanol, ethanol, 1-propanol, 2-propanol, 1-butanol, n-hexane, lauric acid, paradichlorobenzene, phenyl salicylate, and benzoic acid. The probe should not be left in n-pentane for more than 1 hour.
5. The probe can be left in strong basic solutions, such as NaOH, for up to 48 hours, with only minor discoloration. We do not recommend usage in basic solutions that are greater than 3 M in concentration.
6. The chart provides the maximum length of time

Maximum acid exposure time	
1 M HCl	20 min
2 M HCl	10 min
3 M HCl	5 min
1 M H ₂ SO ₄	48 hours
2 M H ₂ SO ₄	20 min
3 M H ₂ SO ₄	10 min
1 M HNO ₃	48 hours
2 M HNO ₃	48 hours
3 M HNO ₃	48 hours
1 M CH ₃ COOH	48 hours
2 M CH ₃ COOH	48 hours
3 M CH ₃ COOH	48 hours
1 M H ₃ PO ₄	48 hours
2 M H ₃ PO ₄	48 hours
3 M H ₃ PO ₄	48 hours

¹ Grade 316 stainless steel has a composition of 0.08% carbon, 2.0% manganese, 0.75% silicon, 0.04% phosphorus, 0.03% sulfur, 16-18% chromium, 10-14% nickel, 2-3% molybdenum, and 0.1% nitrogen.

we recommend for probe exposure to some common acids. Probes left in an acid longer than these times may bubble and/or discolor, but will still be functional. We do not recommend probes be left to soak in any acid longer than 48 hours.

Do I Need to Calibrate This Probe? Probably Not

In most cases, the Stainless Steel Temperature Probe will never need to be calibrated. It is calibrated extremely well before it ships. However, if the need arises to calibrate the sensor, and you are using Logger Pro 3.3 or newer, the sensor can be custom-calibrated. **Note** this can only be done on computers, and can not be done from DataMate or EasyData (calculators) or DataPro (Palm OS).

The process of calibrating a Stainless Steel Temperature probe connected to LabPro and Logger Pro is different than the process for most other sensors. One reason is that this probe uses a thermistor, which has a non-linear response, and you need to calibrate it at three different temperatures.

Obtain a thermometer and three containers of water at three different temperatures.

Choose Experiment/Calibrate and then choose the Stainless Steel Temperature Probe. Choose Calibrate Now. For each of the three water baths, place the temperature probe in the bath with the thermometer. Allow both readings to stabilize, click , and enter the temperature reading measured by the thermometer. Click after the third water bath reading.

If you want the calibration to be saved with the file for later use after you calibrate, you need to do two things:

- Go to the calibration dialog box and choose Calibration Storage from the pull down menu. Then set the Calibration Retrieval Preference to Experiment file.
- Save the Experiment file.

After this, when you open that experiment file, the calibration stored with the experiment file will be used, instead of the normal calibration stored on the computer for this probe.

Warranty

Vernier warrants this product to be free from defects in materials and workmanship for a period of five years from the date of shipment to the customer. This warranty does not cover damage to the product caused by abuse or improper use.



Measure. Analyze. Learn.™

Vernier Software & Technology

13979 S.W. Millikan Way • Beaverton, OR 97005-2886

Toll Free (888) 837-6437 • (503) 277-2299 • FAX (503) 277-2440

info@vernier.com • www.vernier.com

Rev 5/7/07

Logger Pro, Logger Lite, Vernier LabQuest, Vernier LabPro, Go!Link, Vernier EasyLink and other marks shown are our trademarks or registered trademarks in the United States.

CBL 2 and CBL, TI-GRAPH LINK, and TI Connect are trademarks of Texas Instruments.

All other marks not owned by us that appear herein are the property of their respective owners, who may or may not be affiliated with, connected to, or sponsored by us.



Printed on recycled paper.