

Spring Branch Independent School District  
**HEALTH SERVICES**  
**AED (Automated External Defibrillator)**  
**Inspection Readiness Inventory**

Campus/Facility: \_\_\_\_\_ School Year: \_\_\_\_\_

AED Serial # \_\_\_\_\_ AED Battery expires \_\_\_\_\_

AED Adult Pads expire \_\_\_\_\_ AED Pedi Pads expire \_\_\_\_\_

Chest Seals expire \_\_\_\_\_

<b>Month Inspected:</b>	<b>Aug</b>	<b>Sept</b>	<b>Oct</b>	<b>Nov</b>	<b>Dec</b>	<b>Jan</b>	<b>Feb</b>	<b>Mar</b>	<b>Apr</b>	<b>May</b>	<b>June</b>	<b>July</b>
Storage Cabinet Intact :												
AED Exterior Intact:												
Battery Installed and Functional:												
Battery (within expiration date):												
AED Active Test (close and open-active?):												
AED User Guide/Instructions Available:												
Set of Electrodes Adult and Pedi (within expiration date?)												
Ready Kit Available: (mouth barrier, razor, scissors, non-latex gloves & gauze or towelettes)												
AED In-service and Response Plan shared with staff annually?												

Revised 10/1/2019

INSPECTED BY: \_\_\_\_\_ SENT SAFETY AND RM END OF YR \_\_\_\_\_