

April 4, 2023

Stephanie Fisk  
Gateway School District  
12 Littleville Road  
Huntington, MA 01050

Project Location: Chester Elementary School  
Client Job Number:  
Project Number: 2  
Laboratory Work Order Number: 23C1981  
PWSID# 1059012

Enclosed are results of analyses for samples as received by the laboratory on March 17, 2023. If you have any questions concerning this report, please feel free to contact me.

Sincerely,



Kaitlyn A. Feliciano  
Project Manager

# Table of Contents

Sample Summary	3
Case Narrative	4
Sample Results	5
23C1981-01	5
23C1981-02	6
Sample Preparation Information	7
QC Data	8
Semivolatile Organic Compounds by - LC/MS-MS	8
B334620	8
Flag/Qualifier Summary	10
Certifications	11
Chain of Custody/Sample Receipt	12



39 Spruce Street \* East Longmeadow, MA 01028 \* FAX 413/525-6405 \* TEL. 413/525-2332

Gateway School District  
12 Littleville Road  
Huntington, MA 01050  
ATTN: Stephanie Fisk

REPORT DATE: 4/4/2023

PURCHASE ORDER NUMBER:

PROJECT NUMBER: 2

**ANALYTICAL SUMMARY**

WORK ORDER NUMBER: 23C1981

The results of analyses performed on the following samples submitted to Con-Test, a Pace Analytical Laboratory, are found in this report.

PROJECT LOCATION: Chester Elementary School

FIELD SAMPLE #	LAB ID:	MATRIX	SAMPLE DESCRIPTION	TEST	SUB LAB
POE Post	23C1981-01	Drinking Water		EPA 537.1	
POE Post FB	23C1981-02	Drinking Water		EPA 537.1	

**CASE NARRATIVE SUMMARY**

All reported results are within defined laboratory quality control objectives unless listed below or otherwise qualified in this report.

The results of analyses reported only relate to samples submitted to Con-Test, a Pace Analytical Laboratory, for testing. I certify that the analyses listed above, unless specifically listed as subcontracted, if any, were performed under my direction according to the approved methodologies listed in this document, and that based upon my inquiry of those individuals immediately responsible for obtaining the information, the material contained in this report is, to the best of my knowledge and belief, accurate and complete.



Lisa A. Worthington  
Technical Representative

Project Location: Chester Elementary School

Sample Description:

Work Order: 23C1981

Date Received: 3/17/2023

Field Sample #: POE Post

Sampled: 3/17/2023 07:45

Sample ID: 23C1981-01

Sample Matrix: Drinking Water

**Semivolatile Organic Compounds by - LC/MS-MS**

Analyte	Results	RL	MCL/SMCL			DF	Flag/Qual	Method	Date	Date/Time	Analyst
			DL	MA ORSG	Units				Prepared	Analyzed	
Perfluorobutanesulfonic acid (PFBS)	ND	1.9	0.75		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:31	JR2
Perfluorohexanoic acid (PFHxA)	ND	1.9	0.89		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:31	JR2
Perfluorohexanesulfonic acid (PFHxS)	ND	1.9	0.85		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:31	JR2
Perfluoroheptanoic acid (PFHpA)	ND	1.9	0.94		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:31	JR2
Perfluorooctanoic acid (PFOA)	ND	1.9	0.98		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:31	JR2
Perfluorooctanesulfonic acid (PFOS)	1.3	1.9	0.72		ng/L	1	J	EPA 537.1	3/20/23	3/23/23 10:31	JR2
Perfluorononanoic acid (PFNA)	ND	1.9	0.87		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:31	JR2
Perfluorodecanoic acid (PFDA)	ND	1.9	0.91		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:31	JR2
N-EtFOSAA (NEtFOSAA)	ND	1.9	0.63		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:31	JR2
Perfluoroundecanoic acid (PFUnA)	ND	1.9	0.72		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:31	JR2
N-MeFOSAA (NMeFOSAA)	ND	1.9	0.70		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:31	JR2
Perfluorododecanoic acid (PFDoA)	ND	1.9	0.68		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:31	JR2
Perfluorotridecanoic acid (PFTrDA)	ND	1.9	0.69		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:31	JR2
Perfluorotetradecanoic acid (PFTA)	ND	1.9	0.79		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:31	JR2
Hexafluoropropylene oxide dimer acid (HFPO-DA)	ND	1.9	1.2		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:31	JR2
11Cl-PF3OUdS (F53B Major)	ND	1.9	0.63		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:31	JR2
9Cl-PF3ONS (F53B Minor)	ND	1.9	0.77		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:31	JR2
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	ND	1.9	0.83		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:31	JR2
Summation of PFAS 6 for Massachusetts	ND	1.9	1.9		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:31	JR2

Surrogates	% Recovery	Recovery Limits	Flag/Qual
13C-PFHxA	98.9	70-130	3/23/23 10:31
M3HFPO-DA	99.0	70-130	3/23/23 10:31
13C-PFDA	96.2	70-130	3/23/23 10:31
D5-NEtFOSAA	104	70-130	3/23/23 10:31

Project Location: Chester Elementary School

Sample Description:

Work Order: 23C1981

Date Received: 3/17/2023

Field Sample #: POE Post FB

Sampled: 3/17/2023 07:45

Sample ID: 23C1981-02

Sample Matrix: Drinking Water

**Semivolatile Organic Compounds by - LC/MS-MS**

Analyte	Results	RL	MCL/SMCL			DF	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
			DL	MA	ORSG						
Perfluorobutanesulfonic acid (PFBS)	ND	1.9	0.73			1		EPA 537.1	3/20/23	3/23/23 10:38	JR2
Perfluorohexanoic acid (PFHxA)	ND	1.9	0.87			1		EPA 537.1	3/20/23	3/23/23 10:38	JR2
Perfluorohexanesulfonic acid (PFHxS)	ND	1.9	0.83			1		EPA 537.1	3/20/23	3/23/23 10:38	JR2
Perfluoroheptanoic acid (PFHpA)	ND	1.9	0.92			1		EPA 537.1	3/20/23	3/23/23 10:38	JR2
Perfluorooctanoic acid (PFOA)	ND	1.9	0.96			1		EPA 537.1	3/20/23	3/23/23 10:38	JR2
Perfluorooctanesulfonic acid (PFOS)	ND	1.9	0.70			1		EPA 537.1	3/20/23	3/23/23 10:38	JR2
Perfluorononanoic acid (PFNA)	ND	1.9	0.86			1		EPA 537.1	3/20/23	3/23/23 10:38	JR2
Perfluorodecanoic acid (PFDA)	ND	1.9	0.89			1		EPA 537.1	3/20/23	3/23/23 10:38	JR2
N-EtFOSAA (NEtFOSAA)	ND	1.9	0.62			1		EPA 537.1	3/20/23	3/23/23 10:38	JR2
Perfluoroundecanoic acid (PFUnA)	ND	1.9	0.70			1		EPA 537.1	3/20/23	3/23/23 10:38	JR2
N-MeFOSAA (NMeFOSAA)	ND	1.9	0.69			1		EPA 537.1	3/20/23	3/23/23 10:38	JR2
Perfluorododecanoic acid (PFDoA)	ND	1.9	0.67			1		EPA 537.1	3/20/23	3/23/23 10:38	JR2
Perfluorotridecanoic acid (PFTrDA)	ND	1.9	0.68			1		EPA 537.1	3/20/23	3/23/23 10:38	JR2
Perfluorotetradecanoic acid (PFTA)	ND	1.9	0.78			1		EPA 537.1	3/20/23	3/23/23 10:38	JR2
Hexafluoropropylene oxide dimer acid (HFPO-DA)	ND	1.9	1.1			1		EPA 537.1	3/20/23	3/23/23 10:38	JR2
11Cl-PF3OUdS (F53B Major)	ND	1.9	0.62			1		EPA 537.1	3/20/23	3/23/23 10:38	JR2
9Cl-PF3ONS (F53B Minor)	ND	1.9	0.76			1		EPA 537.1	3/20/23	3/23/23 10:38	JR2
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	ND	1.9	0.82			1		EPA 537.1	3/20/23	3/23/23 10:38	JR2
Summation of PFAS 6 for Massachusetts	ND	1.9	1.9			1		EPA 537.1	3/20/23	3/23/23 10:38	JR2

Surrogates	% Recovery	Recovery Limits	Flag/Qual
13C-PFHxA	107	70-130	3/23/23 10:38
M3HFPO-DA	109	70-130	3/23/23 10:38
13C-PFDA	106	70-130	3/23/23 10:38
D5-NEtFOSAA	106	70-130	3/23/23 10:38

**Sample Extraction Data**

Prep Method: EPA 537.1-EPA 537.1

Lab Number [Field ID]	Batch	Initial [mL]	Final [mL]	Date
23C1981-01 [POE Post]	B334620	264	1.00	03/20/23
23C1981-02 [POE Post FB]	B334620	269	1.00	03/20/23

**QUALITY CONTROL**
**Semivolatile Organic Compounds by - LC/MS-MS - Quality Control**

Analyte	Result	Reporting Limit	DL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----------------	----	-------	-------------	---------------	------	-------------	-----	-----------	-------

**Batch B334620 - EPA 537.1**
**Blank (B334620-BLK1)**

Prepared: 03/20/23 Analyzed: 03/23/23

Perfluorobutanesulfonic acid (PFBS)	ND	1.9	0.76	ng/L							
Perfluorohexanoic acid (PFHxA)	ND	1.9	0.90	ng/L							
Perfluorohexanesulfonic acid (PFHxS)	ND	1.9	0.86	ng/L							
Perfluoroheptanoic acid (PFHpA)	ND	1.9	0.96	ng/L							
Perfluorooctanoic acid (PFOA)	ND	1.9	0.99	ng/L							
Perfluorooctanesulfonic acid (PFOS)	ND	1.9	0.73	ng/L							
Perfluorononanoic acid (PFNA)	ND	1.9	0.89	ng/L							
Perfluorodecanoic acid (PFDA)	ND	1.9	0.92	ng/L							
N-EtFOSAA (NEtFOSAA)	ND	1.9	0.64	ng/L							
Perfluoroundecanoic acid (PFUnA)	ND	1.9	0.73	ng/L							
N-MeFOSAA (NMeFOSAA)	ND	1.9	0.71	ng/L							
Perfluorododecanoic acid (PFDoA)	ND	1.9	0.69	ng/L							
Perfluorotridecanoic acid (PFTriDA)	ND	1.9	0.70	ng/L							
Perfluorotetradecanoic acid (PFTA)	ND	1.9	0.80	ng/L							
Hexafluoropropylene oxide dimer acid (HFPO-DA)	ND	1.9	1.2	ng/L							
11Cl-PF3OUdS (F53B Major)	ND	1.9	0.64	ng/L							
9Cl-PF3ONS (F53B Minor)	ND	1.9	0.78	ng/L							
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	ND	1.9	0.85	ng/L							
Summation of PFAS 6 for Massachusetts	ND	1.9	1.9	ng/L							
Surrogate: 13C-PFHxA	32.0			ng/L	38.5		83.1	70-130			
Surrogate: M3HFPO-DA	31.5			ng/L	38.5		82.0	70-130			
Surrogate: 13C-PFDA	32.0			ng/L	38.5		83.1	70-130			
Surrogate: D5-NEtFOSAA	126			ng/L	154		82.1	70-130			

**LCS (B334620-BS1)**

Prepared: 03/20/23 Analyzed: 03/23/23

Perfluorobutanesulfonic acid (PFBS)	6.80	1.9	0.74	ng/L	8.33		81.6	70-130			
Perfluorohexanoic acid (PFHxA)	7.06	1.9	0.88	ng/L	9.39		75.2	70-130			
Perfluorohexanesulfonic acid (PFHxS)	7.47	1.9	0.84	ng/L	8.58		87.1	70-130			
Perfluoroheptanoic acid (PFHpA)	7.63	1.9	0.93	ng/L	9.39		81.3	70-130			
Perfluorooctanoic acid (PFOA)	7.57	1.9	0.97	ng/L	9.39		80.6	70-130			
Perfluorooctanesulfonic acid (PFOS)	7.38	1.9	0.71	ng/L	8.71		84.8	70-130			
Perfluorononanoic acid (PFNA)	7.69	1.9	0.86	ng/L	9.39		81.9	70-130			
Perfluorodecanoic acid (PFDA)	7.09	1.9	0.90	ng/L	9.39		75.6	70-130			
N-EtFOSAA (NEtFOSAA)	7.32	1.9	0.62	ng/L	9.39		77.9	70-130			
Perfluoroundecanoic acid (PFUnA)	7.18	1.9	0.71	ng/L	9.39		76.5	70-130			
N-MeFOSAA (NMeFOSAA)	7.44	1.9	0.70	ng/L	9.39		79.2	70-130			
Perfluorododecanoic acid (PFDoA)	6.92	1.9	0.67	ng/L	9.39		73.7	70-130			
Perfluorotridecanoic acid (PFTriDA)	6.71	1.9	0.68	ng/L	9.39		71.5	70-130			
Perfluorotetradecanoic acid (PFTA)	7.32	1.9	0.78	ng/L	9.39		78.0	70-130			
Hexafluoropropylene oxide dimer acid (HFPO-DA)	7.53	1.9	1.2	ng/L	9.39		80.2	70-130			
11Cl-PF3OUdS (F53B Major)	6.79	1.9	0.62	ng/L	8.85		76.7	70-130			
9Cl-PF3ONS (F53B Minor)	7.31	1.9	0.76	ng/L	8.76		83.4	70-130			
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	7.20	1.9	0.83	ng/L	8.87		81.2	70-130			
Summation of PFAS 6 for Massachusetts	44.8	1.9	1.9	ng/L				0-200			
Surrogate: 13C-PFHxA	32.1			ng/L	37.5		85.5	70-130			
Surrogate: M3HFPO-DA	33.3			ng/L	37.5		88.6	70-130			
Surrogate: 13C-PFDA	32.1			ng/L	37.5		85.5	70-130			



## QUALITY CONTROL

## Semivolatile Organic Compounds by - LC/MS-MS - Quality Control

Analyte	Reporting		DL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
	Result	Limit									

## Batch B334620 - EPA 537.1

## LCS (B334620-BS1)

Prepared: 03/20/23 Analyzed: 03/23/23

Surrogate: D5-NEtFOSAA	134			ng/L	150		89.2	70-130			
------------------------	-----	--	--	------	-----	--	------	--------	--	--	--

## LCS Dup (B334620-BS1)

Prepared: 03/20/23 Analyzed: 03/23/23

Perfluorobutanesulfonic acid (PFBS)	7.60	1.9	0.76	ng/L	8.51		89.3	70-130	11.1	30	
Perfluorohexanoic acid (PFHxA)	8.39	1.9	0.90	ng/L	9.59		87.5	70-130	17.2	30	
Perfluorohexanesulfonic acid (PFHxS)	7.93	1.9	0.86	ng/L	8.76		90.5	70-130	5.90	30	
Perfluoroheptanoic acid (PFHpA)	8.70	1.9	0.95	ng/L	9.59		90.7	70-130	13.1	30	
Perfluorooctanoic acid (PFOA)	8.09	1.9	0.99	ng/L	9.59		84.4	70-130	6.69	30	
Perfluorooctanesulfonic acid (PFOS)	7.37	1.9	0.73	ng/L	8.90		82.8	70-130	0.160	30	
Perfluorononanoic acid (PFNA)	8.57	1.9	0.88	ng/L	9.59		89.4	70-130	10.8	30	
Perfluorodecanoic acid (PFDA)	8.90	1.9	0.92	ng/L	9.59		92.8	70-130	22.6	30	
N-EtFOSAA (NEtFOSAA)	7.66	1.9	0.64	ng/L	9.59		79.9	70-130	4.63	30	
Perfluoroundecanoic acid (PFUnA)	8.28	1.9	0.73	ng/L	9.59		86.3	70-130	14.3	30	
N-MeFOSAA (NMeFOSAA)	8.74	1.9	0.71	ng/L	9.59		91.2	70-130	16.1	30	
Perfluorododecanoic acid (PFDoA)	8.11	1.9	0.69	ng/L	9.59		84.5	70-130	15.8	30	
Perfluorotridecanoic acid (PFTTrDA)	8.17	1.9	0.70	ng/L	9.59		85.2	70-130	19.7	30	
Perfluorotetradecanoic acid (PFTA)	8.27	1.9	0.80	ng/L	9.59		86.3	70-130	12.2	30	
Hexafluoropropylene oxide dimer acid (HFPO-DA)	8.73	1.9	1.2	ng/L	9.59		91.1	70-130	14.8	30	
11Cl-PF3OUdS (F53B Major)	7.72	1.9	0.64	ng/L	9.04		85.4	70-130	12.9	30	
9Cl-PF3ONS (F53B Minor)	8.16	1.9	0.78	ng/L	8.95		91.2	70-130	11.0	30	
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	8.25	1.9	0.84	ng/L	9.06		91.0	70-130	13.5	30	
Summation of PFAS 6 for Massachusetts	49.6	1.9	1.9	ng/L				0-200	10.0		
Surrogate: 13C-PFHxA	39.6			ng/L	38.4		103	70-130			
Surrogate: M3HFPO-DA	40.3			ng/L	38.4		105	70-130			
Surrogate: 13C-PFDA	39.4			ng/L	38.4		103	70-130			
Surrogate: D5-NEtFOSAA	160			ng/L	153		104	70-130			

**FLAG/QUALIFIER SUMMARY**

*	QC result is outside of established limits.
†	Wide recovery limits established for difficult compound.
‡	Wide RPD limits established for difficult compound.
#	Data exceeded client recommended or regulatory level
ND	Not Detected
RL	Reporting Limit
DL	Method Detection Limit
MCL	Maximum Contaminant Level
	Percent recoveries and relative percent differences (RPDs) are determined by the software using values in the calculation which have not been rounded.
	No results have been blank subtracted unless specified in the case narrative section.
J	Detected but below the Reporting Limit (lowest calibration standard); therefore, result is an estimated concentration (CLP J-Flag).

**CERTIFICATIONS**
**Certified Analyses included in this Report**

Analyte	Certifications
<b>EPA 537.1 in Drinking Water</b>	
Perfluorobutanesulfonic acid (PFBS)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Perfluorohexanoic acid (PFHxA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Perfluorohexanesulfonic acid (PFHxS)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Perfluoroheptanoic acid (PFHpA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Perfluorooctanoic acid (PFOA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Perfluorooctanesulfonic acid (PFOS)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Perfluorononanoic acid (PFNA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Perfluorodecanoic acid (PFDA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
N-EtFOSAA (NEtFOSAA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Perfluoroundecanoic acid (PFUnA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
N-MeFOSAA (NMeFOSAA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Perfluorododecanoic acid (PFDoA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Perfluorotridecanoic acid (PFTriDA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Perfluorotetradecanoic acid (PFTA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Hexafluoropropylene oxide dimer acid (HFPO-DA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
11Cl-PF3OUdS (F53B Major)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
9Cl-PF3ONS (F53B Minor)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH

Con-Test, a Pace Environmental Laboratory, operates under the following certifications and accreditations:

Code	Description	Number	Expires
MA	Massachusetts DEP	M-MA100	06/30/2023
CT	Connecticut Department of Public Health	PH-0821	12/31/2024
NY	New York State Department of Health	10899 NELAP	04/1/2024
NH	New Hampshire Environmental Lab	2516 NELAP	02/5/2024
NJ	New Jersey DEP	MA007 NELAP	06/30/2023
VT-DW	Vermont Department of Health Drinking Water	VT-255716	06/12/2023
ME	State of Maine	MA00100	06/9/2023
PA	Commonwealth of Pennsylvania DEP	68-05812	06/30/2023
MI	Dept. of Env, Great Lakes, and Energy	9100	06/30/2023
OH	Ohio Environmental Protection Agency	87781	04/1/2024

23C1981

Phone: 413-525-2332  
Fax: 413-525-6405



Access COC's and Support Requests  
St. Martin's Waterworks  
Address: 25 Russell Rd Huntington, MA-01050  
Phone: 413-207-1164  
Project Name: Gateway Regional  
Project Location: Chester Elementary School  
Project Manager: Norene St. Martin  
Project Number: 2  
Sample ID: N54m  
Invoice Recipient: Gateway Regional School  
Sampled By: N54m

39 Spruce Street  
East Longmeadow, MA 01028

CHAIN OF CUSTODY RECORD

ANALYSIS REQUESTED

Requested Turnaround Time: 7-Day  10-Day  15-Day  30-Day   
 Discolored Metals Samples: Field Filtered  Lab to Filter   
 Orthophosphate Samples: Field Filtered  Lab to Filter   
 Rush/Approval Required: 1-Day  2-Day  3-Day  4-Day   
 Date Delivery: \_\_\_\_\_  
 Format: PDF  EXCEL   
 Other: \_\_\_\_\_  
 CLP Like Data Pkg Required:   
 Email To: KS.midy@grsd.org  
 Fax To #: \_\_\_\_\_

Preservation Code	Counter Use Only	Total Number Of:
VIALS		
GLASS		
PLASTIC		
BACTERIA		
ENCORE		

Glassware in the fridge?  Y/N

Glassware in freezer?  Y/N

Prepackaged Cooler?  Y/N

\*Pace Analytical is not responsible for missing samples from prepacked coolers.

1 Matrix Codes:  
 GW = Ground Water  
 WW = Waste Water  
 DW = Drinking Water  
 A = Air  
 S = Soil  
 SL = Sludge  
 SOL = Solid  
 O = Other (please define)

2 Preservation Codes:  
 I = Iced  
 H = HCL  
 M = Methanol  
 N = Nitric Acid  
 S = Sulfuric Acid  
 B = Sodium Bisulfate  
 X = Sodium Hydroxide  
 T = Sodium Thiosulfate  
 O = Other (please define)

Beginning Date/Time	Ending Date/Time	Matrix Code	Conc Code	VIALS	GLASS	PLASTIC	BACTERIA	ENCORE
3-17-23	7:45am	Grab		4				

Relinquished by: (signature) *Norene St. Martin*  
 Date/Time: 3-17-23 12:44  
 Received by: (signature) *[Signature]*  
 Date/Time: 3/17/23 12:44  
 Relinquished by: (signature) *[Signature]*  
 Date/Time: 3/17/23 16:48  
 Received by: (signature) *[Signature]*  
 Date/Time: 3/17/23 16:48  
 Relinquished by: (signature) *[Signature]*  
 Date/Time: \_\_\_\_\_  
 Received by: (signature) *[Signature]*  
 Date/Time: \_\_\_\_\_  
 Relinquished by: (signature) *[Signature]*  
 Date/Time: \_\_\_\_\_  
 Received by: (signature) *[Signature]*  
 Date/Time: \_\_\_\_\_

Client Comments: \_\_\_\_\_

Detection Limit Requirements: \_\_\_\_\_  
 MA MCP Required   
 MCP Certification Form Required   
 CT RCP Required   
 RCP Certification Form Required   
 MA State DW Required

Special Requirements: \_\_\_\_\_

Project Entity: Government  Federal  City   
 Municipality: 21 J Brownfield  
 MWRA  School  MBTA   
 WRTA   
 Other:  Chromatogram  ALPHA-LAP, LLC

Lab Comments: \_\_\_\_\_

Disclaimer: Pace Analytical is not responsible for any omitted information on the Chain of Custody. The Chain of Custody is a legal document that must be complete and accurate and is used to determine what analyses the laboratory will perform. Any missing information is not the laboratory's responsibility. Pace Analytical values your partnership on each project and will try to assist with missing information, but will not be held accountable.



