

April 4, 2023

Stephanie Fisk
Gateway School District
12 Littleville Road
Huntington, MA 01050

Project Location: 12 Littleville Rd., Huntington, MA
Client Job Number:
Project Number: 1
Laboratory Work Order Number: 23C1986
PWSID# 1143003

Enclosed are results of analyses for samples as received by the laboratory on March 17, 2023. If you have any questions concerning this report, please feel free to contact me.

Sincerely,



Kaitlyn A. Feliciano
Project Manager

Table of Contents

Sample Summary	3
Case Narrative	4
Sample Results	5
23C1986-01	5
23C1986-02	6
Sample Preparation Information	7
QC Data	8
Semivolatile Organic Compounds by - LC/MS-MS	8
B334620	8
Flag/Qualifier Summary	10
Certifications	11
Chain of Custody/Sample Receipt	12

Gateway School District
12 Littleville Road
Huntington, MA 01050
ATTN: Stephanie Fisk

REPORT DATE: 4/4/2023

PURCHASE ORDER NUMBER:

PROJECT NUMBER: 1

ANALYTICAL SUMMARY

WORK ORDER NUMBER: 23C1986

The results of analyses performed on the following samples submitted to Con-Test, a Pace Analytical Laboratory, are found in this report.

PROJECT LOCATION: 12 Littleville Rd., Huntington, MA

FIELD SAMPLE #	LAB ID:	MATRIX	SAMPLE DESCRIPTION	TEST	SUB LAB
POE	23C1986-01	Drinking Water		EPA 537.1	
POE FB	23C1986-02	Drinking Water		EPA 537.1	

CASE NARRATIVE SUMMARY

All reported results are within defined laboratory quality control objectives unless listed below or otherwise qualified in this report.

The results of analyses reported only relate to samples submitted to Con-Test, a Pace Analytical Laboratory, for testing. I certify that the analyses listed above, unless specifically listed as subcontracted, if any, were performed under my direction according to the approved methodologies listed in this document, and that based upon my inquiry of those individuals immediately responsible for obtaining the information, the material contained in this report is, to the best of my knowledge and belief, accurate and complete.



Lisa A. Worthington
Technical Representative



39 Spruce Street * East Longmeadow, MA 01028 * FAX 413/525-6405 * TEL. 413/525-2332

Project Location: 12 Littleville Rd., Huntington, MA

Sample Description:

Work Order: 23C1986

Date Received: 3/17/2023

Field Sample #: POE

Sampled: 3/17/2023 07:15

Sample ID: 23C1986-01

Sample Matrix: Drinking Water

Semivolatile Organic Compounds by - LC/MS-MS

Analyte	Results	RL	MCL/SMCL		Units	DF	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
			DL	MA ORSG							
Perfluorobutanesulfonic acid (PFBS)	ND	1.9	0.74		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:46	JR2
Perfluorohexanoic acid (PFHxA)	ND	1.9	0.88		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:46	JR2
Perfluorohexanesulfonic acid (PFHxS)	ND	1.9	0.85		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:46	JR2
Perfluoroheptanoic acid (PFHpA)	ND	1.9	0.94		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:46	JR2
Perfluorooctanoic acid (PFOA)	ND	1.9	0.98		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:46	JR2
Perfluorooctanesulfonic acid (PFOS)	ND	1.9	0.71		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:46	JR2
Perfluorononanoic acid (PFNA)	ND	1.9	0.87		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:46	JR2
Perfluorodecanoic acid (PFDA)	ND	1.9	0.91		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:46	JR2
N-EtFOSAA (NEtFOSAA)	ND	1.9	0.63		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:46	JR2
Perfluoroundecanoic acid (PFUnA)	ND	1.9	0.72		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:46	JR2
N-MeFOSAA (NMeFOSAA)	ND	1.9	0.70		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:46	JR2
Perfluorododecanoic acid (PFDoA)	ND	1.9	0.68		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:46	JR2
Perfluorotridecanoic acid (PFTrDA)	ND	1.9	0.69		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:46	JR2
Perfluorotetradecanoic acid (PFTA)	ND	1.9	0.79		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:46	JR2
Hexafluoropropylene oxide dimer acid (HFPO-DA)	ND	1.9	1.2		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:46	JR2
11Cl-PF3OUdS (F53B Major)	ND	1.9	0.63		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:46	JR2
9Cl-PF3ONS (F53B Minor)	ND	1.9	0.77		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:46	JR2
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	ND	1.9	0.83		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:46	JR2
Summation of PFAS 6 for Massachusetts	ND	1.9	1.9		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:46	JR2

Surrogates	% Recovery	Recovery Limits	Flag/Qual
13C-PFHxA	98.6	70-130	3/23/23 10:46
M3HFPO-DA	108	70-130	3/23/23 10:46
13C-PFDA	108	70-130	3/23/23 10:46
D5-NEtFOSAA	112	70-130	3/23/23 10:46



39 Spruce Street * East Longmeadow, MA 01028 * FAX 413/525-6405 * TEL. 413/525-2332

Project Location: 12 Littleville Rd., Huntington, MA

Sample Description:

Work Order: 23C1986

Date Received: 3/17/2023

Field Sample #: POE FB

Sampled: 3/17/2023 07:15

Sample ID: 23C1986-02

Sample Matrix: Drinking Water

Semivolatile Organic Compounds by - LC/MS-MS

Analyte	Results	RL	DL	MCL/SMCL		DF	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
				MA	ORSG						
Perfluorobutanesulfonic acid (PFBS)	ND	1.8	0.73		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:53	JR2
Perfluorohexanoic acid (PFHxA)	ND	1.8	0.86		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:53	JR2
Perfluorohexanesulfonic acid (PFHxS)	ND	1.8	0.83		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:53	JR2
Perfluoroheptanoic acid (PFHpA)	ND	1.8	0.92		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:53	JR2
Perfluorooctanoic acid (PFOA)	ND	1.8	0.96		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:53	JR2
Perfluorooctanesulfonic acid (PFOS)	ND	1.8	0.70		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:53	JR2
Perfluorononanoic acid (PFNA)	ND	1.8	0.85		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:53	JR2
Perfluorodecanoic acid (PFDA)	ND	1.8	0.89		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:53	JR2
N-EtFOSAA (NEtFOSAA)	ND	1.8	0.61		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:53	JR2
Perfluoroundecanoic acid (PFUnA)	ND	1.8	0.70		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:53	JR2
N-MeFOSAA (NMeFOSAA)	ND	1.8	0.69		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:53	JR2
Perfluorododecanoic acid (PFDoA)	ND	1.8	0.66		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:53	JR2
Perfluorotridecanoic acid (PFTrDA)	ND	1.8	0.67		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:53	JR2
Perfluorotetradecanoic acid (PFTA)	ND	1.8	0.77		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:53	JR2
Hexafluoropropylene oxide dimer acid (HFPO-DA)	ND	1.8	1.1		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:53	JR2
11Cl-PF3OUdS (F53B Major)	ND	1.8	0.61		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:53	JR2
9Cl-PF3ONS (F53B Minor)	ND	1.8	0.75		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:53	JR2
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	ND	1.8	0.81		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:53	JR2
Summation of PFAS 6 for Massachusetts	ND	1.8	1.8		ng/L	1		EPA 537.1	3/20/23	3/23/23 10:53	JR2

Surrogates	% Recovery	Recovery Limits	Flag/Qual
13C-PFHxA	104	70-130	3/23/23 10:53
M3HFPO-DA	108	70-130	3/23/23 10:53
13C-PFDA	101	70-130	3/23/23 10:53
D5-NEtFOSAA	104	70-130	3/23/23 10:53

Sample Extraction Data**Prep Method: EPA 537.1-EPA 537.1**

Lab Number [Field ID]	Batch	Initial [mL]	Final [mL]	Date
23C1986-01 [POE]	B334620	265	1.00	03/20/23
23C1986-02 [POE FB]	B334620	271	1.00	03/20/23

QUALITY CONTROL
Semivolatile Organic Compounds by - LC/MS-MS - Quality Control

Analyte	Result	Reporting Limit	DL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----------------	----	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch B334620 - EPA 537.1
Blank (B334620-BLK1)

Prepared: 03/20/23 Analyzed: 03/23/23

Perfluorobutanesulfonic acid (PFBS)	ND	1.9	0.76	ng/L							
Perfluorohexanoic acid (PFHxA)	ND	1.9	0.90	ng/L							
Perfluorohexanesulfonic acid (PFHxS)	ND	1.9	0.86	ng/L							
Perfluoroheptanoic acid (PFHpA)	ND	1.9	0.96	ng/L							
Perfluorooctanoic acid (PFOA)	ND	1.9	0.99	ng/L							
Perfluorooctanesulfonic acid (PFOS)	ND	1.9	0.73	ng/L							
Perfluorononanoic acid (PFNA)	ND	1.9	0.89	ng/L							
Perfluorodecanoic acid (PFDA)	ND	1.9	0.92	ng/L							
N-EtFOSAA (NEtFOSAA)	ND	1.9	0.64	ng/L							
Perfluoroundecanoic acid (PFUnA)	ND	1.9	0.73	ng/L							
N-MeFOSAA (NMeFOSAA)	ND	1.9	0.71	ng/L							
Perfluorododecanoic acid (PFDoA)	ND	1.9	0.69	ng/L							
Perfluorotridecanoic acid (PFTriDA)	ND	1.9	0.70	ng/L							
Perfluorotetradecanoic acid (PFTA)	ND	1.9	0.80	ng/L							
Hexafluoropropylene oxide dimer acid (HFPO-DA)	ND	1.9	1.2	ng/L							
11Cl-PF3OUdS (F53B Major)	ND	1.9	0.64	ng/L							
9Cl-PF3ONS (F53B Minor)	ND	1.9	0.78	ng/L							
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	ND	1.9	0.85	ng/L							
Summation of PFAS 6 for Massachusetts	ND	1.9	1.9	ng/L							
Surrogate: 13C-PFHxA	32.0			ng/L	38.5		83.1	70-130			
Surrogate: M3HFPO-DA	31.5			ng/L	38.5		82.0	70-130			
Surrogate: 13C-PFDA	32.0			ng/L	38.5		83.1	70-130			
Surrogate: D5-NEtFOSAA	126			ng/L	154		82.1	70-130			

LCS (B334620-BS1)

Prepared: 03/20/23 Analyzed: 03/23/23

Perfluorobutanesulfonic acid (PFBS)	6.80	1.9	0.74	ng/L	8.33		81.6	70-130			
Perfluorohexanoic acid (PFHxA)	7.06	1.9	0.88	ng/L	9.39		75.2	70-130			
Perfluorohexanesulfonic acid (PFHxS)	7.47	1.9	0.84	ng/L	8.58		87.1	70-130			
Perfluoroheptanoic acid (PFHpA)	7.63	1.9	0.93	ng/L	9.39		81.3	70-130			
Perfluorooctanoic acid (PFOA)	7.57	1.9	0.97	ng/L	9.39		80.6	70-130			
Perfluorooctanesulfonic acid (PFOS)	7.38	1.9	0.71	ng/L	8.71		84.8	70-130			
Perfluorononanoic acid (PFNA)	7.69	1.9	0.86	ng/L	9.39		81.9	70-130			
Perfluorodecanoic acid (PFDA)	7.09	1.9	0.90	ng/L	9.39		75.6	70-130			
N-EtFOSAA (NEtFOSAA)	7.32	1.9	0.62	ng/L	9.39		77.9	70-130			
Perfluoroundecanoic acid (PFUnA)	7.18	1.9	0.71	ng/L	9.39		76.5	70-130			
N-MeFOSAA (NMeFOSAA)	7.44	1.9	0.70	ng/L	9.39		79.2	70-130			
Perfluorododecanoic acid (PFDoA)	6.92	1.9	0.67	ng/L	9.39		73.7	70-130			
Perfluorotridecanoic acid (PFTriDA)	6.71	1.9	0.68	ng/L	9.39		71.5	70-130			
Perfluorotetradecanoic acid (PFTA)	7.32	1.9	0.78	ng/L	9.39		78.0	70-130			
Hexafluoropropylene oxide dimer acid (HFPO-DA)	7.53	1.9	1.2	ng/L	9.39		80.2	70-130			
11Cl-PF3OUdS (F53B Major)	6.79	1.9	0.62	ng/L	8.85		76.7	70-130			
9Cl-PF3ONS (F53B Minor)	7.31	1.9	0.76	ng/L	8.76		83.4	70-130			
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	7.20	1.9	0.83	ng/L	8.87		81.2	70-130			
Summation of PFAS 6 for Massachusetts	44.8	1.9	1.9	ng/L				0-200			
Surrogate: 13C-PFHxA	32.1			ng/L	37.5		85.5	70-130			
Surrogate: M3HFPO-DA	33.3			ng/L	37.5		88.6	70-130			
Surrogate: 13C-PFDA	32.1			ng/L	37.5		85.5	70-130			

QUALITY CONTROL

Semivolatile Organic Compounds by - LC/MS-MS - Quality Control

Analyte	Result	Reporting Limit	DL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
Batch B334620 - EPA 537.1											
LCS (B334620-BS1)											
						Prepared: 03/20/23 Analyzed: 03/23/23					
Surrogate: D5-NEtFOSAA	134			ng/L	150		89.2	70-130			
LCS Dup (B334620-BS1)											
						Prepared: 03/20/23 Analyzed: 03/23/23					
Perfluorobutanesulfonic acid (PFBS)	7.60	1.9	0.76	ng/L	8.51		89.3	70-130	11.1	30	
Perfluorohexanoic acid (PFHxA)	8.39	1.9	0.90	ng/L	9.59		87.5	70-130	17.2	30	
Perfluorohexanesulfonic acid (PFHxS)	7.93	1.9	0.86	ng/L	8.76		90.5	70-130	5.90	30	
Perfluoroheptanoic acid (PFHpA)	8.70	1.9	0.95	ng/L	9.59		90.7	70-130	13.1	30	
Perfluorooctanoic acid (PFOA)	8.09	1.9	0.99	ng/L	9.59		84.4	70-130	6.69	30	
Perfluorooctanesulfonic acid (PFOS)	7.37	1.9	0.73	ng/L	8.90		82.8	70-130	0.160	30	
Perfluorononanoic acid (PFNA)	8.57	1.9	0.88	ng/L	9.59		89.4	70-130	10.8	30	
Perfluorodecanoic acid (PFDA)	8.90	1.9	0.92	ng/L	9.59		92.8	70-130	22.6	30	
N-EtFOSAA (NEtFOSAA)	7.66	1.9	0.64	ng/L	9.59		79.9	70-130	4.63	30	
Perfluoroundecanoic acid (PFUnA)	8.28	1.9	0.73	ng/L	9.59		86.3	70-130	14.3	30	
N-MeFOSAA (NMeFOSAA)	8.74	1.9	0.71	ng/L	9.59		91.2	70-130	16.1	30	
Perfluorododecanoic acid (PFDoA)	8.11	1.9	0.69	ng/L	9.59		84.5	70-130	15.8	30	
Perfluorotridecanoic acid (PFTriDA)	8.17	1.9	0.70	ng/L	9.59		85.2	70-130	19.7	30	
Perfluorotetradecanoic acid (PFTA)	8.27	1.9	0.80	ng/L	9.59		86.3	70-130	12.2	30	
Hexafluoropropylene oxide dimer acid (HFPO-DA)	8.73	1.9	1.2	ng/L	9.59		91.1	70-130	14.8	30	
11Cl-PF3OUdS (F53B Major)	7.72	1.9	0.64	ng/L	9.04		85.4	70-130	12.9	30	
9Cl-PF3ONS (F53B Minor)	8.16	1.9	0.78	ng/L	8.95		91.2	70-130	11.0	30	
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	8.25	1.9	0.84	ng/L	9.06		91.0	70-130	13.5	30	
Summation of PFAS 6 for Massachusetts	49.6	1.9	1.9	ng/L				0-200	10.0		
Surrogate: 13C-PFHxA	39.6			ng/L	38.4		103	70-130			
Surrogate: M3HFPO-DA	40.3			ng/L	38.4		105	70-130			
Surrogate: 13C-PFDA	39.4			ng/L	38.4		103	70-130			
Surrogate: D5-NEtFOSAA	160			ng/L	153		104	70-130			

FLAG/QUALIFIER SUMMARY

*	QC result is outside of established limits.
†	Wide recovery limits established for difficult compound.
‡	Wide RPD limits established for difficult compound.
#	Data exceeded client recommended or regulatory level
ND	Not Detected
RL	Reporting Limit
DL	Method Detection Limit
MCL	Maximum Contaminant Level

Percent recoveries and relative percent differences (RPDs) are determined by the software using values in the calculation which have not been rounded.

No results have been blank subtracted unless specified in the case narrative section.

CERTIFICATIONS
Certified Analyses included in this Report

Analyte	Certifications
EPA 537.1 in Drinking Water	
Perfluorobutanesulfonic acid (PFBS)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Perfluorohexanoic acid (PFHxA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Perfluorohexanesulfonic acid (PFHxS)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Perfluoroheptanoic acid (PFHpA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Perfluorooctanoic acid (PFOA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Perfluorooctanesulfonic acid (PFOS)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Perfluorononanoic acid (PFNA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Perfluorodecanoic acid (PFDA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
N-EtFOSAA (NEtFOSAA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Perfluoroundecanoic acid (PFUnA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
N-MeFOSAA (NMeFOSAA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Perfluorododecanoic acid (PFDoA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Perfluorotridecanoic acid (PFTriDA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Perfluorotetradecanoic acid (PFTA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
Hexafluoropropylene oxide dimer acid (HFPO-DA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
11Cl-PF3OUdS (F53B Major)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
9Cl-PF3ONS (F53B Minor)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	VT-DW,NJ,CT,ME,PA,MI,MA,NY,NH,OH

Con-Test, a Pace Environmental Laboratory, operates under the following certifications and accreditations:

Code	Description	Number	Expires
MA	Massachusetts DEP	M-MA100	06/30/2023
CT	Connecticut Department of Public Health	PH-0821	12/31/2024
NY	New York State Department of Health	10899 NELAP	04/1/2024
NH	New Hampshire Environmental Lab	2516 NELAP	02/5/2024
NJ	New Jersey DEP	MA007 NELAP	06/30/2023
VT-DW	Vermont Department of Health Drinking Water	VT-255716	06/12/2023
ME	State of Maine	MA00100	06/9/2023
PA	Commonwealth of Pennsylvania DEP	68-05812	06/30/2023
MI	Dept. of Env, Great Lakes, and Energy	9100	06/30/2023
OH	Ohio Environmental Protection Agency	87781	04/1/2024

23C1986

ANALYSIS REQUESTED

Requested Turnaround Time: 7-Day 10-Day 14-Day

PFAS 10-Day (std) Due Date: _____

Rush-Approval Required: 3-Day 4-Day

1-Day 2-Day Data Delivery

Format: PDF EXCEL

Other: _____

CLP Like Data Pkg Required:

Email To: **Ksmldy@psd.org**

Fax To #: _____

Disolved Metals Samples: Field Filtered Lab to Filter

Orthophosphate Samples: Field Filtered Lab to Filter

PCB ONLY

SOXHLET

NON SOXHLET

Project Manager: **NORENE ST. MARTIN**

Project Name/Number: _____

Invoice Recipient: **Gateway Regional School**

Sampled By: **NST/M**

Client Sample ID / Description: **1JA POE**

Beginning Date/Time: **3/17/23 7:15am Grab**

Ending Date/Time: _____

Matrix Code: _____

Comp/Grab: _____

City Code: _____

Requested	7-Day	10-Day	14-Day	Field Filtered	Lab to Filter	Orthophosphate	Field Filtered	Lab to Filter	PCB ONLY	SOXHLET	NON SOXHLET	VIALS	GLASS	PLASTIC	BACTERIA	ENCORE

Relinquished by: (signature) **Norene St. Martin**

Received by: (signature) _____

Relinquished by: (signature) _____

Received by: (signature) **Denise White 2-4**

Date/Time: **3/17/23 12:44**

Date/Time: **3/17/23 12:44**

Date/Time: **3/17/23 16:48**

Date/Time: **3/17/23 16:48**

Client Comments: _____

Detection Limit Requirements: _____

Special Requirements: _____

MA MCP Required

MCP Certification Form Required

CT RCP Required

RCP Certification Form Required

MA State DW Required

PWSID # _____

Project Entity: Government Federal City

Municipality Z J Brownfield

MWRA School MBTA

WRTA

Other: Chromatogram AIHA-LAP, LLC

Preservation Codes: I = Iced H = HCL M = Methanol N = Nitric Acid S = Sulfuric Acid B = Sodium Bisulfate X = Sodium Hydroxide T = Sodium Thiosulfate O = Other (please define)

Matrix Codes: GW = Ground Water WW = Waste Water DW = Drinking Water A = Air S = Soil SL = Sludge SOL = Solid O = Other (please define)

Preservation Codes: I = Iced H = HCL M = Methanol N = Nitric Acid S = Sulfuric Acid B = Sodium Bisulfate X = Sodium Hydroxide T = Sodium Thiosulfate O = Other (please define)

Carrier Use Only

Total Number Of: _____

VIALS _____

GLASS _____

PLASTIC _____

BACTERIA _____

ENCORE _____

Glassware in the fridge? Y N

Glassware in freezer? Y N

Prepackaged Cooler? Y N

*Pace Analytical is not responsible for missing samples from prepacked coolers.

*Matrix Codes: GW = Ground Water WW = Waste Water DW = Drinking Water A = Air S = Soil SL = Sludge SOL = Solid O = Other (please define)

Preservation Codes: I = Iced H = HCL M = Methanol N = Nitric Acid S = Sulfuric Acid B = Sodium Bisulfate X = Sodium Hydroxide T = Sodium Thiosulfate O = Other (please define)

Please use the following codes to indicate possible sample concentration within the Conc Code column above:
H - High; M - Medium; L - Low; C - Clean; U - Unknown

MA State DW Required

PWSID # _____

Project Entity: Government Federal City

Municipality Z J Brownfield

MWRA School MBTA

WRTA

Other: Chromatogram AIHA-LAP, LLC

Lab Comments: _____

Disclaimer: Pace Analytical is not responsible for any omitted information on the Chain of Custody. The Chain of Custody is a legal document that must be complete and accurate and is used to determine what analyses the laboratory will perform. Any missing information is not the laboratory's responsibility. Pace Analytical values your partnership on each project and will try to assist with missing information, but will not be held accountable.

