



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

May 09, 2023

Ron Havniar
Medford School District
815 S Oakdale Ave
Medford, OR 97501
TEL: 541-842-3646
FAX 541-842-1160

RE: Jacksonville Elem Lead Testing

Order No.: 23040992

Dear Ron Havniar:

Neilson Research Corporation received 94 sample(s) on 4/26/2023 for the analyses presented in the following report.

The results relate only to the parameters tested or to the sample as received by the laboratory. This report shall not be reproduced except in full, without the written approval of Neilson Research Corporation. If you have any questions regarding these test results, please feel free to call.

Sincerely,
Neilson Research Corporation

Tamra Schmedemann
Senior Project Manager
245 S Grape St
Medford, OR 97501



Original



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Case Narrative

WO#: 23040992
Date: 5/9/2023

CLIENT: Medford School District

Project: Jacksonville Elem Lead Testing

The analyses were performed according to the guidelines in the Neilson Research Corporation Quality Assurance Program. This report contains analytical results for the sample(s) as received by the laboratory.

Neilson Research Corporation certifies that this report is in compliance with the requirements of NELAP. No unusual difficulties were experienced during analysis of this batch except as noted below or qualified with data flags on the reports.

Original



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992
Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-01 **Client Sample ID:** Rm 7-DF
Collection Date: 4/26/2023 6:24:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Rm 7-DF

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	2.26		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-02 **Client Sample ID:** Rm 7-Sink
Collection Date: 4/26/2023 6:24:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Rm 7-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	8.27		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-03 **Client Sample ID:** Rm 7-RR (L) Sink
Collection Date: 4/26/2023 6:25:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Rm 7-RR (L) Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	7.01		0.500	ppb	1	5/4/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992
Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-04 **Client Sample ID:** Rm 7-RR (R) Sink
Collection Date: 4/26/2023 6:25:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Rm 7-RR (R) Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	6.46		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-05 **Client Sample ID:** Rm 6-DF
Collection Date: 4/26/2023 6:26:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Rm 6-DF

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	9.38		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-06 **Client Sample ID:** Rm 6-Sink
Collection Date: 4/26/2023 6:26:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Rm 6-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	11.1		0.500	ppb	1	5/4/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992

Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-07 **Client Sample ID:** Staff Work Rm (by Ofc)-Sink
Collection Date: 4/26/2023 6:27:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Staff Work Rm (by Ofc)-Sink

Trace Metals by EPA 200.8 ICP-MS							Analyst: CJS	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	3.34		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-08 **Client Sample ID:** Health Room-Sink
Collection Date: 4/26/2023 6:27:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Health Room-Sink

Trace Metals by EPA 200.8 ICP-MS							Analyst: CJS	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	7.22		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-09 **Client Sample ID:** Health RR-Sink
Collection Date: 4/26/2023 6:28:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Health RR-Sink

Trace Metals by EPA 200.8 ICP-MS							Analyst: CJS	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	1.16		0.500	ppb	1	5/4/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992
Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-10 **Client Sample ID:** Girls RR by Media Ctr-S 1
Collection Date: 4/26/2023 6:28:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Girls RR by Media Ctr-S 1

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	2.27		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-11 **Client Sample ID:** Girls RR by Media Ctr-S 2
Collection Date: 4/26/2023 6:29:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Girls RR by Media Ctr-S 2

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	2.53		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-12 **Client Sample ID:** Girls RR by Media Ctr-S 3
Collection Date: 4/26/2023 6:29:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Girls RR by Media Ctr-S 3

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	0.766		0.500	ppb	1	5/4/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992

Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-13 **Client Sample ID:** Girls RR by Media Ctr-S 4
Collection Date: 4/26/2023 6:30:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Girls RR by Media Ctr-S 4

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	14.7		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-14 **Client Sample ID:** Staff RR by Media Ctr-Sink
Collection Date: 4/26/2023 6:30:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Staff RR by Media Ctr-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	2.01		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-15 **Client Sample ID:** Boys RR by Media Ctr- (L) S
Collection Date: 4/26/2023 6:31:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Boys RR by Media Ctr- (L) S

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	2.37		0.500	ppb	1	5/4/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992
Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-16 **Client Sample ID:** Boys RR by Media Ctr- (R) S
Collection Date: 4/26/2023 6:31:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Boys RR by Media Ctr- (R) S

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	1.75		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-17 **Client Sample ID:** Outside Rm 7-DF (Outdoor)
Collection Date: 4/26/2023 6:32:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Outside Rm 7-DF (Outdoor)

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	66.0	*	0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-18 **Client Sample ID:** Media Center Work Rm-Sink
Collection Date: 4/26/2023 6:33:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Media Center Work Rm-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	27.3	*	0.500	ppb	1	5/4/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992

Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-19 **Client Sample ID:** Computer Lab-Sink
Collection Date: 4/26/2023 6:34:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Computer Lab-Sink

Trace Metals by EPA 200.8 ICP-MS							Analyst: CJS	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	6.72		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-20 **Client Sample ID:** ERI Rm-Sink
Collection Date: 4/26/2023 6:35:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** ERI Rm-Sink

Trace Metals by EPA 200.8 ICP-MS							Analyst: CJS	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	3.89		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-21 **Client Sample ID:** Btwn Media Ctr/Rm 16-DF
Collection Date: 4/26/2023 6:36:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Btwn Media Ctr/Rm 16-DF

Trace Metals by EPA 200.8 ICP-MS							Analyst: CBB	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	4.86		0.515	ppb	1	5/5/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992
Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-22 **Client Sample ID:** Rm 16-DF
Collection Date: 4/26/2023 6:37:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Rm 16-DF

Trace Metals by EPA 200.8 ICP-MS						Analyst: CBB		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	2.13		0.515	ppb	1	5/5/2023	15.0	A

Lab ID: 23040992-23 **Client Sample ID:** Rm 16-Sink
Collection Date: 4/26/2023 6:37:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Rm 16-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CBB		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	8.76		0.515	ppb	1	5/5/2023	15.0	A

Lab ID: 23040992-24 **Client Sample ID:** Art Supply Rm-Sink
Collection Date: 4/26/2023 6:38:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Art Supply Rm-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CBB		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	26.1	*	0.515	ppb	1	5/5/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992

Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-25 **Client Sample ID:** Rm 17-DF
Collection Date: 4/26/2023 6:39:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Rm 17-DF

Trace Metals by EPA 200.8 ICP-MS						Analyst; CBB		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	2.35		0.515	ppb	1	5/5/2023	15.0	A

Lab ID: 23040992-26 **Client Sample ID:** Rm 17-Sink
Collection Date: 4/26/2023 6:39:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Rm 17-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst; CBB		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	3.63		0.515	ppb	1	5/5/2023	15.0	A

Lab ID: 23040992-27 **Client Sample ID:** Girls RR O/S Rm 17-Sink
Collection Date: 4/26/2023 6:40:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Girls RR O/S Rm 17-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst; CBB		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	1.20		0.515	ppb	1	5/5/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992

Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-28 **Client Sample ID:** Boys RR O/S Rm 17-Sink
Collection Date: 4/26/2023 6:41:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Boys RR O/S Rm 17-Sink

Trace Metals by EPA 200.8 ICP-MS							Analyst: CBB	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	1.41		0.515	ppb	1	5/5/2023	15.0	A

Lab ID: 23040992-29 **Client Sample ID:** Outside Rm 17- (L) DF
Collection Date: 4/26/2023 6:42:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Outside Rm 17- (L) DF

Trace Metals by EPA 200.8 ICP-MS							Analyst: CBB	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	6.45		0.515	ppb	1	5/5/2023	15.0	A

Lab ID: 23040992-30 **Client Sample ID:** Outside Rm 17- (R) DF
Collection Date: 4/26/2023 6:42:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Outside Rm 17- (R) DF

Trace Metals by EPA 200.8 ICP-MS							Analyst: CBB	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	3.17		0.515	ppb	1	5/5/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992
Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-31 **Client Sample ID:** Rm 18-DF
Collection Date: 4/26/2023 6:44:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Rm 18-DF

Trace Metals by EPA 200.8 ICP-MS						Analyst: CBB		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	2.14		0.515	ppb	1	5/5/2023	15.0	A

Lab ID: 23040992-32 **Client Sample ID:** Rm 18-Sink
Collection Date: 4/26/2023 6:45:00 AM **Collected By:** Seth Barry
Matrix: Drinking Water **Sample Location:** Rm 18-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CBB		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	4.93		0.515	ppb	1	5/5/2023	15.0	A

Lab ID: 23040992-33 **Client Sample ID:** Kitchen-Hand Wash Sink
Collection Date: 4/26/2023 6:26:00 AM **Collected By:** Gary Severe
Matrix: Drinking Water **Sample Location:** Kitchen-Hand Wash Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CBB		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	4.48		0.515	ppb	1	5/5/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992
Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-34 Client Sample ID: Kitchen-Veg Prep Sink
Collection Date: 4/26/2023 6:26:00 AM Collected By: Gary Severe
Matrix: Drinking Water Sample Location: Kitchen-Veg Prep Sink

Trace Metals by EPA 200.8 ICP-MS							Analyst: CBB	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	0.985		0.515	ppb	1	5/5/2023	15.0	A

Lab ID: 23040992-35 Client Sample ID: Kitchen-Pre-Rinse Sink (DW)
Collection Date: 4/26/2023 6:27:00 AM Collected By: Gary Severe
Matrix: Drinking Water Sample Location: Kitchen-Pre-Rinse Sink (DW)

Trace Metals by EPA 200.8 ICP-MS							Analyst: CBB	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	1.22		0.515	ppb	1	5/5/2023	15.0	A

Lab ID: 23040992-36 Client Sample ID: By Cafeteria-Elkay Btl Filler-Upr
Collection Date: 4/26/2023 6:28:00 AM Collected By: Gary Severe
Matrix: Drinking Water Sample Location: By Cafeteria-Elkay Btl Filler-Upr

Trace Metals by EPA 200.8 ICP-MS							Analyst: CBB	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	ND		0.515	ppb	1	5/5/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992
Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-37 Client Sample ID: By Cafeteria-Elkay DW-Lwr
Collection Date: 4/26/2023 6:28:00 AM Collected By: Gary Severe
Matrix: Drinking Water Sample Location: By Cafeteria-Elkay DW-Lwr

Trace Metals by EPA 200.8 ICP-MS							Analyst: CBB	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	ND		0.515	ppb	1	5/5/2023	15.0	A

Lab ID: 23040992-38 Client Sample ID: Staff Lounge Rm-Sink
Collection Date: 4/26/2023 6:30:00 AM Collected By: Gary Severe
Matrix: Drinking Water Sample Location: Staff Lounge Rm-Sink

Trace Metals by EPA 200.8 ICP-MS							Analyst: CBB	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	2.60		0.515	ppb	1	5/5/2023	15.0	A

Lab ID: 23040992-39 Client Sample ID: Rm 1-DF
Collection Date: 4/26/2023 6:31:00 AM Collected By: Gary Severe
Matrix: Drinking Water Sample Location: Rm 1-DF

Trace Metals by EPA 200.8 ICP-MS							Analyst: CBB	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	2.93		0.515	ppb	1	5/5/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992
Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-40 **Client Sample ID:** Rm 1-Sink
Collection Date: 4/26/2023 6:32:00 AM **Collected By:** Gary Severe
Matrix: Drinking Water **Sample Location:** Rm 1-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CBB		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	6.71		0.515	ppb	1	5/5/2023	15.0	A

Lab ID: 23040992-41 **Client Sample ID:** Rm 1-RR Sink
Collection Date: 4/26/2023 6:33:00 AM **Collected By:** Gary Severe
Matrix: Drinking Water **Sample Location:** Rm 1-RR Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	4.81		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-42 **Client Sample ID:** Boys RR O/S Rm 1/3-(L) S
Collection Date: 4/26/2023 6:35:00 AM **Collected By:** Gary Severe
Matrix: Drinking Water **Sample Location:** Boys RR O/S Rm 1/3-(L) S

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	0.590		0.500	ppb	1	5/4/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992
Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-43 **Client Sample ID:** Boys RR O/S Rm 1/3-(C) S
Collection Date: 4/26/2023 6:35:00 AM **Collected By:** Gary Severe
Matrix: Drinking Water **Sample Location:** Boys RR O/S Rm 1/3-(C) S

Trace Metals by EPA 200.8 ICP-MS							Analyst: CJS	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	1.12		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-44 **Client Sample ID:** Boys RR O/S Rm 1/3-(R) S
Collection Date: 4/26/2023 6:36:00 AM **Collected By:** Gary Severe
Matrix: Drinking Water **Sample Location:** Boys RR O/S Rm 1/3-(R) S

Trace Metals by EPA 200.8 ICP-MS							Analyst: CJS	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	1.05		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-45 **Client Sample ID:** Girls RR O/S Rm 1/3-(L) S
Collection Date: 4/26/2023 6:37:00 AM **Collected By:** Gary Severe
Matrix: Drinking Water **Sample Location:** Girls RR O/S Rm 1/3-(L) S

Trace Metals by EPA 200.8 ICP-MS							Analyst: CJS	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	5.04		0.500	ppb	1	5/4/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992
Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-46 **Client Sample ID:** Girls RR O/S Rm 1/3-(C) S
Collection Date: 4/26/2023 6:37:00 AM **Collected By:** Gary Severe
Matrix: Drinking Water **Sample Location:** Girls RR O/S Rm 1/3-(C) S

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	4.39		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-47 **Client Sample ID:** Girls RR O/S Rm 1/3-(R) S
Collection Date: 4/26/2023 6:37:00 AM **Collected By:** Gary Severe
Matrix: Drinking Water **Sample Location:** Girls RR O/S Rm 1/3-(R) S

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	4.44		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-48 **Client Sample ID:** Staff RR O/S Rm 1/3-Sink
Collection Date: 4/26/2023 6:38:00 AM **Collected By:** Gary Severe
Matrix: Drinking Water **Sample Location:** Staff RR O/S Rm 1/3-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	3.68		0.500	ppb	1	5/4/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992

Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-49 **Client Sample ID:** Storage Rm Near Rm 3-Sink
Collection Date: 4/26/2023 6:39:00 AM **Collected By:** Gary Severe
Matrix: Drinking Water **Sample Location:** Storage Rm Near Rm 3-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS			
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status	
Lead	19.4	*	0.500	ppb	1	5/4/2023	15.0	A	

Lab ID: 23040992-50 **Client Sample ID:** Storage Rm Near Rm 3-DF
Collection Date: 4/26/2023 6:39:00 AM **Collected By:** Gary Severe
Matrix: Drinking Water **Sample Location:** Storage Rm Near Rm 3-DF

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS			
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status	
Lead	11.1		0.500	ppb	1	5/4/2023	15.0	A	

Lab ID: 23040992-51 **Client Sample ID:** Rm 3-DF
Collection Date: 4/26/2023 6:40:00 AM **Collected By:** Gary Severe
Matrix: Drinking Water **Sample Location:** Rm 3-DF

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS			
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status	
Lead	13.6		0.500	ppb	1	5/4/2023	15.0	A	

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992
Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-52 **Client Sample ID:** Rm 3-Sink
Collection Date: 4/26/2023 6:40:00 AM **Collected By:** Gary Severe
Matrix: Drinking Water **Sample Location:** Rm 3-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	8.96		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-53 **Client Sample ID:** Rm 3-RR Sink (L)
Collection Date: 4/26/2023 6:41:00 AM **Collected By:** Gary Severe
Matrix: Drinking Water **Sample Location:** Rm 3-RR Sink (L)

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	5.13		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-54 **Client Sample ID:** Rm 3-RR Sink (R)
Collection Date: 4/26/2023 6:42:00 AM **Collected By:** Gary Severe
Matrix: Drinking Water **Sample Location:** Rm 3-RR Sink (R)

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	6.69		0.500	ppb	1	5/4/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992

Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-55 **Client Sample ID:** Rm 2-DF
Collection Date: 4/26/2023 6:43:00 AM **Collected By:** Gary Severe
Matrix: Drinking Water **Sample Location:** Rm 2-DF

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	24.3	*	0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-56 **Client Sample ID:** Rm 2-Sink
Collection Date: 4/26/2023 6:43:00 AM **Collected By:** Gary Severe
Matrix: Drinking Water **Sample Location:** Rm 2-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	12.5		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-57 **Client Sample ID:** Rm 5-DF
Collection Date: 4/26/2023 6:45:00 AM **Collected By:** Gary Severe
Matrix: Drinking Water **Sample Location:** Rm 5-DF

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	6.62		0.500	ppb	1	5/4/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992
Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-58 **Client Sample ID:** Rm 5-Sink
Collection Date: 4/26/2023 6:45:00 AM **Collected By:** Gary Severe
Matrix: Drinking Water **Sample Location:** Rm 5-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS			
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status	
Lead	17.6	*	0.500	ppb	1	5/4/2023	15.0	A	

Lab ID: 23040992-59 **Client Sample ID:** Rm 5-RR Sink (L)
Collection Date: 4/26/2023 6:46:00 AM **Collected By:** Gary Severe
Matrix: Drinking Water **Sample Location:** Rm 5-RR Sink (L)

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS			
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status	
Lead	5.84		0.500	ppb	1	5/4/2023	15.0	A	

Lab ID: 23040992-60 **Client Sample ID:** Rm 4-DF
Collection Date: 4/26/2023 6:48:00 AM **Collected By:** Gary Severe
Matrix: Drinking Water **Sample Location:** Rm 4-DF

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS			
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status	
Lead	5.33		0.500	ppb	1	5/4/2023	15.0	A	

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992
Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-61 Client Sample ID: Rm 4-Sink
Collection Date: 4/26/2023 6:49:00 AM Collected By: Gary Severe
Matrix: Drinking Water Sample Location: Rm 4-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	10.8		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-62 Client Sample ID: Rm 19-DF
Collection Date: 4/26/2023 6:25:00 AM Collected By: Mike Medel
Matrix: Drinking Water Sample Location: Rm 19-DF

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	2.26		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-63 Client Sample ID: Rm 19-Sink
Collection Date: 4/26/2023 6:25:00 AM Collected By: Mike Medel
Matrix: Drinking Water Sample Location: Rm 19-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	4.77		0.500	ppb	1	5/4/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992
Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-64 **Client Sample ID:** Rm 20-DF
Collection Date: 4/26/2023 6:28:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Rm 20-DF

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	1.77		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-65 **Client Sample ID:** Rm 20-Sink
Collection Date: 4/26/2023 6:28:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Rm 20-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	6.71		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-66 **Client Sample ID:** Staff RR O/S Rm 20-Sink
Collection Date: 4/26/2023 6:31:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Staff RR O/S Rm 20-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	ND		0.500	ppb	1	5/4/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992

Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-67 **Client Sample ID:** Rm 21-DF
Collection Date: 4/26/2023 6:33:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Rm 21-DF

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	1.72		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-68 **Client Sample ID:** Rm 21-Sink
Collection Date: 4/26/2023 6:33:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Rm 21-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	7.74		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-69 **Client Sample ID:** Rm 8-DF
Collection Date: 4/26/2023 6:36:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Rm 8-DF

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	2.75		0.500	ppb	1	5/4/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992

Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-70 **Client Sample ID:** Rm 8-Sink
Collection Date: 4/26/2023 6:36:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Rm 8-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	3.34		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-71 **Client Sample ID:** Rm 10-DF
Collection Date: 4/26/2023 6:38:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Rm 10-DF

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	1.83		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-72 **Client Sample ID:** Rm 10-Sink
Collection Date: 4/26/2023 6:38:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Rm 10-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	1.75		0.500	ppb	1	5/4/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992
Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-73 **Client Sample ID:** Rm 9-DF
Collection Date: 4/26/2023 6:40:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Rm 9-DF

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	1.76		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-74 **Client Sample ID:** Rm 9-Sink
Collection Date: 4/26/2023 6:40:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Rm 9-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	3.56		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-75 **Client Sample ID:** Rm 11-DF
Collection Date: 4/26/2023 6:42:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Rm 11-DF

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	1.58		0.500	ppb	1	5/4/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992

Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-76 **Client Sample ID:** Rm 11-Sink
Collection Date: 4/26/2023 6:42:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Rm 11-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	2.69		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-77 **Client Sample ID:** Rm 12-DF
Collection Date: 4/26/2023 6:44:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Rm 12-DF

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	1.55		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-78 **Client Sample ID:** Rm 12-Sink
Collection Date: 4/26/2023 6:44:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Rm 12-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	3.29		0.500	ppb	1	5/4/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992
Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-79 **Client Sample ID:** Rm 14-DF
Collection Date: 4/26/2023 6:46:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Rm 14-DF

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	3.66		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-80 **Client Sample ID:** Rm 14-Sink
Collection Date: 4/26/2023 6:46:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Rm 14-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CJS		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	2.02		0.500	ppb	1	5/4/2023	15.0	A

Lab ID: 23040992-81 **Client Sample ID:** Rm 13-DF
Collection Date: 4/26/2023 6:51:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Rm 13-DF

Trace Metals by EPA 200.8 ICP-MS						Analyst: CBB		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	2.23		0.515	ppb	1	5/5/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992
Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-82 Client Sample ID: Rm 13-Sink
Collection Date: 4/26/2023 6:51:00 AM Collected By: Mike Medel
Matrix: Drinking Water Sample Location: Rm 13-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CBB		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	1.56		0.515	ppb	1	5/5/2023	15.0	A

Lab ID: 23040992-83 Client Sample ID: Rm 15-DF
Collection Date: 4/26/2023 6:54:00 AM Collected By: Mike Medel
Matrix: Drinking Water Sample Location: Rm 15-DF

Trace Metals by EPA 200.8 ICP-MS						Analyst: CBB		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	3.15		0.515	ppb	1	5/5/2023	15.0	A

Lab ID: 23040992-84 Client Sample ID: Rm 15-Sink
Collection Date: 4/26/2023 6:54:00 AM Collected By: Mike Medel
Matrix: Drinking Water Sample Location: Rm 15-Sink

Trace Metals by EPA 200.8 ICP-MS						Analyst: CBB		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	2.51		0.515	ppb	1	5/5/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992
Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-85 **Client Sample ID:** Boys (O/S Gym/Rm 14/15)-S (L)
Collection Date: 4/26/2023 6:56:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Boys (O/S Gym/Rm 14/15)-S (L)

Trace Metals by EPA 200.8 ICP-MS							Analyst: CBB	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	6.17		0.515	ppb	1	5/5/2023	15.0	A

Lab ID: 23040992-86 **Client Sample ID:** Boys (O/S Gym/Rm 14/15)-S (R)
Collection Date: 4/26/2023 6:56:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Boys (O/S Gym/Rm 14/15)-S (R)

Trace Metals by EPA 200.8 ICP-MS							Analyst: CBB	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	4.05		0.515	ppb	1	5/5/2023	15.0	A

Lab ID: 23040992-87 **Client Sample ID:** Girls (O/S Gym/Rm 14/15)-S (L)
Collection Date: 4/26/2023 6:57:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Girls (O/S Gym/Rm 14/15)-S (L)

Trace Metals by EPA 200.8 ICP-MS							Analyst: CBB	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	2.30		0.515	ppb	1	5/5/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992
Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-88 Client Sample ID: Girls (O/S Gym/Rm 14/15)-S (R)
Collection Date: 4/26/2023 6:57:00 AM Collected By: Mike Medel
Matrix: Drinking Water Sample Location: Girls (O/S Gym/Rm 14/15)-S (R)

Trace Metals by EPA 200.8 ICP-MS							Analyst: CBB	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	5.12		0.515	ppb	1	5/5/2023	15.0	A

Lab ID: 23040992-89 Client Sample ID: Breezeway (O/S Gym/Rm 14/15)-DF
Collection Date: 4/26/2023 6:58:00 AM Collected By: Mike Medel
Matrix: Drinking Water Sample Location: Breezeway (O/S Gym/Rm 14/15)-DF

Trace Metals by EPA 200.8 ICP-MS							Analyst: CBB	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	1.97		0.515	ppb	1	5/5/2023	15.0	A

Lab ID: 23040992-90 Client Sample ID: Gym-DF by Laundry Rm
Collection Date: 4/26/2023 6:59:00 AM Collected By: Mike Medel
Matrix: Drinking Water Sample Location: Gym-DF by Laundry Rm

Trace Metals by EPA 200.8 ICP-MS							Analyst: CBB	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	2.82		0.515	ppb	1	5/6/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992
Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-91 **Client Sample ID:** Gym-DF (Hall/N Side)
Collection Date: 4/26/2023 7:00:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Gym-DF (Hall/N Side)

Trace Metals by EPA 200.8 ICP-MS							Analyst: CBB	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	6.52		0.515	ppb	1	5/6/2023	15.0	A

Lab ID: 23040992-92 **Client Sample ID:** Gym (Stage)-Utility Sink
Collection Date: 4/26/2023 7:06:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Gym (Stage)-Utility Sink

Trace Metals by EPA 200.8 ICP-MS							Analyst: CBB	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	10.9		0.515	ppb	1	5/6/2023	15.0	A

Lab ID: 23040992-93 **Client Sample ID:** Main Playground- Upper DF
Collection Date: 4/26/2023 7:12:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Main Playground- Upper DF

Trace Metals by EPA 200.8 ICP-MS							Analyst: CBB	
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	3.08		0.515	ppb	1	5/6/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 23040992
Date Reported: 5/9/2023

Medford School District
815 S Oakdale Ave
Medford, OR 97501

Lab Order: 23040992
Received Date: 4/26/2023 8:57:00 AM
Reported Date: 5/9/2023 12:36:01 PM

Sample Information:

Lab ID: 23040992-94 **Client Sample ID:** Main Playground- Lower DF
Collection Date: 4/26/2023 7:12:00 AM **Collected By:** Mike Medel
Matrix: Drinking Water **Sample Location:** Main Playground- Lower DF

Trace Metals by EPA 200.8 ICP-MS						Analyst: CBB		
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	1.39		0.515	ppb	1	5/6/2023	15.0	A

QUALIFIERS	*	Value exceeds Maximum Contaminant Level.	C1	Sample container temperature is out of limit as specified at testcode
	E	Value above quantitation range	H	Holding times for preparation or analysis exceeded
	J	Analyte detected below quantitation limits	MI	Recovery outside control limits due to Matrix Interference
	ND	Not Detected at the Reporting Limit	PL	Permit Limit
	PRE	Percent RE exceeds the Limit	R	RPD outside accepted recovery limits

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original

Results are out of the EPA limits



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

QC SUMMARY REPORT

WO#: 23040992

09-May-23

Client: Medford School District
Project: Jacksonville Elem Lead Testing

TestCode: LEAD_DW

Sample ID: LCS-20969	SampType: LCS	TestCode: LEAD_DW	Units: ppb	Prep Date: 5/4/2023	RunNo: 39157
Client ID: LCSW	Batch ID: 20969	TestNo: E200.8	E200.8	Analysis Date: 5/4/2023	SeqNo: 645651
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Lead	102	0.500	100	0	102	85	115
------	-----	-------	-----	---	-----	----	-----

Sample ID: 23040992-41AMS	SampType: MS	TestCode: LEAD_DW	Units: ppb	Prep Date: 5/4/2023	RunNo: 39157
Client ID: Rm 1-RR Sink	Batch ID: 20969	TestNo: E200.8	E200.8	Analysis Date: 5/4/2023	SeqNo: 645653
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Lead	103	0.500	100	4.81	98.2	70	130
------	-----	-------	-----	------	------	----	-----

Sample ID: 23040992-41AMSD	SampType: MSD	TestCode: LEAD_DW	Units: ppb	Prep Date: 5/4/2023	RunNo: 39157
Client ID: Rm 1-RR Sink	Batch ID: 20969	TestNo: E200.8	E200.8	Analysis Date: 5/4/2023	SeqNo: 645654
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Lead	104	0.500	100	4.81	99.3	70	130	103	1.10	20
------	-----	-------	-----	------	------	----	-----	-----	------	----

Sample ID: MB-20967	SampType: MBLK	TestCode: LEAD_DW	Units: ppb	Prep Date: 5/4/2023	RunNo: 39157
Client ID: PBW	Batch ID: 20967	TestNo: E200.8	E200.8	Analysis Date: 5/4/2023	SeqNo: 645670
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Lead	ND	0.500
------	----	-------

Qualifiers:

* Value exceeds Maximum Contaminant Level.	C1 Sample container temperature is out of limit as specified at testcode	H Holding times for preparation or analysis exceede
MI Recovery outside control limits due to Matrix Interference	ND Not Detected at the Reporting Limit	PL Permit Limit
RL Reporting Detection Limit		

Original



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

QC SUMMARY REPORT

WO#: 23040992

09-May-23

Client: Medford School District
Project: Jacksonville Elem Lead Testing

TestCode: LEAD_DW

Sample ID: LCS-20967	SampType: LCS	TestCode: LEAD_DW	Units: ppb	Prep Date: 5/4/2023	RunNo: 39157
Client ID: LCSW	Batch ID: 20967	TestNo: E200.8	E200.8	Analysis Date: 5/4/2023	SeqNo: 645671
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Lead	106	0.500	100	0	106	85	115
------	-----	-------	-----	---	-----	----	-----

Sample ID: 23040992-01AMS	SampType: MS	TestCode: LEAD_DW	Units: ppb	Prep Date: 5/4/2023	RunNo: 39157
Client ID: Rm 7-DF	Batch ID: 20967	TestNo: E200.8	E200.8	Analysis Date: 5/4/2023	SeqNo: 645673
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Lead	105	0.500	100	2.26	103	70	130
------	-----	-------	-----	------	-----	----	-----

Sample ID: 23040992-01AMSD	SampType: MSD	TestCode: LEAD_DW	Units: ppb	Prep Date: 5/4/2023	RunNo: 39157
Client ID: Rm 7-DF	Batch ID: 20967	TestNo: E200.8	E200.8	Analysis Date: 5/4/2023	SeqNo: 645674
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Lead	106	0.500	100	2.26	103	70	130	105	0.651	20
------	-----	-------	-----	------	-----	----	-----	-----	-------	----

Sample ID: MB-20969	SampType: MBLK	TestCode: LEAD_DW	Units: ppb	Prep Date: 5/4/2023	RunNo: 39157
Client ID: PBW	Batch ID: 20969	TestNo: E200.8	E200.8	Analysis Date: 5/4/2023	SeqNo: 645727
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Lead	ND	0.500
------	----	-------

Qualifiers: * Value exceeds Maximum Contaminant Level.
MI Recovery outside control limits due to Matrix Interference
RL Reporting Detection Limit

C1 Sample container temperature is out of limit as specified at testcode
ND Not Detected at the Reporting Limit

H Holding times for preparation or analysis exceeded
PL Permit Limit

Original



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

QC SUMMARY REPORT

WO#: 23040992

09-May-23

Client: Medford School District
Project: Jacksonville Elem Lead Testing

TestCode: LEAD_DW

Sample ID: MB-20970	SampType: MBLK	TestCode: LEAD_DW	Units: ppb	Prep Date: 5/4/2023	RunNo: 39157
Client ID: PBW	Batch ID: 20970	TestNo: E200.8	E200.8	Analysis Date: 5/4/2023	SeqNo: 645740
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Lead	ND	0.500			
------	----	-------	--	--	--

Sample ID: LCS-20970	SampType: LCS	TestCode: LEAD_DW	Units: ppb	Prep Date: 5/4/2023	RunNo: 39157
Client ID: LCSW	Batch ID: 20970	TestNo: E200.8	E200.8	Analysis Date: 5/4/2023	SeqNo: 645741
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Lead	104	0.500	100	0	104 85 115
------	-----	-------	-----	---	------------

Sample ID: 23040992-61AMS	SampType: MS	TestCode: LEAD_DW	Units: ppb	Prep Date: 5/4/2023	RunNo: 39157
Client ID: Rm 4-Sink	Batch ID: 20970	TestNo: E200.8	E200.8	Analysis Date: 5/4/2023	SeqNo: 645743
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Lead	109	0.500	100	10.8	98.3 70 130
------	-----	-------	-----	------	-------------

Sample ID: 23040992-61AMSD	SampType: MSD	TestCode: LEAD_DW	Units: ppb	Prep Date: 5/4/2023	RunNo: 39157
Client ID: Rm 4-Sink	Batch ID: 20970	TestNo: E200.8	E200.8	Analysis Date: 5/4/2023	SeqNo: 645744
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Lead	110	0.500	100	10.8	99.0 70 130 109 0.655 20
------	-----	-------	-----	------	--------------------------

Qualifiers: * Value exceeds Maximum Contaminant Level. C1 Sample container temperature is out of limit as specified at testcode H Holding times for preparation or analysis exceede
MI Recovery outside control limits due to Matrix Interference ND Not Detected at the Reporting Limit PL Permit Limit
RL Reporting Detection Limit

Original



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

QC SUMMARY REPORT

WO#: 23040992

09-May-23

Client: Medford School District
Project: Jacksonville Elem Lead Testing

TestCode: LEAD_DW

Sample ID: MB-20993	SampType: MBLK	TestCode: LEAD_DW	Units: ppb	Prep Date: 5/5/2023	RunNo: 39208
Client ID: PBW	Batch ID: 20993	TestNo: E200.8	E200.8	Analysis Date: 5/5/2023	SeqNo: 646607
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Lead	ND	0.515			
------	----	-------	--	--	--

Sample ID: LCS-20993	SampType: LCS	TestCode: LEAD_DW	Units: ppb	Prep Date: 5/5/2023	RunNo: 39208
Client ID: LCSW	Batch ID: 20993	TestNo: E200.8	E200.8	Analysis Date: 5/5/2023	SeqNo: 646608
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Lead	102	0.520	100	0	102 85 115
------	-----	-------	-----	---	------------

Sample ID: 23040992-22AMS	SampType: MS	TestCode: LEAD_DW	Units: ppb	Prep Date: 5/5/2023	RunNo: 39208
Client ID: Rm 16-DF	Batch ID: 20993	TestNo: E200.8	E200.8	Analysis Date: 5/5/2023	SeqNo: 646611
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Lead	103	0.520	100	2.14	101 70 130
------	-----	-------	-----	------	------------

Sample ID: 23040992-22AMSD	SampType: MSD	TestCode: LEAD_DW	Units: ppb	Prep Date: 5/5/2023	RunNo: 39208
Client ID: Rm 16-DF	Batch ID: 20993	TestNo: E200.8	E200.8	Analysis Date: 5/5/2023	SeqNo: 646612
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Lead	104	0.520	100	2.14	101 70 130 103 0.767 20
------	-----	-------	-----	------	-------------------------

Qualifiers:

* Value exceeds Maximum Contaminant Level.	C1 Sample container temperature is out of limit as specified at testcode	H Holding times for preparation or analysis exceede
MI Recovery outside control limits due to Matrix Interference	ND Not Detected at the Reporting Limit	PL Permit Limit
RL Reporting Detection Limit		

Original



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

QC SUMMARY REPORT

WO#: 23040992

09-May-23

Client: Medford School District
Project: Jacksonville Elem Lead Testing

TestCode: LEAD_DW

Sample ID: MB-20994	SampType: MBLK	TestCode: LEAD_DW	Units: ppb	Prep Date: 5/5/2023	RunNo: 39208
Client ID: PBW	Batch ID: 20994	TestNo: E200.8	E200.8	Analysis Date: 5/5/2023	SeqNo: 646639
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Lead	ND	0.515			
------	----	-------	--	--	--

Sample ID: LCS-20994	SampType: LCS	TestCode: LEAD_DW	Units: ppb	Prep Date: 5/5/2023	RunNo: 39208
Client ID: LCSW	Batch ID: 20994	TestNo: E200.8	E200.8	Analysis Date: 5/5/2023	SeqNo: 646640
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Lead	103	0.520	100	0	103 85 115
------	-----	-------	-----	---	------------

Sample ID: 23040992-82AMS	SampType: MS	TestCode: LEAD_DW	Units: ppb	Prep Date: 5/5/2023	RunNo: 39208
Client ID: Rm 13-Sink	Batch ID: 20994	TestNo: E200.8	E200.8	Analysis Date: 5/5/2023	SeqNo: 646643
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Lead	102	0.520	100	1.62	101 70 130
------	-----	-------	-----	------	------------

Sample ID: 23040992-82AMSD	SampType: MSD	TestCode: LEAD_DW	Units: ppb	Prep Date: 5/5/2023	RunNo: 39208
Client ID: Rm 13-Sink	Batch ID: 20994	TestNo: E200.8	E200.8	Analysis Date: 5/5/2023	SeqNo: 646644
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC LowLimit HighLimit RPD Ref Val %RPD RPDLimit Qual

Lead	102	0.520	100	1.62	100 70 130 102 0.299 20
------	-----	-------	-----	------	-------------------------

Qualifiers:

* Value exceeds Maximum Contaminant Level.	C1 Sample container temperature is out of limit as specified at testcode	H Holding times for preparation or analysis exceede
MI Recovery outside control limits due to Matrix Interference	ND Not Detected at the Reporting Limit	PL Permit Limit
RL Reporting Detection Limit		

Original

Sample Log-In Check List

Client Name: **MedfordSchoolDist**

Work Order Number: **23040992**

RcptNo: **1**

Logged by: **Sarah Cohen** **4/26/2023 8:57:00 AM**

Sarah Cohen

Completed By: **Jordan Diemer** **4/27/2023 10:50:29 AM**

Jordan Diemer

Reviewed By: **Tamra Schmedemann** **5/9/2023 10:25:20 AM**

Tamra Schmedemann

Chain of Custody

1. Is Chain of Custody complete? Yes ☒ No ☐ Not Present ☐
2. How was the sample delivered? Client

Log In

3. Coolers are present? Yes ☐ No ☐ NA ☒
4. Shipping container/cooler in good condition? Yes ☒ No ☐
Custody seals intact on shipping container/cooler? Yes ☐ No ☐ Not Present ☒
No. Seal Date: Signed By:
5. Was an attempt made to cool the samples? Yes ☐ No ☐ NA ☒
6. Were all samples received at a temperature of >0° C to 6.0°C Yes ☐ No ☐ NA ☒
7. Sample(s) in proper container(s)? Yes ☒ No ☐
8. Sufficient sample volume for indicated test(s)? Yes ☒ No ☐
9. Are samples (except VOA and ONG) properly preserved? Yes ☒ No ☐
10. Was preservative added to bottles? Yes ☒ No ☐ NA ☐
HNO3 pH<2
11. Is the headspace in the VOA vials less than 1/4 inch or 6 mm? Yes ☐ No ☐ No VOA Vials ☒
12. Were any sample containers received broken? Yes ☐ No ☒
13. Does paperwork match bottle labels? Yes ☒ No ☐
(Note discrepancies on chain of custody)
14. Are matrices correctly identified on Chain of Custody? Yes ☒ No ☐
15. Is it clear what analyses were requested? Yes ☒ No ☐
16. Were all holding times able to be met? Yes ☒ No ☐
(If no, notify customer for authorization.)

Special Handling (if applicable)

17. Was client notified of all discrepancies with this order? Yes ☐ No ☐ NA ☒

Person Notified: Date
By Whom: Via: ☐ eMail ☐ Phone ☐ Fax ☐ In Person
Regarding:
Client Instructions:

18. Additional remarks:

Cooler Information

Cooler No	Temp °C	Condition	Seal Intact	Seal No	Seal Date	Signed By
-----------	---------	-----------	-------------	---------	-----------	-----------



Environmental Testing Laboratory
245 South Grape Street * Medford, OR 97501
(541) 770-5678 fax (541) 770-2901

Chain of Custody Record

This Chain of Custody is a LEGAL DOCUMENT and must be filled out accurately.

Page 1 of 3

Erin
B 5/16/22

Section A Required Client Information	Section B Required Project Information	Section C Invoice Information	Section D Rush Status (Subject to Scheduling)
Company: MSD 549C	Project Name: Jacksonville	Attention: Accounting	☐ Standard 10-14 Days
Address: 900 Kenyon St.	Project Number:	Company Name: Medford School District	☐ 5 Business Days (50% surcharge)
Medford, OR 97501	Report andy.chasteen@medford.k12.or.us	Address: 900 Kenyon St.	☐ 3 Business Days (75% surcharge)
Email: seth.barry@medford.k12.or.us	Copy T larry.anderson@medford.k12.or.us	Medford, OR 97501	☐ 24 - 48 hours (100% surcharge)
Phone: 541-601-5183 Fax:		P.O. # P123000	☐ Other _____
Collected By (Print): SETH BARRY			Authorized ☐ Yes ☐ No
Collected By (Sign):			
Email Report ☐ Yes ☐ No Mail Report ☐ Yes ☐ No			
Fax Report ☐ Yes ☐ No			

Section E Sample Information

Sample ID		Bottle	Matrix*	Date Collected	Time	No. of Containers	Analysis Requested										NRC Workorder # (Lab Use Only) 23040992	
																	Remarks/Field Data	NRC Sample # (Lab Use Only)
Room 7 - Drinking Fountain			DW	4/26	624													01
Room 7 - Sink			DW		624													020 ~~020~~
Room 7 Rest Room - Sink (Left)			DW		625													03
Room 7 Rest Room - Sink (Right)			DW		625													04
Room 6 - Drinking Fountain			DW		626													05
Room 6 - Sink			DW		626													06
Staff Work Room (By Office) - Sink			DW		627													07
Health Room - Sink			DW		627													08
Health Room Rest Room - Sink			DW		628													09
Girls Restroom By Media Center - Sink (1 - Left To Right)			DW		628													10
Girls Restroom By Media Center - Sink (2 - Left To Right)			DW		629													11
Girls Restroom By Media Center - Sink (3 - Left To Right)			DW		629													12
Girls Restroom By Media Center - Sink (4 - Left To Right)			DW		630													13
Staff Restroom By Media Center - Sink			DW		630													14
Boys Restroom By Media Center - Sink (Left)			DW		631													15
Boys Restroom By Media Center - Sink (Right)			DW		631													16
Outside Room 7 - Drinking Fountain (Outdoor)			DW		632													17
Media Center Work Room - Sink			DW		633													18
Computer Lab - Sink			DW		634													19
Early Reading Intervention Room (ERI) - Sink			DW		635													20
Between Media Center and Room 16 - Drinking Fountain			DW		636													21
Room 16 - Drinking Fountain			DW		637													22
Room 16 - Sink			DW		638													23
Art Supply Room - Sink			DW		638													24



**NEILSON
RESEARCH
CORPORATION**
Environmental Testing Laboratory
245 South Grape Street * Medford, OR 97501
(541) 770-5678 fax (541) 770-2901

Chain of Custody Record

This Chain of Custody is a LEGAL DOCUMENT and must be filled out accurately.

Page 1 of 3
3 of 6
5/6/23

Section A Required Client Information				Section B Required Project Information				Section C Invoice Information				Section D Rush Status (Subject to Scheduling)			
Company: MSD 549C				Project Name: Jacksonville				Attention: Accounting				☐ Standard 10-14 Days			
Address: 900 Kenyon St.				Project Number:				Company Name: Medford School District				☐ 5 Business Days (50% surcharge)			
Medford, OR 97501				Report andy.chasteen@medford.k12.or.us				Address: 900 Kenyon St.				☐ 3 Business Days (75% surcharge)			
Email: seth.barry@medford.k12.or.us				Copy T larry.anderson@medford.k12.or.us				Medford, OR 97501				☐ 24 - 48 hours (100% surcharge)			
Phone: 541-601-5183 Fax:								P.O. # P123000				☐ Other _____			
Collected By (Print): Gary Severe												Authorized ☐ Yes ☐ No			
Collected By (Sign): <i>[Signature]</i>															
Email Report ☒ Yes ☐ No Mail Report ☐ Yes ☐ No															
Fax Report ☐ Yes ☐ No															

Section E Sample Information					No. of Containers	Analysis Requested										NRC Workorder # (Lab Use Only) 23040992	
Sample ID	Bottle	Matrix*	Date Collected	Time		Remarks/Field Data	NRC Sample # (Lab Use Only)										
Kitchen - Hand Wash Sink		DW	4/28	6:26 AM		33											
Kitchen - Vegetable Prep Sink		DW		6:26		34											
Kitchen - Pre-rinse Sink (Dishwasher Area)		DW		6:27		35											
By Cafeteria In The Hallway - Elkay Bottle Filler (Upper)		DW		6:28		36											
By Cafeteria In The Hallway - Elkay DW (Lower)		DW		6:28		37											
Staff Lounge Room - Sink		DW		6:30		38											
Staff Lounge Room - Refrigerator Ice Maker (Left)	N/A	DW				39											
Staff Lounge Room - Refrigerator Ice Maker (Right)	N/A	DW				40											
Room 1 - Drinking Fountain		DW		6:31		41											
Room 1 - Sink		DW		6:32		42											
Room 1 - Restroom Sink		DW		6:33		43											
Boys Restroom Outside Rooms 1/3 - Sink (Left)		DW		6:35		44											
Boys Restroom Outside Rooms 1/3 - Sink (Center)		DW		6:35		45											
Boys Restroom Outside Rooms 1/3 - Sink (Right)		DW		6:36		46											
Girls Restroom Outside Rooms 1/3 - Sink (Left)		DW		6:37		47											
Girls Restroom Outside Rooms 1/3 - Sink (Center)		DW		6:37		48											
Girls Restroom Outside Rooms 1/3 - Sink (Right)		DW		6:37		49											
Staff Restroom Outside Rooms 1/3 - Sink		DW		6:38		50											
Storage Room Adjacent To Room 3 - Sink		DW		6:39		51											
Storage Room Adjacent To Room 3 - Drinking Fountain		DW		6:39		52											
Room 3 - Drinking Fountain		DW		6:40		53											
Room 3 - Sink		DW		6:40		54											
Room 3 - Restroom Sink (Left)		DW		6:41		55											
Room 3 - Restroom Sink (Right)		DW		6:42		56											



Page 1 of 3
 error → 5/9/2

This Chain of Custody is a LEGAL DOCUMENT and must be filled out accurately.

Section D
Rush Status (Subject to Scheduling)

Company: MSD 549C	Project Name: Jacksonville	Attention: Accounting	Standard 10-14 Days
Address: 900 Kenyon St.	Project Number:	Company Name: Medford School District	5 Business Days (50% surcharge)
Medford, OR 97501	Report andy.chasteen@medford.k12.or.us	Address: 900 Kenyon St.	3 Business Days (75% surcharge)
Email: seth.barry@medford.k12.or.us	Copy T larry.anderson@medford.k12.or.us	Medford, OR 97501	24 - 48 hours (100% surcharge)
Phone: 541-601-5183 Fax:		P.O. # P123000	Other _____
Collected By (Print): MIKE MEDEL			Authorized _____ Yes _____ No

Analysis Requested

[illegible]

-
- A Total Alkalinity and Bicarbonate Alkalinity results are to a pH endpoint of 4.5. Carbonate Alkalinity result is to a pH endpoint of 8.3.
- B Analyte detected in the associated method blank.
- C Sample(s) does not meet NELAP/ORELAP sample acceptance criteria. See Case Narrative.
- C1 Sample(s) does not meet NELAP/ORELAP sample acceptance criteria for temperature.
- CF Results confirmed by re-analysis.
- CU Cleanup performed as specified by method.
- E Estimated value.
- ER Elevated reporting limit due to matrix. Report limits (MDLs, MRLs & PQLs) are adjusted based on variations in sample preparation amounts, analytical dilutions, and percent solids, where applicable.
- FC Fecal Coliforms: Sample(s) received past 40 CFR Part 136 specified holding time. Results reported as estimated values.
- HP Sample re-analysis performed outside of method specified holding time.
- HR Sample received outside of method specified holding time.
- HS Sample analyzed for volatile organics contained headspace.
- HT At the client's request, the sample was analyzed outside of method specified holding time.
- H Analysis performed outside of method specified holding time.
- J Analyte detected below the Minimum Reporting Limit (MRL) and above the Method Detection Limit (MDL). The J flag result is an estimated value and the user should be aware that this data is of limited reliability.
- L Dissolved metals were not filtered within 15 minutes of collection per 40 CFR Part 136.
- MI Surrogate, Duplicate Sample (DUP) or Matrix Spikes recoveries are out of control limits due to matrix interference. Sample results may be biased.
- N See Case Narrative on page 2 of report.
- Q Initial calibration verification (ICV), continuing calibration verification (CCV) or laboratory control sample (LCS) exceeded high recovery limits, but associated samples are non-detect and the sample results are not affected. Data meets EPA/NELAP requirements.
- R Relative percent difference (RPD) is outside of the accepted recovery limits.
- R3 The relative percent difference (RPD) and/or percent recovery for the duplicate (DUP) or matrix spike (MS)/matrix spike duplicate (MSD) cannot be accurately calculated due to the concentration of analyte already present in the sample.
- R4 The Relative percent difference (RPD) is not within control limits because the concentration of the sample result is too low to represent proper statistical error.
- S Surrogate and/or matrix spike recovery is outside of the accepted recovery limits. Sample results may be biased.
- S1 Surrogate or matrix spike recovery is outside of control limits due to dilution necessary for analysis.
- SC Sub-contracted to another laboratory for analysis.
- SP Sample(s) were not collected per EPA Method 5035A protocols. The results are considered minimum values.
- * Value exceeds Maximum Contaminant Level or is outside the acceptable range.