



May 01, 2024

Michael Colangelo
OHM BOCES Oneida-Herkimer-Madison
BOCES
4747 Middle Settlement Rd.
New Hartford, NY 13413

RE: Project: OHM BOCES/LIN/BRO/MVCC 4/17
Pace Project No.: 70295203

Dear Michael Colangelo:

Enclosed are the analytical results for sample(s) received by the laboratory on April 22, 2024. The results relate only to the samples included in this report. Results reported herein conform to the applicable TNI/NELAC Standards and the laboratory's Quality Manual, where applicable, unless otherwise noted in the body of the report.

The test results provided in this final report were generated by each of the following laboratories within the Pace Network:

- Pace Analytical Services - Melville

If you have any questions concerning this report, please feel free to contact me.

Sincerely,

Jack M. Germano
jack.germano@pacelabs.com
516-370-6012
Project Manager

Enclosures

cc: OHM BOCES Safety Services, OHM BOCES Oneida-Herkimer-Madison BOCES



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



CERTIFICATIONS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Pace Analytical Services Long Island

575 Broad Hollow Rd, Melville, NY 11747

Connecticut Certification #: PH-0435

Delaware Certification # NY 10478

Maryland Certification #: 208

Massachusetts Certification #: M-NY026

New Hampshire Certification #: 2987

New Jersey Certification #: NY158

New York Certification #: 10478 Primary Accrediting Body

Pennsylvania Certification #: 68-00350

Rhode Island Certification #: LAO00340

Virginia Certification # 460302

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: OHM 1								
Lab ID: 70295203001								
Collected: 04/17/24 06:18								
Received: 04/22/24 17:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		04/29/24 12:20	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 2		Lab ID: 70295203002	Collected: 04/17/24 06:19	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 12:55	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 3		Lab ID: 70295203003	Collected: 04/17/24 06:20	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 13:04	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 4		Lab ID: 70295203004	Collected: 04/17/24 06:21	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 13:00	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 5		Lab ID: 70295203005	Collected: 04/17/24 06:25	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.8	ug/L	1.0	1		04/29/24 14:41	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 6		Lab ID: 70295203006	Collected: 04/17/24 06:26	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 14:45	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 7		Lab ID: 70295203007	Collected: 04/17/24 06:27	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 14:50	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 8		Lab ID: 70295203008	Collected: 04/17/24 06:28	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 14:52	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 9		Lab ID: 70295203009	Collected: 04/17/24 06:29	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.5	ug/L	1.0	1		04/29/24 14:56	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 10		Lab ID: 70295203010	Collected: 04/17/24 06:30	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 14:58	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 11		Lab ID: 70295203011	Collected: 04/17/24 06:31	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 14:59	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 12		Lab ID: 70295203012	Collected: 04/17/24 06:32	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	4.3	ug/L	1.0	1		04/29/24 15:01	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: OHM 13								
Lab ID: 70295203013								
Collected: 04/17/24 06:34								
Received: 04/22/24 17:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	1.0	ug/L	1.0	1		04/29/24 15:02	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: OHM 14								
Lab ID: 70295203014								
Collected: 04/17/24 06:35								
Received: 04/22/24 17:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	7.0	ug/L	1.0	1		04/29/24 15:04	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 15		Lab ID: 70295203015	Collected: 04/17/24 06:37	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 15:05	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 16		Lab ID: 70295203016	Collected: 04/17/24 06:38	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 15:07	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 17		Lab ID: 70295203017	Collected: 04/17/24 06:40	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 15:09	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 18		Lab ID: 70295203018	Collected: 04/17/24 06:41	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.5	ug/L	1.0	1		04/29/24 15:10	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 19		Lab ID: 70295203019		Collected: 04/17/24 06:42	Received: 04/22/24 17:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.2	ug/L	1.0	1		04/29/24 15:15	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: OHM 20								
Lab ID: 70295203020								
Collected: 04/17/24 06:43								
Received: 04/22/24 17:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		04/29/24 15:16	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: OHM 21								
Lab ID: 70295203021								
Collected: 04/17/24 06:44								
Received: 04/22/24 17:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		04/29/24 15:18	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 22		Lab ID: 70295203022	Collected: 04/17/24 06:45	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 15:19	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 23		Lab ID: 70295203023	Collected: 04/17/24 06:47	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.1	ug/L	1.0	1		04/29/24 15:21	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 24		Lab ID: 70295203024	Collected: 04/17/24 06:49	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 15:22	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 25		Lab ID: 70295203025	Collected: 04/17/24 06:51	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 15:27	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: OHM 26								
Lab ID: 70295203026								
Collected: 04/17/24 06:52								
Received: 04/22/24 17:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		04/29/24 15:35	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 27		Lab ID: 70295203027	Collected: 04/17/24 06:57	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 15:39	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: OHM 28								
Lab ID: 70295203028								
Collected: 04/17/24 06:58								
Received: 04/22/24 17:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		04/29/24 15:41	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 29		Lab ID: 70295203029	Collected: 04/17/24 06:59	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 15:42	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 30		Lab ID: 70295203030	Collected: 04/17/24 07:00	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 15:44	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 31		Lab ID: 70295203031	Collected: 04/17/24 07:01	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 15:45	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: OHM 32								
Lab ID: 70295203032								
Collected: 04/17/24 07:02								
Received: 04/22/24 17:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		04/29/24 15:47	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 33		Lab ID: 70295203033	Collected: 04/17/24 07:03	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 15:51	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 34		Lab ID: 70295203034	Collected: 04/17/24 07:04	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 15:53	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 35		Lab ID: 70295203035	Collected: 04/17/24 07:05	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.9	ug/L	1.0	1		04/29/24 15:54	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: OHM 36								
Lab ID: 70295203036								
Collected: 04/17/24 07:06								
Received: 04/22/24 17:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		04/29/24 15:56	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 37		Lab ID: 70295203037	Collected: 04/17/24 07:07	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 15:58	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 38		Lab ID: 70295203038	Collected: 04/17/24 07:08	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 15:59	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 39		Lab ID: 70295203039	Collected: 04/17/24 07:09	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 16:01	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 40		Lab ID: 70295203040	Collected: 04/17/24 07:10	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 16:02	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 41		Lab ID: 70295203041	Collected: 04/17/24 07:11	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 16:04	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 42		Lab ID: 70295203042	Collected: 04/17/24 07:12	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 16:05	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 43		Lab ID: 70295203043	Collected: 04/17/24 07:13	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.4	ug/L	1.0	1		04/29/24 16:10	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 44		Lab ID: 70295203044	Collected: 04/17/24 07:14	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.8	ug/L	1.0	1		04/29/24 16:12	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: OHM 45								
Lab ID: 70295203045								
Collected: 04/17/24 07:15								
Received: 04/22/24 17:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		04/29/24 16:16	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 46		Lab ID: 70295203046	Collected: 04/17/24 07:16	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 16:21	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 47		Lab ID: 70295203047	Collected: 04/17/24 07:17	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 16:28	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 48		Lab ID: 70295203048	Collected: 04/17/24 07:18	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 16:30	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 49		Lab ID: 70295203049	Collected: 04/17/24 07:19	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 16:31	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 50		Lab ID: 70295203050	Collected: 04/17/24 07:20	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 16:33	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: OHM 51								
Lab ID: 70295203051								
Collected: 04/17/24 07:21								
Received: 04/22/24 17:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		04/29/24 16:34	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 52		Lab ID: 70295203052	Collected: 04/17/24 07:22	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 16:36	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 53		Lab ID: 70295203053		Collected: 04/17/24 07:23	Received: 04/22/24 17:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 16:38	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 54		Lab ID: 70295203054	Collected: 04/17/24 07:24	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 16:39	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: OHM 55								
Lab ID: 70295203055								
Collected: 04/17/24 07:25								
Received: 04/22/24 17:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		04/29/24 16:41	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 56		Lab ID: 70295203056	Collected: 04/17/24 07:26	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 16:42	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: OHM 57								
Lab ID: 70295203057								
Collected: 04/17/24 07:27								
Received: 04/22/24 17:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		04/29/24 16:47	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: OHM 58								
Lab ID: 70295203058								
Collected: 04/17/24 07:28								
Received: 04/22/24 17:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		04/29/24 16:48	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 59		Lab ID: 70295203059	Collected: 04/17/24 07:29	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 16:50	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 60		Lab ID: 70295203060	Collected: 04/17/24 07:30	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 16:51	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 61		Lab ID: 70295203061	Collected: 04/17/24 07:31	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 16:53	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 62		Lab ID: 70295203062	Collected: 04/17/24 07:32	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 16:54	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 63		Lab ID: 70295203063	Collected: 04/17/24 07:33	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 16:56	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 64		Lab ID: 70295203064	Collected: 04/17/24 07:34	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 16:58	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 65		Lab ID: 70295203065	Collected: 04/17/24 07:35	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 17:05	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 66		Lab ID: 70295203066	Collected: 04/17/24 07:36	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 17:10	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 67		Lab ID: 70295203067	Collected: 04/17/24 07:37	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 17:14	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 68		Lab ID: 70295203068	Collected: 04/17/24 07:38	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 17:16	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: OHM 69								
Lab ID: 70295203069								
Collected: 04/17/24 07:39								
Received: 04/22/24 17:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		04/29/24 17:17	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 70		Lab ID: 70295203070	Collected: 04/17/24 07:40	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 17:19	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 71	Lab ID: 70295203071	Collected: 04/17/24 07:41	Received: 04/22/24 17:00	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		04/29/24 17:23	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 72		Lab ID: 70295203072	Collected: 04/17/24 07:42	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 17:25	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 73		Lab ID: 70295203073	Collected: 04/17/24 07:43	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 17:27	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 74		Lab ID: 70295203074	Collected: 04/17/24 07:44	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 17:28	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 75		Lab ID: 70295203075	Collected: 04/17/24 07:45	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 17:30	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: OHM 76								
Lab ID: 70295203076								
Collected: 04/17/24 07:46								
Received: 04/22/24 17:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		04/29/24 17:31	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 77		Lab ID: 70295203077	Collected: 04/17/24 07:47	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 17:33	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 78		Lab ID: 70295203078	Collected: 04/17/24 07:48	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 17:34	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 79		Lab ID: 70295203079	Collected: 04/17/24 07:49	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 17:36	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: OHM 80								
Lab ID: 70295203080								
Collected: 04/17/24 07:50 Received: 04/22/24 17:00 Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		04/29/24 17:37	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 81		Lab ID: 70295203081	Collected: 04/17/24 07:51	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 17:42	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 82		Lab ID: 70295203082	Collected: 04/17/24 07:52	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 17:43	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: OHM 83								
Lab ID: 70295203083								
Collected: 04/17/24 07:53								
Received: 04/22/24 17:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		04/29/24 17:45	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 84		Lab ID: 70295203084	Collected: 04/17/24 07:54	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.0	ug/L	1.0	1		04/29/24 17:47	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 85		Lab ID: 70295203085	Collected: 04/17/24 07:55	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.6	ug/L	1.0	1		04/29/24 17:51	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 86		Lab ID: 70295203086	Collected: 04/17/24 07:56	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.9	ug/L	1.0	1		04/29/24 17:56	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 87		Lab ID: 70295203087	Collected: 04/17/24 07:57	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 18:03	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: OHM 88								
Lab ID: 70295203088								
Collected: 04/17/24 07:58								
Received: 04/22/24 17:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		04/29/24 18:05	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 89		Lab ID: 70295203089	Collected: 04/17/24 07:59	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.4	ug/L	1.0	1		04/29/24 18:06	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 90		Lab ID: 70295203090	Collected: 04/17/24 08:00	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 18:08	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 91		Lab ID: 70295203091	Collected: 04/17/24 08:01	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 18:09	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 92		Lab ID: 70295203092	Collected: 04/17/24 08:02	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 18:11	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 93		Lab ID: 70295203093	Collected: 04/17/24 08:03	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 18:12	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 94		Lab ID: 70295203094	Collected: 04/17/24 08:04	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 18:14	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 95		Lab ID: 70295203095	Collected: 04/17/24 08:05	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 18:19	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 96		Lab ID: 70295203096	Collected: 04/17/24 08:06	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 18:20	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 97		Lab ID: 70295203097	Collected: 04/17/24 08:07	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 18:22	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 98		Lab ID: 70295203098	Collected: 04/17/24 08:08	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 18:23	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 99		Lab ID: 70295203099	Collected: 04/17/24 08:09	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 18:25	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: OHM 100								
Lab ID: 70295203100								
Collected: 04/17/24 08:10								
Received: 04/22/24 17:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		04/29/24 18:26	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 101		Lab ID: 70295203101	Collected: 04/17/24 08:11	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 18:28	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 102		Lab ID: 70295203102	Collected: 04/17/24 08:12	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 18:32	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: OHM 103		Lab ID: 70295203103	Collected: 04/17/24 08:13	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/29/24 18:40	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: LIN 1								
Lab ID: 70295203104								
Collected: 04/17/24 08:29 Received: 04/22/24 17:00 Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		04/29/24 18:45	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: LIN 2		Lab ID: 70295203105	Collected: 04/17/24 08:30	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	33.3	ug/L	1.0	1		04/29/24 18:46	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: LIN 3								
Lab ID: 70295203106								
Collected: 04/17/24 08:31								
Received: 04/22/24 17:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	2.7	ug/L	1.0	1		04/29/24 18:48	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: LIN 4		Lab ID: 70295203107	Collected: 04/17/24 08:32	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	5.1	ug/L	1.0	1		04/29/24 18:49	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: LIN 5		Lab ID: 70295203108	Collected: 04/17/24 08:33	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.3	ug/L	1.0	1		04/29/24 18:51	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: LIN 6								
Lab ID: 70295203109								
Collected: 04/17/24 08:35								
Received: 04/22/24 17:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	10	ug/L	1.0	1		04/29/24 18:55	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: LIN 7		Lab ID: 70295203110	Collected: 04/17/24 08:37	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.4	ug/L	1.0	1		04/29/24 18:57	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: LIN 8								
Lab ID: 70295203111								
Collected: 04/17/24 08:39								
Received: 04/22/24 17:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	18.1	ug/L	1.0	1		04/29/24 18:58	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: LIN 9		Lab ID: 70295203112	Collected: 04/17/24 08:40	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	21.6	ug/L	1.0	1		04/29/24 19:00	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: BRO 1		Lab ID: 70295203113	Collected: 04/17/24 09:32	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/30/24 14:36	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: BRO 2		Lab ID: 70295203114	Collected: 04/17/24 09:33	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/30/24 15:25	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: BRO 3		Lab ID: 70295203115	Collected: 04/17/24 09:34	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/30/24 15:30	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: BRO 4		Lab ID: 70295203116	Collected: 04/17/24 09:35	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/30/24 15:38	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: BRO 5		Lab ID: 70295203117	Collected: 04/17/24 09:39	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.6	ug/L	1.0	1		04/30/24 15:42	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: BRO 6		Lab ID: 70295203118	Collected: 04/17/24 09:40	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/30/24 15:44	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: BRO 7		Lab ID: 70295203119	Collected: 04/17/24 09:41	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.3	ug/L	1.0	1		04/30/24 15:45	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: BRO 8		Lab ID: 70295203120	Collected: 04/17/24 09:46	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/30/24 15:47	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: SUPPORT SERVICES 1 Lab ID: 70295203121 Collected: 04/17/24 10:07 Received: 04/22/24 17:00 Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		04/30/24 15:49	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: CONSERVATION BUILDING 1 **Lab ID: 70295203122** Collected: 04/17/24 00:00 Received: 04/22/24 17:00 Matrix: Drinking Water

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		04/30/24 15:50	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: MVCC 1		Lab ID: 70295203123	Collected: 04/17/24 09:04	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.1	ug/L	1.0	1		04/30/24 15:52	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: MVCC 2		Lab ID: 70295203124	Collected: 04/17/24 09:05	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		04/30/24 15:56	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Sample: MVCC 3		Lab ID: 70295203125	Collected: 04/17/24 09:06	Received: 04/22/24 17:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.4	ug/L	1.0	1		04/30/24 15:58	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

QC Batch: 346054

Analysis Method: EPA 200.8

QC Batch Method: EPA 200.8

Analysis Description: 200.8 MET No Prep Drinking Water

Laboratory: Pace Analytical Services - Melville

Associated Lab Samples: 70295203001

METHOD BLANK: 1787559

Matrix: Water

Associated Lab Samples: 70295203001

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	04/29/24 11:33	

LABORATORY CONTROL SAMPLE: 1787560

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	50.3	101	85-115	

MATRIX SPIKE SAMPLE: 1787562

Parameter	Units	70295341021 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<2.0	100	99.5	99	70-130	

MATRIX SPIKE SAMPLE: 1787564

Parameter	Units	70295341023 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	4.6	100	98.9	94	70-130	

SAMPLE DUPLICATE: 1787561

Parameter	Units	70295341021 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<2.0	<2.0		

SAMPLE DUPLICATE: 1787563

Parameter	Units	70295341023 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	4.6	4.4	4	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

QC Batch: 346055

Analysis Method: EPA 200.8

QC Batch Method: EPA 200.8

Analysis Description: 200.8 MET No Prep Drinking Water

Laboratory: Pace Analytical Services - Melville

Associated Lab Samples: 70295203002, 70295203003, 70295203004

METHOD BLANK: 1787565

Matrix: Water

Associated Lab Samples: 70295203002, 70295203003, 70295203004

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	04/29/24 12:21	

LABORATORY CONTROL SAMPLE: 1787566

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	50.5	101	85-115	

MATRIX SPIKE SAMPLE: 1787572

Parameter	Units	70295203002 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	50.8	101	70-130	

MATRIX SPIKE SAMPLE: 1787574

Parameter	Units	70295203004 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	50.2	100	70-130	

SAMPLE DUPLICATE: 1787571

Parameter	Units	70295203002 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

SAMPLE DUPLICATE: 1787573

Parameter	Units	70295203004 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

QC Batch:	346088	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70295203005, 70295203006, 70295203007, 70295203008, 70295203009, 70295203010, 70295203011, 70295203012, 70295203013, 70295203014, 70295203015, 70295203016, 70295203017, 70295203018, 70295203019, 70295203020, 70295203021, 70295203022, 70295203023, 70295203024		

METHOD BLANK:	1787662	Matrix:	Water
Associated Lab Samples:	70295203005, 70295203006, 70295203007, 70295203008, 70295203009, 70295203010, 70295203011, 70295203012, 70295203013, 70295203014, 70295203015, 70295203016, 70295203017, 70295203018, 70295203019, 70295203020, 70295203021, 70295203022, 70295203023, 70295203024		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	04/29/24 14:38	

LABORATORY CONTROL SAMPLE:	1787663					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	49.6	99	85-115	

MATRIX SPIKE SAMPLE:	1787665						
Parameter	Units	70295203005 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	1.8	50	56.5	109	70-130	

MATRIX SPIKE SAMPLE:	1787667						
Parameter	Units	70295203006 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	54.7	108	70-130	

SAMPLE DUPLICATE:	1787664					
Parameter	Units	70295203005 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	1.8	1.8	0		

SAMPLE DUPLICATE:	1787666					
Parameter	Units	70295203006 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	<1.0	<1.0			

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

QC Batch:	346090	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70295203025, 70295203026, 70295203027, 70295203028, 70295203029, 70295203030, 70295203031, 70295203032, 70295203033, 70295203034, 70295203035, 70295203036, 70295203037, 70295203038, 70295203039, 70295203040, 70295203041, 70295203042, 70295203043, 70295203044		

METHOD BLANK:	1787670	Matrix:	Water
Associated Lab Samples:	70295203025, 70295203026, 70295203027, 70295203028, 70295203029, 70295203030, 70295203031, 70295203032, 70295203033, 70295203034, 70295203035, 70295203036, 70295203037, 70295203038, 70295203039, 70295203040, 70295203041, 70295203042, 70295203043, 70295203044		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	04/29/24 15:24	

LABORATORY CONTROL SAMPLE:	1787671					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	49.3	99	85-115	

MATRIX SPIKE SAMPLE:	1787741						
Parameter	Units	70295203025 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	54.2	107	70-130	

MATRIX SPIKE SAMPLE:	1787743						
Parameter	Units	70295203026 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	58.1	116	70-130	

SAMPLE DUPLICATE:	1787740					
Parameter	Units	70295203025 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	<1.0	<1.0			

SAMPLE DUPLICATE:	1787742					
Parameter	Units	70295203026 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	<1.0	<1.0			

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

QC Batch:	346091	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70295203045, 70295203046, 70295203047, 70295203048, 70295203049, 70295203050, 70295203051, 70295203052, 70295203053, 70295203054, 70295203055, 70295203056, 70295203057, 70295203058, 70295203059, 70295203060, 70295203061, 70295203062, 70295203063, 70295203064		

METHOD BLANK:	1787672	Matrix:	Water
Associated Lab Samples:	70295203045, 70295203046, 70295203047, 70295203048, 70295203049, 70295203050, 70295203051, 70295203052, 70295203053, 70295203054, 70295203055, 70295203056, 70295203057, 70295203058, 70295203059, 70295203060, 70295203061, 70295203062, 70295203063, 70295203064		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	04/29/24 16:13	

LABORATORY CONTROL SAMPLE:	1787673					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	49.0	98	85-115	

MATRIX SPIKE SAMPLE:	1787675						
Parameter	Units	70295203045 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	53.8	107	70-130	

MATRIX SPIKE SAMPLE:	1787677						
Parameter	Units	70295203046 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	56.0	112	70-130	

SAMPLE DUPLICATE:	1787674					
Parameter	Units	70295203045 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	<1.0	<1.0			

SAMPLE DUPLICATE:	1787676					
Parameter	Units	70295203046 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	<1.0	<1.0			

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

QC Batch:	346092	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70295203065, 70295203066, 70295203067, 70295203068, 70295203069, 70295203070, 70295203071, 70295203072, 70295203073, 70295203074, 70295203075, 70295203076, 70295203077, 70295203078, 70295203079, 70295203080, 70295203081, 70295203082, 70295203083, 70295203084		

METHOD BLANK:	1787678	Matrix:	Water
Associated Lab Samples:	70295203065, 70295203066, 70295203067, 70295203068, 70295203069, 70295203070, 70295203071, 70295203072, 70295203073, 70295203074, 70295203075, 70295203076, 70295203077, 70295203078, 70295203079, 70295203080, 70295203081, 70295203082, 70295203083, 70295203084		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	04/29/24 16:59	

LABORATORY CONTROL SAMPLE:	1787679					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	49.8	100	85-115	

MATRIX SPIKE SAMPLE:	1787681						
Parameter	Units	70295203065 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	53.4	107	70-130	

MATRIX SPIKE SAMPLE:	1787683						
Parameter	Units	70295203066 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	54.8	109	70-130	

SAMPLE DUPLICATE:	1787680				
Parameter	Units	70295203065 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

SAMPLE DUPLICATE:	1787682				
Parameter	Units	70295203066 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

QC Batch:	346093	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70295203085, 70295203086, 70295203087, 70295203088, 70295203089, 70295203090, 70295203091, 70295203092, 70295203093, 70295203094, 70295203095, 70295203096, 70295203097, 70295203098, 70295203099, 70295203100, 70295203101		

METHOD BLANK:	1787684	Matrix:	Water
Associated Lab Samples:	70295203085, 70295203086, 70295203087, 70295203088, 70295203089, 70295203090, 70295203091, 70295203092, 70295203093, 70295203094, 70295203095, 70295203096, 70295203097, 70295203098, 70295203099, 70295203100, 70295203101		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	04/29/24 17:48	

LABORATORY CONTROL SAMPLE:	1787685
----------------------------	---------

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	48.6	97	85-115	

MATRIX SPIKE SAMPLE:	1787687
----------------------	---------

Parameter	Units	70295203085 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	1.6	50	53.7	104	70-130	

MATRIX SPIKE SAMPLE:	1787689
----------------------	---------

Parameter	Units	70295203086 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	1.9	50	60.0	116	70-130	

SAMPLE DUPLICATE:	1787686
-------------------	---------

Parameter	Units	70295203085 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	1.6	1.6	1	

SAMPLE DUPLICATE:	1787688
-------------------	---------

Parameter	Units	70295203086 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	1.9	1.9	1	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

QC Batch:	346094	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70295203102, 70295203103, 70295203104, 70295203105, 70295203106, 70295203107, 70295203108, 70295203109, 70295203110, 70295203111, 70295203112		

METHOD BLANK:	1787690	Matrix:	Water
Associated Lab Samples:	70295203102, 70295203103, 70295203104, 70295203105, 70295203106, 70295203107, 70295203108, 70295203109, 70295203110, 70295203111, 70295203112		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	04/29/24 18:29	

LABORATORY CONTROL SAMPLE: 1787691						
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	49.0	98	85-115	

MATRIX SPIKE SAMPLE: 1787693							
Parameter	Units	70295203102 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	58.9	118	70-130	

MATRIX SPIKE SAMPLE: 1787695							
Parameter	Units	70295203103 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	52.4	104	70-130	

SAMPLE DUPLICATE: 1787692					
Parameter	Units	70295203102 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

SAMPLE DUPLICATE: 1787694					
Parameter	Units	70295203103 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

QC Batch: 346186

Analysis Method: EPA 200.8

QC Batch Method: EPA 200.8

Analysis Description: 200.8 MET No Prep Drinking Water

Laboratory: Pace Analytical Services - Melville

Associated Lab Samples: 70295203113

METHOD BLANK: 1788312

Matrix: Water

Associated Lab Samples: 70295203113

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	04/30/24 13:51	

LABORATORY CONTROL SAMPLE: 1788313

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	51.3	103	85-115	

MATRIX SPIKE SAMPLE: 1788315

Parameter	Units	70295439001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	4.5	50	53.5	98	70-130	

MATRIX SPIKE SAMPLE: 1788317

Parameter	Units	70295442001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	1.0	50	49.2	96	70-130	

SAMPLE DUPLICATE: 1788314

Parameter	Units	70295439001 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	4.5	4.5	1	

SAMPLE DUPLICATE: 1788316

Parameter	Units	70295442001 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	1.0	<1.0		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

QC Batch: 346207

Analysis Method: EPA 200.8

QC Batch Method: EPA 200.8

Analysis Description: 200.8 MET No Prep Drinking Water

Laboratory: Pace Analytical Services - Melville

Associated Lab Samples: 70295203114

METHOD BLANK: 1788390

Matrix: Water

Associated Lab Samples: 70295203114

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	04/30/24 14:37	

LABORATORY CONTROL SAMPLE: 1788391

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	49.0	98	85-115	

MATRIX SPIKE SAMPLE: 1788393

Parameter	Units	70295442020 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	51.0	100	70-130	

MATRIX SPIKE SAMPLE: 1788395

Parameter	Units	70295442021 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	46.3	91	70-130	

SAMPLE DUPLICATE: 1788392

Parameter	Units	70295442020 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

SAMPLE DUPLICATE: 1788394

Parameter	Units	70295442021 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

QC Batch:	346208	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70295203115, 70295203116, 70295203117, 70295203118, 70295203119, 70295203120, 70295203121, 70295203122, 70295203123, 70295203124, 70295203125		

METHOD BLANK: 1788396 Matrix: Water
 Associated Lab Samples: 70295203115, 70295203116, 70295203117, 70295203118, 70295203119, 70295203120, 70295203121, 70295203122, 70295203123, 70295203124, 70295203125

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	04/30/24 15:27	

LABORATORY CONTROL SAMPLE: 1788397

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	48.9	98	85-115	

MATRIX SPIKE SAMPLE: 1788399

Parameter	Units	70295203115 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	48.2	96	70-130	

MATRIX SPIKE SAMPLE: 1788401

Parameter	Units	70295203116 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	50.7	101	70-130	

SAMPLE DUPLICATE: 1788398

Parameter	Units	70295203115 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

SAMPLE DUPLICATE: 1788400

Parameter	Units	70295203116 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALIFIERS

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

TNTC - Too Numerous To Count

J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit - The lowest concentration value that meets project requirements for quantitative data with known precision and bias for a specific analyte in a specific matrix.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Reported results are not rounded until the final step prior to reporting. Therefore, calculated parameters that are typically reported as "Total" may vary slightly from the sum of the reported component parameters.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70295203001	OHM 1	EPA 200.8	346054		
70295203002	OHM 2	EPA 200.8	346055		
70295203003	OHM 3	EPA 200.8	346055		
70295203004	OHM 4	EPA 200.8	346055		
70295203005	OHM 5	EPA 200.8	346088		
70295203006	OHM 6	EPA 200.8	346088		
70295203007	OHM 7	EPA 200.8	346088		
70295203008	OHM 8	EPA 200.8	346088		
70295203009	OHM 9	EPA 200.8	346088		
70295203010	OHM 10	EPA 200.8	346088		
70295203011	OHM 11	EPA 200.8	346088		
70295203012	OHM 12	EPA 200.8	346088		
70295203013	OHM 13	EPA 200.8	346088		
70295203014	OHM 14	EPA 200.8	346088		
70295203015	OHM 15	EPA 200.8	346088		
70295203016	OHM 16	EPA 200.8	346088		
70295203017	OHM 17	EPA 200.8	346088		
70295203018	OHM 18	EPA 200.8	346088		
70295203019	OHM 19	EPA 200.8	346088		
70295203020	OHM 20	EPA 200.8	346088		
70295203021	OHM 21	EPA 200.8	346088		
70295203022	OHM 22	EPA 200.8	346088		
70295203023	OHM 23	EPA 200.8	346088		
70295203024	OHM 24	EPA 200.8	346088		
70295203025	OHM 25	EPA 200.8	346090		
70295203026	OHM 26	EPA 200.8	346090		
70295203027	OHM 27	EPA 200.8	346090		
70295203028	OHM 28	EPA 200.8	346090		
70295203029	OHM 29	EPA 200.8	346090		
70295203030	OHM 30	EPA 200.8	346090		
70295203031	OHM 31	EPA 200.8	346090		
70295203032	OHM 32	EPA 200.8	346090		
70295203033	OHM 33	EPA 200.8	346090		
70295203034	OHM 34	EPA 200.8	346090		
70295203035	OHM 35	EPA 200.8	346090		
70295203036	OHM 36	EPA 200.8	346090		
70295203037	OHM 37	EPA 200.8	346090		
70295203038	OHM 38	EPA 200.8	346090		
70295203039	OHM 39	EPA 200.8	346090		
70295203040	OHM 40	EPA 200.8	346090		
70295203041	OHM 41	EPA 200.8	346090		
70295203042	OHM 42	EPA 200.8	346090		
70295203043	OHM 43	EPA 200.8	346090		
70295203044	OHM 44	EPA 200.8	346090		
70295203045	OHM 45	EPA 200.8	346091		
70295203046	OHM 46	EPA 200.8	346091		
70295203047	OHM 47	EPA 200.8	346091		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: OHM BOCES/LIN/BRO/MVCC 4/17

Pace Project No.: 70295203

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70295203048	OHM 48	EPA 200.8	346091		
70295203049	OHM 49	EPA 200.8	346091		
70295203050	OHM 50	EPA 200.8	346091		
70295203051	OHM 51	EPA 200.8	346091		
70295203052	OHM 52	EPA 200.8	346091		
70295203053	OHM 53	EPA 200.8	346091		
70295203054	OHM 54	EPA 200.8	346091		
70295203055	OHM 55	EPA 200.8	346091		
70295203056	OHM 56	EPA 200.8	346091		
70295203057	OHM 57	EPA 200.8	346091		
70295203058	OHM 58	EPA 200.8	346091		
70295203059	OHM 59	EPA 200.8	346091		
70295203060	OHM 60	EPA 200.8	346091		
70295203061	OHM 61	EPA 200.8	346091		
70295203062	OHM 62	EPA 200.8	346091		
70295203063	OHM 63	EPA 200.8	346091		
70295203064	OHM 64	EPA 200.8	346091		
70295203065	OHM 65	EPA 200.8	346092		
70295203066	OHM 66	EPA 200.8	346092		
70295203067	OHM 67	EPA 200.8	346092		
70295203068	OHM 68	EPA 200.8	346092		
70295203069	OHM 69	EPA 200.8	346092		
70295203070	OHM 70	EPA 200.8	346092		
70295203071	OHM 71	EPA 200.8	346092		
70295203072	OHM 72	EPA 200.8	346092		
70295203073	OHM 73	EPA 200.8	346092		
70295203074	OHM 74	EPA 200.8	346092		
70295203075	OHM 75	EPA 200.8	346092		
70295203076	OHM 76	EPA 200.8	346092		
70295203077	OHM 77	EPA 200.8	346092		
70295203078	OHM 78	EPA 200.8	346092		
70295203079	OHM 79	EPA 200.8	346092		
70295203080	OHM 80	EPA 200.8	346092		
70295203081	OHM 81	EPA 200.8	346092		
70295203082	OHM 82	EPA 200.8	346092		
70295203083	OHM 83	EPA 200.8	346092		
70295203084	OHM 84	EPA 200.8	346092		
70295203085	OHM 85	EPA 200.8	346093		
70295203086	OHM 86	EPA 200.8	346093		
70295203087	OHM 87	EPA 200.8	346093		
70295203088	OHM 88	EPA 200.8	346093		
70295203089	OHM 89	EPA 200.8	346093		
70295203090	OHM 90	EPA 200.8	346093		
70295203091	OHM 91	EPA 200.8	346093		
70295203092	OHM 92	EPA 200.8	346093		
70295203093	OHM 93	EPA 200.8	346093		
70295203094	OHM 94	EPA 200.8	346093		
70295203095	OHM 95	EPA 200.8	346093		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: OHM BOCES/LIN/BRO/MVCC 4/17
 Pace Project No.: 70295203

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70295203096	OHM 96	EPA 200.8	346093		
70295203097	OHM 97	EPA 200.8	346093		
70295203098	OHM 98	EPA 200.8	346093		
70295203099	OHM 99	EPA 200.8	346093		
70295203100	OHM 100	EPA 200.8	346093		
70295203101	OHM 101	EPA 200.8	346093		
70295203102	OHM 102	EPA 200.8	346094		
70295203103	OHM 103	EPA 200.8	346094		
70295203104	LIN 1	EPA 200.8	346094		
70295203105	LIN 2	EPA 200.8	346094		
70295203106	LIN 3	EPA 200.8	346094		
70295203107	LIN 4	EPA 200.8	346094		
70295203108	LIN 5	EPA 200.8	346094		
70295203109	LIN 6	EPA 200.8	346094		
70295203110	LIN 7	EPA 200.8	346094		
70295203111	LIN 8	EPA 200.8	346094		
70295203112	LIN 9	EPA 200.8	346094		
70295203113	BRO 1	EPA 200.8	346186		
70295203114	BRO 2	EPA 200.8	346207		
70295203115	BRO 3	EPA 200.8	346208		
70295203116	BRO 4	EPA 200.8	346208		
70295203117	BRO 5	EPA 200.8	346208		
70295203118	BRO 6	EPA 200.8	346208		
70295203119	BRO 7	EPA 200.8	346208		
70295203120	BRO 8	EPA 200.8	346208		
70295203121	SUPPORT SERVICES 1	EPA 200.8	346208		
70295203122	CONCERVATION BUILDING 1	EPA 200.8	346208		
70295203123	MVCC 1	EPA 200.8	346208		
70295203124	MVCC 2	EPA 200.8	346208		
70295203125	MVCC 3	EPA 200.8	346208		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
 without the written consent of Pace Analytical Services, LLC.

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Company Name: OHM BOCES
 Street Address: 4747 Middle Settlement Rd, New Hartford, NY 13413

Contact/Report To: Michael Colangelo

Phone #:
 E-Mail: mcolangelo@oneida-boces.org
 Cc E-Mail:

Invoice To: Michael Colangelo
 Invoice E-Mail: mcolangelo@oneida-boces.org

Purchase Order # (if applicable):
 Quote #:

Time Zone Collected: [] AK [] PT [] MT [] CT [X] ET

County / State origin of sample(s): New York

Regulatory Program (DW, RCRA, etc.) as applicable: NY Lead in School DW
 Rush (Pre-approval required): DW PWSID # or WW Permit # as applicable:
 [] 2 Day [] 3 day [] 5 day [] Other:
 Date Results Requested: Standard 10 business day
 Field Filtered (if applicable): [] Yes [] No
 Analysis:

* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Collected		Res. CL2	Number & Type of Containers	Preservation non-conformance identified for sample
		Date	Time			
OHM 1	DW	4/17/24	06:18		1	200.8 Drinking Water (Pb only)
2			19			
3			20			
4			21			
5			25			
6			26			
7			27			
8			28			
9			29			
10			30			

Customer Remarks / Special Conditions / Possible Hazards:
 Lead
 6 Boxes (one shared with CES)

Collected By:
 Printed Name:
 Signature:
 Received by/Company: (Signature)
 Date/Time: 4-21/19 45
 Received by/Company: (Signature)
 Date/Time: 4/22 1700
 Received by/Company: (Signature)
 Date/Time: 4/23 5:58A
 Received by/Company: (Signature)

LAB USE ONLY - Affix Workorder# onto Label Here

WO#: 70295203



70295203

Specify Container Size **
 Identify Container Preservative Type**
 Analysis Requested

Proj. Mgr: Jack Germano
 AcctNum / Client ID:
 Table #:
 Profile / Template:
 Prelog / Bottle Ord. ID:

Container Size: (1) 1L, (2) 500mL, (3) 250mL, (4) 125mL, (5) 100mL, (6) 40mL Vial, (7) EnCore, (8) TerraCore, (9) Other
 Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sod. Thiosulfate, (9) Ascorbic Acid, (10) MeOH, (11) Other

Additional Instructions from Pace*:
 # Coolers: Thermometer ID: Correction Factor (°C): Obs. Temp. (°C) Corrected Temp. (°C)

Tracking Number:
 Date/Time: 4/22 04:23:24
 Date/Time: 8 - 15
 Date/Time:
 Date/Time:

Delivered by: [] In-Person [] Courier
 [] FedEx [] UPS [] Other

Page: 1 of 13

CHAIN-OF-CUSTODY Analytical Request Document
Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

LAB USE ONLY - Affix Workorder/Login Label Here



Scan QR Code for instructions

Company Name: **OHM BOCES**
Street Address: **4747 Middle Settlement Rd, New Hartford, NY 13413**
Customer Project #: **08215493**
Project Name: **OHM BOCES_BOCES Facility**

Contact/Report To: **Michael Colangelo**
Phone #: _____
E-Mail: **mcolangelo@oneida-boces.org**
Cc E-Mail: _____
Invoice To: **Michael Colangelo**
Invoice E-Mail: **mcolangelo@oneida-boces.org**

Specify Container Size **
Identify Container Preservative Type***
Analysis Requested

** Container Size: (1) 1L, (2) 500mL, (3) 250mL, (4) 125mL, (5) 100mL, (6) 40mL Vial, (7) Encore, (8) TerraCore, (9) Other
*** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sod. Thiosulfate, (9) Ascorbic Acid, (10) MeOH, (11) Other

Time Zone Collected: [] AK [] PT [] MT [] CT [X] ET
Data Deliverables:
[] Level II [] Level III [] Level IV
[] EQUIS
[] Other _____
Regulatory Program (DW, RCRA, etc.) as applicable: **NY Lead in School DW**
County / State origin of sample(s): **New York**
Rush (Pre-approval required):
DW PWSID # or WW Permit # as applicable: _____
Field Filtered (if applicable): [] Yes [] No
Date Results Requested: **Standard 10 business day**
Analysis: _____

Preservation non-conformance identified for sample
Lab Use Only
Proj. Mgr: **Jack Germano**
AcctNum / Client ID: _____
Table #: _____
Profile / Template: _____
X
Preleg / Bottle Ord. ID: _____
Sample Comment

Customer Sample ID	Matrix *	Comp / Grab	Collected		Res. CL2	Number & Type of Containers
			Date	Time		
OHM 11	DW	G	4/17/24	0631		1
12				0632		
13				0634		
14				0635		
15				0637		
16				0638		
17				0640		
18				0641		
19				0642		
20				0643		

Additional Instructions from Pace*:
Coolers: _____ Thermometer ID: _____ Obs. Temp. (°C) _____ Corrected Temp. (°C) _____
Tracking Number: _____
Date/Time: **4/17/24 0630**
Date/Time: **042324**
Date/Time: **8-15**
Date/Time: _____
Delivered by: [] In-Person [] Courier
[] FedEx [] UPS [] Other
Page: **2** of **13**

Customer Remarks / Special Conditions / Possible Hazards:
Lead
Relinquished by/Company: (Signature) _____ Date/Time: _____
Relinquished by/Company: (Signature) _____ Date/Time: **4/17/24 1700**
Relinquished by/Company: (Signature) _____ Date/Time: **4/23/24**
Relinquished by/Company: (Signature) _____ Date/Time: _____

Collected By: **Chris Puns**
Printed Name: _____
Signature: **Chris Puns**
Received by/Company: (Signature) _____ Date/Time: _____
Received by/Company: (Signature) _____ Date/Time: _____
Received by/Company: (Signature) _____ Date/Time: _____
Received by/Company: (Signature) _____ Date/Time: _____

Company Name: **OHM BOCES**
Street Address: **4747 Middle Settlement Rd, New Hartford, NY 13413**

Customer Project #: **08215443**
Project Name: **OHM Boccs_BOCES Facility**

Site Collection Info/Facility ID (as applicable):

Time Zone Collected: [] AK [] PT [] MT [] CT [X] ET
Data Deliverables:

[] Level II [] Level III [] Level IV
[] EQUIS
[] Other:

Rush (Pre-approval required):
[] 2 Day [] 3 day [] 5 day [] Other: _____
Date Results Requested: Standard 10 business day

Field Filtered (if applicable): [] Yes [] No
Analysis:

Regulatory Program (DW, RCRA, etc.) as applicable: NY Lead in School DW
County / State origin of sample(s): **New York**

DW PWSID # or WW Permit # as applicable:

Matrix *
DW

Customer Sample ID
OHM 21
22
23
24
25
26
27
28
29
30

Comp / Grab
G

Collected (or Composite Start) Date Time
4/17/24 0644
0645
0647
0649
0651
0652
0657
0658
0659
0700

Res. CL2
1

Number & Type of Containers
Plastic Glass

Additional Instructions from Pace:

Collected By: *Chris Pace*
Printed Name: *Chris Pace*
Signature: *Chris Pace*

Received by/Company: (Signature)
Received by/Company: (Signature)
Received by/Company: (Signature)
Received by/Company: (Signature)

Date/Time: 4/22 1700
Date/Time: 4/22 0800
Date/Time: 042324
Date/Time: 15423815A

Tracking Number: 042324

Delivered by: [] In-Person [] Courier
[] FedEx [] UPS [] Other

Page: 3 of 13

Subbing a sample via this chain of custody constitutes acknowledgment and acceptance of the Pace Terms and Conditions found at <https://www.pacelabs.com/resource-library/resource/pace-terms-and-conditions/>

CHAIN-OF-CUSTODY Analytical Request Document
Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Contact/Report To: **Michael Colangelo**
Phone #: _____
E-Mail: mcolangelo@oneida-boces.org
Cc E-Mail: _____
Invoice To: **Michael Colangelo**
Invoice E-Mail: mcolangelo@oneida-boces.org

Purchase Order # (if applicable): _____
Quote #: _____

Specify Container Size **
Identify Container Preservative Type***

*** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sod. Thiosulfate, (9) Ascorbic Acid, (10) MeOH, (11) Other

Analysis Requested

Proj. Mgr: **Jack Germano**
AcctNum / Client ID: _____
Table #: _____
Profile / Template: _____
X
Prelog / Bottle Ord. ID: _____

Lab Use Only

Preservation non-conformance identified for sample

Sample Comment

200.8 Drinking Water (Pb only)

X

A

A

A

A

A

A

A

A

A

A

A

A

A

A

A

A

A

A

A

LAB USE ONLY - Affix Workorder/Login Label Here



Scan QR Code for instructions

CHAIN-OF-CUSTODY Analytical Request Document

Scan QR Code for instructions

Company Name: OHM BOCES
Street Address: 4747 Middle Settlement Rd, New Hartford, NY 13413
Customer Project #: 08215443
Project Name: OHM Boces_BOCES Facility

Contact/Report To: Michael Colangelo
Phone #:
E-Mail: mcolangelo@oneida-boces.org
Cc E-Mail:
Invoice To: Michael Colangelo
Invoice E-Mail: mcolangelo@oneida-boces.org

Purchase Order # (if applicable):
Quote #:
County / State origin of sample(s): New York

Regulatory Program (DW, RCRA, etc.) as applicable: NY Lead in School DW
Rush (Pre-approval required): DW PWSID # or WW Permit # as applicable:
 2 Day 3 day 5 day Other: _____
Date Results Requested: Standard 10 business day
Field Filtered (if applicable): Yes No
Analysis:

Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Collected		Res. CLZ	Number & Type of Containers	Sample Comment
		Comp / Grab	Time			
OHM 41 31	DW	G	4/17/14 0701		1	
OHM 42 32			0702			
OHM 43 33			0703			
OHM 44 34			0704			
OHM 45 35			0705			
OHM 46 36			0706			
OHM 47 37			0707			
OHM 48 38			0708			
OHM 49 39			0709			
OHM 50 40			0710			

Additional Instructions from Pace:
 Collected By: Chris Purze
 Printed Name: Chris Purze
 Signature: *Chris Purze*

Received by/Company (Signature): *Chris Purze*
Date/Time: 4/22 17:00
Relinquished by/Company (Signature):
Date/Time:
Relinquished by/Company (Signature):
Date/Time:
Relinquished by/Company (Signature):
Date/Time:

Tracking Number: 0800
Delivered by: [] In-Person [] Courier
Date/Time: 4/22 17:00
Date/Time:
Date/Time:
Date/Time:



LAB USE ONLY- Affix Workorder/Login Label Here

Scan QR Code for instructions

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Pace® Location Requested (City/State):
 Pace Analytical Long Island NY
 575 Broad Hollow Rd, Melville, NY 11747

Company Name: **OHM BOCES**
 Street Address: **4747 Middle Settlement Rd, New Hartford, NY 13413**

Contact/Report To: **Michael Colangelo**
 Phone #: _____
 E-Mail: mcolangelo@oneida-boces.org
 Cc E-Mail: _____

Invoice To: **Michael Colangelo**
 Invoice E-Mail: mcolangelo@oneida-boces.org

Purchase Order # (if applicable): _____
 Quote #: _____

Site Collection Info/Facility ID (as applicable): _____

Time Zone Collected: [] AK [] PT [] MT [] CT [X] ET
 Data Deliverables:
 [] Level II [] Level III [] Level IV
 [] EQUIS
 [] Other: _____

County / State origin of sample(s): **New York**

Rush (Pre-approval required):
 [] 2 Day [] 3 day [] 5 day [] Other: _____
 DW PMSID # or WW Permit # as applicable: _____
 Field Filtered (if applicable): [] Yes [] No
 Date Results Requested: **Standard 10 business day**
 Analysis: _____

* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Blossom (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected (or Composite Start)		Composite End	Res. CL2	Number & Type of Containers		Sample Comment
			Date	Time			Date	Time	
OHM 41	DW	G	4/17/24	0711			1		
42				0712					
43				0713					
44				0714					
45				0715					
46				0716					
47				0717					
48				0718					
49				0719					
50				0720					

200.8 Drinking Water (Pb only)

X

Additional Instructions from Pace:

Coolers: _____ Thermometer ID: _____ Correction Factor (°C): _____ Obs. Temp. (°C) _____ Corrected Temp. (°C) _____

Trading Number: **422 0800**

Date/Time: **4/22 042324**

Delivered by: [] In-Person [] Courier

Date/Time: _____

Received by/Company: (Signature) _____

Received by/Company: (Signature) _____

Received by/Company: (Signature) _____

Received by/Company: (Signature) _____

Received by/Company: (Signature) _____

Pace
 Pace Analytical Long Island NY
 575 Broad Hollow Rd, Melville, NY 11747

CHAIN-OF-CUSTODY Analytical Request Document
 Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

LAB USE ONLY - Affix Workorder/Login Label Here

Scan QR Code for instructions



Specify Container Size **
 Identify Container Preservative Type***
 Analysis Requested

** Container Size: (1) 1L, (2) 500mL, (3) 250mL, (4) 125mL, (5) 100mL, (6) 40mL vial, (7) Encore, (8) TerraCore, (9) Other
 *** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sod. Thiosulfate, (9) Ascorbic Acid, (10) MeOH, (11) Other

Project Name: **OHM BOCES**
 Street Address: **4747 Middle Settlement Rd, New Hartford, NY 13413**
 Customer Project #: **08215443**
 Project Name: **OHM BOCES BOCES Facility**

Contact/Report To: **Michael Colangelo**
 Phone #: **mccolangelo@oneida-boces.org**
 E-Mail: **mccolangelo@oneida-boces.org**
 Cc E-Mail:
 Invoice To: **Michael Colangelo**
 Invoice E-Mail: **mccolangelo@oneida-boces.org**

Purchase Order # (if applicable):
 Quote #:
 County / State origin of sample(s): **New York**

Regulatory Program (DW, RCRA, etc.) as applicable: **NY Lead in School DW**
 Rush (Pre-approval required):
 Date Results Requested: **Standard 10 business day**

Time Zone Collected: [] AK [] PT [] MT [] CT [X] ET
 Data Deliverables:
 [] Level II [] Level III [] Level IV
 [] EQUIS
 [] Other

Site Collection Info/Facility ID (as applicable):
 * Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Collected		Res. CL2	Number & Type of Containers	Sample Comment
		Comp / Grab	Time			
OHM 51	DW	G	4/17/24 0721		1	X
52			0722			
53			0723			
54			0724			
55			0725			
56			0726			
57			0727			
58			0729			
59			0729			
60			0730			

Additional Instructions from Pace:
 Collected By: **Chris Pitzer**
 Printed Name: **Chris Pitzer**
 Signature: *Chris Pitzer*

Received by/Company: (Signature)
 Date/Time: **4/22 0800**
 Tracking Number:
 Delivered by: [] In-Person [] Courier
 Date/Time: **042321**
 [] FedEx [] UPS [] Other
 Date/Time: **8-15**
 Page: **6** of **13**



Scan QR Code for instructions

CHAIN-OF-CUSTODY Analytical Request Document

Company Name: **OHM BOCES** Contact/Report To: **Michael Colangelo**

Street Address: **4747 Middle Settlement Rd, New Hartford, NY 13413**

Customer Project #: **08215443**

Project Name: **OHM Bocas_BOCES Facility**

Site Collection Info/Facility ID (as applicable):

Time Zone Collected: [] AK [] PT [] MT [] CT [X] ET County / State origin of sample(s): **New York**

Data Deliverables: [] Level II [] Level III [] Level IV Regulatory Program (DW, RCRA, etc.) as applicable: NY Lead in School DW

[] EQUIS Rush (Pre-approval required): **DW PWSID # or WW Permit # as applicable:**

[] Other: _____ Date Results Requested: **Standard 10 business day** Field Filtered (if applicable): [] Yes [] No

* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SD), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected		Res. CLZ	Composite End		Number & Type of Containers		Preservation non-conformance identified for sample.
			Date	Time		Date	Time	Plastic	Glass	
OHM 61	DW	G	4/17/24	0731				1		
62				0732						
63				0733						
64				0734						
65				0735						
66				0736						
67				0737						
68				0738						
69				0739						
70				0740						

Additional Instructions from Pace:

Collected By: **Chris Rite**

Printed Name: **Chris Rite**

Signature: *Chris Rite*

Received by/Company: (Signature) *Chris Rite*

Received by/Company: (Signature) _____

Received by/Company: (Signature) _____

Received by/Company: (Signature) _____

Received by/Company: (Signature) _____

Tracking Number: **462 04 2509**

Date/Time: **4/22 1700**

Date/Time: **4/23 85A**

Date/Time: _____

Date/Time: _____

Delivered by: [] In-Person [] Courier

[] FedEx [] UPS [] Other

Page: **7** of **13**

CHAIN-OF-CUSTODY Analytical Request Document
 Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

LAB USE ONLY - Affix Workorder/Login Label Here



Scan QR Code for instructions

Company Name: **OHM BOCES**
 Street Address: **4747 Middle Settlement Rd, New Hartford, NY 13413**

Contact/Report To: **Michael Colangelo**
 Phone #: _____
 E-Mail: **mcolangelo@oneida-boces.org**
 Cc E-Mail: _____

Invoice To: **Michael Colangelo**
 Invoice E-Mail: **mcolangelo@oneida-boces.org**

Purchase Order # (if applicable): _____
 Quote #: _____

Site Collection Info/Facility ID (as applicable): _____

Time Zone Collected: [] AK [] PT [] MT [] CT [X] ET
 County / State origin of sample(s): **New York**

Data Deliverables:
 [] Level II [] Level III [] Level IV
 [] EQUIS
 [] Other: _____

Rush (Pre-approval required):
 [] 2 Day [] 3 day [] 5 day [] Other: _____
 Date Results Requested: **Standard 10 business day**

Field Filtered (if applicable): [] Yes [] No
 Analysis: _____

Customer Sample ID	Matrix *	Comp / Grab	Collected		Res. CL2	Number & Type of Containers	200.8 Drinking Water (Pb only)	Sample Comment
			Date	Time				
OHM 71	DW	G	4/17/24	0741		1	X	
72				0742				
73				0743				
74				0744				
75				0745				
76				0746				
77				0747				
78				0748				
79				0749				
80				0750				

Customer Remarks / Special Conditions / Possible Hazards:
Lead

Collected By: **Chris Pizer**
 Printed Name: _____
 Signature: _____

Received by/Company: (Signature) _____
 Date/Time: **4/22 0800**

Received by/Company: (Signature) _____
 Date/Time: **042324**

Received by/Company: (Signature) _____
 Date/Time: **8-15**

Received by/Company: (Signature) _____
 Date/Time: _____

Lab Use Only

Proj. Mgr: **Jack Germano**
 AcctNum / Client ID: _____
 Table #: _____
 Profile / Template: _____
 Prelog / Bottle Ord. ID: _____

Preservation non-conformance identified for sample: _____

Specify Container Size **	Identify Container Preservative Type***	Analysis Requested
3		
2		

Additional Instructions from Page *:
 # Coolers: _____ Thermometer ID: _____ Obs. Temp. (°C) _____ Corrected Temp. (°C) _____

Tracking Number: _____
 Date/Time: **4/22 0800**

Delivered by: [] In-Person [] Courier
 [] FedEx [] UPS [] Other

Page: **9** of **13**

Scan QR Code for instructions



Specify Container Size **
 Identify Container Preservative Type ***
 Analysis Requested

*** Container Size: (1) 1L, (2) 500mL, (3) 250mL, (4) 125mL, (5) 100mL, (6) 40mL vial, (7) Encore, (8) TerraCore, (9) Other
 *** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sod. Thiosulfate, (9) Ascorbic Acid, (10) MeOH, (11) Other

Lab Use Only
 Proj. Mgr: **Jack Germano**
 AcctNum / Client ID:
 Table #:
 Profile / Template:
 Prelog. / Bottle Ord. ID:
 Sample Comment
 Preservation non-conformance identified for sample

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Contact/Report To: **Michael Colangelo**
 Phone #:
 E-Mail: **mcolangelo@oneida-boces.org**
 Cc E-Mail:
 Invoice To: **Michael Colangelo**
 Invoice E-Mail: **mcolangelo@oneida-boces.org**
 Purchase Order # (if applicable):
 Quote #:

Company Name: **OHM BOCES**
 Street Address: **4747 Middle Settlement Rd, New Hartford, NY 13413**
 Customer Project #: **08215443**
 Project Name: **OHM Boces_BOCES Facility**
 Site Collection Info/Facility ID (as applicable):

County / State origin of sample(s): **New York**
 Regulatory Program (DW, RCRA, etc.) as applicable: **NY Lead in School DW**
 Rush (Pre-approval required):
 () 2 Day () 3 day () 5 day () Other
 Date Results Requested: **Standard 10 business day**
 Field Filtered (if applicable): () Yes () No
 Analysis:

Time Zone Collected: () AK () PT () MT () CT (X) ET
 Data Deliverables:
 () Level II () Level III () Level IV
 () EQUIS
 () Other:
 * Matrix Codes (Inser in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected		Res. CL2	Number & Type of Containers	Sample Comment
			Date	Time			
OHM 81	DW	G	4/17/24	0751		1	X
82				0752			
83				0753			
84				0754			
85				0755			
86				0756			
87				0757			
88				0758			
89				0759			
90				0800			

Additional Instructions from Pace*:
 # Coolers: Thermometer ID: Correction Factor (°C): Obs. Temp. (°C) Corrected Temp. (°C)
 Tracking Number:
 Date/Time: **4/22 0800**
 Date/Time: **042324**
 Date/Time: **8-15**
 Date/Time:
 Delivered by: () In-Person () Courier
 () FedEx () UPS () Other
 Page: **9** of **13**

Collected By: **Chris Peter**
 Printed Name: **Chris Peter**
 Signature: *Chris Peter*
 Received by/Company: (Signature)
 Received by/Company: (Signature)
 Received by/Company: (Signature)
 Received by/Company: (Signature)

Date/Time: **4/22 1700**
 Date/Time: **4/23 815A**
 Date/Time:
 Date/Time:



Scan QR Code for instructions

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Peace* Location Requested (City/State):
 Pace Analytical Long Island NY
 575 Broad Hollow Rd, Melville, NY 11747

Company Name: **OHM BOCES**
 Street Address: **4747 Middle Settlement Rd, New Hartford, NY 13413**
 Contact/Report To: **Michael Colangelo**
 Phone #: _____
 E-Mail: **mcolangelo@oneida-boces.org**
 Cc E-Mail: _____
 Invoice To: **Michael Colangelo**
 Invoice E-Mail: **mcolangelo@oneida-boces.org**
 Purchase Order # (if applicable): _____
 Quote #: _____

Site Collection Info/Facility ID (as applicable): _____
 Time Zone Collected: [] AK [] PT [] MT [] CT [X] ET
 Data Deliverables:
 [] Level II [] Level III [] Level IV
 [] EQUIS
 [] Other: _____
 Rush (Pre-approval required):
 [] 2 Day [] 3 day [] 5 day [] Other: _____
 Date Results Requested: **Standard 10 business day**
 Field Filtered (if applicable): [] Yes [] No
 Analysis: _____
 DW PWSID # or WW Permit # as applicable: _____
 Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected		Res. CL2	Number & Type of Containers	Sample Comment
			Date	Time			
OHM 91	DW	G	4/17/24	0801		1	200.8 Drinking Water (Pb only)
92				0802			
93				0803			
94				0804			
95				0805			
96				0806			
97				0807			
98				0808			
99				0809			
100				0810			

Additional Instructions from Pace*: _____
 Collected By: **Chris Pitzer**
 Printed Name: _____
 Signature: _____
 Received by/Company: (Signature) _____
 Date/Time: _____
 Relinquished by/Company: (Signature) _____
 Date/Time: **4/22 1700**
 Relinquished by/Company: (Signature) _____
 Date/Time: **4/23 815**
 Relinquished by/Company: (Signature) _____
 Date/Time: _____

CHAIN-OF-CUSTODY Analytical Request Document
 Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

LAB USE ONLY - Affix Workorder/Login Label Here



Scan QR Code for instructions

Company Name: OHM BOCES
Street Address: 4747 Middle Settlement Rd, New Hartford, NY 13413

Contact/Report To: Michael Colangelo

Phone #:

E-Mail: mcolangelo@oneida-boces.org

Cc E-Mail:

Customer Project #: 0821543
Project Name: OHM BOCES_BOCES Facility

Invoice To: Michael Colangelo
Invoice E-Mail: mcolangelo@oneida-boces.org

Site Collection Info/Facility ID (as applicable):

Purchase Order # (if applicable):

Quote #:

Time Zone Collected: [] AK [] PT [] MT [] CT [X] ET

County / State origin of sample(s): New York

Data Deliverables:
 Regulatory Program (DW, RCRA, etc.) as applicable: NY Lead in School DW
 Rush (Pre-approval required): DW PWSID # or WW Permit # as applicable:
 [] 2 Day [] 3 day [] 5 day [] Other:
 Date Results Requested: Standard 10 business day
 Field Filtered (if applicable): [] Yes [] No
 Analysis:

* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected		Res. CLZ	Number & Type of Containers	Sample Comment
			Date	Time			
OHM 101	DW	G	4/17/24	0811		1	
102				0812			
103				0813			
LIN 1				0829			
2				0830			
3				0831			
4				0832			
5				0833			
6				0835			
7				0837			

Customer Remarks / Special Conditions / Possible Hazards:
 Lead

Collected By: Chris Perez
Printed Name:
Signature:

Received by/Company: (Signature)
Date/Time: 4/22 0800

Relinquished by/Company: (Signature)
Date/Time: 4/22 1700

Relinquished by/Company: (Signature)
Date/Time: 4/22 8154

Relinquished by/Company: (Signature)
Date/Time:

Relinquished by/Company: (Signature)
Date/Time:

Tracking Number: 042324

Delivered by: [] In-Person [] Courier
 [] FedEx [] UPS [] Other

Date/Time: 8-15

Page: 11 of 13



LAB USE ONLY - Affix Workorder/Login Label Here

Scan QR Code for instructions

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Pace® Location Requested (City/State):
 Pace Analytical Long Island NY
 575 Broad Hollow Rd, Melville, NY 11747

Company Name: **OHM BOCES**
 Street Address: **4747 Middle Settlement Rd, New Hartford, NY 13413**

Customer Project #: **0921 5443**
 Project Name: **OHM Boces_BOCES Facility**

Site Collection Info/Facility ID (as applicable):

Contact/Report To: **Michael Colangelo**
 Phone #: **mccolangelo@oneida-boces.org**
 E-Mail: **mccolangelo@oneida-boces.org**
 Cc E-Mail:

Invoice To: **Michael Colangelo**
 Invoice E-Mail: **mccolangelo@oneida-boces.org**

Purchase Order # (if applicable):
 Quote #:

Specify Container Size **
 Identify Container Preservative Type***
 Analysis Requested

** Container Size: (1) 1L, (2) 500mL, (3) 250mL, (4) 125mL, (5) 100mL, (6) 40mL vial, (7) EnCore, (8) TerraCore, (9) Other
 *** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sod. Thiosulfate, (9) Ascorbic Acid, (10) MeOH, (11) Other

County / State origin of sample(s): **New York**

Regulatory Program (DW, RCRA, etc.) as applicable: **NY Lead in School DW**

DW PWSID # or WW Permit # as applicable:
 Rush (Pre-approval required):
 Field Filtered (if applicable): [] Yes [] No
 Date Results Requested: **Standard 10 business day**
 Analysis:

* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected		Composite End Date	Time	Res. CL2	Number & Type of Containers	Plastic	Glass	Sample Comment
			Date	Time							
LEN 8	DW	G	4/17/24	0839				1			
↓ 9				0840							
Bro 1				0932							
↓ 2				0933							
↓ 3				0934							
↓ 4				0935							
↓ 5				0939							
↓ 6				0940							
↓ 7				0941							
↓ 8				0946							

Additional Instructions from Pace®:

Collected By: **Chris Pitzer**
 Printed Name: **Chris Pitzer**
 Signature: *Chris Pitzer*

Received by/Company: **Chris Pitzer**
 Signature: *Chris Pitzer*

Received by/Company: **Chris Pitzer**
 Signature: *Chris Pitzer*

Received by/Company: **Chris Pitzer**
 Signature: *Chris Pitzer*

Received by/Company: **Chris Pitzer**
 Signature: *Chris Pitzer*

Received by/Company: **Chris Pitzer**
 Signature: *Chris Pitzer*

Tracking Number: **0800**
 Date/Time: **4/22 0800**

Delivered by: [] In-Person [] Courier
 Date/Time: **042324**

[] FedEx [] UPS [] Other
 Date/Time: **8-15**

Page: **12** of **13**

Thermometer ID: _____ Obs. Temp. (°C) _____ Corrected Temp. (°C) _____
 # Coolers: _____ Correction Factor (°C): _____



LAB USE ONLY- Affix Workorder/Login Label Here

Scan QR Code for instructions

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Pace® Location Requested (City/State):
 Pace Analytical Long Island NY
 575 Broad Hollow Rd, Melville, NY 11747

Company Name: **OHM BOCES**
 Street Address: **4747 Middle Settlement Rd, New Hartford, NY 13413**

Contact/Report To: **Michael Colangelo**
 Phone #: _____
 E-Mail: **mcolangelo@oneida-boces.org**
 Cc E-Mail: _____

Invoice To: **Michael Colangelo**
 Invoice E-Mail: **mcolangelo@oneida-boces.org**

Purchase Order # (if applicable): _____
 Quote #: _____

Site Collection Info/Facility ID (as applicable): _____

Time Zone Collected: [] AK [] PT [] MT [] CT [X] ET [] NY

Data Deliverables:
 [] Level II [] Level III [] Level IV
 [] EQUIS
 [] Other: _____

Regulatory Program (DW, RCRA, etc.) as applicable: **NY Lead in School DW**

DW PWSID # or WW Permit # as applicable: _____

Rush (Pre-approval required):
 [] 2 Day [] 3 day [] 5 day [] Other: _____

Date Results Requested: **Standard 10 business day**

Field Filtered (if applicable): [] Yes [] No

Analysis: _____

* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected		Res. CL2	Number & Type of Containers	Sample Comment
			Date	Time			
Support Services 1	DW	G	4/17/24	1007		1	
Concentration Building 1							
NVCC 1				0904			
↓ 2				0905			
↓ 3				0906			

Customer Remarks / Special Conditions / Possible Hazards:
Chris Pizer

Additional Instructions from Pace®:
 # Coolers: _____ Thermometer ID: _____ Obs. Temp. (°C) _____ Corrected Temp. (°C) _____

Tracing Number: _____

Date/Time: **4/22 0800**
 Date/Time: **4/22 1700**
 Date/Time: **4/22 815K**

Received by/Company: (Signature) _____
 Received by/Company: (Signature) _____
 Received by/Company: (Signature) _____
 Received by/Company: (Signature) _____

Delivered by: [] In-Person [] Courier
 [] FedEx [] UPS [] Other

Date/Time: _____

Page: **13** of **13**

Client: **OHM**
 Work ID: **OHM BOCES UN BRD MCLC**
 Profile #: **8879-10067**
 of **10067**
 Add SCLOGED to first sample for field charge

Use Point Number Spreadsheet Multiday Project

Add SCLOGED to first sample for field charge

CC4 Line Item	Matrix	Container	Volume	Material	Notes
1					
2					
3					
4					
5					
6					
7					
8					
9					
10					
11					
12					
13					
14					
15					
16					
17					
18					
19					
20					
21					
22					
23					
24					
25					
26					
27					
28					
29					
30					
31					
32					
33					
34					
35					
36					
37					
38					
39					
40					
41					
42					
43					
44					
45					
46					
47					
48					
49					
50					
51					
52					
53					
54					
55					
56					
57					
58					
59					
60					
61					
62					
63					
64					
65					
66					
67					
68					
69					
70					
71					
72					
73					
74					
75					
76					
77					
78					
79					
80					
81					
82					
83					
84					
85					
86					
87					
88					
89					
90					
91					
92					
93					
94					
95					
96					
97					
98					
99					
100					

Misc.

SP5T	120mL Coliform Na Thio
R	Terracore Kit
WG2U	2oz Unpreserved Jar
WGFU	4oz Unpreserved Jar
WGRU	8oz Unpreserved Jar
WGDU	16oz Unpreserved Jar
ZPLC	Ziplock Bag
TEDL	Teal Jar
BG1H	1L HCL Clear Glass
GN	General
WP	Wipe
LLHG	Low Level Hq Bottles
BG1N	1L HNO3 Clear Glass

Glass

BP4U	125mL unpreserved plastic
BP3U	250mL unpreserved plastic
BP2U	500mL unpreserved plastic
BP1U	1L unpreserved plastic
BP4N	125mL HNO3 plastic
BP3N	250mL HNO3 plastic
BP2N	500mL H2SO4 plastic
BP3S	250mL H2SO4 plastic
BP2S	500mL H2SO4 plastic
BP3T	250mL Trizma
BP3S	250mL Ammonium Acetate
BP3R	250mL NH4SO4-NH4OH
BP1Z	1L NaOH, Zn Acetate
BP1N	1L HNO3 plastic
BP1B	Na Thiosulfate Amber Bottle

Plastic

BP4U	125mL unpreserved plastic
BP3U	250mL unpreserved plastic
BP2U	500mL unpreserved plastic
BP1U	1L unpreserved plastic
BP4N	125mL HNO3 plastic
BP3N	250mL HNO3 plastic
BP2N	500mL H2SO4 plastic
BP3S	250mL H2SO4 plastic
BP2S	500mL H2SO4 plastic
BP3T	250mL Trizma
BP3S	250mL Ammonium Acetate
BP3R	250mL NH4SO4-NH4OH
BP1Z	1L NaOH, Zn Acetate
BP1N	1L HNO3 plastic
BP1B	Na Thiosulfate Amber Bottle

Matrix

WT	Water
SL	Solid
NAL	Non-aqueous Liquid
OL	OIL
WP	Wipe
DW	Drinking Water

IOC

BP1U	1L unpreserved plastic
BP3N	250mL HNO3 plastic
BP3C	250mL Sodium Hydroxide
AG3U	500mL unpres amber glass
BP3U	250mL unpreserved plastic

* Can also be a BP4N

SOC

VG9T	40mL Na Thio amber vial
DG9A	40mL Ascorbic acid/maleic Acid vials
DG6T	Na Thiosulfate 60mL vial
DG6M	MonoChloride/Na Thio 60mL
AG3U	250mL unpres amber glass
AG3T	Na Thiosulfate 250mL bottle
BP1B	Na Thiosulfate Amber bottle
AG1T	Na Thiosulfate 1L Amber
AG1A	525-3 Chemical Blend

Sender Initials _____

Additional Comments _____

WO#: 70295203

PM: JMG Due Date: 05/07/24
 CLIENT: HerkimerMad

Client Name: Herkimer Mad Project # _____
 Courier: Fed Ex UPS USPS Client Commercial Pace Other

Tracking #: _____

Custody Seal on Cooler/Box Present: Yes No Seals intact: Yes No Temperature Blank Present: Yes No
 Packing Material: Bubble Wrap Bubble Bags Ziploc None Other Type of Ice: Wet Blue None
 Thermometer Used: TH211 Correction Factor: -0.4 Samples on ice, cooling process has begun
 Cooler Temperature (°C): 23.0 Cooler Temperature Corrected (°C): 22.6 Date/Time 5035A kits placed in freezer _____
 Temp should be above freezing to 6.0°C

USDA Regulated Soil (N/A, water sample)

Did samples originate in a quarantine zone within the United States: AL, AR, CA, FL, GA, ID, LA, MS, NC, NM, NY, OK, OR, SC, TN, TX, or VA (check map)? Yes No

Did samples originate from a foreign source including Hawaii and Puerto Rico? Yes No

If Yes to either question, fill out a Regulated Soil Checklist (ENV-FRM-MELV-0076) and include with SCUR/COC paperwork.

Date and Initials of person examining contents: 4/23/24 JMG

	COMMENTS:
Chain of Custody Present: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	1.
Chain of Custody Filled Out: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	2.
Chain of Custody Relinquished: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	3.
Sampler Name & Signature on COC: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	4.
Samples Arrived within Hold Time: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	5.
Short Hold Time Analysis (<72hr): <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	6.
Rush Turn Around Time Requested: <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	7.
Sufficient Volume: (Triple volume provided for MS/MSD) <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	8.
Correct Containers Used: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	9.
-Pace Containers Used: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	
Containers Intact: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	10.
Filtered volume received for Dissolved tests <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	11. Note: if sediment is visible in the dissolved container.
Sample Labels match COC: <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	12.
-Includes date/time/ID/Analysis Matrix: SL WT OIL OTHER	

Date and Initials of person checking preservation: _____

All containers needing preservation have been pH paper Lot # <u>211421</u> <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	13. <input type="checkbox"/> HNO ₃ <input type="checkbox"/> H ₂ SO ₄ <input type="checkbox"/> NaOH <input type="checkbox"/> HCl
All containers needing preservation are found to be in compliance with method recommendation? (HNO ₃ , H ₂ SO ₄ , HCl, NaOH>9 Sulfide, NAOH>12 Cyanide) <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	Sample # <u>4/23/24 JMG</u>
Exceptions: VOA, Coliform, TOC/DOC, Oil and Grease, DRO/8015 (water). Per Method, VOA pH is checked after analysis	Initial when completed: _____ Lot # of added preservative: _____ Date/Time preservative added: _____
Samples checked for dechlorination: <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	14. Positive for Res. Chlorine? Y N
KI starch test strips Lot # _____	
SM 4500 CN samples checked for sulf <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	15. Positive for Sulfide? Y N
Lead Acetate Strips Lot # _____	
Headspace in ALK Bottle (>6mm): <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	
Headspace in VOA Vials (>6mm): <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	16.
Trip Blank Present: <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	17.
Trip Blank Custody Seals Present <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	

Client Notification/ Resolution: _____ Field Data Required? Y / N
 Person Contacted: _____ Date/Time: _____
 Comments/ Resolution: _____

* PM (Project Manager) review (which includes the SCUR) is documented electronically in LIMS.