



December 17, 2024

Brian Arias
Jeff-Lewis BOCES_Watertown Central School
District
532 S. Massey Street
Watertown, NY 13601

RE: Project: WILEY SCHOOL 11/21
Pace Project No.: 70324395

Dear Brian Arias:

Enclosed are the analytical results for sample(s) received by the laboratory on November 26, 2024. The results relate only to the samples included in this report. Results reported herein conform to the applicable TNI/NELAC Standards and the laboratory's Quality Manual, where applicable, unless otherwise noted in the body of the report.

The test results provided in this final report were generated by each of the following laboratories within the Pace Network:

- Pace Analytical Services - Melville

If you have any questions concerning this report, please feel free to contact me.

Sincerely,

Michelle Cohen
michelle.cohen@pacelabs.com
516-370-6000
Project Manager

Enclosures

cc: Ray Filley, Jeff-Lewis BOCES_Watertown Central School
District
Tiffany Oconnor, Jeff-Lewis BOCES_Watertown Central
School District
Linda Shaw, Jeff-Lewis BOCES_Watertown Central School
District



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



CERTIFICATIONS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Pace Analytical Services, LLC - Melville, NY

575 Broad Hollow Rd, Melville, NY 11747

Connecticut Certification #: PH-0435

Delaware Certification # NY 10478

Maryland Certification #: 208

Massachusetts Certification #: M-NY026

New Hampshire Certification #: 2987

New Jersey Certification #: NY158

New York Certification #: 10478 Primary Accrediting Body

Pennsylvania Certification #: 68-00350

Rhode Island Certification #: LAO00340

Texas Certification #: T104704582

Florida Certification #: E871198

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 2-POOL LOCKER		Lab ID: 70324395001	Collected: 11/21/24 06:09	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/09/24 11:19	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 3A-DF OUTSIDE CURVE		Lab ID: 70324395002	Collected: 11/21/24 06:10	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/09/24 11:23	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 4-CAFE		Lab ID: 70324395004	Collected: 11/21/24 06:15	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	10.8	ug/L	1.0	1		12/09/24 11:25	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 5-KITCHEN		Lab ID: 70324395005	Collected: 11/21/24 06:45	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.0	ug/L	1.0	1		12/09/24 11:26	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 6-CLINIC		Lab ID: 70324395006	Collected: 11/21/24 06:21	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2220	ug/L	5.0	5		12/09/24 13:12	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 7-CLINIC BACK SINK		Lab ID: 70324395007	Collected: 11/21/24 06:21	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	282	ug/L	1.0	1		12/09/24 11:29	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 8-CLINIC KITCHEN BACK		Lab ID: 70324395008	Collected: 11/21/24 06:21	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	7.9	ug/L	1.0	1		12/09/24 11:30	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 12A-ART		Lab ID: 70324395009	Collected: 11/21/24 06:35	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	9.3	ug/L	1.0	1		12/09/24 11:32	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 12B-ART	Lab ID: 70324395010	Collected: 11/21/24 06:35	Received: 11/26/24 08:00	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	5.9	ug/L	1.0	1		12/09/24 11:33	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 12C-ART	Lab ID: 70324395011	Collected: 11/21/24 06:35	Received: 11/26/24 08:00	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	8.2	ug/L	1.0	1		12/09/24 11:35	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 12D-ART		Lab ID: 70324395012	Collected: 11/21/24 06:35	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.1	ug/L	1.0	1		12/09/24 11:54	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 13-KITCHEN (DENTAL)		Lab ID: 70324395013	Collected: 11/21/24 06:17	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	18.4	ug/L	1.0	1		12/09/24 11:58	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 14A-BAND		Lab ID: 70324395014	Collected: 11/21/24 06:45	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	45.5	ug/L	1.0	1		12/09/24 11:59	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 14B-BAND		Lab ID: 70324395015	Collected: 11/21/24 06:45	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	94.5	ug/L	1.0	1		12/09/24 12:01	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 14C-BAND		Lab ID: 70324395016	Collected: 11/21/24 06:45	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	13.4	ug/L	1.0	1		12/09/24 12:02	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 14D-BAND		Lab ID: 70324395017	Collected: 11/21/24 06:45	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/09/24 12:04	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 15A-DISTRICT DF		Lab ID: 70324395018	Collected: 11/21/24 06:28	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/09/24 12:05	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 15B-DISTRICT DF		Lab ID: 70324395019	Collected: 11/21/24 06:28	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/09/24 12:07	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 17A-DF ACROSS A7		Lab ID: 70324395020	Collected: 11/21/24 06:23	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.3	ug/L	1.0	1		12/09/24 12:08	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 17B-DF ACROSS A7		Lab ID: 70324395021	Collected: 11/21/24 06:23	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.6	ug/L	1.0	1		12/09/24 12:09	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 18-A7		Lab ID: 70324395022	Collected: 11/21/24 06:21	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	4.9	ug/L	1.0	1		12/09/24 12:11	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 19-A6		Lab ID: 70324395023	Collected: 11/21/24 06:21	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.5	ug/L	1.0	1		12/09/24 12:15	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 20-A5		Lab ID: 70324395024	Collected: 11/21/24 06:20	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	9.3	ug/L	1.0	1		12/09/24 12:17	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 21-A4		Lab ID: 70324395025	Collected: 11/21/24 07:15	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	4.2	ug/L	1.0	1		12/09/24 12:18	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 22-A2		Lab ID: 70324395026	Collected: 11/21/24 07:16	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	5.7	ug/L	1.0	1		12/09/24 12:20	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 23-A3	Lab ID: 70324395027	Collected: 11/21/24 07:18	Received: 11/26/24 08:00	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	5.6	ug/L	1.0	1		12/09/24 12:21	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 24-A1	Lab ID: 70324395028	Collected: 11/21/24 07:16	Received: 11/26/24 08:00	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/09/24 12:25	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 25A-DF OUTSIDE A1		Lab ID: 70324395029	Collected: 11/21/24 07:15	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/09/24 12:33	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 25B-DF OUTSIDE A1		Lab ID: 70324395030	Collected: 11/21/24 07:15	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/09/24 12:37	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 26-C3		Lab ID: 70324395031	Collected: 11/21/24 07:36	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.1	ug/L	1.0	1		12/09/24 12:38	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 27-C5		Lab ID: 70324395032	Collected: 11/21/24 07:36	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.8	ug/L	1.0	1		12/09/24 12:40	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 28-C6	Lab ID: 70324395033	Collected: 11/21/24 07:36	Received: 11/26/24 08:00	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	4.3	ug/L	1.0	1		12/09/24 12:41	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 29-C4		Lab ID: 70324395034	Collected: 11/21/24 07:37	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.8	ug/L	1.0	1		12/09/24 12:43	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 30-C7		Lab ID: 70324395035	Collected: 11/21/24 07:38	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	6.6	ug/L	1.0	1		12/09/24 12:44	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 31-C8		Lab ID: 70324395036	Collected: 11/21/24 07:39	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.5	ug/L	1.0	1		12/09/24 12:45	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 32A-DF OUTSIDE C8		Lab ID: 70324395037	Collected: 11/21/24 07:40	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.1	ug/L	1.0	1		12/09/24 12:50	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 32B-DF OUTSIDE C8		Lab ID: 70324395038	Collected: 11/21/24 07:40	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.2	ug/L	1.0	1		12/09/24 12:51	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 33-C9		Lab ID: 70324395039	Collected: 11/21/24 07:41	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.9	ug/L	1.0	1		12/09/24 12:53	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 34-C10		Lab ID: 70324395040	Collected: 11/21/24 07:43	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.4	ug/L	1.0	1		12/09/24 12:54	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 35-C11		Lab ID: 70324395041	Collected: 11/21/24 07:42	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.4	ug/L	1.0	1		12/09/24 12:56	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 36-C12		Lab ID: 70324395042	Collected: 11/21/24 07:43	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.5	ug/L	1.0	1		12/09/24 12:57	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 37-C15		Lab ID: 70324395043	Collected: 11/21/24 07:44	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.7	ug/L	1.0	1		12/09/24 12:59	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 38-C14		Lab ID: 70324395044	Collected: 11/21/24 07:46	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	5.7	ug/L	1.0	1		12/09/24 13:00	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 39-C16		Lab ID: 70324395045	Collected: 11/21/24 07:47	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	7.9	ug/L	1.0	1		12/09/24 13:01	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 40A-DF OUTSIDE 16		Lab ID: 70324395046	Collected: 11/21/24 07:48	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/09/24 13:03	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 40B-DF OUTSIDE 16		Lab ID: 70324395047	Collected: 11/21/24 07:48	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/09/24 13:07	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 42A-DF C2		Lab ID: 70324395048	Collected: 11/21/24 07:37	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/09/24 11:29	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 42B-DF C2		Lab ID: 70324395049	Collected: 11/21/24 07:37	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/09/24 11:35	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 43A-DF OUTSIDE B3		Lab ID: 70324395050	Collected: 11/21/24 06:50	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.8	ug/L	1.0	1		12/09/24 11:39	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 43B-DF OUTSIDE B3		Lab ID: 70324395051	Collected: 11/21/24 06:50	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.7	ug/L	1.0	1		12/09/24 11:44	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 44-B3		Lab ID: 70324395052	Collected: 11/21/24 06:50	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	5.9	ug/L	1.0	1		12/09/24 11:45	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 45-B5		Lab ID: 70324395053	Collected: 11/21/24 06:51	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	14.2	ug/L	1.0	1		12/09/24 11:47	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 46-B4		Lab ID: 70324395054	Collected: 11/21/24 06:55	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	16.3	ug/L	1.0	1		12/09/24 11:49	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 47-B7		Lab ID: 70324395055	Collected: 11/21/24 06:59	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.6	ug/L	1.0	1		12/09/24 11:50	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 48-B8	Lab ID: 70324395056	Collected: 11/21/24 07:01	Received: 11/26/24 08:00	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.4	ug/L	1.0	1		12/09/24 11:52	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 49-B9		Lab ID: 70324395057	Collected: 11/21/24 07:03	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.4	ug/L	1.0	1		12/09/24 11:53	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 50-B10		Lab ID: 70324395058	Collected: 11/21/24 07:05	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.8	ug/L	1.0	1		12/09/24 11:55	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 51A-DF OUTSIDE B10		Lab ID: 70324395059	Collected: 11/21/24 07:03	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/09/24 11:56	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 51B-DF OUTSIDE B10		Lab ID: 70324395060	Collected: 11/21/24 07:03	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/09/24 11:58	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 52-B11		Lab ID: 70324395061	Collected: 11/21/24 07:06	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.2	ug/L	1.0	1		12/09/24 12:03	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 53-B12		Lab ID: 70324395062	Collected: 11/21/24 07:06	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	5.8	ug/L	1.0	1		12/09/24 12:04	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 54-B13		Lab ID: 70324395063	Collected: 11/21/24 07:07	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	5.2	ug/L	1.0	1		12/09/24 12:06	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 55-B14		Lab ID: 70324395064	Collected: 11/21/24 07:08	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	7.1	ug/L	1.0	1		12/09/24 12:07	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 56-B15		Lab ID: 70324395065	Collected: 11/21/24 07:08	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	5.3	ug/L	1.0	1		12/09/24 12:09	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 57-B16		Lab ID: 70324395066	Collected: 11/21/24 07:09	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	4.8	ug/L	1.0	1		12/09/24 12:10	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: 58A-DF OUTSIDE B16 Lab ID: 70324395067 Collected: 11/21/24 07:11 Received: 11/26/24 08:00 Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/09/24 12:12	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 58B-DF OUTSIDE B16		Lab ID: 70324395068	Collected: 11/21/24 07:11	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/09/24 12:17	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 59-FACULTY LOUNGE		Lab ID: 70324395069	Collected: 11/21/24 07:12	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	6.3	ug/L	1.0	1		12/09/24 12:24	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 61-SCIENCE LAB		Lab ID: 70324395070	Collected: 11/21/24 07:30	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	34.5	ug/L	1.0	1		12/09/24 12:28	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Sample: 62-DISTRICT OFFICE		Lab ID: 70324395071	Collected: 11/21/24 07:17	Received: 11/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.1	ug/L	1.0	1		12/09/24 12:30	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

QC Batch:	374394	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70324395001, 70324395002, 70324395004, 70324395005, 70324395006, 70324395007, 70324395008, 70324395009, 70324395010, 70324395011		

METHOD BLANK:	1960675	Matrix:	Water
Associated Lab Samples:	70324395001, 70324395002, 70324395004, 70324395005, 70324395006, 70324395007, 70324395008, 70324395009, 70324395010, 70324395011		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	12/09/24 10:53	

LABORATORY CONTROL SAMPLE:	1960676					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	53.1	106	85-115	

MATRIX SPIKE SAMPLE:	1960678						
Parameter	Units	70324612001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	54.3	108	70-130	

MATRIX SPIKE SAMPLE:	1960680						
Parameter	Units	70324636001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	50.5	101	70-130	

SAMPLE DUPLICATE:	1960677					
Parameter	Units	70324612001 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	<1.0	<1.0			

SAMPLE DUPLICATE:	1960679					
Parameter	Units	70324636001 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	<1.0	<1.0			

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

QC Batch:	374412	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70324395012, 70324395013, 70324395014, 70324395015, 70324395016, 70324395017, 70324395018, 70324395019, 70324395020, 70324395021, 70324395022, 70324395023, 70324395024, 70324395025, 70324395026, 70324395027		

METHOD BLANK:	1960734	Matrix:	Water
Associated Lab Samples:	70324395012, 70324395013, 70324395014, 70324395015, 70324395016, 70324395017, 70324395018, 70324395019, 70324395020, 70324395021, 70324395022, 70324395023, 70324395024, 70324395025, 70324395026, 70324395027		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	12/09/24 11:36	

LABORATORY CONTROL SAMPLE:	1960735					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	50.5	101	85-115	

MATRIX SPIKE SAMPLE:	1960737						
Parameter	Units	70325798002 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	1.7	50	52.9	102	70-130	

MATRIX SPIKE SAMPLE:	1960739						
Parameter	Units	70325798003 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	3.4	50	53.8	101	70-130	

SAMPLE DUPLICATE:	1960736					
Parameter	Units	70325798002 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	1.7	1.7	3		

SAMPLE DUPLICATE:	1960738					
Parameter	Units	70325798003 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	3.4	3.4	2		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

QC Batch:	374413	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70324395028, 70324395029, 70324395030, 70324395031, 70324395032, 70324395033, 70324395034, 70324395035, 70324395036, 70324395037, 70324395038, 70324395039, 70324395040, 70324395041, 70324395042, 70324395043, 70324395044, 70324395045, 70324395046, 70324395047		

METHOD BLANK:	1960740	Matrix:	Water
Associated Lab Samples:	70324395028, 70324395029, 70324395030, 70324395031, 70324395032, 70324395033, 70324395034, 70324395035, 70324395036, 70324395037, 70324395038, 70324395039, 70324395040, 70324395041, 70324395042, 70324395043, 70324395044, 70324395045, 70324395046, 70324395047		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	12/09/24 12:22	

LABORATORY CONTROL SAMPLE:	1960741					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	50.0	100	85-115	

MATRIX SPIKE SAMPLE:	1960743						
Parameter	Units	70324395028 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	50.4	99	70-130	

MATRIX SPIKE SAMPLE:	1960745						
Parameter	Units	70324395029 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	49.6	99	70-130	

SAMPLE DUPLICATE:	1960742					
Parameter	Units	70324395028 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	<1.0	<1.0			

SAMPLE DUPLICATE:	1960744					
Parameter	Units	70324395029 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	<1.0	<1.0			

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

QC Batch:	374435	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70324395048, 70324395049, 70324395050, 70324395051, 70324395052, 70324395053, 70324395054, 70324395055, 70324395056, 70324395057, 70324395058, 70324395059, 70324395060, 70324395061, 70324395062, 70324395063, 70324395064, 70324395065, 70324395066, 70324395067		

METHOD BLANK:	1960853	Matrix:	Water
Associated Lab Samples:	70324395048, 70324395049, 70324395050, 70324395051, 70324395052, 70324395053, 70324395054, 70324395055, 70324395056, 70324395057, 70324395058, 70324395059, 70324395060, 70324395061, 70324395062, 70324395063, 70324395064, 70324395065, 70324395066, 70324395067		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	12/09/24 11:26	

LABORATORY CONTROL SAMPLE:	1960854					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	51.7	103	85-115	

MATRIX SPIKE SAMPLE:	1960857						
Parameter	Units	70324395048 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	51.5	102	70-130	

MATRIX SPIKE SAMPLE:	1960859						
Parameter	Units	70324395049 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	52.1	103	70-130	

SAMPLE DUPLICATE:	1960856					
Parameter	Units	70324395048 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	<1.0	<1.0			

SAMPLE DUPLICATE:	1960858					
Parameter	Units	70324395049 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	<1.0	<1.0			

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

QC Batch:	374436	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville

Associated Lab Samples: 70324395068, 70324395069, 70324395070, 70324395071

METHOD BLANK: 1960861 Matrix: Water
 Associated Lab Samples: 70324395068, 70324395069, 70324395070, 70324395071

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	12/09/24 12:14	

LABORATORY CONTROL SAMPLE: 1960862

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	51.6	103	85-115	

MATRIX SPIKE SAMPLE: 1960864

Parameter	Units	70324395068 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	50.7	100	70-130	

MATRIX SPIKE SAMPLE: 1960866

Parameter	Units	70324395069 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	6.3	50	56.3	100	70-130	

SAMPLE DUPLICATE: 1960863

Parameter	Units	70324395068 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

SAMPLE DUPLICATE: 1960865

Parameter	Units	70324395069 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	6.3	6.0	5	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALIFIERS

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

TNTC - Too Numerous To Count

J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit - The lowest concentration value that meets project requirements for quantitative data with known precision and bias for a specific analyte in a specific matrix.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Reported results are not rounded until the final step prior to reporting. Therefore, calculated parameters that are typically reported as "Total" may vary slightly from the sum of the reported component parameters.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70324395001	2-PPOOL LOCKER	EPA 200.8	374394		
70324395002	3A-DF OUTSIDE CURVE	EPA 200.8	374394		
70324395004	4-CAFE	EPA 200.8	374394		
70324395005	5-KITCHEN	EPA 200.8	374394		
70324395006	6-CLINIC	EPA 200.8	374394		
70324395007	7-CLINIC BACK SINK	EPA 200.8	374394		
70324395008	8-CLINIC KITCHEN BACK	EPA 200.8	374394		
70324395009	12A-ART	EPA 200.8	374394		
70324395010	12B-ART	EPA 200.8	374394		
70324395011	12C-ART	EPA 200.8	374394		
70324395012	12D-ART	EPA 200.8	374412		
70324395013	13-KITCHEN (DENTAL)	EPA 200.8	374412		
70324395014	14A-BAND	EPA 200.8	374412		
70324395015	14B-BAND	EPA 200.8	374412		
70324395016	14C-BAND	EPA 200.8	374412		
70324395017	14D-BAND	EPA 200.8	374412		
70324395018	15A-DISTRICT DF	EPA 200.8	374412		
70324395019	15B-DISTRICT DF	EPA 200.8	374412		
70324395020	17A-DF ACROSS A7	EPA 200.8	374412		
70324395021	17B-DF ACROSS A7	EPA 200.8	374412		
70324395022	18-A7	EPA 200.8	374412		
70324395023	19-A6	EPA 200.8	374412		
70324395024	20-A5	EPA 200.8	374412		
70324395025	21-A4	EPA 200.8	374412		
70324395026	22-A2	EPA 200.8	374412		
70324395027	23-A3	EPA 200.8	374412		
70324395028	24-A1	EPA 200.8	374413		
70324395029	25A-DF OUTSIDE A1	EPA 200.8	374413		
70324395030	25B-DF OUTSIDE A1	EPA 200.8	374413		
70324395031	26-C3	EPA 200.8	374413		
70324395032	27-C5	EPA 200.8	374413		
70324395033	28-C6	EPA 200.8	374413		
70324395034	29-C4	EPA 200.8	374413		
70324395035	30-C7	EPA 200.8	374413		
70324395036	31-C8	EPA 200.8	374413		
70324395037	32A-DF OUTSIDE C8	EPA 200.8	374413		
70324395038	32B-DF OUTSIDE C8	EPA 200.8	374413		
70324395039	33-C9	EPA 200.8	374413		
70324395040	34-C10	EPA 200.8	374413		
70324395041	35-C11	EPA 200.8	374413		
70324395042	36-C12	EPA 200.8	374413		
70324395043	37-C15	EPA 200.8	374413		
70324395044	38-C14	EPA 200.8	374413		
70324395045	39-C16	EPA 200.8	374413		
70324395046	40A-DF OUTSIDE 16	EPA 200.8	374413		
70324395047	40B-DF OUTSIDE 16	EPA 200.8	374413		
70324395048	42A-DF C2	EPA 200.8	374435		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: WILEY SCHOOL 11/21

Pace Project No.: 70324395

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70324395049	42B-DF C2	EPA 200.8	374435		
70324395050	43A-DF OUTSIDE B3	EPA 200.8	374435		
70324395051	43B-DF OUTSIDE B3	EPA 200.8	374435		
70324395052	44-B3	EPA 200.8	374435		
70324395053	45-B5	EPA 200.8	374435		
70324395054	46-B4	EPA 200.8	374435		
70324395055	47-B7	EPA 200.8	374435		
70324395056	48-B8	EPA 200.8	374435		
70324395057	49-B9	EPA 200.8	374435		
70324395058	50-B10	EPA 200.8	374435		
70324395059	51A-DF OUTSIDE B10	EPA 200.8	374435		
70324395060	51B-DF OUTSIDE B10	EPA 200.8	374435		
70324395061	52-B11	EPA 200.8	374435		
70324395062	53-B12	EPA 200.8	374435		
70324395063	54-B13	EPA 200.8	374435		
70324395064	55-B14	EPA 200.8	374435		
70324395065	56-B15	EPA 200.8	374435		
70324395066	57-B16	EPA 200.8	374435		
70324395067	58A-DF OUTSIDE B16	EPA 200.8	374435		
70324395068	58B-DF OUTSIDE B16	EPA 200.8	374436		
70324395069	59-FACULTY LOUNGE	EPA 200.8	374436		
70324395070	61-SCIENCE LAB	EPA 200.8	374436		
70324395071	62-DISTRICT OFFICE	EPA 200.8	374436		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.

Company Name: Jefferson Lewis BOCCS Watertown CSD
 Street Address: 1351 Washington Street, Watertown NY 13601

Customer Project #: 08215607
 Project Name: WILEY SETHL

Site Collection Info/Facility ID (as applicable):

CHAIN-OF-CUSTODY Analytical Request Document
 Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Contact/Report To: Brian Arias
 Phone #: 315-661-4351
 E-Mail: barias@watertowncswd.org
 Cc E-Mail:

Invoice To: Brian Arias
 Invoice E-Mail: barias@watertowncswd.org

Purchase Order # (if applicable):
 Quote #:

County / State origin of sample(s): New York

Regulatory Program (DW, RCRA, etc.) as applicable: NY Lead in School DW
 Rush (Pre-approval required):
 DW PWSID # or WW Permit # as applicable:
 Field Filtered (if applicable): [] Yes [] No

Date Results Requested: Standard 10 business day

* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected		Res. CLZ	Number & Type of Containers	Preservation non-conformance identified for sample
			Date	Time			
2-POOL LOCKER	DW	G	11/21/24	609		1	
3A - DF OUTSIDE CLINIC				610			
3B - DF OUTSIDE CLINIC				610			
4 - CAPE				615			
5 - KITCHEN				615			
6 - clinic				621			
7 - clinic back sink				621			
8 - clinic kitchen back				621			
9A - ART				635			
9B - ART				635			

Customer Remarks / Special Conditions / Possible Hazards:
 Lead

Collected By: Thomas Rankin
 Signature:

Received by/Company (Signature):
 Date/Time: 11/25/24 13:35
 Received by/Company (Signature):
 Date/Time: 11/25 14:00
 Received by/Company (Signature):
 Date/Time: 11/26 09:00
 Received by/Company (Signature):
 Date/Time:

Tracking Number:
 Date/Time: 11/25 13:35
 Delivered by: [] In-Person [] Courier
 Date/Time: 11/26/24 09:00
 Date/Time:

Page: 6 of 18
 Additional Instructions from Pace®:
 # Coolers: Thermometer ID: Correction Factor (°C): Obs. Temp. (°C) Corrected Temp. (°C)

1/7

W0#: 70324395

70324395

Specify Container Size **

Identify Container Preservative Type***

Analysis Requested

Proj. Mgr: Randy Budhu
 AcctNum / Client ID:
 Table #:
 Profile / Template:
 Prelog / Bottle Ord. ID:
 Sample Comment

Lab Use Only

Container Size: (1) 1L, (2) 500mL, (3) 250mL, (4) 125mL, (5) 100mL, (6) 40mL vial, (7) Encore, (8) Terra Core, (9) Other

Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sod. Thiosulfate, (9) Ascorbic Acid, (10) MeOH, (11) Other



Pace* Location Requested (City/State):
 Pace Analytical Long Island NY
 575 Broad Hollow Rd, Melville, NY 11747

Company Name: Jefferson Lewis BOCES, Watertown CSD
 Street Address: 1351 Washington Street, Watertown NY 13601

Contact/Report To: Brian Arias
 Phone #: 315-661-4351
 E-Mail: barias@watertowncscd.org
 Cc E-Mail:

Invoice To: Brian Arias
 Invoice E-Mail: barias@watertowncscd.org

Purchase Order # (if applicable):
 Quote #:

County / State origin of sample(s):
 New York

Regulatory Program (DW, RCRA, etc.) as applicable: NY Lead in School DW

Rush (Pre-approval required):
 DW PWSID # or WW Permit # as applicable:
 [] 2 Day [] 3 day [] 5 day [] Other:
 Date Results Requested: Standard 10 business day
 Field Filtered (if applicable): [] Yes [] No
 Analysis:

* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected		Res. CL2	Number & Type of Containers	200.8 Drinking Water (Pb only)
			Date	Time			
12C - FRET	DW	G	11/21/24	635		1	X
12D - ART				635			
13 - KITCHEN (DEWJAL)				617			
14A - BAMD				645			
14B - BAMD				645			
14C - BAMD				645			
14D - BAMD				645			
15A - DISTRICT DF				628			
15B - DISTRICT DF				628			
17A - DF AROSSA7				623			

Customer Remarks / Special Conditions / Possible Hazards:
 Lead


Collected By: Thomas Rankin
 Printed Name: Thomas Rankin
 Signature:

Relinquished by/Company (Signature):
 Date/Time: 11/25/24 13:35
 Relinquished by/Company (Signature):
 Date/Time: 11/25 14:00
 Relinquished by/Company (Signature):
 Date/Time: 11/26 02:00

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

LAB USE ONLY - Affix Workorder/Login Label Here



Scan QR Code for instructions

Specify Container Size **
 ** Container Size: (1) 1L, (2) 500mL, (3) 250mL, (4) 125mL, (5) 100mL, (6) 40mL vial, (7) Encore, (8) TerraCore, (9) Other
 Identify Container Preservative Type ***
 *** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sod. Thiosulfate, (9) Ascorbic Acid, (10) MeOH, (11) Other

Analysis Requested

Proj. Mgr:
 Randy Budhu
 AcctNum / Client ID:
 Table #:
 Profile / Template:
 K
 Prelog / Bottle Ord. ID:

Lab Use Only
 Preservation non-conformance identified for sample

Additional Instructions from Pace*:
 # Coolers: Thermometer ID: Obs. Temp. (°C) Corrected Temp. (°C)
 Tracking Number:
 Date/Time: 11/25 13:35
 Delivered by: [] In-Person [] Courier
 Date/Time: 11/26 16:00
 [] FedEx [] UPS [] Other
 Date/Time: 11/26 24:00
 Page: 7 of 18

3/7

LAB USE ONLY - Affix Workorder/Login Label Here

Scan QR Code for instructions



CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Pace® Location Requested (City/State):
 Pace Analytical Long Island NY
 575 Broad Hollow Rd, Melville, NY 11747

Company Name: **Jefferson Lewis BOCCES, Watertown CSD**
 Street Address: **1351 Washington Street, Watertown NY 13601**

Contact/Report To: **Brian Arias**
 Phone #: **315-661-4351**
 E-Mail: **barias@watertowncso.org**
 Cc E-Mail:

Invoice To: **Brian Arias**
 Invoice E-Mail: **barias@watertowncso.org**

Purchase Order # (if applicable):
 Quote #:

County / State origin of sample(s): **New York**

Regulatory Program (DW, RCRA, etc.) as applicable: **NY Lead in School DW**

Rush (Pre-approval required):
 2 Day 3 day 5 day Other _____

Date Results Requested: **Standard 10 business day**

Field Filtered (if applicable): Yes No

Analysis:

* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SD), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected		Res. CL2	Number & Type of Containers	Plastic	Glass	Sample Comment
			Date	Time					
17B - DE ACROSS #7	DW	G	11/21/24	6:03		1			
18 - A7				6:21					
19 - A6				6:21					
20 - A5				6:20					
21 - A4				7:15					
22 - A3				7:16					
23 - A3				7:18					
24 - A1				7:16					
25A - DE outside A1				7:15					
25B - DE outside A1				7:15					

Additional Instructions from Pace®:

Coolers: _____ Thermometer ID: _____ Obs. Temp. (°C) _____ Corrected Temp. (°C) _____

Correction Factor (°C): _____

Tracking Number: _____

Delivered by: In-Person Courier FedEx UPS Other

Date/Time: 11/25 13:35

Date/Time: 11/26 14:00

Date/Time: 11/26 9:00

Date/Time: _____


Page: 2 of 18

Subscribing a sample via this chain of custody constitutes acknowledgment and acceptance of the Pace® Terms and Conditions found at <https://www.pacelabs.com/resource-library/resource/pace-terms-and-conditions/>

ENV-FRM-CORQ-0019_v01_082123 ©

4-7

LAB USE ONLY - Affix Workorder/Login Label Here



Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Scan QR Code for instructions

Specify Container Size **

Identify Container Preservative Type ***

Analysis Requested

** Container Size: (1) 1L, (2) 500mL, (3) 250mL, (4) 125mL, (5) 100mL, (6) 40mL via, (7) Encore, (8) TerraCore, (9) Other

*** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sod. Thiosulfate, (9) Ascorbic Acid, (10) MeOH, (11) Other

Proj. Mgr:
Randy Budhu

Account / Client ID:

Table #:

Profile / Template:
X

Prelog / Bottle Ord. ID:

Sample Comment

Preservation non-conformance identified for sample

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Contact/Report To: **Brian Arias**
Phone #: **315-661-4351**
E-Mail: **barias@watertowncscd.org**
Cc E-Mail:

Invoice To: **Brian Arias**
Invoice E-Mail: **barias@watertowncscd.org**

Purchase Order # (if applicable):
Quote #:

County / State origin of sample(s): **New York**

Regulatory Program (DW, RCRA, etc) as applicable: **NY Lead in School DW**

Rush (Pre-approval required):
[] 2 Day [] 3 day [] 5 day [] Other:
Date Results Requested: **Standard 10 business day**

Field Filtered (if applicable): [] Yes [] No
Analysis:

Pace® Location Requested (City/State):
Pace Analytical Long Island NY
575 Broad Hollow Rd, Melville, NY 11747

Company Name: **Jefferson Lewis BOCES, Watertown CSD**
Street Address: **1351 Washington Street, Watertown NY 13601**

Customer Project #: **08215207**
Project Name: **Wiley School**

Site Collection Info/Facility ID (as applicable):

Time Zone Collected: [] AK [] PT [] MT [] CT [] ET [X] ET

Data Deliverables:
[] Level II [] Level III [] Level IV
[] EQUIS
[] Other:

* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix * Grab	Collected Date	Time	Composite End Date	Time	Res. CL	Number & Type of Containers	
							Plastic	Glass
26-C3	DW	11/21/24	736				1	
27-05			736					
28-C6			736					
29-C4			737					
30-C7			738					
31-C8			739					
32A - DF OUTSIDE C8			740					
32B - DF OUTSIDE C8			740					
33 - C9			741					
34 - C10			743					

Additional Instructions from Pace®:

Collected By: **Thomas Rankin**
Printed Name: **Thomas Rankin**
Signature:

Coolers: Thermometer ID: Correction Factor (°C): Obs. Temp. (°C) Corrected Temp. (°C)

Required by/Company: (Signature) **Thomas Rankin** Date/Time: **11/25 13:35**

Received by/Company: (Signature) **Thomas Rankin** Date/Time: **11/25 14:00**

Required by/Company: (Signature) **Thomas Rankin** Date/Time: **11/26 9:00**

Received by/Company: (Signature) **Thomas Rankin** Date/Time: **11/26 9:00**

Tracking Number: **13135**

Delivered by: [] In-Person [] Courier [] FedEx [] UPS [] Other

Page: **9** of **12**

Submitting a sample via this chain of custody constitutes acknowledgment and acceptance of the Pace® Terms and Conditions found at <https://www.pacelabs.com/resource-library/resource/pace-terms-and-conditions/>

5/7

LAB USE ONLY - Affix Workorder/Login Label Here



Scan QR Code for instructions

**Container Size: (1) 1L, (2) 500mL, (3) 250mL, (4) 125mL, (5) 100mL, (6) 40mL vial, (7) Encore, (8) TerraCore, (9) Other

*** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sod. Thiosulfate, (9) Ascorbic Acid, (10) MeOH, (11) Other

Identify Container Preservative Type ***
Analysis Requested

Lab Use Only
Proj. Mgr: Randy Budhu
AcctNum / Client ID:
Table #:
Profile / Template:
Prelog / Bottle Ord. ID:

Preservation non-conformance identified for sample

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Pace® Location Requested (City/State):
Pace Analytical Long Island NY
575 Broad Hollow Rd, Melville, NY 11747

Company Name: Jefferson Lewis BOCES - Watertown CSD
Street Address: 1351 Washington Street, Watertown NY 13601
Contact/Report To: Brian Arias
Phone #: 315-661-4351
E-Mail: barias@watertowncsd.org
Cc E-Mail:
Invoice To: Brian Arias
Invoice E-Mail: barias@watertowncsd.org
Purchase Order # (if applicable):
Quote #:

County / State origin of sample(s): New York

Regulatory Program (DW, RCRA, etc.) as applicable: NY Lead in School DW

Rush (Pre-approval required):
DW PWSID # or WW Permit # as applicable:
Field Filtered (if applicable): [] Yes [] No

Date Results Requested: Standard 10 business day

* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected		Res. CL2	Number & Type of Containers
			Date	Time		
35 - C11	DW	G	11/21/24	742		1
36 - C12				743		
37 - C15				744		
38 - C14				746		
39 - C16				747		
40A - DF outside 16				748		
40B - DF outside 16				748		
42A - DF C2				737		
42B - DF C2				737		
43A - DF outside B3				650		

Additional Instructions from Pace®:
Collected By: Thomas Rankin
Printed Name: Thomas Rankin
Signature:

Delivered by/Company: (Signature)	Date/Time	Tracking Number:
<i>Thomas Rankin</i>	11/25/24 13:35	13135
<i>Thomas Rankin</i>	11/25 14:00	4300
<i>Thomas Rankin</i>	11/25 14:00	4300
<i>Thomas Rankin</i>	11/25 14:00	4300

617

LAB USE ONLY - Affix Workorder/Login Label Here



Scan QR Code for instructions

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Pace®
 Pace Analytical Long Island NY
 575 Broad Hollow Rd, Melville, NY 11747

Company Name: Jefferson Lewis BOCCES_Watertown CSD
Street Address: 1351 Washington Street, Watertown NY 13601

Contact/Report To: Brian Arias
Phone #: 315-661-4351
E-Mail: barias@watertowncsd.org
Cc E-Mail:

Invoice To: Brian Arias
Invoice E-Mail: barias@watertowncsd.org

Purchase Order # (if applicable):
Quote #:

Specify Container Size **
 ** Container Size: (1) 1L, (2) 500ml, (3) 250ml, (4) 125ml, (5) 100ml, (6) 40ml vial, (7) Encore, (8) TerraCore, (9) Other

Identify Container Preservative Type ***
 *** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sod. Thiosulfate, (9) Ascorbic Acid, (10) MeOH, (11) Other

Analysis Requested

County / State origin of sample(s): New York

Regulatory Program (DW, RCRA, etc.) as applicable: NY Lead in School DW

Rush (Pre-approval required): DW PWSID # or WW Permit # as applicable:
 () 2 Day () 3 day () 5 day () Other

Date Results Requested: Standard 10 business day
 Field Filtered (if applicable): () Yes () No
 Analysis:

*** Matrix Codes (Insert in Matrix box below):** Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (O), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Collected		Composite End		Res. CL2	Number & Type of Containers	200.8 Drinking Water (Pb only)	Preservation non-conformance identified for sample
		Comp / Grab	Date	Time	Date				
43B - DF outside B3	DW	G	11/21/24	6:50			1	X	
44 - B3				6:50					
45 - B5				6:51					
46 - B4				6:55					
47 - B7				6:59					
48 - B8				7:01					
49 - B9				7:03					
50 - B10				7:05					
51A - DF outside B10				7:03					
51B - DF outside B10				7:03					

Additional Instructions from Pace®:

Collected By: Thomas Rankin
 Printed Name: Thomas Rankin
 Signature:

# Coolers:	Thermometer ID:	Correction Factor (°C):	Obs. Temp. (°C)	Corrected Temp. (°C)

Relinquished by/Company (Signature)	Date/Time	Received by/Company (Signature)	Date/Time	Tracking Number:
Monsieur PCL	11/26/24 13:35	Thomas Rankin	11/25 13:35	
Thomas Rankin	11/25 14:00	Thomas Rankin	11/26 4:50	
Thomas Rankin	11/26 8:00	Thomas Rankin	11/26 21:00	

Delivered by: () In-Person () Courier
 () FedEx () UPS () Other

Page: 11 of 18

9/7



LAB USE ONLY - Affix Workorder/Login Label Here
Scan QR Code for instructions



CHAIN-OF-CUSTODY Analytical Request Document
Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Company Name: Pace Analytical Long Island NY
575 Broad Hollow Rd, Melville, NY 11747
Company Name: Jefferson Lewis BOCES, Watertown CSD
Street Address: 1351 Washington Street, Watertown NY 13601
Contact/Report To: Brian Arias
Phone #: 315-661-4351
E-Mail: barias@watertowncsd.org
Cc E-Mail:
Invoice To: Brian Arias
Invoice E-Mail: barias@watertowncsd.org
Purchase Order # (if applicable):
Quote #:

Site Collection Info/Facility ID (as applicable):
08215107
Wiley School

Time Zone Collected: [] AK [] PT [] MT [] CT [X] ET
Data Deliverables: Regulatory Program (DW, RCRA, etc.) as applicable: NY Lead in School DW
New York

Rush (Pre-approval required): DW PWSID # or WW Permit # as applicable:
[] 2 Day [] 3 day [] 5 day [] Other:
Date Results Requested: Standard 10 business day
Field Filtered (if applicable): [] Yes [] No
Analysis:

Matrix Codes (insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected		Res. CL2	Number & Type of Containers	Plastic	Glass	Sample Comment
			Date	Time					
B2-B11	DW	G	11/21/24	7:06		1			
B3-B12				7:06					
B4-B13				7:07					
B5-B14				7:08					
B6-B15				7:08					
B7-B16				7:09					
B8A - DF outside B16				7:11					
B8B - DF outside B16				7:11					
B9 - Faculty lounge				7:12					
B1 - Science lab				7:30					
B2 - Disinfectant				7:17					

Additional Instructions from Pace:
Coolers: Thermometer ID: Correction Factor (°C): Obs. Temp. (°C) Corrected Temp. (°C)

Requested by/Company (Signature): Thomas PSE
Date/Time: 11/25/24 13:35
Received by/Company (Signature): Pace
Date/Time: 11/25 14:00
Requested by/Company (Signature): Thomas PSE
Date/Time: 11/26 09:00
Received by/Company (Signature): Pace

Tracking Number: 1125 13:35
Delivered by: [] In-Person [] Courier
[] FedEx [] UPS [] Other
Page: 12 of 18

WO#: 70324395

Client Name: WABERTOWN

Project # **PM: MC1** Due Date: **12/12/24**
CLIENT: WATERTOWN

Courier: Fed Ex UPS USPS Client Commercial Pac Other

Tracking #: _____

Custody Seal on Cooler/Box Present: Yes No Seals intact: Yes No Temperature Blank Present: Yes No
Packing Material: Bubble Wrap Bubble Bags Ziplo Non Other Type of Ice: Wet Blue None

Thermometer Used: TH211 Correction Factor: +0 Samples on ice, cooling process has begun
Cooler Temperature (°C): 16.8 Cooler Temperature Corrected (°C): 16.8 Date/Time 5035A kits placed in freezer _____
Temp should be above freezing to 6.0°C

USDA Regulated Soil (N/A, water sample)

Did samples originate in a quarantine zone within the United States: AL, AR, CA, FL, GA, ID, LA, MS, NC, NM, NY, OK, OR, SC, TN, TX, or VA (check map)? Yes No

Did samples originate from a foreign source including Hawaii and Puerto Rico? Yes No

If Yes to either question, fill out a Regulated Soil Checklist (ENV-FRM-MELV-0076) and include with SCUR/COC paperwork.

Date and Initials of person examining contents: ABB 11/26/24

	COMMENTS:
Chain of Custody Present: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	1.
Chain of Custody Filled Out: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	2.
Chain of Custody Relinquished: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	3.
Sampler Name & Signature on COC: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	4.
Samples Arrived within Hold Time: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	5.
Short Hold Time Analysis (<72hr): <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	6.
Rush Turn Around Time Requested <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	7.
Sufficient Volume: (Triple volume provided for MS/MSD) <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	8.
Correct Containers Used: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	9.
-Pace Containers Used: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	
Containers Intact: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	10.
Filtered volume received for Dissolved tests <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	11. Note: if sediment is visible in the dissolved container.
Sample Labels match COC: <input checked="" type="checkbox"/> Yes <input checked="" type="checkbox"/> No	12. <u>Sample 003 not received</u>
-Includes date/time/ID/Analysis Matrix: SL <input checked="" type="checkbox"/> WT <input checked="" type="checkbox"/> OIL OTHER	

Date and Initials of person checking preservation: ABB 11/26/24

All containers needing preservation have been pH paper Lot # <u>213624</u>	13. <input type="checkbox"/> HNO ₃ <input type="checkbox"/> H ₂ SO ₄ <input type="checkbox"/> NaOH <input type="checkbox"/> HCl
All containers needing preservation are found to be in compliance with method recommendation? (HNO ₃ , H ₂ SO ₄ , HCl, NaOH > 9 Sulfide, > 12 Cyanide) <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	Sample #
Exceptions: VOA, Coliform, TOC/DOC, Oil and Grease, DRO/8015 (water). Per Method, VOA pH is checked after analysis	Initial when completed: Lot # of added preservative: Date/Time preservative added:
Samples checked for dechlorination: <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	14. Positive for Res. Chlorine? Y N
KI starch test strips Lot #	
Residual chlorine strips Lot #	15. Positive for Sulfide? Y N
SM 4500 CN samples checked for sul <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	
Lead Acetate Strips Lot #	
Headspace in ALK Bottle (>6mm): <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	16.
Headspace in VOA Vials (>6mm): <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	
Trip Blank Present: <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	17.
Trip Blank Custody Seals Present <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	

Client Notification/ Resolution:

Field Data Required? Y / N

Person Contacted: _____

Date/Time: _____

Comments/ Resolution:

Received extra bottle empty with no labels

Contacted Brian Arias on 11/27/24 at 12:51pm regarding missing sample and extra bottle received.

* PM (Project Manager) review (which includes the SCUR) is documented electronically in LIMS.

MC 11/27/24