



December 18, 2024

Brian Arias
Jeff-Lewis BOCES_Watertown Central School
District
532 S. Massey Street
Watertown, NY 13601

RE: Project: OHIO ELEMENTARY
Pace Project No.: 70326667

Dear Brian Arias:

Enclosed are the analytical results for sample(s) received by the laboratory on December 10, 2024. The results relate only to the samples included in this report. Results reported herein conform to the applicable TNI/NELAC Standards and the laboratory's Quality Manual, where applicable, unless otherwise noted in the body of the report.

The test results provided in this final report were generated by each of the following laboratories within the Pace Network:

- Pace Analytical Services - Melville

If you have any questions concerning this report, please feel free to contact me.

Sincerely,

Michelle Cohen
michelle.cohen@pacelabs.com
516-370-6000
Project Manager

Enclosures

cc: Ray Filley, Jeff-Lewis BOCES_Watertown Central School
District
Tiffany Oconnor, Jeff-Lewis BOCES_Watertown Central
School District
Linda Shaw, Jeff-Lewis BOCES_Watertown Central School
District



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



CERTIFICATIONS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Pace Analytical Services, LLC - Melville, NY

575 Broad Hollow Rd, Melville, NY 11747

Connecticut Certification #: PH-0435

Delaware Certification # NY 10478

Maryland Certification #: 208

Massachusetts Certification #: M-NY026

New Hampshire Certification #: 2987

New Jersey Certification #: NY158

New York Certification #: 10478 Primary Accrediting Body

Pennsylvania Certification #: 68-00350

Rhode Island Certification #: LAO00340

Texas Certification #: T104704582

Florida Certification #: E871198

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 1 KITCHEN		Lab ID: 70326667001	Collected: 12/06/24 06:31	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.9	ug/L	1.0	1		12/17/24 13:02	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 2 DPK OFFICE		Lab ID: 70326667002	Collected: 12/06/24 06:35	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	63.0	ug/L	1.0	1		12/17/24 13:03	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 3-DF 139		Lab ID: 70326667003	Collected: 12/06/24 06:41	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.6	ug/L	1.0	1		12/17/24 13:05	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: 4-DF MEDIA CENTER								
Lab ID: 70326667004								
Collected: 12/06/24 06:47								
Received: 12/10/24 07:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		12/17/24 13:06	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 5-NURSE		Lab ID: 70326667005	Collected: 12/06/24 06:50	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	37.2	ug/L	1.0	1		12/17/24 13:08	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 6-NURSE SINK		Lab ID: 70326667006	Collected: 12/06/24 06:50	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	12.6	ug/L	1.0	1		12/17/24 13:10	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 7-DF 121		Lab ID: 70326667007	Collected: 12/06/24 07:19	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.6	ug/L	1.0	1		12/17/24 13:11	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: 8-R 114								
Lab ID: 70326667008								
Collected: 12/06/24 06:57								
Received: 12/10/24 07:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	47.9	ug/L	1.0	1		12/17/24 13:13	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 9-R 103		Lab ID: 70326667009	Collected: 12/06/24 07:15	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.4	ug/L	1.0	1		12/17/24 13:14	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 10-R 103 SINK		Lab ID: 70326667010	Collected: 12/06/24 07:15	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.1	ug/L	1.0	1		12/17/24 13:19	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 11-R 112		Lab ID: 70326667011	Collected: 12/06/24 06:58	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/17/24 13:20	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 12-R 112 SINK		Lab ID: 70326667012	Collected: 12/06/24 06:58	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/17/24 13:22	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 13- R 104		Lab ID: 70326667013	Collected: 12/06/24 07:14	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.8	ug/L	1.0	1		12/17/24 13:24	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 14- R 104 SINK		Lab ID: 70326667014	Collected: 12/06/24 07:14	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.3	ug/L	1.0	1		12/17/24 13:25	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 15-R 111		Lab ID: 70326667015	Collected: 12/06/24 07:01	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.6	ug/L	1.0	1		12/17/24 13:27	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 16-R 111 SINK		Lab ID: 70326667016	Collected: 12/06/24 07:01	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.1	ug/L	1.0	1		12/17/24 13:28	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 17-R 105		Lab ID: 70326667017	Collected: 12/06/24 07:12	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.3	ug/L	1.0	1		12/17/24 14:39	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 18-R 105 SINK		Lab ID: 70326667018	Collected: 12/06/24 07:12	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.2	ug/L	1.0	1		12/17/24 14:44	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 19- R 110	Lab ID: 70326667019	Collected: 12/06/24 07:03	Received: 12/10/24 07:00	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	2.1	ug/L	1.0	1		12/17/24 14:48	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 20-R 110 SINK		Lab ID: 70326667020	Collected: 12/06/24 07:03	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	6.4	ug/L	1.0	1		12/17/24 14:50	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 21-R 106		Lab ID: 70326667021	Collected: 12/06/24 07:11	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.0	ug/L	1.0	1		12/17/24 14:54	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 22-R 106 SINK		Lab ID: 70326667022	Collected: 12/06/24 07:11	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.4	ug/L	1.0	1		12/17/24 14:56	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 23- R 109		Lab ID: 70326667023	Collected: 12/06/24 07:05	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.3	ug/L	1.0	1		12/17/24 14:57	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 24- R 109 SINK		Lab ID: 70326667024	Collected: 12/06/24 07:05	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.8	ug/L	1.0	1		12/17/24 14:59	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 25- R 107		Lab ID: 70326667025	Collected: 12/06/24 07:08	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.9	ug/L	1.0	1		12/17/24 15:01	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 26- R 107 SINK		Lab ID: 70326667026	Collected: 12/06/24 07:08	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.3	ug/L	1.0	1		12/17/24 15:02	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 27-R 108		Lab ID: 70326667027	Collected: 12/06/24 07:09	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.1	ug/L	1.0	1		12/17/24 15:04	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: 28-R 108 SINK								
Lab ID: 70326667028								
Collected: 12/06/24 07:09								
Received: 12/10/24 07:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	2.5	ug/L	1.0	1		12/17/24 15:05	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 29- RM 117 SINK		Lab ID: 70326667029	Collected: 12/06/24 06:55	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	9.1	ug/L	1.0	1		12/17/24 15:07	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 31-RM 131		Lab ID: 70326667030	Collected: 12/06/24 07:23	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	11.9	ug/L	1.0	1		12/17/24 15:09	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 32- RM 144 SINK		Lab ID: 70326667031	Collected: 12/06/24 06:38	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	144	ug/L	1.0	1		12/17/24 15:13	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 33-RM 138 SINK		Lab ID: 70326667032	Collected: 12/06/24 06:41	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	100	ug/L	1.0	1		12/17/24 15:15	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: 34-RM 137 SINK								
Lab ID: 70326667033								
Collected: 12/06/24 06:47								
Received: 12/10/24 07:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	71.0	ug/L	1.0	1		12/17/24 15:16	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: 35-RM 143 SINK								
Lab ID: 70326667034								
Collected: 12/06/24 06:36								
Received: 12/10/24 07:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	131	ug/L	1.0	1		12/17/24 15:18	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: 36-RM 142 SINK								
Lab ID: 70326667035								
Collected: 12/06/24 06:37								
Received: 12/10/24 07:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	101	ug/L	1.0	1		12/17/24 15:19	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: 37-RM 139 SINK								
Lab ID: 70326667036								
Collected: 12/06/24 06:40								
Received: 12/10/24 07:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	4.6	ug/L	1.0	1		12/17/24 15:21	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: 38- RM 141 SINK								
Lab ID: 70326667037								
Collected: 12/06/24 06:39								
Received: 12/10/24 07:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	15.1	ug/L	1.0	1		12/17/24 15:25	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: 39- RM 140 SINK								
Lab ID: 70326667038								
Collected: 12/06/24 06:40								
Received: 12/10/24 07:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	29.3	ug/L	1.0	1		12/17/24 15:33	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 40-RM 113		Lab ID: 70326667039		Collected: 12/06/24 06:18	Received: 12/10/24 07:00	Matrix: Drinking Water		
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	12.8	ug/L	1.0	1		12/17/24 15:37	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 41-RM 121		Lab ID: 70326667040	Collected: 12/06/24 07:20	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	16.2	ug/L	1.0	1		12/17/24 15:39	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 42-RM 120		Lab ID: 70326667041	Collected: 12/06/24 07:18	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	17.6	ug/L	1.0	1		12/17/24 15:41	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 43-RM 115		Lab ID: 70326667042	Collected: 12/06/24 07:18	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	7.6	ug/L	1.0	1		12/17/24 15:42	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 44-RM 119		Lab ID: 70326667043	Collected: 12/06/24 07:16	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	26.2	ug/L	1.0	1		12/17/24 15:44	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 45-RM 116		Lab ID: 70326667044	Collected: 12/06/24 06:55	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	11.2	ug/L	1.0	1		12/17/24 15:45	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

Sample: 46-DF OUTSIDE GYM		Lab ID: 70326667045	Collected: 12/06/24 06:35	Received: 12/10/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/17/24 15:50	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

QC Batch:	375733	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70326667001, 70326667002, 70326667003, 70326667004, 70326667005, 70326667006, 70326667007, 70326667008, 70326667009, 70326667010, 70326667011, 70326667012, 70326667013, 70326667014, 70326667015, 70326667016		

METHOD BLANK:	1969011	Matrix:	Water
Associated Lab Samples:	70326667001, 70326667002, 70326667003, 70326667004, 70326667005, 70326667006, 70326667007, 70326667008, 70326667009, 70326667010, 70326667011, 70326667012, 70326667013, 70326667014, 70326667015, 70326667016		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	12/17/24 12:43	

LABORATORY CONTROL SAMPLE:	1969012					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	53.3	107	85-115	

MATRIX SPIKE SAMPLE:	1969014						
Parameter	Units	70327062008 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	55.2	110	70-130	

MATRIX SPIKE SAMPLE:	1969016						
Parameter	Units	70327062009 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	55.8	112	70-130	

SAMPLE DUPLICATE:	1969013					
Parameter	Units	70327062008 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	<1.0	<1.0			

SAMPLE DUPLICATE:	1969015					
Parameter	Units	70327062009 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	<1.0	<1.0			

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

QC Batch:	375782	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70326667017, 70326667018, 70326667019, 70326667020, 70326667021, 70326667022, 70326667023, 70326667024, 70326667025, 70326667026, 70326667027, 70326667028, 70326667029, 70326667030, 70326667031, 70326667032, 70326667033, 70326667034, 70326667035, 70326667036		

METHOD BLANK:	1969167	Matrix:	Water
Associated Lab Samples:	70326667017, 70326667018, 70326667019, 70326667020, 70326667021, 70326667022, 70326667023, 70326667024, 70326667025, 70326667026, 70326667027, 70326667028, 70326667029, 70326667030, 70326667031, 70326667032, 70326667033, 70326667034, 70326667035, 70326667036		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	12/17/24 14:36	

LABORATORY CONTROL SAMPLE:	1969168					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	53.9	108	85-115	

MATRIX SPIKE SAMPLE:	1969170						
Parameter	Units	70326667017 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	2.3	50	57.8	111	70-130	

MATRIX SPIKE SAMPLE:	1969172						
Parameter	Units	70326667018 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	1.2	50	57.2	112	70-130	

SAMPLE DUPLICATE:	1969169					
Parameter	Units	70326667017 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	2.3	2.3	0		

SAMPLE DUPLICATE:	1969171					
Parameter	Units	70326667018 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	1.2	1.2	0		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

QC Batch:	375784	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70326667037, 70326667038, 70326667039, 70326667040, 70326667041, 70326667042, 70326667043, 70326667044, 70326667045		

METHOD BLANK:	1969173	Matrix:	Water
Associated Lab Samples:	70326667037, 70326667038, 70326667039, 70326667040, 70326667041, 70326667042, 70326667043, 70326667044, 70326667045		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	12/17/24 15:22	

LABORATORY CONTROL SAMPLE: 1969174						
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	50.3	101	85-115	

MATRIX SPIKE SAMPLE: 1969176							
Parameter	Units	70326667037 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	15.1	50	80.0	130	70-130	

MATRIX SPIKE SAMPLE: 1969178							
Parameter	Units	70326667038 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	29.3	50	82.0	105	70-130	

SAMPLE DUPLICATE: 1969175						
Parameter	Units	70326667037 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	15.1	15.5	3		

SAMPLE DUPLICATE: 1969177						
Parameter	Units	70326667038 Result	Dup Result	RPD	Qualifiers	
Lead	ug/L	29.3	29.8	2		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALIFIERS

Project: OHIO ELEMENTARY

Pace Project No.: 70326667

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

TNTC - Too Numerous To Count

J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit - The lowest concentration value that meets project requirements for quantitative data with known precision and bias for a specific analyte in a specific matrix.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Reported results are not rounded until the final step prior to reporting. Therefore, calculated parameters that are typically reported as "Total" may vary slightly from the sum of the reported component parameters.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: OHIO ELEMENTARY

Pace Project No.: 70326667


Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70326667001	1 KITCHEN	EPA 200.8	375733		
70326667002	2 DPK OFFICE	EPA 200.8	375733		
70326667003	3-DF 139	EPA 200.8	375733		
70326667004	4-DF MEDIA CENTER	EPA 200.8	375733		
70326667005	5-NURSE	EPA 200.8	375733		
70326667006	6-NURSE SINK	EPA 200.8	375733		
70326667007	7-DF 121	EPA 200.8	375733		
70326667008	8-R 114	EPA 200.8	375733		
70326667009	9-R 103	EPA 200.8	375733		
70326667010	10-R 103 SINK	EPA 200.8	375733		
70326667011	11-R 112	EPA 200.8	375733		
70326667012	12-R 112 SINK	EPA 200.8	375733		
70326667013	13- R 104	EPA 200.8	375733		
70326667014	14- R 104 SINK	EPA 200.8	375733		
70326667015	15-R 111	EPA 200.8	375733		
70326667016	16-R 111 SINK	EPA 200.8	375733		
70326667017	17-R 105	EPA 200.8	375782		
70326667018	18-R 105 SINK	EPA 200.8	375782		
70326667019	19- R 110	EPA 200.8	375782		
70326667020	20-R 110 SINK	EPA 200.8	375782		
70326667021	21-R 106	EPA 200.8	375782		
70326667022	22-R 106 SINK	EPA 200.8	375782		
70326667023	23- R 109	EPA 200.8	375782		
70326667024	24- R 109 SINK	EPA 200.8	375782		
70326667025	25- R 107	EPA 200.8	375782		
70326667026	26- R 107 SINK	EPA 200.8	375782		
70326667027	27-R 108	EPA 200.8	375782		
70326667028	28-R 108 SINK	EPA 200.8	375782		
70326667029	29- RM 117 SINK	EPA 200.8	375782		
70326667030	31-RM 131	EPA 200.8	375782		
70326667031	32- RM 144 SINK	EPA 200.8	375782		
70326667032	33-RM 138 SINK	EPA 200.8	375782		
70326667033	34-RM 137 SINK	EPA 200.8	375782		
70326667034	35-RM 143 SINK	EPA 200.8	375782		
70326667035	36-RM 142 SINK	EPA 200.8	375782		
70326667036	37-RM 139 SINK	EPA 200.8	375782		
70326667037	38- RM 141 SINK	EPA 200.8	375784		
70326667038	39- RM 140 SINK	EPA 200.8	375784		
70326667039	40-RM 113	EPA 200.8	375784		
70326667040	41-RM 121	EPA 200.8	375784		
70326667041	42-RM 120	EPA 200.8	375784		
70326667042	43-RM 115	EPA 200.8	375784		
70326667043	44-RM 119	EPA 200.8	375784		
70326667044	45-RM 116	EPA 200.8	375784		
70326667045	46-DF OUTSIDE GYM	EPA 200.8	375784		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.

LAB USE ONLY - Affix WorkOrder/Login Label Here

WO#: 70326667



CHAIN-OF-CUSTODY Analytical Request Document
 Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Pace® Location Requested (City/State): Pace Analytical Long Island NY, 575 Broad Hollow Rd, Melville, NY 11747
 Company Name: Jefferson Lewis BOCES, Watertown CSD
 Street Address: 1351 Washington Street, Watertown NY 13601
 Contact/Report To: Brian Arias
 Phone #: 315-661-4351
 E-Mail: barias@watertowncso.org
 Cc E-Mail:
 Invoice To: Brian Arias
 Invoice E-Mail: barias@watertowncso.org
 Purchase Order # (if applicable):
 Quote #:

Customer Project #: 08215607
 Project Name: Ohio Elementary
 Site Collection Info/Facility ID (as applicable):
 Time Zone Collected: [] AK [] PT [] MT [] CT [] ET
 Data Deliverables: Regulatory Program (DW, RCRA, etc.) as applicable: NY Lead in School DW

Specify Container Size: **
 Identify Container Preservative Type: ***
 Analysis Requested:
 Proj. Mgr: Randy Budhu
 AcctNum / Client ID:
 Table #:
 Profile / Template:
 Prelog / Bottle Ord. ID:
 Sample Comment:
 Preservation non-conformance identified for sample:

** Container Size: (1) 1L, (2) 500mL, (3) 250mL, (4) 125mL, (5) 100mL, (6) 40mL vial, (7) Encore, (8) TerraCore, (9) Other
 *** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sod. Thiosulfate, (9) Ascorbic Acid, (10) MeDH, (11) Other

Customer Sample ID	Matrix *	Comp / Grab	Collected		Res. CL2	Number & Type of Containers	Field Filtered (if applicable): [] Yes [] No	Analysis:
			Date	Time				
1 - KITCHEN	DW	G	12/6/24	6:31		1	X	200.8 Drinking Water (Pb only)
2 - UPK OFFICE				6:35				
3 - DF 139				6:41				
4 - DF media Center				6:47				
5 - NURSE				6:50				
6 - NURSE SINK				7:19				
7 - DF 121				6:57				
8 - R 114				7:15				
9 - R 103				7:15				
10 - R 103 SINK								

Additional Instructions from Pace®:
 Collected By: Thomas Rankin
 Printed Name: Thomas Rankin
 Signature:
 Received by/Company (Signature):
 Date/Time: 12/5 11:15
 Tracking Number:
 Delivered by: [] In-Person [] Courier
 () FedEx [] UPS [] Other
 Page: of

LAB USE ONLY - Affix Workorder/Login Label Here

WO#: 70326667
PM: MC1 Due Date: **12/24/24**
CLIENT: WATERTOWN



CHAIN-OF-CUSTODY Analytical Request Document
 Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Specify Container Size **
 125mL (5) 100mL (6) 160mL Vial (7) Encore (8) TerraCore (9) Other
 Identify Container Preservative Type***
 *** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sod. Thiosulfate, (9) Ascorbic Acid, (10) MeOH, (11) Other
 Analysis Requested

Project Name: **05215607 Ohio Elk waterway**
 Site Collection Info/Facility ID (as applicable):
 Time Zone Collected: [] AK [] PT [] MT [] CT [] ET [X] ET
 Regulatory Program (DW, RCRA, etc.) as applicable: **NY Lead in School DW**

Proj. Mgr: **Randy Budhu**
 AcctNum / Client ID:
 Table #:
 Profile / Template:
 Prelog / Bottle Ord. ID:

Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk
 Date Results Requested: Standard 10 business day
 Field Filtered (if applicable): [] Yes [] No
 Analysis:

Customer Sample ID	Matrix *	Collected		Res. CL2	Number & Type of Containers	Sample Comment
		Comp / Grab	Time			
11 - R112	DW	G	12/6/24 6:58		1	X
12 - R112 sink			6:58			
13 - R104			7:14			
14 - R104 sink			7:14			
15 - R111			7:01			
16 - R111 sink			7:01			
17 - R105			7:12			
18 - R105 sink			7:12			
19 - R110			7:03			
20 - R110 sink			7:03			

Additional Instructions from Pace*:
 Collected By: **Thomas Rankin**
 Printed Name: **Thomas Rankin**
 Signature:
 Received by/Company: **James E. Cady**
 Date/Time: **12/9 12:00**
 Received by/Company: **James E. Cady**
 Date/Time: **12/9 12:00**
 Received by/Company: **James E. Cady**
 Date/Time: **12/10 7c**
 Received by/Company: **James E. Cady**
 Date/Time: **12/10 7c**

Company Name: Jefferson Lewis BOCES Watertown CSD
Street Address: 1351 Washington Street, Watertown NY 13601

Contact/Report To: Brian Arias
Phone #: 315-661-4351
E-Mail: barias@watertowncsd.org
Cc E-Mail:

Invoice To: Brian Arias
Invoice E-Mail: barias@watertowncsd.org

Customer Project #: 08215607
Project Name: CN10 elementary

Site Collection Info/Facility ID (as applicable):

Time Zone Collected: [] AK [] MT [] CT [X] ET
County / State origin of sample(s): New York

Data Deliverables:
 [] Level II [] Level III [] Level IV
 [] EQUIS
 [] Other

Regulatory Program (DW, RCRA, etc.) as applicable: NY Lead in School DW
Rush (Pre-approval required): DW PWSID # or WW Permit # as applicable:
 [] 2 Day [] 3 day [] 5 day [] Other

Date Results Requested: Standard 10 business day
 Field Filtered (if applicable): [] Yes [] No
 Analysis:

*** Matrix Codes (Insert in Matrix box below):** Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Blossay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Collected (or Composite Start)		Composite End	Res. CL2	Number & Type of Containers	
		Date	Time			Date	Time
32 - RM 144 SINK	DW	G	12/6/24	638			1
33 - RM 138 SINK			641				
34 - RM 137 SINK			647				
35 - RM 143 SINK			636				
36 - RM 142 SINK			637				
37 - RM 139 SINK			640				
38 - RM 141 SINK			639				
39 - RM 140 SINK			640				
40 - RM 113			618				
41 - RM 121			720				

Customer Remarks / Special Conditions / Possible Hazards:
 Lead

Additional Instructions from Pace®:

Coolers: Thermometer ID: Correction Factor (°C): Obs. Temp (°C) Corrected Temp. (°C)

Tracking Number: 1219 11:45

Delivered by: [] In-Person [] Courier
 [] FedEx [] UPS [] Other

Date/Time: 12/19/24 12:00
 Date/Time: 12/19/24 12:00
 Date/Time: 12/19/24 7:00

Received by/Company (Signature):
 Received by/Company (Signature):
 Received by/Company (Signature):
 Received by/Company (Signature):

Page 56 of 59

LAB USE ONLY - Affix Workorder/Login Label Here

Scan QR Code for instructions

** Container Size: (1) 1L, (2) 500mL, (3) 250mL, (4) 125mL, (5) 100mL, (6) 40mL Vial, (7) Encore, (8) TerraCore, (9) Other

*** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sod. Thiosulfate, (9) Ascorbic Acid, (10) MeOH, (11) Other

Specify Container Size **

Identify Container Preservative Type ***

Analysis Requested

Proj. Mgr: Randy Budhu
 AcctNum / Client ID:
 Table #:
 Profile / Template:
 Preleg / Bottle Ord. ID:
 MeOH:

Lab Use Only

Preservation non-conformance identified for

Sample Comment

CHAIN-OF-CUSTODY Analytical Request Document
 Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

LAB USE ONLY- Affix Workorder/Login Label Here



Scan QR Code for instructions

Company Name: **Jefferson Lewis BOCES, Watertown CSD**
 Street Address: **1351 Washington Street, Watertown NY 13601**

Contact/Report To: **Brian Arias**
 Phone #: **315-661-4351**
 E-Mail: **barias@watertowncsd.org**
 Cc E-Mail:

Invoice To: **Brian Arias**
 Invoice E-Mail: **barias@watertowncsd.org**

Purchase Order # (if applicable):
 Quote #:

County / State origin of sample(s): **New York**

Regulatory Program (DW, RCRA, etc.) as applicable: **NY Lead in School DW**
 Rush (Pre-approval required):
 DW PWSID # or WW Permit # as applicable:
 Field Filtered (if applicable): [] Yes [] No

Date Results Requested: **Standard 10 business day**

* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected (or Composite Start)		Composite End		Res. CL2	Number & Type of Containers		Sample Comment
			Date	Time	Date	Time		Plastic	Glass	
42 - Rm 120	DW	G	12/16/24	7:18				1		X
43 - Rm 115				7:18						
44 - Rm 119				7:16						
45 - Rm 116				6:50						
Ag - DF OUTSIDE ROOM				6:33						

Additional Instructions from Pace*:

Collected By: _____
 Printed Name: Thomas Rankin
 Signature: _____

Received by/Company: (Signature) *James E. Kelly*
 Date/Time: 12/19/24 11:45

Received by/Company: (Signature) _____
 Date/Time: 12/19/24 12:00

Received by/Company: (Signature) *James E. Kelly*
 Date/Time: 12/10/24 9:00

Received by/Company: (Signature) _____
 Date/Time: _____

Trucking Number: _____
 Delivered by: [] In-Person [] Courier
 [] FedEx [] UPS [] Other

Page: _____ of _____

WO#: 70326667

Client Name: Watertown

Project #

PM: MC1

Due Date: 12/24/24

Courier: Fed Ex UPS USPS Client Commercial Pac Other

CLIENT: WATERTOWN

Tracking #:

Custody Seal on Cooler/Box Present: Yes No Seals intact: Yes No Temperature Blank Present: Yes No
 Packing Material: Bubble Wrap Bubble Bags Ziplo Non Other Type of Ice: Wet Blue None

Thermometer Used: TH211 Correction Factor: 0.0 Samples on ice, cooling process has begun
 Cooler Temperature (°C): 19.0 Cooler Temperature Corrected (°C): 19.0 Date/Time 5035A kits placed in freezer _____

Temp should be above freezing to 5.0°C

USDA Regulated Soil (N/A, water sample)

Did samples originate in a quarantine zone within the United States: AL, AR, CA, FL, GA, ID, LA, MS, NC, NM, NY, OK, OR, SC, TN, TX, or VA (check map)? Yes No

Did samples originate from a foreign source including Hawaii and Puerto Rico? Yes No

If Yes to either question, fill out a Regulated Soil Checklist (ENV-FRM-MELV-0076) and include with SCUR/COC paperwork.

Date and Initials of person examining contents: 12/12/24 PD

		COMMENTS:
Chain of Custody Present:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	1.
Chain of Custody Filled Out:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	2.
Chain of Custody Relinquished:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	3.
Sampler Name & Signature on COC:	<input type="checkbox"/> Yes <input type="checkbox"/> No <u>N/A</u>	4.
Samples Arrived within Hold Time:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	5.
Short Hold Time Analysis (<72hr):	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	6.
Rush Turn Around Time Requested	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	7.
Sufficient Volume: (Triple volume provided for MS/MSD)	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	8.
Correct Containers Used:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	9.
-Pace Containers Used:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	10.
Containers Intact:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	11.
Filtered volume received for Dissolved tests	<input type="checkbox"/> Yes <input type="checkbox"/> No <u>N/A</u>	11. Note: if sediment is visible in the dissolved container.
Sample Labels match COC:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	12.
-Includes date/time/ID/Analysis Matrix:	SL WT OIL OTHER	

Date and Initials of person checking preservation: 12/12/24 PD

All containers needing preservation have been	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	13. <input type="checkbox"/> HNO ₃ <input type="checkbox"/> H ₂ SO ₄ <input type="checkbox"/> NaOH <input type="checkbox"/> HCl
pH paper Lot # <u>213624</u>		Sample #
All containers needing preservation are found to be in compliance with method recommendation? (HNO ₃ , H ₂ SO ₄ , HCl, NaOH > 9 Sulfide, NAOH > 12 Cyanide)	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	
Exceptions: VOA, Coliform, TOC/DOC, Oil and Grease, DRO/8015 (water).		Initial when completed: Lot # of added preservative: Date/Time preservative added:
Per Method, VOA pH is checked after analysis		
Samples checked for dechlorination:	<input type="checkbox"/> Yes <input type="checkbox"/> No <u>N/A</u>	14.
KI starch test strips Lot #		Positive for Res. Chlorine? Y N
Residual chlorine strips Lot #		15.
SM 4500 CN samples checked for sul	<input type="checkbox"/> Yes <input type="checkbox"/> No <u>N/A</u>	Positive for Sulfide? Y N
Lead Acetate Strips Lot #		
Headspace in ALK Bottle (>6mm):	<input type="checkbox"/> Yes <input type="checkbox"/> No <u>N/A</u>	16.
Headspace in VOA Vials (>6mm):	<input type="checkbox"/> Yes <input type="checkbox"/> No <u>N/A</u>	17.
Trip Blank Present:	<input type="checkbox"/> Yes <input type="checkbox"/> No <u>N/A</u>	
Trip Blank Custody Seals Present	<input type="checkbox"/> Yes <input type="checkbox"/> No <u>N/A</u>	

Client Notification/ Resolution:

Field Data Required? Y / N

Person Contacted:

Date/Time:

Comments/ Resolution:

* PM (Project Manager) review (which includes the SCUR) is documented electronically in LIMS.