



January 30, 2025

**Request for Proposal Clarification
25.06 AED Equipment & AED Storage Cabinets**

Addendum 1

QUESTIONS:

1. If needed, who would be the designated POC to address our proposal/quote to?
Answer: It is not necessary, but Rochester Community Schools Purchasing Department would be an appropriate option.
2. Is this line required to be fulfilled for all vendors? If so, please describe what will be needed to do for the in-person setup, locations, etc...?
Answer: You can decide to leave blank if you choose not to provide this service.
3. Do you still require a surety bond or certified check even though this bid is for providing equipment? If so, since this is an online bid, do we make a copy of the check and send the copy of the check online, or do we need to mail it?
Answer: Yes. A photocopy of the check or bond is acceptable, which can be uploaded with your bid documents.
4. Can you please provide more details re: what you're looking for re: "In person set-up for WiFi and AED"? Are you looking for someone to install the indoor and outdoor AED cabinets in the walls and connect the school WiFi to all 33 AEDs?
Answer: We are looking to have the units set in place in the cabinets and to test that they connect to the wi-fi and that they report to a system that alerts when a unit needs servicing. Our current vendor of all our existing AED's provides this service for us and is very beneficial for the district so we would like to have the same level of service provided for the newly purchased units.

END ADDENDUM