

ANALYTICAL REPORT

PREPARED FOR

Attn: Ian St. John
St. John Environmental Consulting
5515 Grace Ave.
Saint Louis, Missouri 63116

Generated 7/4/2024 2:59:25 AM

JOB DESCRIPTION

St. Joseph School District - Layfayette

JOB NUMBER

810-109594-1

Eurofins Eaton Analytical South Bend

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and will meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Eaton Analytical, LLC Project Manager.

Authorization



Generated
7/4/2024 2:59:25 AM

Authorized for release by
Pamela Brown, Project Manager
Pamela.Brown@et.eurofinsus.com
(574)233-4777



Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	12
QC Sample Results	22
QC Association Summary	26
Lab Chronicle	29
Certification Summary	39
Method Summary	40
Sample Summary	41
Chain of Custody	43
Receipt Checklists	50

Definitions/Glossary

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Qualifiers

Metals

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: St. John Environmental Consulting
Project: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Job ID: 810-109594-1

Eurofins Eaton Analytical South Bend

Job Narrative 810-109594-1

Analytical test results meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page unless otherwise noted under the individual analysis. Data qualifiers are applied to indicate exceptions. Noncompliant quality control (QC) is further explained in narrative comments.

- Matrix QC may not be reported if insufficient sample or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD may be performed, unless otherwise specified in the method.
- Surrogate and/or isotope dilution analyte recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in the narrative.

Regulated compliance samples (e.g. SDWA, NPDES) must comply with the associated agency requirements/permits.

Receipt

The samples were received on 6/26/2024 9:45 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice.

Receipt Exceptions

One or more containers for the following sample(s) was received empty:

SJL-25 WF
SJL-54 SINK
SJL-66 WF (Field House)

Metals

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Eurofins Eaton Analytical South Bend

Detection Summary

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Layfayette

Job ID: 810-109594-1

Client Sample ID: SJL - 1 Sink

Lab Sample ID: 810-109594-1

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	2.4		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 2 Sink

Lab Sample ID: 810-109594-2

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	0.88		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 3 WF

Lab Sample ID: 810-109594-3

No Detections.

Client Sample ID: SJL - 4 Sink

Lab Sample ID: 810-109594-4

No Detections.

Client Sample ID: SJL - 5 Sink

Lab Sample ID: 810-109594-5

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	1.1		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 6 Sink

Lab Sample ID: 810-109594-6

No Detections.

Client Sample ID: SJL - 7 Sink

Lab Sample ID: 810-109594-7

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	6.8		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 8 Sink

Lab Sample ID: 810-109594-8

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	19		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 9 Sink

Lab Sample ID: 810-109594-9

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	9.0		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 10 WF

Lab Sample ID: 810-109594-10

No Detections.

Client Sample ID: SJL - 11 Sink

Lab Sample ID: 810-109594-11

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	13		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 12 Sink

Lab Sample ID: 810-109594-12

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	29		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 13 Sink

Lab Sample ID: 810-109594-13

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	3.6		0.50		ug/L	1		200.8	Total/NA

This Detection Summary does not include radiochemical test results.

Eurofins Eaton Analytical South Bend

Detection Summary

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Layfayette

Job ID: 810-109594-1

Client Sample ID: SJL - 14 Sink

Lab Sample ID: 810-109594-14

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	1.3		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 15 Sink

Lab Sample ID: 810-109594-15

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	4.5		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 16 Sink

Lab Sample ID: 810-109594-16

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	87		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 17 Sink

Lab Sample ID: 810-109594-17

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	11		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 18 WF

Lab Sample ID: 810-109594-18

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	0.99		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 19 Sink

Lab Sample ID: 810-109594-19

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	2.6		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 20 Sink

Lab Sample ID: 810-109594-20

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	4.8		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 21 Sink

Lab Sample ID: 810-109594-21

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	4.2		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 22 WF

Lab Sample ID: 810-109594-22

No Detections.

Client Sample ID: SJL - 23 Sink

Lab Sample ID: 810-109594-23

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	4.8		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 24 WF (ROTC)

Lab Sample ID: 810-109594-24

No Detections.

Client Sample ID: SJL - 26 WF

Lab Sample ID: 810-109594-25

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	1.0		0.50		ug/L	1		200.8	Total/NA

This Detection Summary does not include radiochemical test results.

Eurofins Eaton Analytical South Bend

Detection Summary

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Client Sample ID: SJL - 27 WF

Lab Sample ID: 810-109594-26

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	0.86		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 28 WF

Lab Sample ID: 810-109594-27

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	0.63		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 29 Sink

Lab Sample ID: 810-109594-28

No Detections.

Client Sample ID: SJL - 30 WF

Lab Sample ID: 810-109594-29

No Detections.

Client Sample ID: SJL - 31 WF

Lab Sample ID: 810-109594-30

No Detections.

Client Sample ID: SJL - 32 WF

Lab Sample ID: 810-109594-31

No Detections.

Client Sample ID: SJL - 33 Sink

Lab Sample ID: 810-109594-32

No Detections.

Client Sample ID: SJL - 34 Sink

Lab Sample ID: 810-109594-33

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	2.8		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 35 Sink

Lab Sample ID: 810-109594-34

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	42		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 36 Sink

Lab Sample ID: 810-109594-35

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	2.9		0.50		ug/L	1		200.8	Total Recoverable

Client Sample ID: SJL - 37 Sink

Lab Sample ID: 810-109594-36

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	7.4		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 38 Sink

Lab Sample ID: 810-109594-37

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	24		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 39 Sink

Lab Sample ID: 810-109594-38

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	3.4		0.50		ug/L	1		200.8	Total/NA

This Detection Summary does not include radiochemical test results.

Eurofins Eaton Analytical South Bend

Detection Summary

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Client Sample ID: SJL - 40 Sink

Lab Sample ID: 810-109594-39

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	15		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 41 Sink

Lab Sample ID: 810-109594-40

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	70		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 42 Sink

Lab Sample ID: 810-109594-41

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	150		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 43 Sink

Lab Sample ID: 810-109594-42

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	140		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 44 WF

Lab Sample ID: 810-109594-43

No Detections.

Client Sample ID: SJL - 45 Sink

Lab Sample ID: 810-109594-44

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	170		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 46 Sink

Lab Sample ID: 810-109594-45

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	32		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 47 Sink

Lab Sample ID: 810-109594-46

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	30		0.50		ug/L	1		200.8	Total Recoverable

Client Sample ID: SJL - 48 Sink

Lab Sample ID: 810-109594-47

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	23		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 49 Sink

Lab Sample ID: 810-109594-48

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	32		0.50		ug/L	1		200.8	Total Recoverable

Client Sample ID: SJL - 50 Sink

Lab Sample ID: 810-109594-49

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	46		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 51 Sink

Lab Sample ID: 810-109594-50

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	6.2		0.50		ug/L	1		200.8	Total/NA

This Detection Summary does not include radiochemical test results.

Eurofins Eaton Analytical South Bend

Detection Summary

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Layfayette

Job ID: 810-109594-1

Client Sample ID: SJL - 52 Sink

Lab Sample ID: 810-109594-51

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	14		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 53 Sink

Lab Sample ID: 810-109594-52

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	4.8		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 55 Sink

Lab Sample ID: 810-109594-53

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	12		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 56 WF

Lab Sample ID: 810-109594-54

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	2.4		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 57 Sink

Lab Sample ID: 810-109594-55

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	85		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 58 WF

Lab Sample ID: 810-109594-56

No Detections.

Client Sample ID: SJL - 59 Sink

Lab Sample ID: 810-109594-57

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	2.2		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 60 ICE

Lab Sample ID: 810-109594-58

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	2.6		0.50		ug/L	1		200.8	Total Recoverable

Client Sample ID: SJL - 61 ICE

Lab Sample ID: 810-109594-59

No Detections.

Client Sample ID: SJL - 62 ICE

Lab Sample ID: 810-109594-60

No Detections.

Client Sample ID: SJL - 63 WF

Lab Sample ID: 810-109594-61

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	0.53		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL - 64 WF

Lab Sample ID: 810-109594-62

No Detections.

Client Sample ID: SJL - 65 WF (Field House)

Lab Sample ID: 810-109594-63

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Eaton Analytical South Bend

Detection Summary

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Layfayette

Job ID: 810-109594-1

Client Sample ID: SJL - 67 ICE (Main)

Lab Sample ID: 810-109594-64

No Detections.

1

2

3

4

5

6

7

8

9

10

11

12

13

14

This Detection Summary does not include radiochemical test results.

Eurofins Eaton Analytical South Bend

Client Sample Results

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Client Sample ID: SJL - 1 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-1

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	2.4		0.50		ug/L			07/02/24 11:45	1

Client Sample ID: SJL - 2 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-2

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	0.88		0.50		ug/L			07/02/24 11:48	1

Client Sample ID: SJL - 3 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-3

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 11:51	1

Client Sample ID: SJL - 4 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-4

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 11:53	1

Client Sample ID: SJL - 5 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-5

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	1.1		0.50		ug/L			07/02/24 11:56	1

Client Sample ID: SJL - 6 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-6

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 11:59	1

Client Sample ID: SJL - 7 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-7

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	6.8		0.50		ug/L			07/02/24 12:55	1

Client Sample Results

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Client Sample ID: SJL - 8 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-8

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	19		0.50		ug/L			07/02/24 13:00	1

Client Sample ID: SJL - 9 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-9

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	9.0		0.50		ug/L			07/02/24 13:02	1

Client Sample ID: SJL - 10 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-10

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 13:03	1

Client Sample ID: SJL - 11 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-11

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	13		0.50		ug/L			07/02/24 13:05	1

Client Sample ID: SJL - 12 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-12

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	29		0.50		ug/L			07/02/24 13:06	1

Client Sample ID: SJL - 13 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-13

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	3.6		0.50		ug/L			07/02/24 13:08	1

Client Sample ID: SJL - 14 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-14

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	1.3		0.50		ug/L			07/02/24 13:10	1

Eurofins Eaton Analytical South Bend

Client Sample Results

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Client Sample ID: SJL - 15 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-15

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	4.5		0.50		ug/L			07/02/24 13:11	1

Client Sample ID: SJL - 16 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-16

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	87		0.50		ug/L			07/02/24 13:13	1

Client Sample ID: SJL - 17 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-17

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	11		0.50		ug/L			07/02/24 13:18	1

Client Sample ID: SJL - 18 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-18

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	0.99		0.50		ug/L			07/02/24 13:20	1

Client Sample ID: SJL - 19 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-19

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	2.6		0.50		ug/L			07/02/24 13:21	1

Client Sample ID: SJL - 20 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-20

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	4.8		0.50		ug/L			07/02/24 13:26	1

Client Sample ID: SJL - 21 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-21

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	4.2		0.50		ug/L			07/02/24 13:28	1

Eurofins Eaton Analytical South Bend

Client Sample Results

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Client Sample ID: SJL - 22 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-22

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 13:29	1

Client Sample ID: SJL - 23 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-23

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	4.8		0.50		ug/L			07/02/24 13:31	1

Client Sample ID: SJL - 24 WF (ROTC)

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-24

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 13:33	1

Client Sample ID: SJL - 26 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-25

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	1.0		0.50		ug/L			07/02/24 13:34	1

Client Sample ID: SJL - 27 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-26

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	0.86		0.50		ug/L			07/02/24 13:36	1

Client Sample ID: SJL - 28 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-27

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	0.63		0.50		ug/L			07/02/24 13:37	1

Client Sample ID: SJL - 29 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-28

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 13:45	1

Client Sample Results

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Client Sample ID: SJL - 30 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-29

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 13:47	1

Client Sample ID: SJL - 31 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-30

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 13:50	1

Client Sample ID: SJL - 32 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-31

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 13:53	1

Client Sample ID: SJL - 33 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-32

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 13:55	1

Client Sample ID: SJL - 34 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-33

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	2.8		0.50		ug/L			07/02/24 13:58	1

Client Sample ID: SJL - 35 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-34

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	42		0.50		ug/L			07/02/24 14:01	1

Client Sample ID: SJL - 36 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-35

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS) - Total Recoverable

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	2.9		0.50		ug/L		07/02/24 12:38	07/03/24 13:25	1

Eurofins Eaton Analytical South Bend

Client Sample Results

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Client Sample ID: SJL - 37 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-36

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	7.4		0.50		ug/L			07/02/24 14:04	1

Client Sample ID: SJL - 38 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-37

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	24		0.50		ug/L			07/02/24 14:07	1

Client Sample ID: SJL - 39 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-38

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	3.4		0.50		ug/L			07/02/24 14:54	1

Client Sample ID: SJL - 40 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-39

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	15		0.50		ug/L			07/02/24 14:18	1

Client Sample ID: SJL - 41 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-40

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	70		0.50		ug/L			07/02/24 14:26	1

Client Sample ID: SJL - 42 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-41

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	150		0.50		ug/L			07/02/24 14:29	1

Client Sample ID: SJL - 43 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-42

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	140		0.50		ug/L			07/02/24 14:32	1

Eurofins Eaton Analytical South Bend

Client Sample Results

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Client Sample ID: SJL - 44 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-43

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 14:35	1

Client Sample ID: SJL - 45 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-44

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	170		0.50		ug/L			07/02/24 14:38	1

Client Sample ID: SJL - 46 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-45

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	32		0.50		ug/L			07/02/24 14:40	1

Client Sample ID: SJL - 47 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-46

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS) - Total Recoverable

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	30		0.50		ug/L		07/02/24 12:38	07/03/24 13:27	1

Client Sample ID: SJL - 48 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-47

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	23		0.50		ug/L			07/02/24 14:43	1

Client Sample ID: SJL - 49 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-48

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS) - Total Recoverable

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	32		0.50		ug/L		07/02/24 12:38	07/03/24 13:29	1

Client Sample ID: SJL - 50 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-49

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	46		0.50		ug/L			07/02/24 14:46	1

Eurofins Eaton Analytical South Bend

Client Sample Results

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Client Sample ID: SJL - 51 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-50

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	6.2		0.50		ug/L			07/02/24 15:08	1

Client Sample ID: SJL - 52 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-51

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	14		0.50		ug/L			07/02/24 15:11	1

Client Sample ID: SJL - 53 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-52

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	4.8		0.50		ug/L			07/02/24 15:14	1

Client Sample ID: SJL - 55 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-53

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	12		0.50		ug/L			07/02/24 15:23	1

Client Sample ID: SJL - 56 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-54

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	2.4		0.50		ug/L			07/02/24 15:25	1

Client Sample ID: SJL - 57 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-55

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	85		0.50		ug/L			07/02/24 15:28	1

Client Sample ID: SJL - 58 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-56

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 15:31	1

Client Sample Results

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Client Sample ID: SJL - 59 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-57

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	2.2		0.50		ug/L			07/02/24 16:28	1

Client Sample ID: SJL - 60 ICE

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-58

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS) - Total Recoverable

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	2.6		0.50		ug/L		07/02/24 12:38	07/03/24 13:31	1

Client Sample ID: SJL - 61 ICE

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-59

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 15:37	1

Client Sample ID: SJL - 62 ICE

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-60

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 15:40	1

Client Sample ID: SJL - 63 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-61

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	0.53		0.50		ug/L			07/02/24 15:22	1

Client Sample ID: SJL - 64 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-62

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 15:27	1

Client Sample ID: SJL - 65 WF (Field House)

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-63

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 15:29	1

Client Sample Results

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Layfayette

Job ID: 810-109594-1

Client Sample ID: SJL - 67 ICE (Main)

Lab Sample ID: 810-109594-64

Date Collected: 06/25/24 08:00

Matrix: Drinking Water

Date Received: 06/26/24 09:45

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 15:31	1

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

QC Sample Results

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Method: 200.8 - Metals (ICP/MS)

Lab Sample ID: MB 810-104642/13

Matrix: Drinking Water

Analysis Batch: 104642

Client Sample ID: Method Blank

Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 09:26	1

Lab Sample ID: MB 810-104642/43

Matrix: Drinking Water

Analysis Batch: 104642

Client Sample ID: Method Blank

Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 10:47	1

Lab Sample ID: LCS 810-104642/44

Matrix: Drinking Water

Analysis Batch: 104642

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Lead	50.0	49.4		ug/L		99	85 - 115

Lab Sample ID: LLCS 810-104642/11

Matrix: Drinking Water

Analysis Batch: 104642

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Analyte	Spike Added	LLCS Result	LLCS Qualifier	Unit	D	%Rec	%Rec Limits
Lead	0.300	0.272	J	ug/L		91	50 - 150

Lab Sample ID: MB 810-104689/11

Matrix: Drinking Water

Analysis Batch: 104689

Client Sample ID: Method Blank

Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 12:03	1

Lab Sample ID: MB 810-104689/131

Matrix: Drinking Water

Analysis Batch: 104689

Client Sample ID: Method Blank

Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 15:19	1

Lab Sample ID: MB 810-104689/41

Matrix: Drinking Water

Analysis Batch: 104689

Client Sample ID: Method Blank

Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 12:52	1

Lab Sample ID: LCS 810-104689/132

Matrix: Drinking Water

Analysis Batch: 104689

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Lead	50.0	49.7		ug/L		99	85 - 115

Eurofins Eaton Analytical South Bend

QC Sample Results

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Layfayette

Job ID: 810-109594-1

Method: 200.8 - Metals (ICP/MS)

Lab Sample ID: LCS 810-104689/42

Matrix: Drinking Water

Analysis Batch: 104689

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Lead	50.0	48.7		ug/L		97	85 - 115

Lab Sample ID: LLCS 810-104689/10

Matrix: Drinking Water

Analysis Batch: 104689

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Analyte	Spike Added	LLCS Result	LLCS Qualifier	Unit	D	%Rec	%Rec Limits
Lead	0.300	0.275	J	ug/L		92	50 - 150

Lab Sample ID: 810-109594-7 MS

Matrix: Drinking Water

Analysis Batch: 104689

Client Sample ID: SJL - 7 Sink

Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Lead	6.8		50.0	57.7		ug/L		102	70 - 130

Lab Sample ID: 810-109594-7 MSD

Matrix: Drinking Water

Analysis Batch: 104689

Client Sample ID: SJL - 7 Sink

Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Lead	6.8		50.0	57.2		ug/L		101	70 - 130	1	20

Lab Sample ID: 810-109594-19 MS

Matrix: Drinking Water

Analysis Batch: 104689

Client Sample ID: SJL - 19 Sink

Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Lead	2.6		50.0	54.8		ug/L		104	70 - 130

Lab Sample ID: 810-109594-19 MSD

Matrix: Drinking Water

Analysis Batch: 104689

Client Sample ID: SJL - 19 Sink

Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Lead	2.6		50.0	55.0		ug/L		105	70 - 130	0	20

Lab Sample ID: 810-109594-61 MS

Matrix: Drinking Water

Analysis Batch: 104689

Client Sample ID: SJL - 63 WF

Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Lead	0.53		50.0	54.2		ug/L		107	70 - 130

Lab Sample ID: 810-109594-61 MSD

Matrix: Drinking Water

Analysis Batch: 104689

Client Sample ID: SJL - 63 WF

Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Lead	0.53		50.0	53.7		ug/L		106	70 - 130	1	20

Eurofins Eaton Analytical South Bend

QC Sample Results

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Method: 200.8 - Metals (ICP/MS)

Lab Sample ID: MB 810-104696/11
Matrix: Drinking Water
Analysis Batch: 104696

Client Sample ID: Method Blank
Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 13:29	1

Lab Sample ID: MB 810-104696/45
Matrix: Drinking Water
Analysis Batch: 104696

Client Sample ID: Method Blank
Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			07/02/24 15:03	1

Lab Sample ID: LCS 810-104696/13
Matrix: Drinking Water
Analysis Batch: 104696

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Lead	50.0	48.7		ug/L		97	85 - 115

Lab Sample ID: LCS 810-104696/46
Matrix: Drinking Water
Analysis Batch: 104696

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Lead	50.0	49.0		ug/L		98	85 - 115

Lab Sample ID: LLCS 810-104696/12
Matrix: Drinking Water
Analysis Batch: 104696

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	LLCS Result	LLCS Qualifier	Unit	D	%Rec	%Rec Limits
Lead	0.300	0.281	J	ug/L		94	50 - 150

Lab Sample ID: 810-109594-27 MS
Matrix: Drinking Water
Analysis Batch: 104696

Client Sample ID: SJL - 28 WF
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Lead	0.63		50.0	48.7		ug/L		96	70 - 130

Lab Sample ID: 810-109594-27 MSD
Matrix: Drinking Water
Analysis Batch: 104696

Client Sample ID: SJL - 28 WF
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD Limit
Lead	0.63		50.0	48.5		ug/L		96	70 - 130	0 20

Lab Sample ID: 810-109594-39 MS
Matrix: Drinking Water
Analysis Batch: 104696

Client Sample ID: SJL - 40 Sink
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Lead	15		50.0	62.7		ug/L		95	70 - 130

Eurofins Eaton Analytical South Bend

QC Sample Results

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Layfayette

Job ID: 810-109594-1

Method: 200.8 - Metals (ICP/MS)

Lab Sample ID: 810-109594-39 MSD

Matrix: Drinking Water

Analysis Batch: 104696

Client Sample ID: SJL - 40 Sink

Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Lead	15		50.0	64.2		ug/L		98	70 - 130	2	20

Lab Sample ID: 810-109594-52 MS

Matrix: Drinking Water

Analysis Batch: 104696

Client Sample ID: SJL - 53 Sink

Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits		
Lead	4.8		50.0	53.4		ug/L		97	70 - 130		

Lab Sample ID: 810-109594-52 MSD

Matrix: Drinking Water

Analysis Batch: 104696

Client Sample ID: SJL - 53 Sink

Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Lead	4.8		50.0	53.7		ug/L		98	70 - 130	1	20

Lab Sample ID: MB 810-104650/1-A

Matrix: Drinking Water

Analysis Batch: 104792

Client Sample ID: Method Blank

Prep Type: Total Recoverable

Prep Batch: 104650

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L		07/02/24 12:38	07/03/24 13:10	1

Lab Sample ID: LCS 810-104650/3-A

Matrix: Drinking Water

Analysis Batch: 104792

Client Sample ID: Lab Control Sample

Prep Type: Total Recoverable

Prep Batch: 104650

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits		
Lead	50.0	53.2		ug/L		106	85 - 115		

Lab Sample ID: LLCS 810-104650/2-A

Matrix: Drinking Water

Analysis Batch: 104792

Client Sample ID: Lab Control Sample

Prep Type: Total Recoverable

Prep Batch: 104650

Analyte	Spike Added	LLCS Result	LLCS Qualifier	Unit	D	%Rec	%Rec Limits		
Lead	0.300	0.275	J	ug/L		92	50 - 150		

QC Association Summary

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Metals

Analysis Batch: 104642

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
810-109594-1	SJL - 1 Sink	Total/NA	Drinking Water	200.8	
810-109594-2	SJL - 2 Sink	Total/NA	Drinking Water	200.8	
810-109594-3	SJL - 3 WF	Total/NA	Drinking Water	200.8	
810-109594-4	SJL - 4 Sink	Total/NA	Drinking Water	200.8	
810-109594-5	SJL - 5 Sink	Total/NA	Drinking Water	200.8	
810-109594-6	SJL - 6 Sink	Total/NA	Drinking Water	200.8	
MB 810-104642/13	Method Blank	Total/NA	Drinking Water	200.8	
MB 810-104642/43	Method Blank	Total/NA	Drinking Water	200.8	
LCS 810-104642/44	Lab Control Sample	Total/NA	Drinking Water	200.8	
LLCS 810-104642/11	Lab Control Sample	Total/NA	Drinking Water	200.8	

Prep Batch: 104650

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
810-109594-35	SJL - 36 Sink	Total Recoverable	Drinking Water	200.8	
810-109594-46	SJL - 47 Sink	Total Recoverable	Drinking Water	200.8	
810-109594-48	SJL - 49 Sink	Total Recoverable	Drinking Water	200.8	
810-109594-58	SJL - 60 ICE	Total Recoverable	Drinking Water	200.8	
MB 810-104650/1-A	Method Blank	Total Recoverable	Drinking Water	200.8	
LCS 810-104650/3-A	Lab Control Sample	Total Recoverable	Drinking Water	200.8	
LLCS 810-104650/2-A	Lab Control Sample	Total Recoverable	Drinking Water	200.8	

Analysis Batch: 104689

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
810-109594-7	SJL - 7 Sink	Total/NA	Drinking Water	200.8	
810-109594-8	SJL - 8 Sink	Total/NA	Drinking Water	200.8	
810-109594-9	SJL - 9 Sink	Total/NA	Drinking Water	200.8	
810-109594-10	SJL - 10 WF	Total/NA	Drinking Water	200.8	
810-109594-11	SJL - 11 Sink	Total/NA	Drinking Water	200.8	
810-109594-12	SJL - 12 Sink	Total/NA	Drinking Water	200.8	
810-109594-13	SJL - 13 Sink	Total/NA	Drinking Water	200.8	
810-109594-14	SJL - 14 Sink	Total/NA	Drinking Water	200.8	
810-109594-15	SJL - 15 Sink	Total/NA	Drinking Water	200.8	
810-109594-16	SJL - 16 Sink	Total/NA	Drinking Water	200.8	
810-109594-17	SJL - 17 Sink	Total/NA	Drinking Water	200.8	
810-109594-18	SJL - 18 WF	Total/NA	Drinking Water	200.8	
810-109594-19	SJL - 19 Sink	Total/NA	Drinking Water	200.8	
810-109594-20	SJL - 20 Sink	Total/NA	Drinking Water	200.8	
810-109594-21	SJL - 21 Sink	Total/NA	Drinking Water	200.8	
810-109594-22	SJL - 22 WF	Total/NA	Drinking Water	200.8	
810-109594-23	SJL - 23 Sink	Total/NA	Drinking Water	200.8	
810-109594-24	SJL - 24 WF (ROTC)	Total/NA	Drinking Water	200.8	
810-109594-25	SJL - 26 WF	Total/NA	Drinking Water	200.8	
810-109594-26	SJL - 27 WF	Total/NA	Drinking Water	200.8	
810-109594-61	SJL - 63 WF	Total/NA	Drinking Water	200.8	
810-109594-62	SJL - 64 WF	Total/NA	Drinking Water	200.8	
810-109594-63	SJL - 65 WF (Field House)	Total/NA	Drinking Water	200.8	
810-109594-64	SJL - 67 ICE (Main)	Total/NA	Drinking Water	200.8	
MB 810-104689/11	Method Blank	Total/NA	Drinking Water	200.8	
MB 810-104689/131	Method Blank	Total/NA	Drinking Water	200.8	
MB 810-104689/41	Method Blank	Total/NA	Drinking Water	200.8	
LCS 810-104689/132	Lab Control Sample	Total/NA	Drinking Water	200.8	

Eurofins Eaton Analytical South Bend

QC Association Summary

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Metals (Continued)

Analysis Batch: 104689 (Continued)

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
LCS 810-104689/42	Lab Control Sample	Total/NA	Drinking Water	200.8	
LLCS 810-104689/10	Lab Control Sample	Total/NA	Drinking Water	200.8	
810-109594-7 MS	SJL - 7 Sink	Total/NA	Drinking Water	200.8	
810-109594-7 MSD	SJL - 7 Sink	Total/NA	Drinking Water	200.8	
810-109594-19 MS	SJL - 19 Sink	Total/NA	Drinking Water	200.8	
810-109594-19 MSD	SJL - 19 Sink	Total/NA	Drinking Water	200.8	
810-109594-61 MS	SJL - 63 WF	Total/NA	Drinking Water	200.8	
810-109594-61 MSD	SJL - 63 WF	Total/NA	Drinking Water	200.8	

Analysis Batch: 104696

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
810-109594-27	SJL - 28 WF	Total/NA	Drinking Water	200.8	
810-109594-28	SJL - 29 Sink	Total/NA	Drinking Water	200.8	
810-109594-29	SJL - 30 WF	Total/NA	Drinking Water	200.8	
810-109594-30	SJL - 31 WF	Total/NA	Drinking Water	200.8	
810-109594-31	SJL - 32 WF	Total/NA	Drinking Water	200.8	
810-109594-32	SJL - 33 Sink	Total/NA	Drinking Water	200.8	
810-109594-33	SJL - 34 Sink	Total/NA	Drinking Water	200.8	
810-109594-34	SJL - 35 Sink	Total/NA	Drinking Water	200.8	
810-109594-36	SJL - 37 Sink	Total/NA	Drinking Water	200.8	
810-109594-37	SJL - 38 Sink	Total/NA	Drinking Water	200.8	
810-109594-38	SJL - 39 Sink	Total/NA	Drinking Water	200.8	
810-109594-39	SJL - 40 Sink	Total/NA	Drinking Water	200.8	
810-109594-40	SJL - 41 Sink	Total/NA	Drinking Water	200.8	
810-109594-41	SJL - 42 Sink	Total/NA	Drinking Water	200.8	
810-109594-42	SJL - 43 Sink	Total/NA	Drinking Water	200.8	
810-109594-43	SJL - 44 WF	Total/NA	Drinking Water	200.8	
810-109594-44	SJL - 45 Sink	Total/NA	Drinking Water	200.8	
810-109594-45	SJL - 46 Sink	Total/NA	Drinking Water	200.8	
810-109594-47	SJL - 48 Sink	Total/NA	Drinking Water	200.8	
810-109594-49	SJL - 50 Sink	Total/NA	Drinking Water	200.8	
810-109594-50	SJL - 51 Sink	Total/NA	Drinking Water	200.8	
810-109594-51	SJL - 52 Sink	Total/NA	Drinking Water	200.8	
810-109594-52	SJL - 53 Sink	Total/NA	Drinking Water	200.8	
810-109594-53	SJL - 55 Sink	Total/NA	Drinking Water	200.8	
810-109594-54	SJL - 56 WF	Total/NA	Drinking Water	200.8	
810-109594-55	SJL - 57 Sink	Total/NA	Drinking Water	200.8	
810-109594-56	SJL - 58 WF	Total/NA	Drinking Water	200.8	
810-109594-57	SJL - 59 Sink	Total/NA	Drinking Water	200.8	
810-109594-59	SJL - 61 ICE	Total/NA	Drinking Water	200.8	
810-109594-60	SJL - 62 ICE	Total/NA	Drinking Water	200.8	
MB 810-104696/11	Method Blank	Total/NA	Drinking Water	200.8	
MB 810-104696/45	Method Blank	Total/NA	Drinking Water	200.8	
LCS 810-104696/13	Lab Control Sample	Total/NA	Drinking Water	200.8	
LCS 810-104696/46	Lab Control Sample	Total/NA	Drinking Water	200.8	
LLCS 810-104696/12	Lab Control Sample	Total/NA	Drinking Water	200.8	
810-109594-27 MS	SJL - 28 WF	Total/NA	Drinking Water	200.8	
810-109594-27 MSD	SJL - 28 WF	Total/NA	Drinking Water	200.8	
810-109594-39 MS	SJL - 40 Sink	Total/NA	Drinking Water	200.8	
810-109594-39 MSD	SJL - 40 Sink	Total/NA	Drinking Water	200.8	
810-109594-52 MS	SJL - 53 Sink	Total/NA	Drinking Water	200.8	

Eurofins Eaton Analytical South Bend

QC Association Summary

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Layfayette

Job ID: 810-109594-1

Metals (Continued)

Analysis Batch: 104696 (Continued)

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
810-109594-52 MSD	SJL - 53 Sink	Total/NA	Drinking Water	200.8	

Analysis Batch: 104792

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
810-109594-35	SJL - 36 Sink	Total Recoverable	Drinking Water	200.8	104650
810-109594-46	SJL - 47 Sink	Total Recoverable	Drinking Water	200.8	104650
810-109594-48	SJL - 49 Sink	Total Recoverable	Drinking Water	200.8	104650
810-109594-58	SJL - 60 ICE	Total Recoverable	Drinking Water	200.8	104650
MB 810-104650/1-A	Method Blank	Total Recoverable	Drinking Water	200.8	104650
LCS 810-104650/3-A	Lab Control Sample	Total Recoverable	Drinking Water	200.8	104650
LLCS 810-104650/2-A	Lab Control Sample	Total Recoverable	Drinking Water	200.8	104650

Lab Chronicle

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Client Sample ID: SJL - 1 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-1

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104642	CA	EA SB	07/02/24 11:45

Client Sample ID: SJL - 2 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-2

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104642	CA	EA SB	07/02/24 11:48

Client Sample ID: SJL - 3 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-3

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104642	CA	EA SB	07/02/24 11:51

Client Sample ID: SJL - 4 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-4

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104642	CA	EA SB	07/02/24 11:53

Client Sample ID: SJL - 5 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-5

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104642	CA	EA SB	07/02/24 11:56

Client Sample ID: SJL - 6 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-6

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104642	CA	EA SB	07/02/24 11:59

Client Sample ID: SJL - 7 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-7

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 12:55

Eurofins Eaton Analytical South Bend

Lab Chronicle

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Client Sample ID: SJL - 8 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-8

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 13:00

Client Sample ID: SJL - 9 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-9

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 13:02

Client Sample ID: SJL - 10 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-10

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 13:03

Client Sample ID: SJL - 11 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-11

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 13:05

Client Sample ID: SJL - 12 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-12

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 13:06

Client Sample ID: SJL - 13 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-13

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 13:08

Client Sample ID: SJL - 14 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-14

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 13:10

Eurofins Eaton Analytical South Bend

Lab Chronicle

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Client Sample ID: SJL - 15 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-15

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 13:11

Client Sample ID: SJL - 16 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-16

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 13:13

Client Sample ID: SJL - 17 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-17

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 13:18

Client Sample ID: SJL - 18 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-18

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 13:20

Client Sample ID: SJL - 19 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-19

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 13:21

Client Sample ID: SJL - 20 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-20

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 13:26

Client Sample ID: SJL - 21 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-21

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 13:28

Eurofins Eaton Analytical South Bend

Lab Chronicle

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Layfayette

Job ID: 810-109594-1

Client Sample ID: SJL - 22 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-22

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 13:29

Client Sample ID: SJL - 23 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-23

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 13:31

Client Sample ID: SJL - 24 WF (ROTC)

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-24

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 13:33

Client Sample ID: SJL - 26 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-25

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 13:34

Client Sample ID: SJL - 27 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-26

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 13:36

Client Sample ID: SJL - 28 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-27

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 13:37

Client Sample ID: SJL - 29 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-28

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 13:45

Eurofins Eaton Analytical South Bend

Lab Chronicle

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Client Sample ID: SJL - 30 WF

Lab Sample ID: 810-109594-29

Date Collected: 06/25/24 08:00

Matrix: Drinking Water

Date Received: 06/26/24 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 13:47

Client Sample ID: SJL - 31 WF

Lab Sample ID: 810-109594-30

Date Collected: 06/25/24 08:00

Matrix: Drinking Water

Date Received: 06/26/24 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 13:50

Client Sample ID: SJL - 32 WF

Lab Sample ID: 810-109594-31

Date Collected: 06/25/24 08:00

Matrix: Drinking Water

Date Received: 06/26/24 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 13:53

Client Sample ID: SJL - 33 Sink

Lab Sample ID: 810-109594-32

Date Collected: 06/25/24 08:00

Matrix: Drinking Water

Date Received: 06/26/24 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 13:55

Client Sample ID: SJL - 34 Sink

Lab Sample ID: 810-109594-33

Date Collected: 06/25/24 08:00

Matrix: Drinking Water

Date Received: 06/26/24 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 13:58

Client Sample ID: SJL - 35 Sink

Lab Sample ID: 810-109594-34

Date Collected: 06/25/24 08:00

Matrix: Drinking Water

Date Received: 06/26/24 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 14:01

Client Sample ID: SJL - 36 Sink

Lab Sample ID: 810-109594-35

Date Collected: 06/25/24 08:00

Matrix: Drinking Water

Date Received: 06/26/24 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total Recoverable	Prep	200.8			104650	NB	EA SB	07/02/24 12:38
Total Recoverable	Analysis	200.8		1	104792	NB	EA SB	07/03/24 13:25

Eurofins Eaton Analytical South Bend

Lab Chronicle

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Layfayette

Job ID: 810-109594-1

Client Sample ID: SJL - 37 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-36

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 14:04

Client Sample ID: SJL - 38 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-37

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 14:07

Client Sample ID: SJL - 39 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-38

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 14:54

Client Sample ID: SJL - 40 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-39

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 14:18

Client Sample ID: SJL - 41 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-40

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 14:26

Client Sample ID: SJL - 42 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-41

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 14:29

Client Sample ID: SJL - 43 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-42

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 14:32

Eurofins Eaton Analytical South Bend

Lab Chronicle

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Layfayette

Job ID: 810-109594-1

Client Sample ID: SJL - 44 WF

Lab Sample ID: 810-109594-43

Date Collected: 06/25/24 08:00

Matrix: Drinking Water

Date Received: 06/26/24 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 14:35

Client Sample ID: SJL - 45 Sink

Lab Sample ID: 810-109594-44

Date Collected: 06/25/24 08:00

Matrix: Drinking Water

Date Received: 06/26/24 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 14:38

Client Sample ID: SJL - 46 Sink

Lab Sample ID: 810-109594-45

Date Collected: 06/25/24 08:00

Matrix: Drinking Water

Date Received: 06/26/24 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 14:40

Client Sample ID: SJL - 47 Sink

Lab Sample ID: 810-109594-46

Date Collected: 06/25/24 08:00

Matrix: Drinking Water

Date Received: 06/26/24 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total Recoverable	Prep	200.8			104650	NB	EA SB	07/02/24 12:38
Total Recoverable	Analysis	200.8		1	104792	NB	EA SB	07/03/24 13:27

Client Sample ID: SJL - 48 Sink

Lab Sample ID: 810-109594-47

Date Collected: 06/25/24 08:00

Matrix: Drinking Water

Date Received: 06/26/24 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 14:43

Client Sample ID: SJL - 49 Sink

Lab Sample ID: 810-109594-48

Date Collected: 06/25/24 08:00

Matrix: Drinking Water

Date Received: 06/26/24 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total Recoverable	Prep	200.8			104650	NB	EA SB	07/02/24 12:38
Total Recoverable	Analysis	200.8		1	104792	NB	EA SB	07/03/24 13:29

Client Sample ID: SJL - 50 Sink

Lab Sample ID: 810-109594-49

Date Collected: 06/25/24 08:00

Matrix: Drinking Water

Date Received: 06/26/24 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 14:46

Eurofins Eaton Analytical South Bend

Lab Chronicle

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Client Sample ID: SJL - 51 Sink

Lab Sample ID: 810-109594-50

Date Collected: 06/25/24 08:00

Matrix: Drinking Water

Date Received: 06/26/24 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 15:08

Client Sample ID: SJL - 52 Sink

Lab Sample ID: 810-109594-51

Date Collected: 06/25/24 08:00

Matrix: Drinking Water

Date Received: 06/26/24 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 15:11

Client Sample ID: SJL - 53 Sink

Lab Sample ID: 810-109594-52

Date Collected: 06/25/24 08:00

Matrix: Drinking Water

Date Received: 06/26/24 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 15:14

Client Sample ID: SJL - 55 Sink

Lab Sample ID: 810-109594-53

Date Collected: 06/25/24 08:00

Matrix: Drinking Water

Date Received: 06/26/24 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 15:23

Client Sample ID: SJL - 56 WF

Lab Sample ID: 810-109594-54

Date Collected: 06/25/24 08:00

Matrix: Drinking Water

Date Received: 06/26/24 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 15:25

Client Sample ID: SJL - 57 Sink

Lab Sample ID: 810-109594-55

Date Collected: 06/25/24 08:00

Matrix: Drinking Water

Date Received: 06/26/24 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 15:28

Client Sample ID: SJL - 58 WF

Lab Sample ID: 810-109594-56

Date Collected: 06/25/24 08:00

Matrix: Drinking Water

Date Received: 06/26/24 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 15:31

Eurofins Eaton Analytical South Bend

Lab Chronicle

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Client Sample ID: SJL - 59 Sink

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-57

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 16:28

Client Sample ID: SJL - 60 ICE

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-58

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total Recoverable	Prep	200.8			104650	NB	EA SB	07/02/24 12:38
Total Recoverable	Analysis	200.8		1	104792	NB	EA SB	07/03/24 13:31

Client Sample ID: SJL - 61 ICE

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-59

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 15:37

Client Sample ID: SJL - 62 ICE

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-60

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104696	NB	EA SB	07/02/24 15:40

Client Sample ID: SJL - 63 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-61

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 15:22

Client Sample ID: SJL - 64 WF

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-62

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 15:27

Client Sample ID: SJL - 65 WF (Field House)

Date Collected: 06/25/24 08:00

Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-63

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 15:29

Eurofins Eaton Analytical South Bend

Lab Chronicle

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Layfayette

Job ID: 810-109594-1

Client Sample ID: SJL - 67 ICE (Main)
Date Collected: 06/25/24 08:00
Date Received: 06/26/24 09:45

Lab Sample ID: 810-109594-64
Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	104689	CA	EA SB	07/02/24 15:31

Laboratory References:
EA SB = Eurofins Eaton Analytical South Bend, 110 S Hill Street, South Bend, IN 46617, TEL (574)233-4777



Accreditation/Certification Summary

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Layfayette

Job ID: 810-109594-1

Laboratory: Eurofins Eaton Analytical South Bend

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Missouri	State	880	09-30-24

1
2
3
4
5
6
7
8
9
10
11
12
13
14

Method Summary

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Layfayette

Job ID: 810-109594-1

Method	Method Description	Protocol	Laboratory
200.8	Metals (ICP/MS)	EPA	EA SB
200.8	Preparation, Total Recoverable Metals	EPA	EA SB

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA SB = Eurofins Eaton Analytical South Bend, 110 S Hill Street, South Bend, IN 46617, TEL (574)233-4777

Sample Summary

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-109594-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
810-109594-1	SJL - 1 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-2	SJL - 2 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-3	SJL - 3 WF	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-4	SJL - 4 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-5	SJL - 5 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-6	SJL - 6 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-7	SJL - 7 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-8	SJL - 8 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-9	SJL - 9 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-10	SJL - 10 WF	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-11	SJL - 11 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-12	SJL - 12 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-13	SJL - 13 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-14	SJL - 14 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-15	SJL - 15 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-16	SJL - 16 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-17	SJL - 17 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-18	SJL - 18 WF	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-19	SJL - 19 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-20	SJL - 20 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-21	SJL - 21 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-22	SJL - 22 WF	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-23	SJL - 23 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-24	SJL - 24 WF (ROTC)	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-25	SJL - 26 WF	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-26	SJL - 27 WF	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-27	SJL - 28 WF	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-28	SJL - 29 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-29	SJL - 30 WF	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-30	SJL - 31 WF	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-31	SJL - 32 WF	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-32	SJL - 33 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-33	SJL - 34 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-34	SJL - 35 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-35	SJL - 36 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-36	SJL - 37 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-37	SJL - 38 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-38	SJL - 39 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-39	SJL - 40 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-40	SJL - 41 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-41	SJL - 42 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-42	SJL - 43 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-43	SJL - 44 WF	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-44	SJL - 45 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-45	SJL - 46 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-46	SJL - 47 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-47	SJL - 48 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-48	SJL - 49 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-49	SJL - 50 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-50	SJL - 51 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-51	SJL - 52 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-52	SJL - 53 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-53	SJL - 55 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-54	SJL - 56 WF	Drinking Water	06/25/24 08:00	06/26/24 09:45

Sample Summary

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Layfayette

Job ID: 810-109594-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
810-109594-55	SJL - 57 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-56	SJL - 58 WF	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-57	SJL - 59 Sink	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-58	SJL - 60 ICE	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-59	SJL - 61 ICE	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-60	SJL - 62 ICE	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-61	SJL - 63 WF	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-62	SJL - 64 WF	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-63	SJL - 65 WF (Field House)	Drinking Water	06/25/24 08:00	06/26/24 09:45
810-109594-64	SJL - 67 ICE (Main)	Drinking Water	06/25/24 08:00	06/26/24 09:45

Eurofins Eaton Analytical: South Bend
110 South Hill Street
South Bend, IN 46617
Phone (574) 233-4777

Chain of Custody Record



Environment Testing
America

Client Information		Sampler: Ian St. John		Lab PM:		Carrier Tracking No(s):		COC No:	
Client Contact: Casey Housman		Phone: 314-853-4668		E-Mail: ian@stjohnenv.com		State of Origin: MO		Page: 1 of 7	
Company: St. Joseph School District		PWSID:		Analysis Requested		Job #:			
Address: 1415 North 26th Street		Due Date Requested:		Field Filtered Sample (Yes or No)		Perform MS/MSD (Yes or No)		Total number of containers	
City: St. Joseph		TAT Requested (days): Standard							
State, Zip: Missouri 64506		Compliance Project: ☐ Yes ☒ No							
Phone: 816-671-4000		PO #:							
Email: CaseyHousman@sjsd.k12.mo.us		WO #:							
Project Name: St. Joseph School District		Project #:							
Site: LAYFAYETTE		SSOW#:							
Sample Identification		Sample Date	Sample Time	Sample Type (C=Comp, G=grab)	Matrix (W=water, S=solid, O=soil, L=liquid, T=tissue, A=air)	Preservation Code:			
ST L-1 Spk	6/25/24 0800	G	W						
ST L-2 Spk									
ST L-3 WF									
ST L-4 Spk									
ST L-5 "									
ST L-6 "									
ST L-7 "									
ST L-8 "									
ST L-9 "									
ST L-10 WF									
ST L-11 Spk									
Possible Hazard Identification						Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)			
☐ Non-Hazard ☐ Flammable ☐ Skin Irritant ☐ Poison B ☒ Unknown ☐ Radiological						☐ Return To Client ☒ Disposal By Lab ☐ Archive For Months			
Deliverable Requested: I, II, III, IV, Other (specify)						Special Instructions/QC Requirements:			
Empty Kit Relinquished by:						Method of Shipment:			
Relinquished by: [Signature]						Received by: [Signature]			
Relinquished by:						Received by:			
Relinquished by:						Received by:			
Custody Seals Intact: ☐ Yes ☒ No						Cooler Temperature(s) °C and Other Remarks: Ambient			

Chain of Custody Record



Environment Testing
America

2-64

Client Information		Sampler: Ian St. John		Lab PM:		Carrier Tracking No(s):		COC No:	
Client Contact: Casey Housman		Phone: 314-853-4668		E-Mail: ian@stjohnenv.com		State of Origin: MO		Page: 2 of 7	
Company: St. Joseph School District		PWSID:		Analysis Requested		Total Number of Containers		Job #:	
Address: 1415 North 26th Street		Due Date Requested:		Perform MS/MSD (Yes or No)		Field Filtered Sample (Yes or No)		Preservation Codes:	
City: St. Joseph		TAT Requested (days): Standard		Sample Date		Sample Type (C=Comp, G=grab)		Matrix (W=water, S=solid, O=on-site, BT=trace, A=lab)	
State, Zip: Missouri 64506		Compliance Project: Δ Yes Δ No		Sample Time		Preservation Code: G W		A-HCL B-NaOH C-Zn Acetate D-Nitric Acid E-NaHSO4 F-MeOH G-Amchlor H-Ascorbic Acid I-Ice J-DI Water K-EDTA L-EDA Other:	
Phone: 816-671-4000		PO #:		Sample Date		Sample Time		M-Hexane N-None O-AsNaO2 P-Na2O4S Q-Na2SO3 R-NaHSO4 S-H2SO4 T-TSP Dodecahydrate U-Acetone V-MCAA W-pH 4-5 Y-Tritama Z-other (specify)	
Email: CaseyHousman@sjsd.k12.mo.us		WO #:		Sample Date		Sample Time		Special Instructions/Note:	
Project Name: St. Joseph School District		Project #:		Sample Date		Sample Time			
Site: LA FAYETTE		SSOW#:		Sample Date		Sample Time			
Sample Identification		Sample Date		Sample Time		Sample Type		Matrix	
ST L-12 Sink		6/25/24 0800		G		W			
ST L-13 "									
ST L-14 "									
ST L-15 "									
ST L-16 "									
ST L-17 "									
ST L-18 WF									
ST L-19 Sink									
ST L-20 Sink									
ST L-21 Sink									
ST L-22 WF									
Possible Hazard Identification		Sample Date		Sample Time		Sample Type		Matrix	
☐ Non-Hazard ☐ Flammable ☐ Skin Irritant ☐ Poison B ☒ Unknown ☐ Radiological		Sample Date		Sample Time		Sample Type		Matrix	
Deliverable Requested: I, II, III, IV, Other (specify)		Sample Date		Sample Time		Sample Type		Matrix	
Empty Kit Relinquished by:		Date:		Time:		Method of Shipment:		Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)	
Relinquished by: [Signature]		Date/Time: April 14, 2024		Company: [Signature]		Received by: [Signature]		☐ Return To Client ☒ Disposal By Lab ☐ Archive For Months	
Relinquished by:		Date/Time:		Company:		Received by:		Special Instructions/QC Requirements:	
Relinquished by:		Date/Time:		Company:		Received by:			
Custody Seals Intact: Δ Yes Δ No		Custody Seal No.:		Cooler Temperature(s) °C and Other Remarks:					

Chain of Custody Record



Environment Testing
America

Client Information		Sampler: Ian St. John		Lab PM:		Carrier Tracking No(s):		COC No:	
Client Contact: Casey Housman		Phone: 314-853-4668		E-Mail: ian@stjohnenv.com		State of Origin: MO		Page: 1 of 3 of 7	
Company: St. Joseph School District		PWSID:		Analysis Requested		Total Number of Containers		Preservation Codes:	
Address: 1415 North 26th Street		Due Date Requested:		Field Filtered Sample (Yes or No)		Perform MS/MSD (Yes or No)		A - HCL B - NaOH C - Zn Acetate D - Nitric Acid E - NaHSO4 F - MeOH G - Anchlor H - Ascorbic Acid I - Ice J - DI Water K - EDTA L - EDA Other:	
City: St. Joseph		TAT Requested (days): Standard		Sample Date		Sample Time		M - Hexane N - None O - AsNaO2 P - Na2O4S Q - Na2SO3 R - Na2S2O3 S - H2SO4 T - TSP Dodecahydrate U - Acetone V - MCAA W - pH 4-5 Y - Trizma Z - other (specify)	
State, Zip: Missouri 64506		Compliance Project: Δ Yes Δ No		Sample Type (C=Comp, G=grab)		Matrix (W=water, S=solid, O=wastewater, L=tissue, A=air)		Special Instructions/Note:	
Phone: 816-671-4000		PO #:		Preservation Code:		W		EMPTY	
Email: CaseyHousman@sjsd.k12.mo.us		WO #:		Sample Date		6/25/24 0800			
Project Name: St. Joseph School District		Project #:		Sample Time					
Site: LAFAYETTE		SSOW#:		Sample Date					
Sample Identification		Sample Type		Sample Time		Sample Date		Sample Time	
SJ L-23 SINK		WF		(Rorc)					
SJ L-24 WF									
SJ L-25 WF									
SJ L-26 WF									
SJ L-27 WF									
SJ L-28 WF									
SJ L-29 SINK									
SJ L-30 WF									
SJ L-31 WF									
SJ L-32 WF									
SJ L-33 SINK									
Possible Hazard Identification		Non-Hazard		Flammable		Skin Irritant		Deliverable Requested: I, II, III, IV, Other (specify)	
Empty Kit Relinquished by:		Date:		Time:		Method of Shipment:		Return To Client	
Relinquished by: [Signature]		Date/Time: April 14, 2024		Company: [Signature]		Date/Time: 6/25/24 0815		Company: [Signature]	
Relinquished by:		Date/Time:		Company:		Date/Time:		Company:	
Relinquished by:		Date/Time:		Company:		Date/Time:		Company:	
Custody Seals Intact: Δ Yes Δ No		Custody Seal No.:		Cooler Temperature(s) °C and Other Remarks:					

Chain of Custody Record



Environment Testing
America

Client Information		Sampler: Ian St. John		Lab PM:		Carrier Tracking No(s):		COC No:	
Client Contact: Casey Housman		Phone: 314-853-4668		E-Mail: ian@stjohnenv.com		State of Origin: MO		Page: 4 of 7	
Company: St. Joseph School District		PWSID:		Analysis Requested				Job #:	
Address: 1415 North 26th Street		Due Date Requested:							
City: St. Joseph		TAT Requested (days): Standard							
State, Zip: Missouri 64506		Compliance Project: ☐ Yes ☒ No							
Phone: 816-671-4000		PO #:							
Email: CaseyHousman@sjsd.k12.mo.us		WO #:							
Project Name: St. Joseph School District		Project #:							
Site: LAFAYETTE		SSOW#:							
Sample Identification		Sample Date	Sample Time	Sample Type (C=Comp, G=grab)	Matrix (W=water, S=solid, O=soil, T=tissue, A=air)	Field Filtered Sample (Yes or No)	Perform MS/MSD (Yes or No)	Total Number of Containers	
ST L-34 Spk		6/25/24	0800	G	W				
ST L-35 Spk									
ST L-36 Spk									
ST L-37 Spk									
ST L-38 Spk									
ST L-39 Spk									
ST L-40 Spk									
ST L-41 "									
ST L-42 "									
ST L-43 "									
ST L-44 WF									
Possible Hazard Identification		☒ Poison B ☐ Unknown ☐ Radiological							
☐ Non-Hazard ☐ Flammable ☐ Skin Irritant									
Deliverable Requested: I, II, III, IV, Other (specify)									
Empty Kit Relinquished by:		Date:		Time:		Method of Shipment:			
Relinquished by: [Signature]		Date/Time: April 14, 2024				Received by: [Signature]		Date/Time: 6/26/24 0945	
Relinquished by:		Date/Time:				Received by:		Date/Time:	
Relinquished by:		Date/Time:				Received by:		Date/Time:	
Custody Seals Intact: ☐ Yes ☐ No		Custody Seal No.:				Cooler Temperature(s) °C and Other Remarks:			

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

Chain of Custody Record



Client Information		Sampler:	Lab PM:	Carrier Tracking No(s):	COC No:			
Client Contact: Casey Housman		Phone: 314-853-4668	Lab PM: Ian St. John	State of Origin: MO	Page: 5 of 7			
Company: St. Joseph School District		PWSID:			Job #:			
Address: 1415 North 26th Street		Due Date Requested:		Analysis Requested				
City: St. Joseph		TAT Requested (days): Standard						
State, Zip: Missouri 64506		Compliance Project: ☐ Yes ☐ No						
Phone: 816-671-4000		PO #:						
Email: CaseyHousman@sjsd.k12.mo.us		WO #:						
Project Name: St. Joseph School District		Project #:						
Site: LAFAYETTE		SSOW#:						
Sample Identification	Sample Date	Sample Time	Sample Type (C=Comp, G=grab)	Matrix (W=water, S=solid, O=soil, G=grass, L=leaves, A=air)	Field Filtered Sample (Yes or No)	Perform MS/MSD (Yes or No)	Total Number of Containers	Special Instructions/Note:
SJ L-45 Spark	6/25/24	0800	G	W				
SJ L-46 "								
SJ L-47 "								
SJ L-48 "								
SJ L-49 "								
SJ L-50 "								
SJ L-51 "								
SJ L-52 "								
SJ L-53 "								
SJ L-54 "								
SJ L-55 "								EMPTY
Possible Hazard Identification ☐ Non-Hazard ☐ Flammable ☐ Skin Irritant ☐ Poison B ☒ Unknown ☐ Radiological Deliverable Requested: I, II, III, IV, Other (specify)								
Sample Disposal (A fee may be assessed if samples are retained longer than 1 month) ☐ Return To Client ☒ Disposal By Lab ☐ Archive For _____ Months								
Special Instructions/QC Requirements:								
Empty Kit Relinquished by: _____ Date: _____								
Relinquished by: _____ Date/Time: _____ Company: _____								
Relinquished by: _____ Date/Time: _____ Company: _____								
Relinquished by: _____ Date/Time: _____ Company: _____								
Custody Seals Intact: ☐ Yes ☐ No Custody Seal No.: _____								

Chain of Custody Record



Client Information		Sampler:	Lab PM:	Carrier Tracking No(s):	COC No:			
Client Contact: Casey Housman		Phone: 314-853-4668	E-Mail: ian@stjohnenv.com	State of Origin: MO	Page: 6 of 7			
Company: St. Joseph School District		Analysis Requested						
Address: 1415 North 26th Street		Due Date Requested:						
City: St. Joseph		TAT Requested (days): Standard						
State, Zip: Missouri 64506		Compliance Project: ☐ Yes ☐ No						
Phone: 816-671-4000		PO #:						
Email: CaseyHousman@sjsd.k12.mo.us		WO #:						
Project Name: St. Joseph School District		Project #:						
Site: LAFAYETTE		SSOW#:						
Sample Identification	Sample Date	Sample Time	Sample Type (C=Comp, G=grab)	Matrix (W=water, S=solid, O=soil, BT=biota, A=air)	Field Filtered Sample (Yes or No)	Perform MS/MSD (Yes or No)	Total Number of Containers	Special Instructions/Note:
ST L-56 WF	6/25/24	0800	G	W				
ST L-57 Snrk								
ST L-58 WF								
ST L-59 Snrk								
ST L-60 LOT (Field House) DE								Visiting Locker
ST L-61 WF (Field House) DE								Visiting Locker
ST L-62 CE (Near Est)								Home Locker
ST L-63 WF								
ST L-64 WF								
ST L-65 WF (Field House)								Visiting Locker
ST L-66 WF (Field House)								Empty 21
Possible Hazard Identification ☐ Non-Hazard ☐ Flammable ☐ Skin Irritant ☐ Poison B ☒ Unknown ☐ Radiological					Sample Disposal (A fee may be assessed if samples are retained longer than 1 month) ☐ Return To Client ☒ Disposal By Lab ☐ Archive For _____ Months			
Deliverable Requested: I, II, III, IV, Other (specify)					Special Instructions/QC Requirements:			
Empty Kit Relinquished by: _____ Date: _____					Method of Shipment: _____			
Relinquished by: _____ Date/Time: _____ Company: _____					Received by: _____ Date/Time: _____ Company: _____			
Relinquished by: _____ Date/Time: _____ Company: _____					Received by: _____ Date/Time: _____ Company: _____			
Relinquished by: _____ Date/Time: _____ Company: _____					Received by: _____ Date/Time: _____ Company: _____			
Custody Seals Intact: ☐ Yes ☐ No					Cooler Temperature(s) °C and Other Remarks:			

Environment Testing America

Chain of Custody Record

Ver. 02/24/23

Login Sample Receipt Checklist

Client: St. John Environmental Consulting

Job Number: 810-109594-1

Login Number: 109594

List Source: Eurofins Eaton Analytical South Bend

List Number: 1

Creator: Williams, Kameron

Question	Answer	Comment
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
Samples were received on ice.	False	Thermal preservation not required.
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	False	Thermal preservation not required.
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	False	Refer to Job Narrative for details.
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	