

# ANALYTICAL REPORT

## PREPARED FOR

Attn: Ian St. John  
St. John Environmental Consulting  
5515 Grace Ave.  
Saint Louis, Missouri 63116

Generated 8/13/2024 6:48:57 AM

## JOB DESCRIPTION

St. Joseph School District - Webster

## JOB NUMBER

810-114326-1

# Eurofins Eaton Analytical South Bend

## Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and will meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Eaton Analytical, LLC Project Manager.

## Authorization



Generated  
8/13/2024 6:48:57 AM

Authorized for release by  
Pamela Brown, Project Manager  
[Pamela.Brown@et.eurofinsus.com](mailto:Pamela.Brown@et.eurofinsus.com)  
(574)233-4777

# Table of Contents

Cover Page . . . . .	1
Table of Contents . . . . .	3
Definitions/Glossary . . . . .	4
Case Narrative . . . . .	5
Detection Summary . . . . .	6
Client Sample Results . . . . .	7
QC Sample Results . . . . .	9
QC Association Summary . . . . .	10
Lab Chronicle . . . . .	11
Certification Summary . . . . .	13
Method Summary . . . . .	14
Sample Summary . . . . .	15
Chain of Custody . . . . .	16
Receipt Checklists . . . . .	17



## Definitions/Glossary

Client: St. John Environmental Consulting  
Project/Site: St. Joseph School District - Webster

Job ID: 810-114326-1

### Qualifiers

#### Metals

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

### Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

## Case Narrative

Client: St. John Environmental Consulting  
Project: St. Joseph School District - Webster

Job ID: 810-114326-1

**Job ID: 810-114326-1**

**Eurofins Eaton Analytical South Bend**

### **Job Narrative 810-114326-1**

Analytical test results meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page unless otherwise noted under the individual analysis. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable.

- Matrix QC may not be reported if insufficient sample is provided or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD may be performed, unless otherwise specified in the method.
- Surrogate and/or isotope dilution analyte recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in the narrative.

Regulated compliance samples (e.g. SDWA, NPDES) must comply with the associated agency requirements/permits.

#### **Receipt**

The samples were received on 8/6/2024 9:15 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice.

#### **Metals**

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Eurofins Eaton Analytical South Bend

## Detection Summary

Client: St. John Environmental Consulting  
Project/Site: St. Joseph School District - Webster

Job ID: 810-114326-1

### Client Sample ID: SJW - 7 Sink

Lab Sample ID: 810-114326-1

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	63		0.50		ug/L	1		200.8	Total/NA

### Client Sample ID: SJW - 8 Sink

Lab Sample ID: 810-114326-2

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	98		0.50		ug/L	1		200.8	Total/NA

### Client Sample ID: SJW - 9 Sink

Lab Sample ID: 810-114326-3

No Detections.

### Client Sample ID: SJW - 11 Sink

Lab Sample ID: 810-114326-4

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	7.4		0.50		ug/L	1		200.8	Total/NA

### Client Sample ID: SJW - 14 Sink

Lab Sample ID: 810-114326-5

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	3.3		0.50		ug/L	1		200.8	Total/NA

### Client Sample ID: SJW - 16 Sink

Lab Sample ID: 810-114326-6

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	3.8		0.50		ug/L	1		200.8	Total/NA

### Client Sample ID: SJW - 18 Sink

Lab Sample ID: 810-114326-7

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	1.5		0.50		ug/L	1		200.8	Total/NA

### Client Sample ID: SJW - 20 Sink

Lab Sample ID: 810-114326-8

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	1.9		0.50		ug/L	1		200.8	Total/NA

This Detection Summary does not include radiochemical test results.

Eurofins Eaton Analytical South Bend

## Client Sample Results

Client: St. John Environmental Consulting  
Project/Site: St. Joseph School District - Webster

Job ID: 810-114326-1

### Client Sample ID: SJW - 7 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/06/24 09:15

Lab Sample ID: 810-114326-1

Matrix: Drinking Water

#### Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	63		0.50		ug/L			08/12/24 21:32	1

### Client Sample ID: SJW - 8 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/06/24 09:15

Lab Sample ID: 810-114326-2

Matrix: Drinking Water

#### Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	98		0.50		ug/L			08/12/24 21:41	1

### Client Sample ID: SJW - 9 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/06/24 09:15

Lab Sample ID: 810-114326-3

Matrix: Drinking Water

#### Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			08/12/24 21:44	1

### Client Sample ID: SJW - 11 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/06/24 09:15

Lab Sample ID: 810-114326-4

Matrix: Drinking Water

#### Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	7.4		0.50		ug/L			08/12/24 21:47	1

### Client Sample ID: SJW - 14 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/06/24 09:15

Lab Sample ID: 810-114326-5

Matrix: Drinking Water

#### Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	3.3		0.50		ug/L			08/12/24 21:49	1

### Client Sample ID: SJW - 16 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/06/24 09:15

Lab Sample ID: 810-114326-6

Matrix: Drinking Water

#### Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	3.8		0.50		ug/L			08/12/24 21:52	1

### Client Sample ID: SJW - 18 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/06/24 09:15

Lab Sample ID: 810-114326-7

Matrix: Drinking Water

#### Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	1.5		0.50		ug/L			08/12/24 22:01	1

Eurofins Eaton Analytical South Bend

Client Sample Results

Client: St. John Environmental Consulting  
Project/Site: St. Joseph School District - Webster

Job ID: 810-114326-1

Client Sample ID: SJW - 20 Sink

Lab Sample ID: 810-114326-8

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/06/24 09:15

Method: EPA 200.8 - Metals (ICP/MS)									
Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	1.9		0.50		ug/L			08/12/24 22:04	1



# QC Sample Results

Client: St. John Environmental Consulting  
Project/Site: St. Joseph School District - Webster

Job ID: 810-114326-1

## Method: 200.8 - Metals (ICP/MS)

Lab Sample ID: MB 810-109657/13

Matrix: Drinking Water

Analysis Batch: 109657

Client Sample ID: Method Blank

Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			08/12/24 18:06	1

Lab Sample ID: MB 810-109657/74

Matrix: Drinking Water

Analysis Batch: 109657

Client Sample ID: Method Blank

Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			08/12/24 20:59	1

Lab Sample ID: LCS 810-109657/75

Matrix: Drinking Water

Analysis Batch: 109657

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Lead	50.0	48.6		ug/L		97	85 - 115

Lab Sample ID: LLCS 810-109657/11

Matrix: Drinking Water

Analysis Batch: 109657

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Analyte	Spike Added	LLCS Result	LLCS Qualifier	Unit	D	%Rec	%Rec Limits
Lead	0.300	0.291	J	ug/L		97	50 - 150

Lab Sample ID: 810-114326-1 MS

Matrix: Drinking Water

Analysis Batch: 109657

Client Sample ID: SJW - 7 Sink

Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Lead	63		50.0	110		ug/L		93	70 - 130

Lab Sample ID: 810-114326-1 MSD

Matrix: Drinking Water

Analysis Batch: 109657

Client Sample ID: SJW - 7 Sink

Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Lead	63		50.0	108		ug/L		89	70 - 130	2	20

## QC Association Summary

Client: St. John Environmental Consulting  
Project/Site: St. Joseph School District - Webster

Job ID: 810-114326-1

### Metals

#### Analysis Batch: 109657

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
810-114326-1	SJW - 7 Sink	Total/NA	Drinking Water	200.8	
810-114326-2	SJW - 8 Sink	Total/NA	Drinking Water	200.8	
810-114326-3	SJW - 9 Sink	Total/NA	Drinking Water	200.8	
810-114326-4	SJW - 11 Sink	Total/NA	Drinking Water	200.8	
810-114326-5	SJW - 14 Sink	Total/NA	Drinking Water	200.8	
810-114326-6	SJW - 16 Sink	Total/NA	Drinking Water	200.8	
810-114326-7	SJW - 18 Sink	Total/NA	Drinking Water	200.8	
810-114326-8	SJW - 20 Sink	Total/NA	Drinking Water	200.8	
MB 810-109657/13	Method Blank	Total/NA	Drinking Water	200.8	
MB 810-109657/74	Method Blank	Total/NA	Drinking Water	200.8	
LCS 810-109657/75	Lab Control Sample	Total/NA	Drinking Water	200.8	
LLCS 810-109657/11	Lab Control Sample	Total/NA	Drinking Water	200.8	
810-114326-1 MS	SJW - 7 Sink	Total/NA	Drinking Water	200.8	
810-114326-1 MSD	SJW - 7 Sink	Total/NA	Drinking Water	200.8	

# Lab Chronicle

Client: St. John Environmental Consulting  
Project/Site: St. Joseph School District - Webster

Job ID: 810-114326-1

**Client Sample ID: SJW - 7 Sink**

**Lab Sample ID: 810-114326-1**

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/06/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109657	NB	EA SB	08/12/24 21:32

**Client Sample ID: SJW - 8 Sink**

**Lab Sample ID: 810-114326-2**

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/06/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109657	NB	EA SB	08/12/24 21:41

**Client Sample ID: SJW - 9 Sink**

**Lab Sample ID: 810-114326-3**

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/06/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109657	NB	EA SB	08/12/24 21:44

**Client Sample ID: SJW - 11 Sink**

**Lab Sample ID: 810-114326-4**

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/06/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109657	NB	EA SB	08/12/24 21:47

**Client Sample ID: SJW - 14 Sink**

**Lab Sample ID: 810-114326-5**

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/06/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109657	NB	EA SB	08/12/24 21:49

**Client Sample ID: SJW - 16 Sink**

**Lab Sample ID: 810-114326-6**

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/06/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109657	NB	EA SB	08/12/24 21:52

**Client Sample ID: SJW - 18 Sink**

**Lab Sample ID: 810-114326-7**

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/06/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109657	NB	EA SB	08/12/24 22:01

Eurofins Eaton Analytical South Bend

Lab Chronicle

Client: St. John Environmental Consulting  
Project/Site: St. Joseph School District - Webster

Job ID: 810-114326-1

**Client Sample ID: SJW - 20 Sink**  
**Date Collected: 08/01/24 08:00**  
**Date Received: 08/06/24 09:15**

**Lab Sample ID: 810-114326-8**  
**Matrix: Drinking Water**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109657	NB	EA SB	08/12/24 22:04

**Laboratory References:**  
EA SB = Eurofins Eaton Analytical South Bend, 110 S Hill Street, South Bend, IN 46617, TEL (574)233-4777

Accreditation/Certification Summary

Client: St. John Environmental Consulting  
Project/Site: St. Joseph School District - Webster

Job ID: 810-114326-1

Laboratory: Eurofins Eaton Analytical South Bend

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Missouri	State	880	09-30-24

1
2
3
4
5
6
7
8
9
10
11
12
13
14

Method Summary

Client: St. John Environmental Consulting  
Project/Site: St. Joseph School District - Webster

Job ID: 810-114326-1

Method	Method Description	Protocol	Laboratory
200.8	Metals (ICP/MS)	EPA	EA SB

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA SB = Eurofins Eaton Analytical South Bend, 110 S Hill Street, South Bend, IN 46617, TEL (574)233-4777

## Sample Summary

Client: St. John Environmental Consulting  
Project/Site: St. Joseph School District - Webster

Job ID: 810-114326-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
810-114326-1	SJW - 7 Sink	Drinking Water	08/01/24 08:00	08/06/24 09:15
810-114326-2	SJW - 8 Sink	Drinking Water	08/01/24 08:00	08/06/24 09:15
810-114326-3	SJW - 9 Sink	Drinking Water	08/01/24 08:00	08/06/24 09:15
810-114326-4	SJW - 11 Sink	Drinking Water	08/01/24 08:00	08/06/24 09:15
810-114326-5	SJW - 14 Sink	Drinking Water	08/01/24 08:00	08/06/24 09:15
810-114326-6	SJW - 16 Sink	Drinking Water	08/01/24 08:00	08/06/24 09:15
810-114326-7	SJW - 18 Sink	Drinking Water	08/01/24 08:00	08/06/24 09:15
810-114326-8	SJW - 20 Sink	Drinking Water	08/01/24 08:00	08/06/24 09:15

LABORATORY ANALYTICAL: South Bend  
110 South Hill Street  
South Bend, IN 46817  
Phone (574) 233-4777

## Chain of Custody Record

eurofins  
Environmental Testing  
America

Client Information		Sample: Ian St. John		Lab File	
Client Contact: Casey Housman		Phone: 314-853-4668		E-Mail: lan@stjohnenv.com	
Company: St. Joseph School District		State of Origin: MO		Page: Page 1 of 1	
Address: 1415 North 26th Street		City: St. Joseph		Lab #:	
State, Zip: Missouri 64506		Phone: 816-671-4000		Analysis Requested: 42	
Email: CaseyHousman@sjsd.k12.mo.us		Project #: 3809W		Preservation Codes:	
Project Name: St. Joseph School District		Matrix (Pre-Test, Sample, Post-Test):		A-HCL B-NaOH C-Zn Acetate D-Nitric Acid E-NaHCO3 F-NaOH G-Arsonic Acid H-Ascorbic Acid I-Ice J-DI Water K-EDTA L-EDA Other:	
Site: Webster		Sample Date: 8/1/24		Sample Time: 0800	
Sample Identification		Sample Type (C=Comp, G=Grab):		Preservation Code: G W	
STW - 7 sink					
STW - 8 sink					
STW - 9 sink					
STW - 11 sink					
STW - 14 sink					
STW - 16 sink					
STW - 18 sink					
STW - 20 sink					
ST					
ST					
ST					
Possible Hazard Identification		Polison B <input checked="" type="checkbox"/> Unknown <input type="checkbox"/> Radiological		Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)	
<input type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Other (specify)				<input type="checkbox"/> Return To Client <input checked="" type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For Months	
Deliverable Requested: I, II, III, IV, Other (specify)				Special Instructions/Notes:	
Empty Kit Relinquished by: [Signature]		Date/Time: 8/1/24		Method of Shipment:	
Relinquished by: [Signature]		Date/Time: 8/1/24		Company:	
Relinquished by: [Signature]		Date/Time: 8/1/24		Company:	
Relinquished by: [Signature]		Date/Time: 8/1/24		Company:	
Custody Seal Intact: [Signature]		Custody Seal No.:		Cooler Temperature: 41	



## Login Sample Receipt Checklist

Client: St. John Environmental Consulting

Job Number: 810-114326-1

Login Number: 114326

List Source: Eurofins Eaton Analytical South Bend

List Number: 1

Creator: Alfred, Robbin

Question	Answer	Comment
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
Samples were received on ice.	False	Thermal preservation not required.
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	False	Thermal preservation not required.
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	