

ANALYTICAL REPORT

PREPARED FOR

Attn: Ian St. John
St. John Environmental Consulting
5515 Grace Ave.
Saint Louis, Missouri 63116

Generated 8/12/2024 11:53:53 PM

JOB DESCRIPTION

St. Joseph School District - Lafayette

JOB NUMBER

810-114270-1

Eurofins Eaton Analytical South Bend

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and will meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Eaton Analytical, LLC Project Manager.

Authorization



Generated
8/12/2024 11:53:53 PM

Authorized for release by
Pamela Brown, Project Manager
Pamela.Brown@et.eurofinsus.com
(574)233-4777

Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	8
QC Sample Results	12
QC Association Summary	14
Lab Chronicle	15
Certification Summary	19
Method Summary	20
Sample Summary	21
Chain of Custody	22
Receipt Checklists	25



Definitions/Glossary

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-114270-1

Qualifiers

Metals

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: St. John Environmental Consulting
Project: St. Joseph School District - Lafayette

Job ID: 810-114270-1

Job ID: 810-114270-1

Eurofins Eaton Analytical South Bend

Job Narrative 810-114270-1

Analytical test results meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page unless otherwise noted under the individual analysis. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable.

- Matrix QC may not be reported if insufficient sample is provided or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD may be performed, unless otherwise specified in the method.
- Surrogate and/or isotope dilution analyte recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in the narrative.

Regulated compliance samples (e.g. SDWA, NPDES) must comply with the associated agency requirements/permits.

Receipt

The samples were received on 8/5/2024 9:15 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice.

Metals

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Eurofins Eaton Analytical South Bend

Detection Summary

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-114270-1

Client Sample ID: SJL-7 Sink

Lab Sample ID: 810-114270-1

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	0.83		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL-8 Sink

Lab Sample ID: 810-114270-2

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	6.0		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL-9 Sink

Lab Sample ID: 810-114270-3

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	1.8		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL-11 Sink

Lab Sample ID: 810-114270-4

No Detections.

Client Sample ID: SJL-12 Sink

Lab Sample ID: 810-114270-5

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	20		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL-16 Sink

Lab Sample ID: 810-114270-6

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	0.99		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL-17 Sink

Lab Sample ID: 810-114270-7

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	1.1		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL-35 Sink

Lab Sample ID: 810-114270-8

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	14		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL-37 Sink

Lab Sample ID: 810-114270-9

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	3.0		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL-38 Sink

Lab Sample ID: 810-114270-10

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	3.4		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL-40 Sink

Lab Sample ID: 810-114270-11

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	0.96		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL-41 Sink

Lab Sample ID: 810-114270-12

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	7.0		0.50		ug/L	1		200.8	Total/NA

This Detection Summary does not include radiochemical test results.

Eurofins Eaton Analytical South Bend

Detection Summary

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-114270-1

Client Sample ID: SJL-42 Sink

Lab Sample ID: 810-114270-13

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	330		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL-43 Sink

Lab Sample ID: 810-114270-14

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	18		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL-45 Sink

Lab Sample ID: 810-114270-15

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	5.5		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL-46 Sink

Lab Sample ID: 810-114270-16

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	4.2		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL-47 Sink

Lab Sample ID: 810-114270-17

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	57		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL-48 Sink

Lab Sample ID: 810-114270-18

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	7.7		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL-49 Sink

Lab Sample ID: 810-114270-19

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	2.8		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL-50 Sink

Lab Sample ID: 810-114270-20

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	3.1		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL-51 Sink

Lab Sample ID: 810-114270-21

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	2.5		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL-52 Sink

Lab Sample ID: 810-114270-22

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	2.0		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL-55 Sink

Lab Sample ID: 810-114270-23

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	1.4		0.50		ug/L	1		200.8	Total/NA

Client Sample ID: SJL-57 Sink

Lab Sample ID: 810-114270-24

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Lead	17		0.50		ug/L	1		200.8	Total/NA

This Detection Summary does not include radiochemical test results.

Eurofins Eaton Analytical South Bend

Client Sample Results

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-114270-1

Client Sample ID: SJL-7 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-1

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	0.83		0.50		ug/L			08/09/24 18:33	1

Client Sample ID: SJL-8 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-2

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	6.0		0.50		ug/L			08/09/24 18:35	1

Client Sample ID: SJL-9 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-3

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	1.8		0.50		ug/L			08/09/24 18:38	1

Client Sample ID: SJL-11 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-4

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			08/09/24 18:41	1

Client Sample ID: SJL-12 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-5

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	20		0.50		ug/L			08/09/24 18:49	1

Client Sample ID: SJL-16 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-6

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	0.99		0.50		ug/L			08/09/24 18:52	1

Client Sample ID: SJL-17 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-7

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	1.1		0.50		ug/L			08/09/24 18:55	1

Client Sample Results

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-114270-1

Client Sample ID: SJL-35 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-8

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	14		0.50		ug/L			08/09/24 18:58	1

Client Sample ID: SJL-37 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-9

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	3.0		0.50		ug/L			08/12/24 12:42	1

Client Sample ID: SJL-38 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-10

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	3.4		0.50		ug/L			08/12/24 12:47	1

Client Sample ID: SJL-40 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-11

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	0.96		0.50		ug/L			08/12/24 12:49	1

Client Sample ID: SJL-41 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-12

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	7.0		0.50		ug/L			08/12/24 12:50	1

Client Sample ID: SJL-42 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-13

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	330		0.50		ug/L			08/12/24 12:52	1

Client Sample ID: SJL-43 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-14

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	18		0.50		ug/L			08/12/24 12:53	1

Eurofins Eaton Analytical South Bend

Client Sample Results

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-114270-1

Client Sample ID: SJL-45 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-15

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	5.5		0.50		ug/L			08/12/24 12:55	1

Client Sample ID: SJL-46 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-16

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	4.2		0.50		ug/L			08/12/24 12:57	1

Client Sample ID: SJL-47 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-17

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	57		0.50		ug/L			08/12/24 12:58	1

Client Sample ID: SJL-48 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-18

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	7.7		0.50		ug/L			08/12/24 13:00	1

Client Sample ID: SJL-49 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-19

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	2.8		0.50		ug/L			08/12/24 13:05	1

Client Sample ID: SJL-50 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-20

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	3.1		0.50		ug/L			08/12/24 13:07	1

Client Sample ID: SJL-51 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-21

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	2.5		0.50		ug/L			08/12/24 13:12	1

Eurofins Eaton Analytical South Bend

Client Sample Results

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-114270-1

Client Sample ID: SJL-52 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-22

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	2.0		0.50		ug/L			08/12/24 13:14	1

Client Sample ID: SJL-55 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-23

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	1.4		0.50		ug/L			08/12/24 13:15	1

Client Sample ID: SJL-57 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-24

Matrix: Drinking Water

Method: EPA 200.8 - Metals (ICP/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	17		0.50		ug/L			08/12/24 13:17	1

QC Sample Results

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-114270-1

Method: 200.8 - Metals (ICP/MS)

Lab Sample ID: MB 810-109356/11
Matrix: Drinking Water
Analysis Batch: 109356

Client Sample ID: Method Blank
Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			08/09/24 09:14	1

Lab Sample ID: MB 810-109356/191
Matrix: Drinking Water
Analysis Batch: 109356

Client Sample ID: Method Blank
Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			08/09/24 17:43	1

Lab Sample ID: LCS 810-109356/192
Matrix: Drinking Water
Analysis Batch: 109356

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Lead	50.0	49.3		ug/L		99	85 - 115

Lab Sample ID: LLCS 810-109356/10
Matrix: Drinking Water
Analysis Batch: 109356

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	LLCS Result	LLCS Qualifier	Unit	D	%Rec	%Rec Limits
Lead	0.300	0.266	J	ug/L		89	50 - 150

Lab Sample ID: 810-114270-4 MS
Matrix: Drinking Water
Analysis Batch: 109356

Client Sample ID: SJL-11 Sink
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Lead	<0.50		50.0	49.1		ug/L		98	70 - 130

Lab Sample ID: 810-114270-4 MSD
Matrix: Drinking Water
Analysis Batch: 109356

Client Sample ID: SJL-11 Sink
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Lead	<0.50		50.0	49.5		ug/L		99	70 - 130	1	20

Lab Sample ID: MB 810-109645/11
Matrix: Drinking Water
Analysis Batch: 109645

Client Sample ID: Method Blank
Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Lead	<0.50		0.50		ug/L			08/12/24 12:39	1

Lab Sample ID: LCS 810-109645/12
Matrix: Drinking Water
Analysis Batch: 109645

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Lead	50.0	51.1		ug/L		102	85 - 115

Eurofins Eaton Analytical South Bend

QC Sample Results

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-114270-1

Method: 200.8 - Metals (ICP/MS)

Lab Sample ID: LLCS 810-109645/10
Matrix: Drinking Water
Analysis Batch: 109645

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte			Spike Added	LLCS Result	LLCS Qualifier	Unit	D	%Rec	%Rec Limits		
Lead			0.300	0.321	J	ug/L		107	50 - 150		

Lab Sample ID: 810-114270-9 MS
Matrix: Drinking Water
Analysis Batch: 109645

Client Sample ID: SJL-37 Sink
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits		
Lead	3.0		50.0	59.1		ug/L		112	70 - 130		

Lab Sample ID: 810-114270-9 MSD
Matrix: Drinking Water
Analysis Batch: 109645

Client Sample ID: SJL-37 Sink
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Lead	3.0		50.0	59.4		ug/L		113	70 - 130	1	20

Lab Sample ID: 810-114270-20 MS
Matrix: Drinking Water
Analysis Batch: 109645

Client Sample ID: SJL-50 Sink
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits		
Lead	3.1		50.0	55.6		ug/L		105	70 - 130		

Lab Sample ID: 810-114270-20 MSD
Matrix: Drinking Water
Analysis Batch: 109645

Client Sample ID: SJL-50 Sink
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Lead	3.1		50.0	57.1		ug/L		108	70 - 130	3	20

QC Association Summary

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-114270-1

Metals

Analysis Batch: 109356

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
810-114270-1	SJL-7 Sink	Total/NA	Drinking Water	200.8	
810-114270-2	SJL-8 Sink	Total/NA	Drinking Water	200.8	
810-114270-3	SJL-9 Sink	Total/NA	Drinking Water	200.8	
810-114270-4	SJL-11 Sink	Total/NA	Drinking Water	200.8	
810-114270-5	SJL-12 Sink	Total/NA	Drinking Water	200.8	
810-114270-6	SJL-16 Sink	Total/NA	Drinking Water	200.8	
810-114270-7	SJL-17 Sink	Total/NA	Drinking Water	200.8	
810-114270-8	SJL-35 Sink	Total/NA	Drinking Water	200.8	
MB 810-109356/11	Method Blank	Total/NA	Drinking Water	200.8	
MB 810-109356/191	Method Blank	Total/NA	Drinking Water	200.8	
LCS 810-109356/192	Lab Control Sample	Total/NA	Drinking Water	200.8	
LLCS 810-109356/10	Lab Control Sample	Total/NA	Drinking Water	200.8	
810-114270-4 MS	SJL-11 Sink	Total/NA	Drinking Water	200.8	
810-114270-4 MSD	SJL-11 Sink	Total/NA	Drinking Water	200.8	

Analysis Batch: 109645

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
810-114270-9	SJL-37 Sink	Total/NA	Drinking Water	200.8	
810-114270-10	SJL-38 Sink	Total/NA	Drinking Water	200.8	
810-114270-11	SJL-40 Sink	Total/NA	Drinking Water	200.8	
810-114270-12	SJL-41 Sink	Total/NA	Drinking Water	200.8	
810-114270-13	SJL-42 Sink	Total/NA	Drinking Water	200.8	
810-114270-14	SJL-43 Sink	Total/NA	Drinking Water	200.8	
810-114270-15	SJL-45 Sink	Total/NA	Drinking Water	200.8	
810-114270-16	SJL-46 Sink	Total/NA	Drinking Water	200.8	
810-114270-17	SJL-47 Sink	Total/NA	Drinking Water	200.8	
810-114270-18	SJL-48 Sink	Total/NA	Drinking Water	200.8	
810-114270-19	SJL-49 Sink	Total/NA	Drinking Water	200.8	
810-114270-20	SJL-50 Sink	Total/NA	Drinking Water	200.8	
810-114270-21	SJL-51 Sink	Total/NA	Drinking Water	200.8	
810-114270-22	SJL-52 Sink	Total/NA	Drinking Water	200.8	
810-114270-23	SJL-55 Sink	Total/NA	Drinking Water	200.8	
810-114270-24	SJL-57 Sink	Total/NA	Drinking Water	200.8	
MB 810-109645/11	Method Blank	Total/NA	Drinking Water	200.8	
LCS 810-109645/12	Lab Control Sample	Total/NA	Drinking Water	200.8	
LLCS 810-109645/10	Lab Control Sample	Total/NA	Drinking Water	200.8	
810-114270-9 MS	SJL-37 Sink	Total/NA	Drinking Water	200.8	
810-114270-9 MSD	SJL-37 Sink	Total/NA	Drinking Water	200.8	
810-114270-20 MS	SJL-50 Sink	Total/NA	Drinking Water	200.8	
810-114270-20 MSD	SJL-50 Sink	Total/NA	Drinking Water	200.8	

Lab Chronicle

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-114270-1

Client Sample ID: SJL-7 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-1

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109356	NB	EA SB	08/09/24 18:33

Client Sample ID: SJL-8 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-2

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109356	NB	EA SB	08/09/24 18:35

Client Sample ID: SJL-9 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-3

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109356	NB	EA SB	08/09/24 18:38

Client Sample ID: SJL-11 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-4

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109356	NB	EA SB	08/09/24 18:41

Client Sample ID: SJL-12 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-5

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109356	NB	EA SB	08/09/24 18:49

Client Sample ID: SJL-16 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-6

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109356	NB	EA SB	08/09/24 18:52

Client Sample ID: SJL-17 Sink

Date Collected: 08/01/24 08:00

Date Received: 08/05/24 09:15

Lab Sample ID: 810-114270-7

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109356	NB	EA SB	08/09/24 18:55

Lab Chronicle

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-114270-1

Client Sample ID: SJL-35 Sink

Lab Sample ID: 810-114270-8

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/05/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109356	NB	EA SB	08/09/24 18:58

Client Sample ID: SJL-37 Sink

Lab Sample ID: 810-114270-9

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/05/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109645	NB	EA SB	08/12/24 12:42

Client Sample ID: SJL-38 Sink

Lab Sample ID: 810-114270-10

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/05/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109645	NB	EA SB	08/12/24 12:47

Client Sample ID: SJL-40 Sink

Lab Sample ID: 810-114270-11

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/05/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109645	NB	EA SB	08/12/24 12:49

Client Sample ID: SJL-41 Sink

Lab Sample ID: 810-114270-12

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/05/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109645	NB	EA SB	08/12/24 12:50

Client Sample ID: SJL-42 Sink

Lab Sample ID: 810-114270-13

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/05/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109645	NB	EA SB	08/12/24 12:52

Client Sample ID: SJL-43 Sink

Lab Sample ID: 810-114270-14

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/05/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109645	NB	EA SB	08/12/24 12:53

Eurofins Eaton Analytical South Bend

Lab Chronicle

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-114270-1

Client Sample ID: SJL-45 Sink

Lab Sample ID: 810-114270-15

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/05/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109645	NB	EA SB	08/12/24 12:55

Client Sample ID: SJL-46 Sink

Lab Sample ID: 810-114270-16

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/05/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109645	NB	EA SB	08/12/24 12:57

Client Sample ID: SJL-47 Sink

Lab Sample ID: 810-114270-17

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/05/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109645	NB	EA SB	08/12/24 12:58

Client Sample ID: SJL-48 Sink

Lab Sample ID: 810-114270-18

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/05/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109645	NB	EA SB	08/12/24 13:00

Client Sample ID: SJL-49 Sink

Lab Sample ID: 810-114270-19

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/05/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109645	NB	EA SB	08/12/24 13:05

Client Sample ID: SJL-50 Sink

Lab Sample ID: 810-114270-20

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/05/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109645	NB	EA SB	08/12/24 13:07

Client Sample ID: SJL-51 Sink

Lab Sample ID: 810-114270-21

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/05/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109645	NB	EA SB	08/12/24 13:12

Eurofins Eaton Analytical South Bend

Lab Chronicle

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-114270-1

Client Sample ID: SJL-52 Sink

Lab Sample ID: 810-114270-22

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/05/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109645	NB	EA SB	08/12/24 13:14

Client Sample ID: SJL-55 Sink

Lab Sample ID: 810-114270-23

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/05/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109645	NB	EA SB	08/12/24 13:15

Client Sample ID: SJL-57 Sink

Lab Sample ID: 810-114270-24

Date Collected: 08/01/24 08:00

Matrix: Drinking Water

Date Received: 08/05/24 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	200.8		1	109645	NB	EA SB	08/12/24 13:17

Laboratory References:
EA SB = Eurofins Eaton Analytical South Bend, 110 S Hill Street, South Bend, IN 46617, TEL (574)233-4777

Accreditation/Certification Summary

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-114270-1

Laboratory: Eurofins Eaton Analytical South Bend

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Missouri	State	880	09-30-24

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

Method Summary

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-114270-1

Method	Method Description	Protocol	Laboratory
200.8	Metals (ICP/MS)	EPA	EA SB

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA SB = Eurofins Eaton Analytical South Bend, 110 S Hill Street, South Bend, IN 46617, TEL (574)233-4777

Sample Summary

Client: St. John Environmental Consulting
Project/Site: St. Joseph School District - Lafayette

Job ID: 810-114270-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
810-114270-1	SJL-7 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-2	SJL-8 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-3	SJL-9 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-4	SJL-11 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-5	SJL-12 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-6	SJL-16 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-7	SJL-17 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-8	SJL-35 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-9	SJL-37 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-10	SJL-38 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-11	SJL-40 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-12	SJL-41 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-13	SJL-42 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-14	SJL-43 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-15	SJL-45 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-16	SJL-46 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-17	SJL-47 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-18	SJL-48 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-19	SJL-49 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-20	SJL-50 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-21	SJL-51 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-22	SJL-52 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-23	SJL-55 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15
810-114270-24	SJL-57 Sink	Drinking Water	08/01/24 08:00	08/05/24 09:15

Environmental Analytical: South Bend
110 South Hill Street
South Bend, IN 46617
Phone (574) 233-4777

Chain of Custody Record

eurofins
Environment Testing
America

2nd Dean

Client Information		Sampler: Ian St. John		Lab P/N: _____		Carrier Tracking No(s): _____		COC No: _____	
Client Contact: Casey Housman		Phone: 314-853-4668		E-Mail: ian@stjohnenv.com		State of Origin: MO		Page: 1 of 3 * KW 8/5/24	
Company: St. Joseph School District		Due Date Requested: _____		PWSD: _____		Analysis Requested		Job #: _____	
Address: 1415 North 26th Street		TAT Requested (days): Standard		Compliance Project: ☐ Yes ☐ No		Barcode: 810-114270 Chain of Custody		Preservation Codes:	
City: St. Joseph		PO #: 816-671-4000		WO #: _____		Matrix: (W=water, S=solid, O=oil, A=air)		A - HCL B - NaOH C - Zn Acetate D - Nitric Acid E - NaHSO4 F - MeOH G - Amchlor H - Ascorbic Acid I - Ice J - DI Water K - EDTA L - EDA Other: _____	
State, Zip: Missouri 64506		Project #: _____		SSOW#: _____		Field Filtered Sample (Yes or No)		M - Hexane N - None O - AsNaO2 P - Na2O4S Q - Na2SO3 R - Na2SO3 S - H2SO4 T - TSP Dodecahydrate U - Acetone V - MCAA W - pH 4-5 Y - Trizma Z - other (specify)	
Phone: _____		Sample Date: 8/1/24 0800		Sample Time: 6		Perform MS/MSD (Yes or No)		Total Number of containers	
Email: CaseyHousman@sjsd.k12.mo.us		Sample Type: (C=Comp, G=grab)		Preservation Code: W		Field Filtered Sample (Yes or No)		Special Instructions/Note:	
Project Name: St. Joseph School District		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)		*Rec'd unpreserved. Will preserve upon receipt * KW 8/5/24	
Site: Lafayette		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)		8/5/24	
Sample Identification		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)			
STL-7 sink		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)			
STL-8 sink		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)			
STL-9 sink		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)			
STL-11 sink		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)			
STL-12 sink		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)			
STL-16 sink		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)			
STL-17 sink		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)			
STL-35 sink		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)			
STL-37 sink		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)			
STL-38 sink		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)			
STL-40 sink		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)			
Possible Hazard Identification		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)			
☐ Non-Hazard ☐ Flammable ☐ Skin Irritant ☐ Poison B ☒ Unknown ☐ Radiological		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)			
Deliverable Requested: I, II, III, IV, Other (specify)		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)			
Empty Kit Relinquished by: _____		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)			
Relinquished by: _____		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)			
Relinquished by: _____		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)			
Custody Seals Intact: ☐ Yes ☐ No		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)			
Custody Seal No.: _____		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)			
Cooler Temperature(s) °C and Other Remarks: Ambient		Sample Date: 8/1/24 0800		Sample Time: 6		Field Filtered Sample (Yes or No)			

Environmental Analytical: South Bend
110 South Hill Street
South Bend, IN 46617
Phone (574) 233-4777

Chain of Custody Record

eurofins
Environment Testing
America

2nd Dean

Client Information		Sampler: Ian St. John		Lab PMT:		Carrier Tracking No(s):		COC No:	
Client Contact:		Phone: 314-853-4668		E-Mail: ian@stjohnenv.com		State of Origin: MO		Page 3 of 3 # KW 8/5/24	
Company: Casey Housman		IPWSID:		Analysis Requested		Total Number of Containers		Special Instructions/Note:	
Address: St. Joseph School District		Due Date Requested:		Field Filtered Sample (Yes or No)		Perform MS/MSD (Yes or No)		Preservation Codes:	
City: 1415 North 26th Street		TAT Requested (days): Standard		Sample Date		Sample Type (C=Comp, G=grab)		Matrix (W=water, S=solid, O=oil, A=air)	
State, Zip: Missouri 64506		Compliance Project: ☐ Yes ☐ No		Sample Time		Sample Code		Preservation Code: W	
Phone: 816-671-4000		PO #:		Sample Date		Sample Type		Matrix	
Email: CaseyHousman@sjsd.k12.mo.us		WO #:		Sample Date		Sample Type		Matrix	
Project Name: St. Joseph School District		Project #:		Sample Date		Sample Type		Matrix	
Site: Lafayette		SSOW#:		Sample Date		Sample Type		Matrix	
Sample Identification		Sample Date		Sample Time		Sample Code		Preservation Code	
SJL- 41 sink		8/1/24 0800		G		W			
SJL- 42 sink									
SJL- 43 sink									
SJL- 45 sink									
SJL- 46 sink									
SJL- 47 sink									
SJL- 48 sink									
SJL- 49 sink									
SJL- 50 sink									
SJL- 51 sink									
SJL- 52 sink									
Possible Hazard Identification		☐ Non-Hazard ☐ Flammable ☐ Skin Irritant ☐ Poison B ☒ Unknown ☐ Radiological		Deliverable Requested: I, II, III, IV, Other (specify)		Empty Kit Relinquished by:		Relinquished by:	
Relinquished by:		Date: 8/1/24		Company:		Relinquished by:		Date: 8/1/24	
Relinquished by:		Date:		Company:		Relinquished by:		Date:	
Relinquished by:		Date:		Company:		Relinquished by:		Date:	
Custody Seals Intact: ☐ Yes ☐ No		Custody Seal No.:		Cooler Temperature(s) °C and Other Remarks: Ambient		Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)		Return To Client ☐ Disposal By Lab ☐ Archive For ☐ Months	

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

Login Sample Receipt Checklist

Client: St. John Environmental Consulting

Job Number: 810-114270-1

Login Number: 114270

List Source: Eurofins Eaton Analytical South Bend

List Number: 1

Creator: Williams, Kameron

Question	Answer	Comment
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
Samples were received on ice.	False	Thermal preservation not required.
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	False	Thermal preservation not required.
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	