



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

December 10, 2021

John Harding
Eagle Point School District #9
PO Box 548
Eagle Point, OR 97524
TEL: (541) 830-1240
FAX (541) 830-6375

RE: Shady Cove School Lead and Copper

Order No.: 21111210

Dear John Harding:

Neilson Research Corporation received 8 sample(s) on 11/30/2021 for the analyses presented in the following report.


The results relate only to the parameters tested or to the sample as received by the laboratory. This report shall not be reproduced except in full, without the written approval of Neilson Research Corporation. If you have any questions regarding these test results, please feel free to call.

Sincerely,
Neilson Research Corporation

Tamra Schmedemann
Senior Project Manager
245 S Grape St
Medford, OR 97501



Original



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Case Narrative

WO#: 21111210
Date: 12/10/2021

CLIENT: Eagle Point School District #9

Project: Shady Cove School Lead and Copper

The analyses were performed according to the guidelines in the Neilson Research Corporation Quality Assurance Program. This report contains analytical results for the sample(s) as received by the laboratory.

Neilson Research Corporation certifies that this report is in compliance with the requirements of NELAP. No unusual difficulties were experienced during analysis of this batch except as noted below or qualified with data flags on the reports.

Original



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 2111210

Date Reported: 12/10/2021

CLIENT: Eagle Point School District #9
Lab ID: 2111210-01A
Client Sample ID Bottle #65860
Project: Shady Cove School Lead and Copper
Sample Location: Staff Room Sink
Sample Address:

Collection Date: 11/30/2021 7:47:00 AM
Received Date: 11/30/2021 10:31:00 AM
Matrix: DRINKING WATER
PWS #: 41-91511
Source ID: DIST-A
Sample Collector: STEVE LAMBERT

Analyses	Code	Method	NELAP		Qual	DF	RL Units	Date		Analyst
			Status	Result				MCL	Analyzed	
Copper	1022	E200.8	A	ND	1		0.00206 mg/L	1.30	12/02/21 21:12	SJS
Lead	1030	E200.8	A	ND	1		0.000515 mg/L	0.0150	12/02/21 21:12	SJS

QUALIFIERS

C1	Sample container temperature is out of limit as specified at testcode	H	Holding times for preparation or analysis exceeded
MI	Recovery outside control limits due to Matrix Interference	ND	Not Detected at the Reporting Limit
PL	Permit Limit		

Original

NELAP

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 2111210
Date Reported: 12/10/2021

CLIENT: Eagle Point School District #9
Lab ID: 2111210-02A
Client Sample ID: Bottle #64284
Project: Shady Cove School Lead and Copper
Sample Location: Room 1
Sample Address:

Collection Date: 11/30/2021 7:45:00 AM
Received Date: 11/30/2021 10:31:00 AM
Matrix: DRINKING WATER
PWS #: 41-91511
Source ID: DIST-A
Sample Collector: STEVE LAMBERT

Analyses	Code	Method	NELAP		Qual	RL Units	Date		Analyst
			Status	Result	DF		MCL	Analyzed	
Copper	1022	E200.8	A	ND	1	0.00206 mg/L	1.30	12/02/21 21:16	SJS
Lead	1030	E200.8	A	0.00274	1	0.000515 mg/L	0.0150	12/02/21 21:16	SJS

QUALIFIERS

C1	Sample container temperature is out of limit as specified at testcode	H	Holding times for preparation or analysis exceeded
MI	Recovery outside control limits due to Matrix Interference	ND	Not Detected at the Reporting Limit
PL	Permit Limit		

Original

NELAP

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 2111210

Date Reported: 12/10/2021

CLIENT: Eagle Point School District #9
Lab ID: 2111210-03A
Client Sample ID Bottle #61949
Project: Shady Cove School Lead and Copper
Sample Location: Kitchen Hand Wash Sink
Sample Address:

Collection Date: 11/30/2021 8:23:00 AM
Received Date: 11/30/2021 10:31:00 AM
Matrix: DRINKING WATER
PWS #: 41-91511
Source ID: DIST-A
Sample Collector: STEVE LAMBERT

Analyses	Code	Method	NELAP		Qual	RL Units	Date		Analyst
			Status	Result	DF		MCL	Analyzed	
Copper	1022	E200.8	A	0.0150	1	0.00206 mg/L	1.30	12/02/21 21:20	SJS
Lead	1030	E200.8	A	ND	1	0.000515 mg/L	0.0150	12/02/21 21:20	SJS

QUALIFIERS

C1	Sample container temperature is out of limit as specified at testcode	H	Holding times for preparation or analysis exceeded
MI	Recovery outside control limits due to Matrix Interference	ND	Not Detected at the Reporting Limit
PL	Permit Limit		

Original

NELAP

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 2111210
Date Reported: 12/10/2021

CLIENT: Eagle Point School District #9
Lab ID: 2111210-04A
Client Sample ID Bottle #61948
Project: Shady Cove School Lead and Copper
Sample Location: Room 10
Sample Address:

Collection Date: 11/30/2021 7:50:00 AM
Received Date: 11/30/2021 10:31:00 AM
Matrix: DRINKING WATER
PWS #: 41-91511
Source ID: DIST-A
Sample Collector: STEVE LAMBERT

Analyses	Code	Method	NELAP		Qual		RL Units	Date		Analyst
			Status	Result	DF	MCL		Analyzed		
Copper	1022	E200.8	A	ND	1	0.00206 mg/L	1.30	12/02/21 21:25	SJS	
Lead	1030	E200.8	A	ND	1	0.000515 mg/L	0.0150	12/02/21 21:25	SJS	

QUALIFIERS

C1	Sample container temperature is out of limit as specified at testcode	H	Holding times for preparation or analysis exceeded
MI	Recovery outside control limits due to Matrix Interference	ND	Not Detected at the Reporting Limit
PL	Permit Limit		

Original

NELAP

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 2111210

Date Reported: 12/10/2021

CLIENT: Eagle Point School District #9
Lab ID: 2111210-05A
Client Sample ID Bottle #03508
Project: Shady Cove School Lead and Copper
Sample Location: Wellhead 1
Sample Address:

Collection Date: 11/30/2021 7:56:00 AM
Received Date: 11/30/2021 10:31:00 AM
Matrix: DRINKING WATER
PWS #: 41-91511
Source ID: EP-A
Sample Collector: STEVE LAMBERT

Analyses	Code	Method	NELAP		Qual	DF	RL Units	Date		Analyst
			Status	Result				MCL	Analyzed	
Copper	1022	E200.8	A	ND	1		0.00206 mg/L	1.30	12/02/21 21:29	SJS
Lead	1030	E200.8	A	ND	1		0.000515 mg/L	0.0150	12/02/21 21:29	SJS

QUALIFIERS

C1	Sample container temperature is out of limit as specified at testcode	H	Holding times for preparation or analysis exceeded
MI	Recovery outside control limits due to Matrix Interference	ND	Not Detected at the Reporting Limit
PL	Permit Limit		

Original

NELAP

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 2111210

Date Reported: 12/10/2021

CLIENT: Eagle Point School District #9
Lab ID: 2111210-06A
Client Sample ID Bottle #65850
Project: Shady Cove School Lead and Copper
Sample Location: Portable RR
Sample Address:

Collection Date: 11/30/2021 8:05:00 AM
Received Date: 11/30/2021 10:31:00 AM
Matrix: DRINKING WATER
PWS #: 41-91511
Source ID: DIST-A
Sample Collector: STEVE LAMBERT

Analyses	Code	Method	NELAP		Qual	RL Units	Date		Analyst
			Status	Result	DF		MCL	Analyzed	
Copper	1022	E200.8	A	0.0225	1	0.00206 mg/L	1.30	12/02/21 21:33	SJS
Lead	1030	E200.8	A	ND	1	0.000515 mg/L	0.0150	12/02/21 21:33	SJS

QUALIFIERS

C1	Sample container temperature is out of limit as specified at testcode	H	Holding times for preparation or analysis exceeded
MI	Recovery outside control limits due to Matrix Interference	ND	Not Detected at the Reporting Limit
PL	Permit Limit		

Original

NELAP

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 2111210
Date Reported: 12/10/2021

CLIENT: Eagle Point School District #9
Lab ID: 2111210-07A
Client Sample ID: Bottle #65854
Project: Shady Cove School Lead and Copper
Sample Location: MS RR
Sample Address:

Collection Date: 11/30/2021 8:09:00 AM
Received Date: 11/30/2021 10:31:00 AM
Matrix: DRINKING WATER
PWS #: 41-91511
Source ID: DIST-A
Sample Collector: STEVE LAMBERT

Analyses	Code	Method	NELAP		Qual		RL Units	Date		Analyst
			Status	Result	DF	MCL		Analyzed		
Copper	1022	E200.8	A	0.122	1	0.00206 mg/L	1.30	12/02/21 21:38	SJS	
Lead	1030	E200.8	A	ND	1	0.000515 mg/L	0.0150	12/02/21 21:38	SJS	

QUALIFIERS

C1	Sample container temperature is out of limit as specified at testcode	H	Holding times for preparation or analysis exceeded
MI	Recovery outside control limits due to Matrix Interference	ND	Not Detected at the Reporting Limit
PL	Permit Limit		

Original

NELAP

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Analytical Report

WO#: 2111210
Date Reported: 12/10/2021

CLIENT: Eagle Point School District #9
Lab ID: 2111210-08A
Client Sample ID Bottle #77312
Project: Shady Cove School Lead and Copper
Sample Location: Well 2
Sample Address:

Collection Date: 11/30/2021 8:15:00 AM
Received Date: 11/30/2021 10:31:00 AM
Matrix: DRINKING WATER
PWS #: 41-91511
Source ID: EP-B
Sample Collector: STEVE LAMBERT

Analyses	Code	Method	NELAP		Qual	RL Units	Date		Analyst
			Status	Result	DF		MCL	Analyzed	
Copper	1022	E200.8	A	ND	1	0.00206 mg/L	1.30	12/02/21 21:42	SJS
Lead	1030	E200.8	A	ND	1	0.000515 mg/L	0.0150	12/02/21 21:42	SJS

QUALIFIERS

C1	Sample container temperature is out of limit as specified at testcode	H	Holding times for preparation or analysis exceeded
MI	Recovery outside control limits due to Matrix Interference	ND	Not Detected at the Reporting Limit
PL	Permit Limit		

Original

NELAP

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

QC SUMMARY REPORT

WO#: 21111210
10-Dec-21

Client: Eagle Point School District #9
Project: Shady Cove School Lead and Copper

TestCode: ICPMS_200.8_DW

Sample ID: MB-15007	SampType: MBLK	TestCode: ICPMS_200.8	Units: mg/L	Prep Date: 12/1/2021	RunNo: 26370						
Client ID: PBW	Batch ID: 15007	TestNo: E200.8	E200.8	Analysis Date: 12/2/2021	SeqNo: 423693						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Copper	ND	0.00206									
Lead	ND	0.000515									

Sample ID: LCS-15007	SampType: LCS	TestCode: ICPMS_200.8	Units: mg/L	Prep Date: 12/1/2021	RunNo: 26370						
Client ID: LCSW	Batch ID: 15007	TestNo: E200.8	E200.8	Analysis Date: 12/2/2021	SeqNo: 423694						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Copper	0.100	0.00208	0.1000	0	100	85	115				
Lead	0.101	0.000520	0.1000	0	101	85	115				

Sample ID: 21111188-01AMS	SampType: MS	TestCode: ICPMS_200.8	Units: mg/L	Prep Date: 12/1/2021	RunNo: 26370						
Client ID: BatchQC	Batch ID: 15007	TestNo: E200.8	E200.8	Analysis Date: 12/2/2021	SeqNo: 423696						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Copper	0.171	0.00208	0.1000	0.08207	88.5	70	130				
Lead	0.0961	0.000520	0.1000	0.0001215	96.0	70	130				

Sample ID: 21111188-01AMSD	SampType: MSD	TestCode: ICPMS_200.8	Units: mg/L	Prep Date: 12/1/2021	RunNo: 26370						
Client ID: BatchQC	Batch ID: 15007	TestNo: E200.8	E200.8	Analysis Date: 12/2/2021	SeqNo: 423697						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Copper	0.172	0.00208	0.1000	0.08207	89.8	70	130	0.1706	0.739	20	

Qualifiers: CI Sample container temperature is out of limit as specified at testcode H Holding times for preparation or analysis exceeded MI Recovery outside control limits due to Matrix Int
ND Not Detected at the Reporting Limit PL Permit Limit RL Reporting Detection Limit

Original



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

QC SUMMARY REPORT

WO#: 21111210
10-Dec-21

Client: Eagle Point School District #9

Project: Shady Cove School Lead and Copper

TestCode: ICPMS_200.8_DW

Sample ID: 21111188-01AMSD	SampType: MSD	TestCode: ICPMS_200.8	Units: mg/L	Prep Date: 12/1/2021	RunNo: 26370						
Client ID: BatchQC	Batch ID: 15007	TestNo: E200.8	E200.8	Analysis Date: 12/2/2021	SeqNo: 423697						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Lead	0.0968	0.000520	0.1000	0.0001215	96.7	70	130	0.09609	0.773	20	

Qualifiers: CI Sample container temperature is out of limit as specified at testcode
ND Not Detected at the Reporting Limit

H Holding times for preparation or analysis exceeded
PL Permit Limit

MI Recovery outside control limits due to Matrix In
RL Reporting Detection Limit

Original

Sample Log-In Check List

Client Name: **EAGLEPTSCHOOL**

Work Order Number: **21111210**

RcptNo: **1**

Logged by: **Haylee Crowe** **11/30/2021 10:31:00 AM**

Completed By: **Krizzle Calip** **12/1/2021 10:37:03 AM**

Reviewed By: **Dorie Maier** **12/10/2021 11:01:36 AM**

Haylee Crowe
Krizzle Calip
Dorie Maier

Chain of Custody

1. Is Chain of Custody complete? Yes ☒ No ☐ Not Present ☐
2. How was the sample delivered? Client

Log In

3. Coolers are present? Yes ☐ No ☐ NA ☒
4. Shipping container/cooler in good condition? Yes ☒ No ☐
Custody seals intact on shipping container/cooler? Yes ☐ No ☐ Not Present ☒
No. Seal Date: Signed By:
5. Was an attempt made to cool the samples? Yes ☐ No ☐ NA ☒
6. Were all samples received at a temperature of >0° C to 6.0°C Yes ☐ No ☐ NA ☒
7. Sample(s) in proper container(s)? Yes ☒ No ☐
8. Sufficient sample volume for indicated test(s)? Yes ☒ No ☐
9. Are samples (except VOA and ONG) properly preserved? Yes ☒ No ☐
10. Was preservative added to bottles? Yes ☒ No ☐ NA ☐
11. Is the headspace in the VOA vials less than 1/4 inch or 6 mm? Yes ☐ No ☐ HNO₃ pH<2
No VOA Vials ☒
12. Were any sample containers received broken? Yes ☐ No ☒
13. Does paperwork match bottle labels? Yes ☒ No ☐
(Note discrepancies on chain of custody)
14. Are matrices correctly identified on Chain of Custody? Yes ☒ No ☐
15. Is it clear what analyses were requested? Yes ☒ No ☐
16. Were all holding times able to be met? Yes ☒ No ☐
(If no, notify customer for authorization.)

Special Handling (if applicable)

17. Was client notified of all discrepancies with this order? Yes ☐ No ☐ NA ☒

Person Notified: Date:
By Whom: Via: ☐ eMail ☐ Phone ☐ Fax ☐ In Person
Regarding:
Client Instructions:

18. Additional remarks:

The samples submitted for Sample IDs 21111210-05A and -08A contained visible sediments.

Cooler Information

Cooler No	Temp °C	Condition	Seal Intact	Seal No	Seal Date	Signed By
-----------	---------	-----------	-------------	---------	-----------	-----------

Lab Sample ID
Date Received
Time Received
Received By

2111210 01A
11/30/21
10:31
HMC

Lead & Copper First Draw Sample Collection Procedures

These samples are being collected to determine lead and copper levels in your tap water. This sampling effort is required by the U.S. Environmental Protection Agency and your state, and is being accomplished through the cooperation of homeowners and residents.

Collect all water samples before the facility opens and before any water is used. Ideally, the water should sit in the pipes unused for at least 6 hours but not more than 18 hours before a sample is taken. Make sure that no water is withdrawn from the taps or fountains from which the samples are to be collected prior to their sampling.

Unless specifically directed to do so, do not collect samples in the morning after vacations, weekends, or holidays because the water will have remained stagnant for too long and would not represent the water used for drinking during most of the days of the week.

1. Prior arrangement will be made with customer to coordinate the sample collection event. Dates will be set for a sample kit delivery and pick-up by the water department staff.
2. A kitchen or bathroom cold water faucet is to be used for sampling. Place the open sample bottle below the faucet and gently open the cold water tap. Fill the sample bottle to the neck and turn off the water.
3. Tightly cap the sample bottle. Please carefully complete this form.
4. IF ANY PLUMBING REPAIRS OR REPLACEMENT HAVE BEEN DONE IN THE HOME SINCE THE PREVIOUS SAMPLING EVENT, NOTE THIS INFORMATION ON THIS FORM BELOW.
5. Place the sample with form attached outside of the residence in the location of the delivery for pick up.
6. Results from this monitoring effort will be provided to participating customers when reports are generated for the State unless excessive lead and/or copper levels are found. In those cases, immediate notification will be provided, usually 10 working days from the time of sample collection.

If you have any questions please call: _____

TO BE COMPLETED BY RESIDENT

Water was last used:

Time 6:00 (am/pm) Date 11/29/21

Sample was collected:

Time 7:47 (am/pm) Date 11/30/21

Name of Water System

Shady Cove School

PWS ID 41-

91511

Sample Collected by

Steven Lambert

Bottle #

65860

Address

37 School house Ln Shady Cove

Space #

Faucet Location

Staff room Sink

Note any plumbing repairs or replacements made since last sampling event:

I have read the above directions and have taken a tap sample in accordance with these directions.

Signature

Steven Lambert

Date 11/30/21

Lab Sample ID

2111210 02A

Date Received

11 / 30 / 21

Time Received

10 : 31

Received By

HMC

Lead & Copper First Draw Sample Collection Procedures

These samples are being collected to determine lead and copper levels in your tap water. This sampling effort is required by the U.S. Environmental Protection Agency and your state, and is being accomplished through the cooperation of homeowners and residents.

Collect all water samples before the facility opens and before any water is used. Ideally, the water should sit in the pipes unused for at least 6 hours but not more than 18 hours before a sample is taken. Make sure that no water is withdrawn from the taps or fountains from which the samples are to be collected prior to their sampling.

Unless specifically directed to do so, do not collect samples in the morning after vacations, weekends, or holidays because the water will have remained stagnant for too long and would not represent the water used for drinking during most of the days of the week.

1. Prior arrangement will be made with customer to coordinate the sample collection event. Dates will be set for a sample kit delivery and pick-up by the water department staff.
2. A kitchen or bathroom cold water faucet is to be used for sampling. Place the open sample bottle below the faucet and gently open the cold water tap. Fill the sample bottle to the neck and turn off the water.
3. Tightly cap the sample bottle. Please carefully complete this form.
4. IF ANY PLUMBING REPAIRS OR REPLACEMENT HAVE BEEN DONE IN THE HOME SINCE THE PREVIOUS SAMPLING EVENT, NOTE THIS INFORMATION ON THIS FORM BELOW.
5. Place the sample with form attached outside of the residence in the location of the delivery for pick up.
6. Results from this monitoring effort will be provided to participating customers when reports are generated for the State unless excessive lead and/or copper levels are found. In those cases, immediate notification will be provided, usually 10 working days from the time of sample collection.

If you have any questions please call: _____.

TO BE COMPLETED BY RESIDENT

Water was last used:

Time 6:00 (am/pm) Date 11/29/21

Sample was collected:

Time 7:45 (am/pm) Date 11/30/21

Name of Water System

Shady Cove School

PWS ID 41-

91511

Sample Collected by

Steven Lambert

Bottle #

64284

Address

37 Schoolhouse Ln Shady Cove

Space #


Faucet Location

Room 1

Note any plumbing repairs or replacements made since last sampling event:

I have read the above directions and have taken a tap sample in accordance with these directions.

Signature



Date

11/30/21

Lab Sample ID 2111210 03A
Date Received 11/30/21
Time Received 10:31
Received By HMC

Lead & Copper First Draw Sample Collection Procedures

These samples are being collected to determine lead and copper levels in your tap water. This sampling effort is required by the U.S. Environmental Protection Agency and your state, and is being accomplished through the cooperation of homeowners and residents.

Collect all water samples before the facility opens and before any water is used. Ideally, the water should sit in the pipes unused for at least 6 hours but not more than 18 hours before a sample is taken. Make sure that no water is withdrawn from the taps or fountains from which the samples are to be collected prior to their sampling.

Unless specifically directed to do so, do not collect samples in the morning after vacations, weekends, or holidays because the water will have remained stagnant for too long and would not represent the water used for drinking during most of the days of the week.

1. Prior arrangement will be made with customer to coordinate the sample collection event. Dates will be set for a sample kit delivery and pick-up by the water department staff.
2. A kitchen or bathroom cold water faucet is to be used for sampling. Place the open sample bottle below the faucet and gently open the cold water tap. Fill the sample bottle to the neck and turn off the water.
3. Tightly cap the sample bottle. Please carefully complete this form.
4. IF ANY PLUMBING REPAIRS OR REPLACEMENT HAVE BEEN DONE IN THE HOME SINCE THE PREVIOUS SAMPLING EVENT, NOTE THIS INFORMATION ON THIS FORM BELOW.
5. Place the sample with form attached outside of the residence in the location of the delivery for pick up.
6. Results from this monitoring effort will be provided to participating customers when reports are generated for the State unless excessive lead and/or copper levels are found. In those cases, immediate notification will be provided, usually 10 working days from the time of sample collection.

If you have any questions please call: _____.

TO BE COMPLETED BY RESIDENT

Water was last used: Time 5:30 (am/pm) Date 11/29/21
Sample was collected: Time 8:23 (am/pm) Date 11/30/21
Name of Water System Shady Cove School PWS ID 41- 91511
Sample Collected by Steven Lambert Bottle # 61949
Address 37 School house in Shady Cove Space # _____
Faucet Location Kitchen hand wash sink
Note any plumbing repairs or replacements made since last sampling event:

I have read the above directions and have taken a tap sample in accordance with these directions.

Signature Steven Lambert Date 11/30/21

Lab Sample ID
Date Received
Time Received
Received By

2111210 04A
11 / 30 / 21
10 : 31
HMC

Lead & Copper First Draw Sample Collection Procedures

These samples are being collected to determine lead and copper levels in your tap water. This sampling effort is required by the U.S. Environmental Protection Agency and your state, and is being accomplished through the cooperation of homeowners and residents.

Collect all water samples before the facility opens and before any water is used. Ideally, the water should sit in the pipes unused for at least 6 hours but not more than 18 hours before a sample is taken. Make sure that no water is withdrawn from the taps or fountains from which the samples are to be collected prior to their sampling.

Unless specifically directed to do so, do not collect samples in the morning after vacations, weekends, or holidays because the water will have remained stagnant for too long and would not represent the water used for drinking during most of the days of the week.

1. Prior arrangement will be made with customer to coordinate the sample collection event. Dates will be set for a sample kit delivery and pick-up by the water department staff.
2. A kitchen or bathroom cold water faucet is to be used for sampling. Place the open sample bottle below the faucet and gently open the cold water tap. Fill the sample bottle to the neck and turn off the water.
3. Tightly cap the sample bottle. Please carefully complete this form.
4. IF ANY PLUMBING REPAIRS OR REPLACEMENT HAVE BEEN DONE IN THE HOME SINCE THE PREVIOUS SAMPLING EVENT, NOTE THIS INFORMATION ON THIS FORM BELOW.
5. Place the sample with form attached outside of the residence in the location of the delivery for pick up.
6. Results from this monitoring effort will be provided to participating customers when reports are generated for the State unless excessive lead and/or copper levels are found. In those cases, immediate notification will be provided, usually 10 working days from the time of sample collection.

If you have any questions please call: _____.

TO BE COMPLETED BY RESIDENT	
Water was last used:	Time <u>7:00</u> (am/pm) Date <u>11/29/21</u>
Sample was collected:	Time <u>7:50</u> (am/pm) Date <u>11/30/21</u>
Name of Water System <u>ShadyCove School</u>	FWS ID 41- <u>91511</u>
Sample Collected by <u>Steven Lambert</u>	Bottle # <u>61948</u>
Address <u>37 School house Ln ShadyCove</u>	Space # _____
Faucet Location <u>RM 10</u>	
Note any plumbing repairs or replacements made since last sampling event:	
I have read the above directions and have taken a tap sample in accordance with these directions.	
Signature <u>Steven Lambert</u>	Date <u>11/30/21</u>

Lab Sample ID

2111210 OSA

Date Received

11/30/21

Time Received

10:31

Received By

HMC

Lead & Copper First Draw Sample Collection Procedures

These samples are being collected to determine lead and copper levels in your tap water. This sampling effort is required by the U.S. Environmental Protection Agency and your state, and is being accomplished through the cooperation of homeowners and residents.

Collect all water samples before the facility opens and before any water is used. Ideally, the water should sit in the pipes unused for at least 6 hours but not more than 18 hours before a sample is taken. Make sure that no water is withdrawn from the taps or fountains from which the samples are to be collected prior to their sampling.

Unless specifically directed to do so, do not collect samples in the morning after vacations, weekends, or holidays because the water will have remained stagnant for too long and would not represent the water used for drinking during most of the days of the week.

1. Prior arrangement will be made with customer to coordinate the sample collection event. Dates will be set for a sample kit delivery and pick-up by the water department staff.
2. A kitchen or bathroom cold water faucet is to be used for sampling. Place the open sample bottle below the faucet and gently open the cold water tap. Fill the sample bottle to the neck and turn off the water.
3. Tightly cap the sample bottle. Please carefully complete this form.
4. IF ANY PLUMBING REPAIRS OR REPLACEMENT HAVE BEEN DONE IN THE HOME SINCE THE PREVIOUS SAMPLING EVENT, NOTE THIS INFORMATION ON THIS FORM BELOW.
5. Place the sample with form attached outside of the residence in the location of the delivery for pick up.
6. Results from this monitoring effort will be provided to participating customers when reports are generated for the State unless excessive lead and/or copper levels are found. In those cases, immediate notification will be provided, usually 10 working days from the time of sample collection.

If you have any questions please call: _____.

TO BE COMPLETED BY RESIDENT

Water was last used:

Time 6:00 (am/pm) Date 11/29/21

Sample was collected:

Time 7:56 (am/pm) Date 11/30/21

Name of Water System

Shady Cove SchoolPWS ID 41- 91511

Sample Collected by

Steven LambertBottle # 03508

Address

37 School house in Shady Cove

Space # _____

Faucet Location

Well head 1

Note any plumbing repairs or replacements made since last sampling event:

I have read the above directions and have taken a tap sample in accordance with these directions.

Signature

Steven LambertDate 11/30/21

Lab Sample ID

21111210 06A

Date Received

11 / 30 / 21

Time Received

10 : 51

Received By

HMC

Lead & Copper First Draw Sample Collection Procedures

These samples are being collected to determine lead and copper levels in your tap water. This sampling effort is required by the U.S. Environmental Protection Agency and your state, and is being accomplished through the cooperation of homeowners and residents.

Collect all water samples before the facility opens and before any water is used. Ideally, the water should sit in the pipes unused for at least 6 hours but not more than 18 hours before a sample is taken. Make sure that no water is withdrawn from the taps or fountains from which the samples are to be collected prior to their sampling.

Unless specifically directed to do so, do not collect samples in the morning after vacations, weekends, or holidays because the water will have remained stagnant for too long and would not represent the water used for drinking during most of the days of the week.

1. Prior arrangement will be made with customer to coordinate the sample collection event. Dates will be set for a sample kit delivery and pick-up by the water department staff.
2. A kitchen or bathroom cold water faucet is to be used for sampling. Place the open sample bottle below the faucet and gently open the cold water tap. Fill the sample bottle to the neck and turn off the water.
3. Tightly cap the sample bottle. Please carefully complete this form.
4. IF ANY PLUMBING REPAIRS OR REPLACEMENT HAVE BEEN DONE IN THE HOME SINCE THE PREVIOUS SAMPLING EVENT, NOTE THIS INFORMATION ON THIS FORM BELOW.
5. Place the sample with form attached outside of the residence in the location of the delivery for pick up.
6. Results from this monitoring effort will be provided to participating customers when reports are generated for the State unless excessive lead and/or copper levels are found. In those cases, immediate notification will be provided, usually 10 working days from the time of sample collection.

If you have any questions please call: _____.

TO BE COMPLETED BY RESIDENT

Water was last used:

Time 5:30 (am/pm) Date 11/29/21

Sample was collected:

Time 8:05 (am/pm) Date 11/30/21

Name of Water System

ShadyCove School

PWS ID 41-

91511

Sample Collected by

Steven Lambert

Bottle #

65850

Address

37 School house Ln ShadyCove

Space #

Faucet Location

portable R.R.

Note any plumbing repairs or replacements made since last sampling event:

I have read the above directions and have taken a tap sample in accordance with these directions.

Signature

Steven Lambert

Date

11/30/21

Lab Sample ID
Date Received
Time Received
Received By

2111210 07A
11 / 30 / 21
10 : 31
ITMC

Lead & Copper First Draw Sample Collection Procedures

These samples are being collected to determine lead and copper levels in your tap water. This sampling effort is required by the U.S. Environmental Protection Agency and your state, and is being accomplished through the cooperation of homeowners and residents.

Collect all water samples before the facility opens and before any water is used. Ideally, the water should sit in the pipes unused for at least 6 hours but not more than 18 hours before a sample is taken. Make sure that no water is withdrawn from the taps or fountains from which the samples are to be collected prior to their sampling.

Unless specifically directed to do so, do not collect samples in the morning after vacations, weekends, or holidays because the water will have remained stagnant for too long and would not represent the water used for drinking during most of the days of the week.

1. Prior arrangement will be made with customer to coordinate the sample collection event. Dates will be set for a sample kit delivery and pick-up by the water department staff.
2. A kitchen or bathroom cold water faucet is to be used for sampling. Place the open sample bottle below the faucet and gently open the cold water tap. Fill the sample bottle to the neck and turn off the water.
3. Tightly cap the sample bottle. Please carefully complete this form.
4. IF ANY PLUMBING REPAIRS OR REPLACEMENT HAVE BEEN DONE IN THE HOME SINCE THE PREVIOUS SAMPLING EVENT, NOTE THIS INFORMATION ON THIS FORM BELOW.
5. Place the sample with form attached outside of the residence in the location of the delivery for pick up.
6. Results from this monitoring effort will be provided to participating customers when reports are generated for the State unless excessive lead and/or copper levels are found. In those cases, immediate notification will be provided, usually 10 working days from the time of sample collection.

If you have any questions please call: _____

TO BE COMPLETED BY RESIDENT

Water was last used:

Time 5:30 (am/pm) Date 11/29/21

Sample was collected:

Time 8:09 (am/pm) Date 11/30/21

Name of Water System

ShadyCove School

PWS ID 41-

91511

Sample Collected by

Steven Lambert

Bottle #

65854

Address

37 School house in ShadyCove

Space #

Faucet Location

MS RR

Note any plumbing repairs or replacements made since last sampling event:

I have read the above directions and have taken a tap sample in accordance with these directions.

Signature

Steven Lambert

Date

Lab Sample ID
Date Received
Time Received
Received By

2111210 08A
11 / 30 / 21
10:31
HMC

Lead & Copper First Draw Sample Collection Procedures

These samples are being collected to determine lead and copper levels in your tap water. This sampling effort is required by the U.S. Environmental Protection Agency and your state, and is being accomplished through the cooperation of homeowners and residents.

Collect all water samples before the facility opens and before any water is used. Ideally, the water should sit in the pipes unused for at least 6 hours but not more than 18 hours before a sample is taken. Make sure that no water is withdrawn from the taps or fountains from which the samples are to be collected prior to their sampling.

Unless specifically directed to do so, do not collect samples in the morning after vacations, weekends, or holidays because the water will have remained stagnant for too long and would not represent the water used for drinking during most of the days of the week.

1. Prior arrangement will be made with customer to coordinate the sample collection event. Dates will be set for a sample kit delivery and pick-up by the water department staff.
2. A kitchen or bathroom cold water faucet is to be used for sampling. Place the open sample bottle below the faucet and gently open the cold water tap. Fill the sample bottle to the neck and turn off the water.
3. Tightly cap the sample bottle. Please carefully complete this form.
4. IF ANY PLUMBING REPAIRS OR REPLACEMENT HAVE BEEN DONE IN THE HOME SINCE THE PREVIOUS SAMPLING EVENT, NOTE THIS INFORMATION ON THIS FORM BELOW.
5. Place the sample with form attached outside of the residence in the location of the delivery for pick up.
6. Results from this monitoring effort will be provided to participating customers when reports are generated for the State unless excessive lead and/or copper levels are found. In those cases, immediate notification will be provided, usually 10 working days from the time of sample collection.

If you have any questions please call: _____.

TO BE COMPLETED BY RESIDENT

Water was last used: Time 5:30 (am/pm) Date 11/29/21
Sample was collected: Time 8:15 (am/pm) Date 11/30/21
Name of Water System Shady Cove School PWS ID 41- 91511
Sample Collected by Steven Lambert Bottle # 77312
Address 37 School house Ln Shady Cove Space # _____
Faucet Location Well 2

Note any plumbing repairs or replacements made since last sampling event:

I have read the above directions and have taken a tap sample in accordance with these directions.

Signature Steven Lambert Date 11/30/21