July 02, 2025 Service Request No:R2507356

Mr. Aaron Koss
Rush-Henrietta Central School District
2034 Lehigh Station Road
Henrietta, NY 14467

Laboratory Results for: Good Shepherd

Dear Mr.Koss,

Enclosed are the results of the sample(s) submitted to our laboratory June 20, 2025
For your reference, these analyses have been assigned our service request number **R2507356.**

All testing was performed according to our laboratory's quality assurance program and met the requirements of the TNI standards except as noted in the case narrative report. Any testing not included in the lab's accreditation is identified on a Non-Certified Analytes report. All results are intended to be considered in their entirety. ALS Environmental is not responsible for use of less than the complete report. Results apply only to the individual samples submitted to the lab for analysis, as listed in the report. The measurement uncertainty of the results included in this report is within that expected when using the prescribed method(s), and represented by Laboratory Control Sample control limits. Any events, such as QC failures or Holding Time exceedances, which may add to the uncertainty are explained in the report narrative or are flagged with qualifiers. The flags are explained in the Report Qualifiers and Definitions page of this report.

Please contact me if you have any questions. My extension is 7476. You may also contact me via email at Chris.Leavy@alsglobal.com.

Respectfully submitted,

ALS Group USA, Corp. dba ALS Environmental

Christopher Leavy
Project Manager

ADDRESS 1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623
PHONE +1 585 288 5380 | FAX +1 585 288 8475
ALS Group USA, Corp.
dba ALS Environmental

Narrative Documents

ALS Environmental—Rochester Laboratory
1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623
Phone (585) 288-5380 Fax (585) 288-8475
www.alsglobal.com

Client: Rush-Henrietta Central School District
Project: Good Shepherd
Sample Matrix: Drinking Water

Service Request: R2507356
Date Received: 06/20/2025

CASE NARRATIVE

All analyses were performed consistent with the quality assurance program of ALS Environmental. This report contains analytical results for samples for the Tier II level requested by the client.

Sample Receipt:

Six drinking water samples were received for analysis at ALS Environmental on 06/20/2025. Any discrepancies upon initial sample inspection are annotated on the sample receipt and preservation form included within this report. The samples were stored at minimum in accordance with the analytical method requirements.

Metals:

No significant anomalies were noted with this analysis.

Approved by _____ Date 07/02/2025

SAMPLE DETECTION SUMMARY

This form includes only detections above the reporting levels. For a full listing of sample results, continue to the Sample Results section of this Report.

CLIENT ID: Main Office Converence Room			Lab ID: R2507356-003			
Analyte	**Results**	**Flag**	**MDL**	**MRL**	**Units**	**Method**
Lead, Total	5.0			1.0	ug/L	200.8

CLIENT ID: Main Enterance Drinking Fountain			Lab ID: R2507356-004			
Analyte	**Results**	**Flag**	**MDL**	**MRL**	**Units**	**Method**
Lead, Total	6.3			1.0	ug/L	200.8

CLIENT ID: Health Office			Lab ID: R2507356-006			
Analyte	**Results**	**Flag**	**MDL**	**MRL**	**Units**	**Method**
Lead, Total	1.0			1.0	ug/L	200.8

Sample Receipt Information

ALS Environmental—Rochester Laboratory
1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623
Phone (585) 288-5380 Fax (585) 288-8475
www.alsglobal.com

Client: Rush-Henrietta Central School District **Service Request:**R2507356
Project: Good Shepherd

SAMPLE CROSS-REFERENCE

SAMPLE #	CLIENT SAMPLE ID	DATE	TIME
R2507356-001	Drinking Fountain Opposite 115	6/20/2025	1110
R2507356-002	Drinking Fountain Opposite 111	6/20/2025	1110
R2507356-003	Main Office Converence Room	6/20/2025	1100
R2507356-004	Main Enterance Drinking Fountain	6/20/2025	1100
R2507356-005	Rm. 108	6/20/2025	1110
R2507356-006	Health Office	6/20/2025	1110

Chain of Custody / Analytical Request Form

1565 Jefferson Road, Building 300, Suite 360 • Rochester, NY 14623 • +1 585 288 5380 • alsglobal.com

SR#:

Page of

Report To:

Company: Rush-Henrietta CSD
Contact: Ken Nelson
Email: knelson@rhnet.org
Phone: 585-738-8070
Address: 2034 Lehigh Station Road
Henrietta NY 14467

ALL SHADED AREAS MUST BE COMPLETED BY THE CLIENT / SAMPLER

Project Name: Good Shepherd
Project Number:
ALS Quote #: 38456
Sampler's Signature:
Email CC: slacey@rhnet.org
Email CC: jlschultz@rhnet.org
State Samples Collected: NY

Preservative: 0. None; 1. HCl; 2. HNO3; 3. H2SO4; 4. NAOH; 5. Zn Acet.; 6. MeOH; 7. NaHSO4; 8. Other

Lab ID (ALS)	Sample Collection Information: Sample ID:	Date	Time	Matrix (GW WW SW DW S L NA)	Number of Containers	MS/MSD?	200.8 Lead	Notes:
1	1 Drinking Fountain Opposite 115	6/20/25	11:10	DW			X	
2	2 Drinking Fountain Opposite 111	6/20/25	11:10	DW			X	
3	3 Main Office Conference Room	6/20/25	11:00	DW			X	
4	4 Main Entrance Drinking Fountain	6/20/25	11:00	DW			X	
5	5 Rm. 108	6/20/25	11:10	DW			X	
6	6 Health Office	6/20/25	11:10	DW			X	
7				DW			X	
8				DW			X	
9				DW			X	
10				DW			X	

Special Instructions / Comments:

Turnaround Requirements

____Rush (Surcharges Apply)
Subject to Availability
Please Check with your PM
x Standard (10 Business Days)
Date Required:

Report Requirements

__X__Tier II/Cat A -Results/QC
____Tier IV/Cat B - Data Validation Report w/. Data
EDD: ____Yes ____No
EDD Type:

Metals: RCRA 8•PP 13•TAL 23•TCLP•Other (List)

VOA/SVOA Report List: TCL • BTEX • TCLP • CP-51/Stars •THM • Other: _____

Invoice To: (□ Same as Report To)

PO #: 7225-01793
Company: Rush-Henrietta CSD
Contact: Ken Nelson
Email: knelson@rhnet.org
Phone: 585-738-8070
Address: 2034 Lehigh Station Rd. Henrietta, NY 14467

	Relinquished By:	Received By:	Relinquished By:	Received By:	Relinquished By:	Received By:
Signature						
Printed Name	Jake Durfee	Molli Woods				
Company	Rush-Henrietta CSD	ALS				
Date/Time	6/20/25 12:50	6/20/25 1250				

R2507356
Rush-Henrietta Central School District
Good Shepherd
5

Distribution: White - Lab Copy; Yellow - Return to Originator

© 2012 by ALS Group

R2507356 5
Rush-Henrietta Central School District
Good Shepherd

Cooler Receipt and Preserva

Project/Client Rush-Henrietta Folder Number

Cooler received on 6/20/25 by: MW

COURIER: ALS UPS FEDEX VELOCITY CLIENT

1	Were **Custody seals** on outside of cooler?	Y (N)	5a	Did **VOA** vials have sig* **bubbles?**	Y N (NA)
2	**Custody papers** properly completed (ink, signed)?	(Y) N	5b	Sig* bubbles: **Alk?** Y N (NA) **Sulfide?**	Y N (NA)
3	Did all bottles arrive in good **condition** (unbroken)?	(Y) N	6	Where did the bottles originate?	(ALS/ROC) CLIENT
4	Circle: **Wet Ice** **Dry Ice** **Gel packs** present?	Y (N)	7	Soil VOA received as: Bulk Encore 5035set	(NA)

8. Temperature Readings Date: 6/20 Time: 13:04 ID: IR#12 (IR#11) From: Temp Blank (Sample Bottle)

Temp (°C)	23.5						
Within 0-6°C?	Y (N)	Y N	Y N	Y N	Y N	Y N	Y N
If <0°C, were samples frozen?	Y N	Y N	Y N	Y N	Y N	Y N	Y N

If out of Temperature, note packing/ice condition: _____ Ice melted Poorly Packed (described below) (Same Day Rule)

&Client Approval to Run Samples: _____ Standing Approval Client aware at drop-off Client notified by: _____

All samples held in storage location: SMO by MW on 6/20 at 13:02
5035 samples placed in storage location: _____ by _____ on _____ at _____ within 48 hours of sampling? Y N

Cooler Breakdown/Preservation Check**: Date: 6/20/25 Time: 1628 by: MW

9. Were all bottle labels complete (*i.e.* analysis, preservation, etc.)? (YES) NO
10. Did all bottle labels and tags agree with custody papers? (YES) NO
11. Were correct containers used for the tests indicated? (YES) NO
12. Were 5035 vials acceptable (no extra labels, not leaking)? YES NO (N/A)
13. Were dissolved metals filtered in the field? YES NO (N/A)
14. Air Samples: Cassettes / Tubes Intact Y / N with MS Y / N Canisters Pressurized Tedlar® Bags Inflated (N/A)

pH	Lot of test paper	Reagent	Preserved? Yes	Preserved? No	Lot Received	Exp	Sample ID Adjusted	Vol. Added	Lot Added	Final pH
≥12		NaOH								
≤2		HNO_3								
≤2		H_2SO_4								
<4		$NaHSO_4$								
5-9		For 608pest			No=Notify for 3day					
Residual Chlorine (-)		For CN, Phenol, 625, 608pest, 522			If +, contact PM to add $Na_2S_2O_3$ (625, 608, CN), ascorbic (phenol).					
		$Na_2S_2O_3$								
		ZnAcetate	-	-			**VOAs and 1664 Not to be tested before analysis. Otherwise, all bottles of all samples with chemical preservatives are checked (not just representatives).			
		HCl	**	**						

Bottle lot numbers: 110424-7ADD

Explain all Discrepancies/ Other Comments:

HPROD	BULK
HTR	FLDT
SUB	HGFB
ALS	LL3541

Labels secondary reviewed by: MW

*significant air bubbles: VOA > 5-6 mm : WC >1 in. diameter

Miscellaneous Forms

ALS Environmental—Rochester Laboratory
1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623
Phone (585) 288-5380 Fax (585) 288-8475
www.alsglobal.com

REPORT QUALIFIERS AND DEFINITIONS

U	Analyte was analyzed for but not detected. The sample quantitation limit has been corrected for dilution and for percent moisture, unless otherwise noted in the case narrative.
J	Estimated value due to either being a Tentatively Identified Compound (TIC) or that the concentration is between the MRL and the MDL. Concentrations are not verified within the linear range of the calibration. For DoD: concentration >40% difference between two GC columns (pesticides/Arclors).
B	Analyte was also detected in the associated method blank at a concentration that may have contributed to the sample result.
E	Inorganics- Concentration is estimated due to the serial dilution was outside control limits.
E	Organics- Concentration has exceeded the calibration range for that specific analysis.
D	Concentration is a result of a dilution, typically a secondary analysis of the sample due to exceeding the calibration range or that a surrogate has been diluted out of the sample and cannot be assessed.
*	Indicates that a quality control parameter has exceeded laboratory limits. Under the "Notes" column of the Form I, this qualifier denotes analysis was performed out of Holding Time.
H	Analysis was performed out of hold time for tests that have an "immediate" hold time criteria.
#	Spike was diluted out.
+	Correlation coefficient for MSA is <0.995.
N	Inorganics- Matrix spike recovery was outside laboratory limits.
N	Organics- Presumptive evidence of a compound (reported as a TIC) based on the MS library search.
S	Concentration has been determined using Method of Standard Additions (MSA).
W	Post-Digestion Spike recovery is outside control limits and the sample absorbance is <50% of the spike absorbance.
P	Concentration >40% difference between the two GC columns.
C	Confirmed by GC/MS
Q	DoD reports: indicates a pesticide/Aroclor is not confirmed (≥100% Difference between two GC columns).
X	See Case Narrative for discussion.
MRL LOQ	Method Reporting Limit. Also known as: Limit of Quantitation (LOQ) The lowest concentration at which the method analyte may be reliably quantified under the method conditions.
MDL	Method Detection Limit. A statistical value derived from a study designed to provide the lowest concentration that will be detected 99% of the time. Values between the MDL and MRL are estimated (see J qualifier).
LOD	Limit of Detection. A value at or above the MDL which has been verified to be detectable.
ND	Non-Detect. Analyte was not detected at the concentration listed. Same as U qualifier.

Rochester Lab ID # for State Accreditations[1]

NELAP States
Florida ID # E87674
New Hampshire ID # 2941
New York ID # 10145
Pennsylvania ID# 68-786
Texas ID#T104704581
Virginia #460167

Non-NELAP States
Connecticut ID #PH0556
Delaware Approved
Maine ID #NY01587
North Carolina #36701
North Carolina #676
Rhode Island LAO00333

[1] Analyses were performed according to our laboratory's NELAP-approved quality assurance program and any applicable state or agency requirements. The test results meet requirements of the current NELAP/TNI standards or state or agency requirements, where applicable, except as noted in the case narrative. Since not all analyte/method/matrix combinations are offered for state/NELAC accreditation, this report may contain results which are not accredited. For a specific list of accredited analytes, contact the laboratory. To verify NH accredited analytes, go to https://www4.des.state.nh.us/CertifiedLabs/Certified-Method.aspx.

ALS Laboratory Group

Acronyms

ASTM	American Society for Testing and Materials
A2LA	American Association for Laboratory Accreditation
CARB	California Air Resources Board
CAS Number	Chemical Abstract Service registry Number
CFC	Chlorofluorocarbon
CFU	Colony-Forming Unit
DEC	Department of Environmental Conservation
DEQ	Department of Environmental Quality
DHS	Department of Health Services
DOE	Department of Ecology
DOH	Department of Health
EPA	U. S. Environmental Protection Agency
ELAP	Environmental Laboratory Accreditation Program
GC	Gas Chromatography
GC/MS	Gas Chromatography/Mass Spectrometry
LUFT	Leaking Underground Fuel Tank
M	Modified
MCL	Maximum Contaminant Level is the highest permissible concentration of a substance allowed in drinking water as established by the USEPA.
MDL	Method Detection Limit
MPN	Most Probable Number
MRL	Method Reporting Limit
NA	Not Applicable
NC	Not Calculated
NCASI	National Council of the Paper Industry for Air and Stream Improvement
ND	Not Detected
NIOSH	National Institute for Occupational Safety and Health
PQL	Practical Quantitation Limit
RCRA	Resource Conservation and Recovery Act
SIM	Selected Ion Monitoring
TPH	Total Petroleum Hydrocarbons
tr	Trace level is the concentration of an analyte that is less than the PQL but greater than or equal to the MDL.

Client: Rush-Henrietta Central School District **Service Request:** R2507356
Project: Good Shepherd/

Sample Name: Drinking Fountain Opposite 115 **Date Collected:** 06/20/25
Lab Code: R2507356-001 **Date Received:** 06/20/25
Sample Matrix: Drinking Water

Analysis Method	Extracted/Digested By	Analyzed By
200.8		NMANSEN

Sample Name: Drinking Fountain Opposite 111 **Date Collected:** 06/20/25
Lab Code: R2507356-002 **Date Received:** 06/20/25
Sample Matrix: Drinking Water

Analysis Method	Extracted/Digested By	Analyzed By
200.8		NMANSEN

Sample Name: Main Office Converence Room **Date Collected:** 06/20/25
Lab Code: R2507356-003 **Date Received:** 06/20/25
Sample Matrix: Drinking Water

Analysis Method	Extracted/Digested By	Analyzed By
200.8		NMANSEN

Sample Name: Main Enterance Drinking Fountain **Date Collected:** 06/20/25
Lab Code: R2507356-004 **Date Received:** 06/20/25
Sample Matrix: Drinking Water

Analysis Method	Extracted/Digested By	Analyzed By
200.8		NMANSEN

Sample Name: Rm. 108 **Date Collected:** 06/20/25
Lab Code: R2507356-005 **Date Received:** 06/20/25
Sample Matrix: Drinking Water

Analysis Method	Extracted/Digested By	Analyzed By
200.8		NMANSEN

Client:	Rush-Henrietta Central School District	**Service Request:**	R2507356
Project:	Good Shepherd/		

Sample Name:	Health Office	**Date Collected:**	06/20/25
Lab Code:	R2507356-006	**Date Received:**	06/20/25
Sample Matrix:	Drinking Water		

Analysis Method	Extracted/Digested By	Analyzed By
200.8		NMANSEN

PREPARATION METHODS

The preparation methods associated with this report are found in these tables unless discussed in the case narrative.

INORGANIC

Water/Liquid Matrix

Analytical Method	Preparation Method
200.7	200.2
200.8	200.2
6010C or 6010D	3005A/3010A
6020A or 6020B	ILM05.3
9034 Sulfide Acid Soluble	9030B
SM 4500-CN-N-2016 Amenable and Residual Cyanide	SM 4500-CN-G and SM 4500-CN-B,C-2016
SM 4500-CN-E WAD Cyanide	SM 4500-CN-I

Solid/Soil/Non-Aqueous Matrix

Analytical Method	Preparation Method
6010C or 6010D	3050B
6020A or 6020B	3050B
6010C or 6010D TCLP (1311) extract	3005A/3010A
6010C or 6010D SPLP (1312) extract	3005A/3010A
7199	3060A
300.0 Anions/ 350.1/ 353.2/ SM 2320B/ SM 5210B/ 9056A Anions	DI extraction
For analytical methods not listed, the preparation method is the same as the analytical method reference.	

ORGANIC

Preparation Methods for Organic methods are listed in the header of the Results pages.

Regarding "Bulk/5035A":
For soil/solid samples submitted in soil jars for Volatiles analysis, the prep method is listed as "Bulk/5035A". The lab follows the closed-system EPA 5035A protocols once the sample is transferred to a sealed vial, but collection in bulk in soil jars does not follow the collection protocols listed in EPA 5035A. In accordance with the NYSDOH technical notice of October 2012, all results or reporting limits <200 ug/kg are to be considered estimated due to potential low bias.

Sample Results

ALS Environmental—Rochester Laboratory
1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623
Phone (585) 288-5380 Fax (585) 288-8475
www.alsglobal.com

Metals

ALS Environmental—Rochester Laboratory
1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623
Phone (585) 288-5380 Fax (585) 288-8475
www.alsglobal.com

Analytical Report

Client:	Rush-Henrietta Central School District	**Service Request:**	R2507356
Project:	Good Shepherd	**Date Collected:**	06/20/25 11:10
Sample Matrix:	Drinking Water	**Date Received:**	06/20/25 12:50
Sample Name:	Drinking Fountain Opposite 115	**Basis:**	NA
Lab Code:	R2507356-001		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Lead, Total	200.8	1.0 U	ug/L	1.0	1	06/30/25 14:31	

Analytical Report

Client:	Rush-Henrietta Central School District	**Service Request:**	R2507356
Project:	Good Shepherd	**Date Collected:**	06/20/25 11:10
Sample Matrix:	Drinking Water	**Date Received:**	06/20/25 12:50
Sample Name:	Drinking Fountain Opposite 111	**Basis:**	NA
Lab Code:	R2507356-002		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Lead, Total	200.8	1.0 U	ug/L	1.0	1	06/30/25 14:33	

Analytical Report

Client:	Rush-Henrietta Central School District	**Service Request:**	R2507356
Project:	Good Shepherd	**Date Collected:**	06/20/25 11:00
Sample Matrix:	Drinking Water	**Date Received:**	06/20/25 12:50
Sample Name:	Main Office Converence Room	**Basis:**	NA
Lab Code:	R2507356-003		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Lead, Total	200.8	**5.0**	ug/L	1.0	1	06/30/25 14:34	

Analytical Report

Client:	Rush-Henrietta Central School District	**Service Request:**	R2507356
Project:	Good Shepherd	**Date Collected:**	06/20/25 11:00
Sample Matrix:	Drinking Water	**Date Received:**	06/20/25 12:50
Sample Name:	Main Enterance Drinking Fountain	**Basis:**	NA
Lab Code:	R2507356-004		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Lead, Total	200.8	**6.3**	ug/L	1.0	1	06/30/25 14:44	

Analytical Report

Client:	Rush-Henrietta Central School District	**Service Request:**	R2507356
Project:	Good Shepherd	**Date Collected:**	06/20/25 11:10
Sample Matrix:	Drinking Water	**Date Received:**	06/20/25 12:50
Sample Name:	Rm. 108	**Basis:**	NA
Lab Code:	R2507356-005		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Lead, Total	200.8	1.0 U	ug/L	1.0	1	06/30/25 14:48	

Analytical Report

Client:	Rush-Henrietta Central School District	**Service Request:**	R2507356
Project:	Good Shepherd	**Date Collected:**	06/20/25 11:10
Sample Matrix:	Drinking Water	**Date Received:**	06/20/25 12:50
Sample Name:	Health Office	**Basis:**	NA
Lab Code:	R2507356-006		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Lead, Total	200.8	**1.0**	ug/L	1.0	1	06/30/25 14:49	

QC Summary Forms

ALS Environmental—Rochester Laboratory
1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623
Phone (585) 288-5380 Fax (585) 288-8475
www.alsglobal.com

Metals

ALS Environmental—Rochester Laboratory
1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623
Phone (585) 288-5380 Fax (585) 288-8475
www.alsglobal.com

Analytical Report

Client:	Rush-Henrietta Central School District	**Service Request:**	R2507356
Project:	Good Shepherd	**Date Collected:**	NA
Sample Matrix:	Drinking Water	**Date Received:**	NA
Sample Name:	Method Blank	**Basis:**	NA
Lab Code:	R2507356-MB1		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Lead, Total	200.8	1.0 U	ug/L	1.0	1	06/30/25 13:57	

Analytical Report

Client:	Rush-Henrietta Central School District	**Service Request:**	R2507356
Project:	Good Shepherd	**Date Collected:**	NA
Sample Matrix:	Drinking Water	**Date Received:**	NA
Sample Name:	Method Blank	**Basis:**	NA
Lab Code:	R2507356-MB2		

Inorganic Parameters

Analyte Name	Analysis Method	Result	Units	MRL	Dil.	Date Analyzed	Q
Lead, Total	200.8	1.0 U	ug/L	1.0	1	06/30/25 14:41	

Client:	Rush-Henrietta Central School District	**Service Request:**	R2507356
Project:	Good Shepherd	**Date Collected:**	06/20/25
Sample Matrix:	Drinking Water	**Date Received:**	06/20/25
		Date Analyzed:	06/30/25

Duplicate Matrix Spike Summary
Inorganic Parameters

Sample Name:	Main Office Converence Room	**Units:**	ug/L
Lab Code:	R2507356-003	**Basis:**	NA
Analysis Method:	200.8		

		Matrix Spike			Duplicate Matrix Spike					
		R2507356-003MS			R2507356-003DMS					
Analyte Name	**Sample Result**	**Result**	**Spike Amount**	**% Rec**	**Result**	**Spike Amount**	**% Rec**	**% Rec Limits**	**RPD**	**RPD Limit**
Lead, Total	5.0	23.2	20.0	91	23.4	20.0	92	70-130	<1	20

Results flagged with an asterisk (*) indicate values outside control criteria.

Results flagged with a pound (#) indicate the control criteria is not applicable.

Percent recoveries and relative percent differences (RPD) are determined by the software using values in the calculation which have not been rounded.

Matrix Spike and Matrix Spike Duplicate Data is presented for information purposes only. The matrix may or may not be relevant to samples reported in this report. The laboratory evaluates system performance based on the LCS and LCSD control limits.

Client:	Rush-Henrietta Central School District	**Service Request:**	R2507356
Project:	Good Shepherd	**Date Collected:**	06/20/25
Sample Matrix:	Drinking Water	**Date Received:**	06/20/25
		Date Analyzed:	06/30/25

Duplicate Matrix Spike Summary
Inorganic Parameters

Sample Name:	Main Enterance Drinking Fountain	**Units:**	ug/L
Lab Code:	R2507356-004	**Basis:**	NA
Analysis Method:	200.8		

		Matrix Spike			Duplicate Matrix Spike					
		R2507356-004MS			R2507356-004DMS					
Analyte Name	**Sample Result**	**Result**	**Spike Amount**	**% Rec**	**Result**	**Spike Amount**	**% Rec**	**% Rec Limits**	**RPD**	**RPD Limit**
Lead, Total	6.3	23.7	20.0	87	25.2	20.0	95	70-130	6	20

Results flagged with an asterisk (*) indicate values outside control criteria.

Results flagged with a pound (#) indicate the control criteria is not applicable.

Percent recoveries and relative percent differences (RPD) are determined by the software using values in the calculation which have not been rounded.

Matrix Spike and Matrix Spike Duplicate Data is presented for information purposes only. The matrix may or may not be relevant to samples reported in this report. The laboratory evaluates system performance based on the LCS and LCSD control limits.

ALS Group USA, Corp.
dba ALS Environmental

QA/QC Report

Client: Rush-Henrietta Central School District
Project: Good Shepherd
Sample Matrix: Drinking Water

Service Request: R2507356
Date Analyzed: 06/30/25

Lab Control Sample Summary
Inorganic Parameters

Units:ug/L
Basis:NA

Lab Control Sample
R2507356-LCS1

Analyte Name	Analytical Method	Result	Spike Amount	% Rec	% Rec Limits
Lead, Total	200.8	19.0	20.0	95	85-115

ALS Group USA, Corp.
dba ALS Environmental

QA/QC Report

Client: Rush-Henrietta Central School District
Project: Good Shepherd
Sample Matrix: Drinking Water

Service Request: R2507356
Date Analyzed: 06/30/25

Lab Control Sample Summary
Inorganic Parameters

Units:ug/L
Basis:NA

Lab Control Sample
R2507356-LCS2

Analyte Name	Analytical Method	Result	Spike Amount	% Rec	% Rec Limits
Lead, Total	200.8	19.8	20.0	99	85-115