



May 22, 2025

Service Request No:R2505176

Mr. Ken Nelson  
Rush-Henrietta Central School District  
2034 Lehigh Station Road  
Henrietta, NY 14467

**Laboratory Results for: Sperry Senior High School**

Dear Mr.Nelson,

Enclosed are the results of the sample(s) submitted to our laboratory May 08, 2025  
For your reference, these analyses have been assigned our service request number **R2505176**.

All testing was performed according to our laboratory's quality assurance program and met the requirements of the TNI standards except as noted in the case narrative report. Any testing not included in the lab's accreditation is identified on a Non-Certified Analytes report. All results are intended to be considered in their entirety. ALS Environmental is not responsible for use of less than the complete report. Results apply only to the individual samples submitted to the lab for analysis, as listed in the report. The measurement uncertainty of the results included in this report is within that expected when using the prescribed method(s), and represented by Laboratory Control Sample control limits. Any events, such as QC failures or Holding Time exceedances, which may add to the uncertainty are explained in the report narrative or are flagged with qualifiers. The flags are explained in the Report Qualifiers and Definitions page of this report.

Please contact me if you have any questions. My extension is 7476. You may also contact me via email at [Chris.Leavy@alsglobal.com](mailto:Chris.Leavy@alsglobal.com).

Respectfully submitted,

**ALS Group USA, Corp. dba ALS Environmental**

Christopher Leavy  
Project Manager

**ADDRESS**

1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623

**PHONE**

+1 585 288 5380

**FAX**

+1 585 288 8475

ALS Group USA, Corp.  
dba ALS Environmental



## Narrative Documents

**ALS Environmental—Rochester Laboratory**

1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623

Phone (585) 288-5380 Fax (585) 288-8475

[www.alsglobal.com](http://www.alsglobal.com)



**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Received:** 05/08/2025

#### CASE NARRATIVE

All analyses were performed consistent with the quality assurance program of ALS Environmental. This report contains analytical results for samples for the Tier II level requested by the client.

#### Sample Receipt:

Thirty four drinking water samples were received for analysis at ALS Environmental on 05/08/2025. Any discrepancies upon initial sample inspection are annotated on the sample receipt and preservation form included within this report. The samples were stored at minimum in accordance with the analytical method requirements.

#### Metals:

No significant anomalies were noted with this analysis.

Approved by   
\_\_\_\_\_

Date 05/22/2025



### SAMPLE DETECTION SUMMARY

This form includes only detections above the reporting levels. For a full listing of sample results, continue to the Sample Results section of this Report.

| CLIENT ID: Sperry-53 B-106 - Drinking Fountain                |         |      |     | Lab ID: R2505176-003 |       |        |
|---------------------------------------------------------------|---------|------|-----|----------------------|-------|--------|
| Analyte                                                       | Results | Flag | MDL | MRL                  | Units | Method |
| Lead, Total                                                   | 15.2    |      |     | 1.0                  | ug/L  | 200.8  |
| CLIENT ID: Sperry-55 B-104 - Drinking Fountain                |         |      |     | Lab ID: R2505176-005 |       |        |
| Analyte                                                       | Results | Flag | MDL | MRL                  | Units | Method |
| Lead, Total                                                   | 9.4     |      |     | 1.0                  | ug/L  | 200.8  |
| CLIENT ID: Sperry-56 B-101 - Health Office Exam Room Sink     |         |      |     | Lab ID: R2505176-006 |       |        |
| Analyte                                                       | Results | Flag | MDL | MRL                  | Units | Method |
| Lead, Total                                                   | 3.1     |      |     | 1.0                  | ug/L  | 200.8  |
| CLIENT ID: Sperry-57 B-101 - Health Office Main Sink Faucet   |         |      |     | Lab ID: R2505176-007 |       |        |
| Analyte                                                       | Results | Flag | MDL | MRL                  | Units | Method |
| Lead, Total                                                   | 6.4     |      |     | 1.0                  | ug/L  | 200.8  |
| CLIENT ID: Sperry-58 B-101 - Health Office Ice Maker          |         |      |     | Lab ID: R2505176-008 |       |        |
| Analyte                                                       | Results | Flag | MDL | MRL                  | Units | Method |
| Lead, Total                                                   | 4.0     |      |     | 1.0                  | ug/L  | 200.8  |
| CLIENT ID: Sperry-59 Drinking Fountain Lower A-B Wing         |         |      |     | Lab ID: R2505176-009 |       |        |
| Analyte                                                       | Results | Flag | MDL | MRL                  | Units | Method |
| Lead, Total                                                   | 2.8     |      |     | 1.0                  | ug/L  | 200.8  |
| CLIENT ID: Sperry-60 Drinking Fountain Lower A-Wing LEFT      |         |      |     | Lab ID: R2505176-010 |       |        |
| Analyte                                                       | Results | Flag | MDL | MRL                  | Units | Method |
| Lead, Total                                                   | 2.0     |      |     | 1.0                  | ug/L  | 200.8  |
| CLIENT ID: Sperry-61 Drinking Fountain Lower A-Wing RIGHT     |         |      |     | Lab ID: R2505176-011 |       |        |
| Analyte                                                       | Results | Flag | MDL | MRL                  | Units | Method |
| Lead, Total                                                   | 2.1     |      |     | 1.0                  | ug/L  | 200.8  |
| CLIENT ID: Sperry-62 Main Office Breakroom Sink Faucet        |         |      |     | Lab ID: R2505176-012 |       |        |
| Analyte                                                       | Results | Flag | MDL | MRL                  | Units | Method |
| Lead, Total                                                   | 18.3    |      |     | 1.0                  | ug/L  | 200.8  |
| CLIENT ID: Sperry-63 Drinking Fountain Adjacent to Auditorium |         |      |     | Lab ID: R2505176-013 |       |        |
| Analyte                                                       | Results | Flag | MDL | MRL                  | Units | Method |
| Lead, Total                                                   | 1.7     |      |     | 1.0                  | ug/L  | 200.8  |
| CLIENT ID: Sperry-64 D-109 - Music Room Office Sink Faucet    |         |      |     | Lab ID: R2505176-014 |       |        |
| Analyte                                                       | Results | Flag | MDL | MRL                  | Units | Method |
| Lead, Total                                                   | 139     |      |     | 1.0                  | ug/L  | 200.8  |
| CLIENT ID: Sperry-65 D-203 - Senior Lounge Sink Faucet        |         |      |     | Lab ID: R2505176-015 |       |        |
| Analyte                                                       | Results | Flag | MDL | MRL                  | Units | Method |
| Lead, Total                                                   | 33.4    |      |     | 1.0                  | ug/L  | 200.8  |



### SAMPLE DETECTION SUMMARY

This form includes only detections above the reporting levels. For a full listing of sample results, continue to the Sample Results section of this Report.

|                                                                      |                             |
|----------------------------------------------------------------------|-----------------------------|
| <b>CLIENT ID: Sperry-66 D-213 - Literacy Center Kitchenette Sink</b> | <b>Lab ID: R2505176-016</b> |
|----------------------------------------------------------------------|-----------------------------|

| Analyte     | Results | Flag | MDL | MRL | Units | Method |
|-------------|---------|------|-----|-----|-------|--------|
| Lead, Total | 29.9    |      |     | 1.0 | ug/L  | 200.8  |

|                                                            |                             |
|------------------------------------------------------------|-----------------------------|
| <b>CLIENT ID: Sperry-68 Drinking Fountain Upper A-Wing</b> | <b>Lab ID: R2505176-018</b> |
|------------------------------------------------------------|-----------------------------|

| Analyte     | Results | Flag | MDL | MRL | Units | Method |
|-------------|---------|------|-----|-----|-------|--------|
| Lead, Total | 3.8     |      |     | 1.0 | ug/L  | 200.8  |

|                                                                            |                         |
|----------------------------------------------------------------------------|-------------------------|
| <b>CLIENT ID: Sperry-69 Adj. A-226 - Drinking Fountain Adjacent to Lab</b> | <b>ID: R2505176-019</b> |
|----------------------------------------------------------------------------|-------------------------|

| Analyte     | Results | Flag | MDL | MRL | Units | Method |
|-------------|---------|------|-----|-----|-------|--------|
| Lead, Total | 1.6     |      |     | 1.0 | ug/L  | 200.8  |

|                                                                     |                             |
|---------------------------------------------------------------------|-----------------------------|
| <b>CLIENT ID: Sperry-70 Drinking Fountain Upper A&amp;B Hallway</b> | <b>Lab ID: R2505176-020</b> |
|---------------------------------------------------------------------|-----------------------------|

| Analyte     | Results | Flag | MDL | MRL | Units | Method |
|-------------|---------|------|-----|-----|-------|--------|
| Lead, Total | 1.9     |      |     | 1.0 | ug/L  | 200.8  |

|                                                              |                             |
|--------------------------------------------------------------|-----------------------------|
| <b>CLIENT ID: Sperry-71 Drinking Fountain Opposite B-205</b> | <b>Lab ID: R2505176-021</b> |
|--------------------------------------------------------------|-----------------------------|

| Analyte     | Results | Flag | MDL | MRL | Units | Method |
|-------------|---------|------|-----|-----|-------|--------|
| Lead, Total | 2.4     |      |     | 1.0 | ug/L  | 200.8  |

|                                                                 |                             |
|-----------------------------------------------------------------|-----------------------------|
| <b>CLIENT ID: Sperry-72 Drinking Fountain Adjacent to Boy's</b> | <b>Lab ID: R2505176-022</b> |
|-----------------------------------------------------------------|-----------------------------|

| Analyte     | Results | Flag | MDL | MRL | Units | Method |
|-------------|---------|------|-----|-----|-------|--------|
| Lead, Total | 1.2     |      |     | 1.0 | ug/L  | 200.8  |

|                                                                    |                             |
|--------------------------------------------------------------------|-----------------------------|
| <b>CLIENT ID: Sperry-75 Drinking Fountain Gordon Stadium Green</b> | <b>Lab ID: R2505176-025</b> |
|--------------------------------------------------------------------|-----------------------------|

| Analyte     | Results | Flag | MDL | MRL | Units | Method |
|-------------|---------|------|-----|-----|-------|--------|
| Lead, Total | 1.5     |      |     | 1.0 | ug/L  | 200.8  |

|                                                                      |                             |
|----------------------------------------------------------------------|-----------------------------|
| <b>CLIENT ID: Sperry-79 Drinking Fountain Comet Cafe Sink Faucet</b> | <b>Lab ID: R2505176-028</b> |
|----------------------------------------------------------------------|-----------------------------|

| Analyte     | Results | Flag | MDL | MRL | Units | Method |
|-------------|---------|------|-----|-----|-------|--------|
| Lead, Total | 5.0     |      |     | 1.0 | ug/L  | 200.8  |

|                                                               |                             |
|---------------------------------------------------------------|-----------------------------|
| <b>CLIENT ID: Sperry-80 Drinking Fountain Music Wing LEFT</b> | <b>Lab ID: R2505176-029</b> |
|---------------------------------------------------------------|-----------------------------|

| Analyte     | Results | Flag | MDL | MRL | Units | Method |
|-------------|---------|------|-----|-----|-------|--------|
| Lead, Total | 11.3    |      |     | 1.0 | ug/L  | 200.8  |

|                                                                |                             |
|----------------------------------------------------------------|-----------------------------|
| <b>CLIENT ID: Sperry-81 Drinking Fountain Music Wing RIGHT</b> | <b>Lab ID: R2505176-030</b> |
|----------------------------------------------------------------|-----------------------------|

| Analyte     | Results | Flag | MDL | MRL | Units | Method |
|-------------|---------|------|-----|-----|-------|--------|
| Lead, Total | 12.2    |      |     | 1.0 | ug/L  | 200.8  |

|                                                              |                             |
|--------------------------------------------------------------|-----------------------------|
| <b>CLIENT ID: Sperry-82 Kleeman Stadium Concession Stand</b> | <b>Lab ID: R2505176-031</b> |
|--------------------------------------------------------------|-----------------------------|

| Analyte     | Results | Flag | MDL | MRL | Units | Method |
|-------------|---------|------|-----|-----|-------|--------|
| Lead, Total | 1.1     |      |     | 1.0 | ug/L  | 200.8  |



## Sample Receipt Information

**ALS Environmental—Rochester Laboratory**

1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623

Phone (585) 288-5380 Fax (585) 288-8475

[www.alsglobal.com](http://www.alsglobal.com)

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School

**Service Request:**R2505176

**SAMPLE CROSS-REFERENCE**

| <u>SAMPLE #</u> | <u>CLIENT SAMPLE ID</u>                                    | <u>DATE</u> | <u>TIME</u> |
|-----------------|------------------------------------------------------------|-------------|-------------|
| R2505176-001    | Sperry-51 B-133 - Kitchenette Slnk Faucet                  | 5/8/2025    | 0430        |
| R2505176-002    | Sperry-52 Adj Mop Sink - Lower B-Wing Drinking Fountain    | 5/8/2025    | 0440        |
| R2505176-003    | Sperry-53 B-106 - Drinking Fountain                        | 5/8/2025    | 0442        |
| R2505176-004    | Sperry-54 Drinking Fountain Adj to Lower B-Wing Boy's Lav  | 5/8/2025    | 0445        |
| R2505176-005    | Sperry-55 B-104 - Drinking Fountain                        | 5/8/2025    | 0442        |
| R2505176-006    | Sperry-56 B-101 - Health Office Exam Room Sink Faucet      | 5/8/2025    | 0434        |
| R2505176-007    | Sperry-57 B-101 - Health Office Main Sink Faucet           | 5/8/2025    | 0436        |
| R2505176-008    | Sperry-58 B-101 - Health Office Ice Maker                  | 5/8/2025    | 0436        |
| R2505176-009    | Sperry-59 Drinking Fountain Lower A-B Wing Hallway         | 5/8/2025    | 0440        |
| R2505176-010    | Sperry-60 Drinking Fountain Lower A-Wing LEFT              | 5/8/2025    | 0445        |
| R2505176-011    | Sperry-61 Drinking Fountain Lower A-Wing RIGHT             | 5/8/2025    | 0448        |
| R2505176-012    | Sperry-62 Main Office Breakroom Sink Faucet                | 5/8/2025    | 0452        |
| R2505176-013    | Sperry-63 Drinking Fountain Adjacent to Auditorium         | 5/8/2025    | 0458        |
| R2505176-014    | Sperry-64 D-109 - Music Room Office Sink Faucet            | 5/8/2025    | 0456        |
| R2505176-015    | Sperry-65 D-203 - Senior Lounge Slnk Faucet                | 5/8/2025    | 0459        |
| R2505176-016    | Sperry-66 D-213 - Literacy Center Kitchenette Sink Faucet  | 5/8/2025    | 0506        |
| R2505176-017    | Sperry-67 A-220 - Teacher's Lounge Sink Faucet             | 5/8/2025    | 0507        |
| R2505176-018    | Sperry-68 Drinking Fountain Upper A-Wing                   | 5/8/2025    | 0514        |
| R2505176-019    | Sperry-69 Adj. A-226 - Drinking Fountain Adjacent to A-226 | 5/8/2025    | 0512        |
| R2505176-020    | Sperry-70 Drinking Fountain Upper A&B Hallway              | 5/8/2025    | 0515        |
| R2505176-021    | Sperry-71 Drinking Fountain Opposite B-205                 | 5/8/2025    | 0524        |
| R2505176-022    | Sperry-72 Drinking Fountain Adjacent to Boy's Lavatory     | 5/8/2025    | 0521        |
| R2505176-023    | Sperry-73 Computer Services Conference Room Sink           | 5/8/2025    | 0540        |
| R2505176-024    | Sperry-74 Gordon Stadium Concession Stand Sink Faucet      | 5/8/2025    | 0550        |
| R2505176-025    | Sperry-75 Drinking Fountain Gordon Stadium Green           | 5/8/2025    | 0550        |
| R2505176-026    | Sperry-76 Bottle Filler Gordon Stadium Green Monster       | 5/8/2025    | 0551        |
| R2505176-028    | Sperry-79 Drinking Fountain Comet Cafe Sink Faucet         | 5/8/2025    | 0502        |
| R2505176-029    | Sperry-80 Drinking Fountain Music Wing LEFT                | 5/8/2025    | 0500        |
| R2505176-030    | Sperry-81 Drinking Fountain Music Wing RIGHT               | 5/8/2025    | 0500        |
| R2505176-031    | Sperry-82 Kleeman Stadium Concession Stand Single Sink     | 5/8/2025    | 0545        |
| R2505176-032    | Sperry-83 Kleeman Stadium Concession Stand Double Sink     | 5/8/2025    | 0545        |
| R2505176-033    | Sperry-84 Kleeman Stadium Concession Drinking Fountain     | 5/8/2025    | 0547        |
| R2505176-034    | Sperry-85 Kleeman Stadium Concession Bottle Filler Green   | 5/8/2025    | 0547        |



# Chain of Custody / Analytical Request Form

1565 Jefferson Road, Building 300, Suite 360 • Rochester, NY 14623 • +1 585 288 5380 • alsglobal.com

SR#:

Page of

## Report To:

ALL SHADED AREAS MUST BE COMPLETED BY THE CLIENT / SAMPLER

Preservative

Company: Rush-Henrietta CSD  
Contact: Ken Nelson  
Email: knelson@rhnet.org  
Phone: 585-738-8070  
Address: 2034 Lehigh Station Road  
Henrietta NY 14467

Project Name: Sperry Senior High School  
Project Number:  
ALS Quote #: 38456  
Sampler's Signature:  
Email CC: slacey@rhnet.org  
Email CC: jlschultz@rhnet.org  
State Samples Collected NY

GW  
WW  
SW  
DW  
S  
L  
NA

Matrix  
Number of Containers  
MS/MSD?

200.8 Lead

- 0. None
- 1. HCl
- 2. HNO3
- 3. H2SO4
- 4. NaOH
- 5. Zn Acet.
- 6. MeOH
- 7. NaHSO4
- 8. Other

Notes:

| Lab ID<br>(ALS) | Sample Collection Information:                            |          |      |
|-----------------|-----------------------------------------------------------|----------|------|
|                 | Sample ID:                                                | Date     | Time |
| 51              | B-133 - Kitchenette Sink Faucet                           | 05/08/25 | 4:30 |
| 52              | Adj Mop Sink - Lower B-Wing Drinking Fountain             | 05/08/25 | 4:40 |
| 53              | B-106 - Drinking Fountain                                 | 05/08/25 | 4:42 |
| 54              | Drinking Fountain Adjacent to Lower B-Wing Boy's Lavatory | 05/08/25 | 4:45 |
| 55              | B-104 - Drinking Fountain                                 | 05/08/25 | 4:42 |
| 56              | B-101 - Health Office Exam Room Sink Faucet               | 05/08/25 | 4:34 |
| 57              | B-101 - Health Office Main Sink Faucet                    | 05/08/25 | 4:36 |
| 58              | B-101 - Health Office Ice Maker                           | 05/08/25 | 4:36 |
| 59              | Drinking Fountain Lower A-B Wing Hallway                  | 05/08/25 | 4:40 |
| 60              | Drinking Fountain Lower A-Wing LEFT                       | 05/08/25 | 4:45 |

Special Instructions / Comments:

Turnaround Requirements

Report Requirements

Metals: RCRA 8•PP 13•TAL 23•TCLP•Other (List)

\_\_\_\_ Rush (Surcharges Apply)  
\*Subject to Availability\*  
\*Please Check with your PM\*  
X \_\_\_\_ Standard (10 Business Days)  
Date Required:

\_\_\_\_ X \_\_\_\_ Tier II/Cat A - Results/QC  
\_\_\_\_ Tier IV/Cat B - Data  
Validation Report w/. Data  
EDD: \_\_\_\_ Yes \_\_\_\_ No  
EDD Type:

VOA/SVOA Report List: TCL • BTEX • TCLP •  
CP-51/Stars • THM • Other: \_\_\_\_\_

Invoice To: ( ☐ Same as Report To )

PO #: 7225-01793

Company: Rush-Henrietta CSD

Contact: Ken Nelson

Email: knelson@rhnet.org

Phone: 585-738-8070

Address: Rush-Henrietta Central School District  
Sperry Senior High School

203

| Relinquished By:                | Received By:                 | Relinquished By: | Received By: | Relinquished By: | Received By: |
|---------------------------------|------------------------------|------------------|--------------|------------------|--------------|
| Signature: <i>James Schultz</i> | Signature: <i>Ken Nelson</i> |                  |              |                  |              |
| Printed Name: James Schultz     | Printed Name: Ken Nelson     |                  |              |                  |              |
| Company: Rush-Henrietta CSD     | Company: Rush-Henrietta CSD  |                  |              |                  |              |
| Date/Time: 5-8-25 8:15          | Date/Time: 5/8/25 8:15       |                  |              |                  |              |

R2505176 5

Rush-Henrietta Central School District  
Sperry Senior High School





# Chain of Custody / Analytical Request Form

1565 Jefferson Road, Building 300, Suite 360 • Rochester, NY 14623 • +1 585 288 5380 • alsglobal.com

SR#: \_\_\_\_\_  
Page \_\_\_\_\_ of \_\_\_\_\_

## Report To:

ALL SHADED AREAS MUST BE COMPLETED BY THE  
CLIENT / SAMPLER

Preservative

|                                   |                                         |
|-----------------------------------|-----------------------------------------|
| Company: Rush-Henrietta CSD       | Project Name: Sperry Senior High School |
| Contact: Ken Nelson               | Project Number:                         |
| Email: knelson@rhnet.org          | ALS Quote #: 38456                      |
| Phone: 585-738-8070               | Sampler's Signature:                    |
| Address: 2034 Lehigh Station Road | Email CC: slacey@rhnet.org              |
| Henrietta NY 14467                | Email CC: jlschultz@rhnet.org           |
| State Samples Collected NY        |                                         |

|        |                      |         |            |                                                                                                                                                         |  |  |  |  |  |  |  |  |  |  |  |  |  |  |  |        |
|--------|----------------------|---------|------------|---------------------------------------------------------------------------------------------------------------------------------------------------------|--|--|--|--|--|--|--|--|--|--|--|--|--|--|--|--------|
| Matrix | Number of Containers | MS/MSD? | 200.8 Lead | Preservative                                                                                                                                            |  |  |  |  |  |  |  |  |  |  |  |  |  |  |  | Notes: |
|        |                      |         |            | 0. None<br>1. HCl<br>2. HNO <sub>3</sub><br>3. H <sub>2</sub> SO <sub>4</sub><br>4. NaOH<br>5. Zn Acet.<br>6. MeOH<br>7. NaHSO <sub>4</sub><br>8. Other |  |  |  |  |  |  |  |  |  |  |  |  |  |  |  |        |

| Lab ID<br>(ALS) | Sample Collection Information:                   |          |      |
|-----------------|--------------------------------------------------|----------|------|
|                 | Sample ID:                                       | Date     | Time |
| 61              | Drinking Fountain Lower A-Wing RIGHT             | 05/08/25 | 4:48 |
| 62              | Main Office Breakroom Sink Faucet                | 05/08/25 | 4:52 |
| 63              | Drinking Fountain Adjacent to Auditorium         | 05/08/25 | 4:58 |
| 64              | D-109 - Music Room Office Sink Faucet            | 05/08/25 | 4:56 |
| 65              | D-203 - Senior Lounge Sink Faucet                | 05/08/25 | 4:59 |
| 66              | D-213 - Literacy Center Kitchenette Sink Faucet  | 05/08/25 | 5:06 |
| 67              | A-220 - Teacher's Lounge Sink Faucet             | 05/08/25 | 5:07 |
| 68              | Drinking Fountain Upper A-Wing                   | 05/08/25 | 5:14 |
| 69              | Adj. A-226 - Drinking Fountain Adjacent to A-226 | 05/08/25 | 5:12 |
| 70              | Drinking Fountain Upper A&B Hallway              | 05/08/25 | 5:15 |

|                                  |                                                                                                                                                                                 |                                                                                                                                               |                                                                               |
|----------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------|
| Special Instructions / Comments: | Turnaround Requirements                                                                                                                                                         | Report Requirements                                                                                                                           | Metals: RCRA 8•PP 13•TAL 23•TCLP•Other (List)                                 |
|                                  | <input type="checkbox"/> Rush (Surcharges Apply)<br>*Subject to Availability*<br>*Please Check with your PM*<br><input checked="" type="checkbox"/> Standard (10 Business Days) | <input checked="" type="checkbox"/> Tier II/Cat A - Results/QC<br><input type="checkbox"/> Tier IV/Cat B - Data<br>Validation Report w/. Data | VOA/SVOA Report List: TCL • BTEX • TCLP •<br>CP-51/Stars • THM • Other: _____ |
|                                  | Date Required:                                                                                                                                                                  | EDD: <input type="checkbox"/> Yes <input type="checkbox"/> No<br>EDD Type:                                                                    | Invoice To: ( <input type="checkbox"/> Same as Report To )                    |
|                                  |                                                                                                                                                                                 |                                                                                                                                               | PO #: 7225-01793<br>Company: Rush-Henrietta CSD                               |

|                                 |                   |                  |              |                  |              |                                             |
|---------------------------------|-------------------|------------------|--------------|------------------|--------------|---------------------------------------------|
| Relinquished By:                | Received By:      | Relinquished By: | Received By: | Relinquished By: | Received By: | Contact: Ken Nelson                         |
| Signature: <i>James Schultz</i> | <i>Ken Nelson</i> |                  |              |                  |              | Email: knelson@rhnet.org                    |
| Printed Name: James Schultz     | <i>Ken Nelson</i> |                  |              |                  |              | Phone: 585-738-8070                         |
| Company: Rush-Henrietta CSD     | <i>Ken Nelson</i> |                  |              |                  |              | Address:                                    |
| Date/Time: 5-8-25 8:15          | 5/8/25 08:15      |                  |              |                  |              | 2034 Lehigh Station Rd. Henrietta, NY 14467 |



## Chain of Custody / Analytical Request Form

1565 Jefferson Road, Building 300, Suite 360 • Rochester, NY 14623 • +1 585 288 5380 • alsglobal.com

SR#:

Page of

## Report To:

ALL SHADED AREAS MUST BE COMPLETED BY THE  
CLIENT / SAMPLER

Preservative

Company: Rush-Henrietta CSD

Project Name: Sperry Senior High School

Contact: Ken Nelson

Project Number:

Email: knelson@rhnet.org

ALS Quote #: 38456

Phone: 585-738-8070

Sampler's Signature:

Address: 2034 Lehigh Station Road

Email CC: slacey@rhnet.org

Henrietta NY 14667

Email CC: jlschultz@rhnet.org

State Samples Collected NY

| Lab ID<br>(ALS) | Sample Collection Information:                 |          |         |
|-----------------|------------------------------------------------|----------|---------|
|                 | Sample ID:                                     | Date     | Time    |
| 71              | Drinking Fountain Opposite B-205               | 05/08/25 | 5:24    |
| 72              | Drinking Fountain Adjacent to Boy's Lavatory   | 05/08/25 | 5:21    |
| 73              | Computer Services Conference Room Sink Faucet  | 05/08/25 | 5:40    |
| 74              | Gordon Stadium Concession Stand Sink Faucet    | 05/08/25 | 5:50    |
| 75              | Drinking Fountain Gordon Stadium Green Monster | 05/08/25 | 5:50    |
| 76              | Bottle Filler Gordon Stadium Green Monster     | 05/08/25 | 5:51    |
| 77              | Drinking Fountain Computer Services LEFT       | 05/08/25 | 5:45    |
| 78              | Drinking Fountain Computer Services RIGHT      | 05/08/25 | Removed |
| 79              | Comet Café Sink Faucet                         | 05/08/25 | 5:02    |
| 80              | Drinking Fountain Music Wing LEFT              | 05/08/25 | 5:00    |

GW  
WW  
SW  
DW  
S  
L  
NA

Matrix

Number of Containers

MS/MSD?

200.8 Lead

0. None  
1. HCl  
2. HNO<sub>3</sub>  
3. H<sub>2</sub>SO<sub>4</sub>  
4. NaOH  
5. Zn Acet.  
6. MeOH  
7. NaHSO<sub>4</sub>  
8. Other

Notes:

## Special Instructions / Comments:

## Turnaround Requirements

\_\_\_\_ Rush (Surcharges Apply)  
\*Subject to Availability\*  
\*Please Check with your PM\*  
X \_\_\_\_ Standard (10 Business Days)

Date Required:

## Report Requirements

\_\_\_\_ X \_\_\_\_ Tier II/Cat A - Results/QC  
\_\_\_\_ Tier IV/Cat B - Data  
Validation Report w/ Data

EDD: \_\_\_\_ Yes \_\_\_\_ No

EDD Type:

Metals: RCRA 8•PP 13•TAL 23•TCLP•Other (List)

VOA/SVOA Report List: TCL • BTEX • TCLP •  
CP-51/Stars • THM • Other: \_\_\_\_\_

Invoice To: (□ Same as Report To)

PO #: 7225-01793

Company: Rush-Henrietta CSD

Contact: Ken Nelson

Email: knelson@rhnet.org

Phone: 585-738-8070

Address:

2034 Lehigh Station Rd. Henrietta, NY 14667

|              | Relinquished By:     | Received By:      | Relinquished By: | Received By: | Relinquished By: | Received By: |
|--------------|----------------------|-------------------|------------------|--------------|------------------|--------------|
| Signature    | <i>James Schultz</i> | <i>Ken Nelson</i> |                  |              |                  |              |
| Printed Name | James Schultz        | Ken Nelson        |                  |              |                  |              |
| Company      | Rush-Henrietta CSD   | ALS               |                  |              |                  |              |
| Date/Time    | 5-8-25 8:15          | 5/8/25 08:15      |                  |              |                  |              |



## Chain of Custody / Analytical Request Form

1565 Jefferson Road, Building 300, Suite 360 • Rochester, NY 14623 • +1 585 288 5380 • alsglobal.com

SR#:

Page of

## Report To:

ALL SHADED AREAS MUST BE COMPLETED BY THE  
CLIENT / SAMPLER

Preservative

|                                                         |                                         |
|---------------------------------------------------------|-----------------------------------------|
| Company: Rush-Henrietta CSD                             | Project Name: Sperry Senior High School |
| Contact: Ken Nelson                                     | Project Number:                         |
| Email: knelson@rhnet.org                                | ALS Quote #: 38456                      |
| Phone: 585-738-8070                                     | Sampler's Signature:                    |
| Address: 2034 Lehigh Station Road<br>Henrietta NY 14467 | Email CC: slacey@rhnet.org              |
|                                                         | Email CC: jlschultz@rhnet.org           |
|                                                         | State Samples Collected NY              |

GW  
WW  
SW  
DW  
S  
L  
NA

Matrix

Number of Containers

MS/MSD?

200.8 Lead

0. None  
1. HCl  
2. HNO<sub>3</sub>  
3. H<sub>2</sub>SO<sub>4</sub>  
4. NaOH  
5. Zn Acet.  
6. MeOH  
7. NaHSO<sub>4</sub>  
8. Other

Notes:

| Lab ID<br>(ALS) | Sample Collection Information:                             |          |      | Matrix | Number of Containers | MS/MSD? | 200.8 Lead |  |  |  |  |  |  |  |  |  |  |  |  |
|-----------------|------------------------------------------------------------|----------|------|--------|----------------------|---------|------------|--|--|--|--|--|--|--|--|--|--|--|--|
|                 | Sample ID:                                                 | Date     | Time |        |                      |         |            |  |  |  |  |  |  |  |  |  |  |  |  |
| 81              | Drinking Fountain Music Wing RIGHT                         | 05/08/25 | 5:00 | DW     |                      |         | X          |  |  |  |  |  |  |  |  |  |  |  |  |
| 82              | Kleeman Stadium Concession Stand Single Sink               | 05/08/25 | 5:45 | DW     |                      |         | X          |  |  |  |  |  |  |  |  |  |  |  |  |
| 83              | Kleeman Stadium Concession Stand Double Sink               | 05/08/25 | 5:45 | DW     |                      |         | X          |  |  |  |  |  |  |  |  |  |  |  |  |
| 84              | Kleeman Stadium Concession Drinking Fountain Green Monster | 05/08/25 | 5:47 | DW     |                      |         | X          |  |  |  |  |  |  |  |  |  |  |  |  |
| 85              | Kleeman Stadium Concession Bottle Filler Green Monster     | 05/08/25 | 5:47 | DW     |                      |         | X          |  |  |  |  |  |  |  |  |  |  |  |  |
|                 |                                                            |          |      | DW     |                      |         | X          |  |  |  |  |  |  |  |  |  |  |  |  |
|                 |                                                            |          |      | DW     |                      |         | X          |  |  |  |  |  |  |  |  |  |  |  |  |
|                 |                                                            |          |      | DW     |                      |         | X          |  |  |  |  |  |  |  |  |  |  |  |  |
|                 |                                                            |          |      | DW     |                      |         | X          |  |  |  |  |  |  |  |  |  |  |  |  |
|                 |                                                            |          |      | DW     |                      |         | X          |  |  |  |  |  |  |  |  |  |  |  |  |

Special Instructions / Comments:

## Turnaround Requirements

Rush (Surcharges Apply)  
\*Subject to Availability\*  
\*Please Check with your PM\*  
X Standard (10 Business Days)  
Date Required:

## Report Requirements

X Tier II/Cat A - Results/QC  
Tier IV/Cat B - Data  
Validation Report w/ Data  
EDD: Yes No  
EDD Type:

Metals: RCRA 8•PP 13•TAL 23•TCLP•Other (List)

VOA/SVOA Report List: TCL • BTEX • TCLP •  
CP-51/Stars • THM • Other:

Invoice To: (Same as Report To)

PO #: 7225-01793

Company: Rush-Henrietta CSD

Contact: Ken Nelson

Email: knelson@rhnet.org

Phone: 585-738-8070

Address:

2034 Lehigh Station Rd. Henrietta, NY 14467

| Relinquished By:                | Received By:                 | Relinquished By: | Received By: | Relinquished By: | Received By: |
|---------------------------------|------------------------------|------------------|--------------|------------------|--------------|
| Signature: <i>James Schmitt</i> | Signature: <i>Ken Nelson</i> |                  |              |                  |              |
| Printed Name: James Schmitt     | Printed Name: Ken Nelson     |                  |              |                  |              |
| Company: Rush-Henrietta CSD     | Company: ALS                 |                  |              |                  |              |
| Date/Time: 5-8-25 8:15          | Date/Time: 5/8/25 08:15      |                  |              |                  |              |



## Cooler Receipt and Preservation Check

R2505176

5

Rush-Henrietta Central School District  
Sperry Senior High SchoolProject/Client Rush Henrietta CSD Folder Number \_\_\_\_\_Cooler received on 5/8/25 by: RECOURIER: ALS UPS FEDEX VELOCITY CLIENT

|   |                                                      |            |
|---|------------------------------------------------------|------------|
| 1 | Were Custody seals on outside of cooler?             | Y <u>N</u> |
| 2 | Custody papers properly completed (ink, signed)?     | Y <u>N</u> |
| 3 | Did all bottles arrive in good condition (unbroken)? | Y <u>N</u> |
| 4 | Circle: Wet Ice Dry Ice Gel packs present?           | Y <u>N</u> |

|    |                                                                       |                              |
|----|-----------------------------------------------------------------------|------------------------------|
| 5a | Did VOA vials have sig* bubbles?                                      | Y <u>N</u> <u>NA</u>         |
| 5b | Sig* bubbles: Alk? Y <u>N</u> <u>NA</u> Sulfide? Y <u>N</u> <u>NA</u> |                              |
| 6  | Where did the bottles originate?                                      | <u>ALS/ROC</u> <u>CLIENT</u> |
| 7  | Soil VOA received as: Bulk Encore 5035set                             | <u>NA</u>                    |

8. Temperature Readings Date: 5/8/25 Time: 08152 ID: IR#12 IR#11 From: Temp Blank Sample Bottle

|                               |             |            |            |            |            |            |            |
|-------------------------------|-------------|------------|------------|------------|------------|------------|------------|
| Temp (°C)                     | <u>17.5</u> |            |            |            |            |            |            |
| Within 0-6°C?                 | Y <u>N</u>  | Y <u>N</u> | Y <u>N</u> | Y <u>N</u> | Y <u>N</u> | Y <u>N</u> | Y <u>N</u> |
| If <0°C, were samples frozen? | Y <u>N</u>  | Y <u>N</u> | Y <u>N</u> | Y <u>N</u> | Y <u>N</u> | Y <u>N</u> | Y <u>N</u> |

If out of Temperature, note packing/ice condition: \_\_\_\_\_ Ice melted Poorly Packed (described below) Same Day Rule  
& Client Approval to Run Samples: \_\_\_\_\_ Standing Approval Client aware at drop-off Client notified by: \_\_\_\_\_

All samples held in storage location: SMD by RE on 5/8/25 at 08:57  
5035 samples placed in storage location: \_\_\_\_\_ by \_\_\_\_\_ on \_\_\_\_\_ at \_\_\_\_\_ within 48 hours of sampling? Y N

Cooler Breakdown/Preservation Check\*\*: Date: 5/13/25 Time: 08:00 by: RE

9. Were all bottle labels complete (i.e. analysis, preservation, etc.)? YES NO  
10. Did all bottle labels and tags agree with custody papers? YES NO  
11. Were correct containers used for the tests indicated? YES NO  
12. Were 5035 vials acceptable (no extra labels, not leaking)? YES NO N/A  
13. Were dissolved metals filtered in the field? YES NO N/A  
14. Air Samples: Cassettes / Tubes Intact Y / N with MS Y / N Canisters Pressurized Tedlar® Bags Inflated N/A

| pH                    | Lot of test paper | Reagent                                       | Preserved? |    | Lot Received                                                                                             | Exp | Sample ID Adjusted | Vol. Added      | Lot Added     | Final pH         |
|-----------------------|-------------------|-----------------------------------------------|------------|----|----------------------------------------------------------------------------------------------------------|-----|--------------------|-----------------|---------------|------------------|
| ≥12                   |                   | NaOH                                          |            |    |                                                                                                          |     |                    |                 |               |                  |
| ≤2                    | <u>226322</u>     | HNO <sub>3</sub>                              | <u>X</u>   |    | <u>None</u>                                                                                              |     | <u>All</u>         | <u>4ml each</u> | <u>241523</u> | <u>All &lt;2</u> |
| ≤2                    |                   | H <sub>2</sub> SO <sub>4</sub>                |            |    |                                                                                                          |     |                    |                 |               |                  |
| <4                    |                   | NaHSO <sub>4</sub>                            |            |    |                                                                                                          |     |                    |                 |               |                  |
| 5-9                   |                   | For 608pest                                   |            |    | No=Notify for 3day                                                                                       |     |                    |                 |               |                  |
| Residual Chlorine (-) |                   | For CN, Phenol, 625, 608pest, 522             |            |    | If +, contact PM to add Na <sub>2</sub> S <sub>2</sub> O <sub>3</sub> (625, 608, CN), ascorbic (phenol). |     |                    |                 |               |                  |
|                       |                   | Na <sub>2</sub> S <sub>2</sub> O <sub>3</sub> |            |    |                                                                                                          |     |                    |                 |               |                  |
|                       |                   | ZnAcetate                                     | -          | -  |                                                                                                          |     |                    |                 |               |                  |
|                       |                   | HCl                                           | **         | ** |                                                                                                          |     |                    |                 |               |                  |

\*\*VOAs and 1664 Not to be tested before analysis.  
Otherwise, all bottles of all samples with chemical preservatives are checked (not just representatives).

Bottle lot numbers: 80419-05  
Explain all Discrepancies/ Other Comments:

\* Missing Sperry 77 (027) Location cancelled in LIMS

|       |        |
|-------|--------|
| HPROD | BULK   |
| HTR   | FLDT   |
| SUB   | HGFB   |
| ALS   | LL3541 |

Labels secondary reviewed by: RE

\*significant air bubbles: VOA &gt; 5-6 mm : WC &gt; 1 in. diameter



## Miscellaneous Forms

**ALS Environmental—Rochester Laboratory**

1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623

Phone (585) 288-5380 Fax (585) 288-8475

[www.alsglobal.com](http://www.alsglobal.com)



## REPORT QUALIFIERS AND DEFINITIONS

|   |                                                                                                                                                                                                                                                                                                        |     |                                                                                                                                                                                                                      |
|---|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| U | Analyte was analyzed for but not detected. The sample quantitation limit has been corrected for dilution and for percent moisture, unless otherwise noted in the case narrative.                                                                                                                       | +   | Correlation coefficient for MSA is <0.995.                                                                                                                                                                           |
| J | Estimated value due to either being a Tentatively Identified Compound (TIC) or that the concentration is between the MRL and the MDL. Concentrations are not verified within the linear range of the calibration. For DoD: concentration >40% difference between two GC columns (pesticides/Aroclors). | N   | Inorganics- Matrix spike recovery was outside laboratory limits.                                                                                                                                                     |
| B | Analyte was also detected in the associated method blank at a concentration that may have contributed to the sample result.                                                                                                                                                                            | N   | Organics- Presumptive evidence of a compound (reported as a TIC) based on the MS library search.                                                                                                                     |
| E | Inorganics- Concentration is estimated due to the serial dilution was outside control limits.                                                                                                                                                                                                          | S   | Concentration has been determined using Method of Standard Additions (MSA).                                                                                                                                          |
| E | Organics- Concentration has exceeded the calibration range for that specific analysis.                                                                                                                                                                                                                 | W   | Post-Digestion Spike recovery is outside control limits and the sample absorbance is <50% of the spike absorbance.                                                                                                   |
| D | Concentration is a result of a dilution, typically a secondary analysis of the sample due to exceeding the calibration range or that a surrogate has been diluted out of the sample and cannot be assessed.                                                                                            | P   | Concentration >40% difference between the two GC columns.                                                                                                                                                            |
| * | Indicates that a quality control parameter has exceeded laboratory limits. Under the "Notes" column of the Form I, this qualifier denotes analysis was performed out of Holding Time.                                                                                                                  | C   | Confirmed by GC/MS                                                                                                                                                                                                   |
| H | Analysis was performed out of hold time for tests that have an "immediate" hold time criteria.                                                                                                                                                                                                         | Q   | DoD reports: indicates a pesticide/Aroclor is not confirmed ( $\geq 100\%$ Difference between two GC columns).                                                                                                       |
| # | Spike was diluted out.                                                                                                                                                                                                                                                                                 | X   | See Case Narrative for discussion.                                                                                                                                                                                   |
|   |                                                                                                                                                                                                                                                                                                        | MRL | Method Reporting Limit. Also known as:                                                                                                                                                                               |
|   |                                                                                                                                                                                                                                                                                                        | LOQ | Limit of Quantitation (LOQ)<br>The lowest concentration at which the method analyte may be reliably quantified under the method conditions.                                                                          |
|   |                                                                                                                                                                                                                                                                                                        | MDL | Method Detection Limit. A statistical value derived from a study designed to provide the lowest concentration that will be detected 99% of the time. Values between the MDL and MRL are estimated (see J qualifier). |
|   |                                                                                                                                                                                                                                                                                                        | LOD | Limit of Detection. A value at or above the MDL which has been verified to be detectable.                                                                                                                            |
|   |                                                                                                                                                                                                                                                                                                        | ND  | Non-Detect. Analyte was not detected at the concentration listed. Same as U qualifier.                                                                                                                               |

### Rochester Lab ID # for State Accreditations<sup>1</sup>



| NELAP States            |
|-------------------------|
| Florida ID # E87674     |
| New Hampshire ID # 2941 |
| New York ID # 10145     |
| Pennsylvania ID# 68-786 |
| Texas ID#T104704581     |
| Virginia #460167        |

| Non-NELAP States       |
|------------------------|
| Connecticut ID #PH0556 |
| Delaware Approved      |
| Maine ID #NY01587      |
| North Carolina #36701  |
| North Carolina #676    |
| Rhode Island LAO00333  |

<sup>1</sup> Analyses were performed according to our laboratory's NELAP-approved quality assurance program and any applicable state or agency requirements. The test results meet requirements of the current NELAP/TNI standards or state or agency requirements, where applicable, except as noted in the case narrative. Since not all analyte/method/matrix combinations are offered for state/NELAC accreditation, this report may contain results which are not accredited. For a specific list of accredited analytes, contact the laboratory. To verify NH accredited analytes, go to <https://www4.des.state.nh.us/CertifiedLabs/Certified-Method.aspx>.

# ALS Laboratory Group

---

## Acronyms

|            |                                                                                                                                          |
|------------|------------------------------------------------------------------------------------------------------------------------------------------|
| ASTM       | American Society for Testing and Materials                                                                                               |
| A2LA       | American Association for Laboratory Accreditation                                                                                        |
| CARB       | California Air Resources Board                                                                                                           |
| CAS Number | Chemical Abstract Service registry Number                                                                                                |
| CFC        | Chlorofluorocarbon                                                                                                                       |
| CFU        | Colony-Forming Unit                                                                                                                      |
| DEC        | Department of Environmental Conservation                                                                                                 |
| DEQ        | Department of Environmental Quality                                                                                                      |
| DHS        | Department of Health Services                                                                                                            |
| DOE        | Department of Ecology                                                                                                                    |
| DOH        | Department of Health                                                                                                                     |
| EPA        | U. S. Environmental Protection Agency                                                                                                    |
| ELAP       | Environmental Laboratory Accreditation Program                                                                                           |
| GC         | Gas Chromatography                                                                                                                       |
| GC/MS      | Gas Chromatography/Mass Spectrometry                                                                                                     |
| LUFT       | Leaking Underground Fuel Tank                                                                                                            |
| M          | Modified                                                                                                                                 |
| MCL        | Maximum Contaminant Level is the highest permissible concentration of a substance allowed in drinking water as established by the USEPA. |
| MDL        | Method Detection Limit                                                                                                                   |
| MPN        | Most Probable Number                                                                                                                     |
| MRL        | Method Reporting Limit                                                                                                                   |
| NA         | Not Applicable                                                                                                                           |
| NC         | Not Calculated                                                                                                                           |
| NCASI      | National Council of the Paper Industry for Air and Stream Improvement                                                                    |
| ND         | Not Detected                                                                                                                             |
| NIOSH      | National Institute for Occupational Safety and Health                                                                                    |
| PQL        | Practical Quantitation Limit                                                                                                             |
| RCRA       | Resource Conservation and Recovery Act                                                                                                   |
| SIM        | Selected Ion Monitoring                                                                                                                  |
| TPH        | Total Petroleum Hydrocarbons                                                                                                             |
| tr         | Trace level is the concentration of an analyte that is less than the PQL but greater than or equal to the MDL.                           |

ALS Group USA, Corp.  
dba ALS Environmental

Analyst Summary report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School/

**Service Request:** R2505176

**Sample Name:** Sperry-51 B-133 - Kitchenette Sink Faucet  
**Lab Code:** R2505176-001  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-52 Adj Mop Sink - Lower B-Wing  
**Lab Code:** R2505176-002  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-53 B-106 - Drinking Fountain  
**Lab Code:** R2505176-003  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-54 Drinking Fountain Adj to Lower  
**Lab Code:** R2505176-004  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-55 B-104 - Drinking Fountain  
**Lab Code:** R2505176-005  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

ALS Group USA, Corp.  
dba ALS Environmental

Analyst Summary report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School/

**Service Request:** R2505176

**Sample Name:** Sperry-56 B-101 - Health Office Exam  
**Lab Code:** R2505176-006  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-57 B-101 - Health Office Main  
**Lab Code:** R2505176-007  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-58 B-101 - Health Office Ice Maker  
**Lab Code:** R2505176-008  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-59 Drinking Fountain Lower A-B  
**Lab Code:** R2505176-009  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-60 Drinking Fountain Lower A-  
**Lab Code:** R2505176-010  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

ALS Group USA, Corp.  
dba ALS Environmental

Analyst Summary report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School/

**Service Request:** R2505176

**Sample Name:** Sperry-61 Drinking Fountain Lower A-  
**Lab Code:** R2505176-011  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-62 Main Office Breakroom Sink  
**Lab Code:** R2505176-012  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-63 Drinking Fountain Adjacent to  
**Lab Code:** R2505176-013  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-64 D-109 - Music Room Office  
**Lab Code:** R2505176-014  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-65 D-203 - Senior Lounge SInk  
**Lab Code:** R2505176-015  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

ALS Group USA, Corp.  
dba ALS Environmental

Analyst Summary report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School/

**Service Request:** R2505176

**Sample Name:** Sperry-66 D-213 - Literacy Center  
**Lab Code:** R2505176-016  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-67 A-220 - Teacher's Lounge Sink  
**Lab Code:** R2505176-017  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-68 Drinking Fountain Upper A-  
**Lab Code:** R2505176-018  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-69 Adj. A-226 - Drinking Fountain  
**Lab Code:** R2505176-019  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-70 Drinking Fountain Upper A&B  
**Lab Code:** R2505176-020  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

ALS Group USA, Corp.  
dba ALS Environmental

Analyst Summary report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School/

**Service Request:** R2505176

**Sample Name:** Sperry-71 Drinking Fountain Opposite B-  
**Lab Code:** R2505176-021  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-72 Drinking Fountain Adjacent to  
**Lab Code:** R2505176-022  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-73 Computer Services Conference  
**Lab Code:** R2505176-023  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-74 Gordon Stadium Concession  
**Lab Code:** R2505176-024  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-75 Drinking Fountain Gordon  
**Lab Code:** R2505176-025  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

ALS Group USA, Corp.  
dba ALS Environmental

Analyst Summary report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School/

**Service Request:** R2505176

**Sample Name:** Sperry-76 Bottle Filler Gordon Stadium  
**Lab Code:** R2505176-026  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-79 Drinking Fountain Comet Cafe  
**Lab Code:** R2505176-028  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-80 Drinking Fountain Music Wing  
**Lab Code:** R2505176-029  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-81 Drinking Fountain Music Wing  
**Lab Code:** R2505176-030  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
MKASTAN

**Sample Name:** Sperry-82 Kleeman Stadium Concession  
**Lab Code:** R2505176-031  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
NMANSEN

ALS Group USA, Corp.  
dba ALS Environmental

Analyst Summary report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School/

**Service Request:** R2505176

**Sample Name:** Sperry-83 Kleeman Stadium Concession  
**Lab Code:** R2505176-032  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
NMANSEN

**Sample Name:** Sperry-84 Kleeman Stadium Concession  
**Lab Code:** R2505176-033  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
NMANSEN

**Sample Name:** Sperry-85 Kleeman Stadium Concession  
**Lab Code:** R2505176-034  
**Sample Matrix:** Drinking Water

**Date Collected:** 05/8/25  
**Date Received:** 05/8/25

**Analysis Method**  
200.8

**Extracted/Digested By**

**Analyzed By**  
NMANSEN



## PREPARATION METHODS

The preparation methods associated with this report are found in these tables unless discussed in the case narrative.

### INORGANIC

#### Water/Liquid Matrix

| Analytical Method                                     | Preparation Method                      |
|-------------------------------------------------------|-----------------------------------------|
| 200.7                                                 | 200.2                                   |
| 200.8                                                 | 200.2                                   |
| 6010C or 6010D                                        | 3005A/3010A                             |
| 6020A or 6020B                                        | ILM05.3                                 |
| 9034 Sulfide Acid Soluble                             | 9030B                                   |
| SM 4500-CN-N-2016<br>Amenable and Residual<br>Cyanide | SM 4500-CN-G and<br>SM 4500-CN-B,C-2016 |
| SM 4500-CN-E WAD<br>Cyanide                           | SM 4500-CN-I                            |

#### Solid/Soil/Non-Aqueous Matrix

| Analytical Method                                                                                         | Preparation Method |
|-----------------------------------------------------------------------------------------------------------|--------------------|
| 6010C or 6010D                                                                                            | 3050B              |
| 6020A or 6020B                                                                                            | 3050B              |
| 6010C or 6010D TCLP<br>(1311) extract                                                                     | 3005A/3010A        |
| 6010C or 6010D SPLP<br>(1312) extract                                                                     | 3005A/3010A        |
| 7199                                                                                                      | 3060A              |
| 300.0 Anions/ 350.1/ 353.2/<br>SM 2320B/ SM 5210B/<br>9056A Anions                                        | DI extraction      |
| For analytical methods not listed, the preparation method is the same as the analytical method reference. |                    |

### ORGANIC

**Preparation Methods for Organic methods are listed in the header of the Results pages.**

#### Regarding "Bulk/5035A":

For soil/solid samples submitted in soil jars for Volatiles analysis, the prep method is listed as "Bulk/5035A". The lab follows the closed-system EPA 5035A protocols once the sample is transferred to a sealed vial, but collection in bulk in soil jars does not follow the collection protocols listed in EPA 5035A. In accordance with the NYSDOH technical notice of October 2012, all results or reporting limits <200 ug/kg are to be considered estimated due to potential low bias.



## Sample Results

**ALS Environmental—Rochester Laboratory**

1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623

Phone (585) 288-5380 Fax (585) 288-8475

[www.alsglobal.com](http://www.alsglobal.com)



## Metals

**ALS Environmental—Rochester Laboratory**

1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623

Phone (585) 288-5380 Fax (585) 288-8475

[www.alsglobal.com](http://www.alsglobal.com)

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water  
**Sample Name:** Sperry-51 B-133 - Kitchenette Sink Faucet  
**Lab Code:** R2505176-001

**Service Request:** R2505176  
**Date Collected:** 05/08/25 04:30  
**Date Received:** 05/08/25 08:15  
**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 1.0 U  | ug/L  | 1.0 | 1    | 05/19/25 19:54 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 04:40  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-52 Adj Mop Sink - Lower B-Wing Drinking Fountain  
**Lab Code:** R2505176-002

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 1.0 U  | ug/L  | 1.0 | 1    | 05/19/25 19:55 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water  
**Sample Name:** Sperry-53 B-106 - Drinking Fountain  
**Lab Code:** R2505176-003

**Service Request:** R2505176  
**Date Collected:** 05/08/25 04:42  
**Date Received:** 05/08/25 08:15  
**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 15.2   | ug/L  | 1.0 | 1    | 05/19/25 19:57 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District

**Project:** Sperry Senior High School

**Sample Matrix:** Drinking Water

**Service Request:** R2505176

**Date Collected:** 05/08/25 04:45

**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-54 Drinking Fountain Adj to Lower B-Wing Boy's Lav

**Basis:** NA

**Lab Code:** R2505176-004

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 1.0 U  | ug/L  | 1.0 | 1    | 05/19/25 19:58 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water  
**Sample Name:** Sperry-55 B-104 - Drinking Fountain  
**Lab Code:** R2505176-005

**Service Request:** R2505176  
**Date Collected:** 05/08/25 04:42  
**Date Received:** 05/08/25 08:15  
**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 9.4    | ug/L  | 1.0 | 1    | 05/19/25 20:00 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 04:34  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-56 B-101 - Health Office Exam Room Sink Faucet  
**Lab Code:** R2505176-006

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 3.1    | ug/L  | 1.0 | 1    | 05/19/25 20:04 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 04:36  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-57 B-101 - Health Office Main Sink Faucet  
**Lab Code:** R2505176-007

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 6.4    | ug/L  | 1.0 | 1    | 05/19/25 20:06 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water  
**Sample Name:** Sperry-58 B-101 - Health Office Ice Maker  
**Lab Code:** R2505176-008

**Service Request:** R2505176  
**Date Collected:** 05/08/25 04:36  
**Date Received:** 05/08/25 08:15  
**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 4.0    | ug/L  | 1.0 | 1    | 05/19/25 20:07 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 04:40  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-59 Drinking Fountain Lower A-B Wing Hallway  
**Lab Code:** R2505176-009

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 2.8    | ug/L  | 1.0 | 1    | 05/19/25 20:09 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 04:45  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-60 Drinking Fountain Lower A-Wing LEFT  
**Lab Code:** R2505176-010

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 2.0    | ug/L  | 1.0 | 1    | 05/19/25 20:19 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 04:48  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-61 Drinking Fountain Lower A-Wing RIGHT  
**Lab Code:** R2505176-011

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 2.1    | ug/L  | 1.0 | 1    | 05/19/25 20:23 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District

**Project:** Sperry Senior High School

**Sample Matrix:** Drinking Water

**Service Request:** R2505176

**Date Collected:** 05/08/25 04:52

**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-62 Main Office Breakroom Sink Faucet

**Basis:** NA

**Lab Code:** R2505176-012

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 18.3   | ug/L  | 1.0 | 1    | 05/19/25 20:25 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 04:58  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-63 Drinking Fountain Adjacent to Auditorium  
**Lab Code:** R2505176-013

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 1.7    | ug/L  | 1.0 | 1    | 05/19/25 20:26 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 04:56  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-64 D-109 - Music Room Office Sink Faucet  
**Lab Code:** R2505176-014

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 139    | ug/L  | 1.0 | 1    | 05/19/25 20:28 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District

**Project:** Sperry Senior High School

**Sample Matrix:** Drinking Water

**Sample Name:** Sperry-65 D-203 - Senior Lounge SInk Faucet

**Lab Code:** R2505176-015

**Service Request:** R2505176

**Date Collected:** 05/08/25 04:59

**Date Received:** 05/08/25 08:15

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 33.4   | ug/L  | 1.0 | 1    | 05/19/25 20:29 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 05:06  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-66 D-213 - Literacy Center Kitchenette Sink Faucet  
**Lab Code:** R2505176-016

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 29.9   | ug/L  | 1.0 | 1    | 05/19/25 20:34 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 05:07  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-67 A-220 - Teacher's Lounge Sink Faucet  
**Lab Code:** R2505176-017

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 1.0 U  | ug/L  | 1.0 | 1    | 05/19/25 20:35 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District

**Project:** Sperry Senior High School

**Sample Matrix:** Drinking Water

**Service Request:** R2505176

**Date Collected:** 05/08/25 05:14

**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-68 Drinking Fountain Upper A-Wing

**Basis:** NA

**Lab Code:** R2505176-018

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 3.8    | ug/L  | 1.0 | 1    | 05/19/25 20:37 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District

**Project:** Sperry Senior High School

**Sample Matrix:** Drinking Water

**Service Request:** R2505176

**Date Collected:** 05/08/25 05:12

**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-69 Adj. A-226 - Drinking Fountain Adjacent to A-226

**Basis:** NA

**Lab Code:** R2505176-019

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 1.6    | ug/L  | 1.0 | 1    | 05/19/25 20:38 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 05:15  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-70 Drinking Fountain Upper A&B Hallway  
**Lab Code:** R2505176-020

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 1.9    | ug/L  | 1.0 | 1    | 05/19/25 20:39 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District

**Project:** Sperry Senior High School

**Sample Matrix:** Drinking Water

**Sample Name:** Sperry-71 Drinking Fountain Opposite B-205

**Lab Code:** R2505176-021

**Service Request:** R2505176

**Date Collected:** 05/08/25 05:24

**Date Received:** 05/08/25 08:15

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 2.4    | ug/L  | 1.0 | 1    | 05/19/25 20:41 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 05:21  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-72 Drinking Fountain Adjacent to Boy's Lavatory  
**Lab Code:** R2505176-022

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 1.2    | ug/L  | 1.0 | 1    | 05/19/25 20:42 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 05:40  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-73 Computer Services Conference Room Sink Faucet  
**Lab Code:** R2505176-023

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 1.0 U  | ug/L  | 1.0 | 1    | 05/19/25 20:44 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 05:50  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-74 Gordon Stadium Concession Stand Sink Faucet  
**Lab Code:** R2505176-024

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 1.0 U  | ug/L  | 1.0 | 1    | 05/19/25 20:45 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 05:50  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-75 Drinking Fountain Gordon Stadium Green Monster  
**Lab Code:** R2505176-025

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 1.5    | ug/L  | 1.0 | 1    | 05/19/25 20:47 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 05:51  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-76 Bottle Filler Gordon Stadium Green Monster  
**Lab Code:** R2505176-026

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 1.0 U  | ug/L  | 1.0 | 1    | 05/19/25 20:51 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 05:02  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-79 Drinking Fountain Comet Cafe Sink Faucet  
**Lab Code:** R2505176-028

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 5.0    | ug/L  | 1.0 | 1    | 05/19/25 20:53 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 05:00  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-80 Drinking Fountain Music Wing LEFT  
**Lab Code:** R2505176-029

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 11.3   | ug/L  | 1.0 | 1    | 05/19/25 20:54 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 05:00  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-81 Drinking Fountain Music Wing RIGHT  
**Lab Code:** R2505176-030

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 12.2   | ug/L  | 1.0 | 1    | 05/19/25 20:56 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 05:45  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-82 Kleeman Stadium Concession Stand Single Sink  
**Lab Code:** R2505176-031

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 1.1    | ug/L  | 1.0 | 1    | 05/20/25 21:41 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 05:45  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-83 Kleeman Stadium Concession Stand Double Sink  
**Lab Code:** R2505176-032

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 1.0 U  | ug/L  | 1.0 | 1    | 05/20/25 21:43 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 05:47  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-84 Kleeman Stadium Concession Drinking Fountain Green  
**Lab Code:** R2505176-033

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 1.0 U  | ug/L  | 1.0 | 1    | 05/20/25 21:44 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25 05:47  
**Date Received:** 05/08/25 08:15

**Sample Name:** Sperry-85 Kleeman Stadium Concession Bottle Filler Green Mon  
**Lab Code:** R2505176-034

**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 1.0 U  | ug/L  | 1.0 | 1    | 05/20/25 21:45 |   |



## QC Summary Forms

**ALS Environmental—Rochester Laboratory**

1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623

Phone (585) 288-5380 Fax (585) 288-8475

[www.alsglobal.com](http://www.alsglobal.com)



## Metals

**ALS Environmental—Rochester Laboratory**

1565 Jefferson Road, Building 300, Suite 360, Rochester, NY 14623

Phone (585) 288-5380 Fax (585) 288-8475

[www.alsglobal.com](http://www.alsglobal.com)

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water  
**Sample Name:** Method Blank  
**Lab Code:** R2505176-MB1

**Service Request:** R2505176  
**Date Collected:** NA  
**Date Received:** NA  
**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 1.0 U  | ug/L  | 1.0 | 1    | 05/19/25 19:29 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water  
**Sample Name:** Method Blank  
**Lab Code:** R2505176-MB2

**Service Request:** R2505176  
**Date Collected:** NA  
**Date Received:** NA  
**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 1.0 U  | ug/L  | 1.0 | 1    | 05/19/25 20:16 |   |

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water  
**Sample Name:** Method Blank  
**Lab Code:** R2505176-MB3

**Service Request:** R2505176  
**Date Collected:** NA  
**Date Received:** NA  
**Basis:** NA

Inorganic Parameters

| Analyte Name | Analysis Method | Result | Units | MRL | Dil. | Date Analyzed  | Q |
|--------------|-----------------|--------|-------|-----|------|----------------|---|
| Lead, Total  | 200.8           | 1.0 U  | ug/L  | 1.0 | 1    | 05/20/25 21:08 |   |

ALS Group USA, Corp.  
dba ALS Environmental

QA/QC Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25  
**Date Received:** 05/08/25  
**Date Analyzed:** 05/19/25

**Duplicate Matrix Spike Summary**  
**Inorganic Parameters**

**Sample Name:** Sperry-59 Drinking Fountain Lower A-B Wing Hallway  
**Lab Code:** R2505176-009  
**Analysis Method:** 200.8  
**Units:** ug/L  
**Basis:** NA

| Analyte Name | Sample Result | Result | Matrix Spike<br>R2505176-009MS |       | Result | Duplicate Matrix Spike<br>R2505176-009DMS |       | % Rec Limits | RPD | RPD Limit |
|--------------|---------------|--------|--------------------------------|-------|--------|-------------------------------------------|-------|--------------|-----|-----------|
|              |               |        | Spike Amount                   | % Rec |        | Spike Amount                              | % Rec |              |     |           |
| Lead, Total  | 2.8           | 23.2   | 20.0                           | 102   | 23.4   | 20.0                                      | 103   | 70-130       | <1  | 20        |

Results flagged with an asterisk (\*) indicate values outside control criteria.

Results flagged with a pound (#) indicate the control criteria is not applicable.

Percent recoveries and relative percent differences (RPD) are determined by the software using values in the calculation which have not been rounded.

Matrix Spike and Matrix Spike Duplicate Data is presented for information purposes only. The matrix may or may not be relevant to samples reported in this report. The laboratory evaluates system performance based on the LCS and LCSD control limits.

ALS Group USA, Corp.  
dba ALS Environmental

QA/QC Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25  
**Date Received:** 05/08/25  
**Date Analyzed:** 05/19/25

**Duplicate Matrix Spike Summary**  
**Inorganic Parameters**

**Sample Name:** Sperry-60 Drinking Fountain Lower A-Wing  
LEFT  
**Lab Code:** R2505176-010  
**Analysis Method:** 200.8

**Units:** ug/L  
**Basis:** NA

| Analyte Name | Sample Result | Result | Matrix Spike<br>R2505176-010MS |       | Result | Duplicate Matrix Spike<br>R2505176-010DMS |       | % Rec Limits | RPD | RPD Limit |
|--------------|---------------|--------|--------------------------------|-------|--------|-------------------------------------------|-------|--------------|-----|-----------|
|              |               |        | Spike Amount                   | % Rec |        | Spike Amount                              | % Rec |              |     |           |
| Lead, Total  | 2.0           | 21.5   | 20.0                           | 97    | 21.6   | 20.0                                      | 98    | 70-130       | <1  | 20        |

Results flagged with an asterisk (\*) indicate values outside control criteria.

Results flagged with a pound (#) indicate the control criteria is not applicable.

Percent recoveries and relative percent differences (RPD) are determined by the software using values in the calculation which have not been rounded.

Matrix Spike and Matrix Spike Duplicate Data is presented for information purposes only. The matrix may or may not be relevant to samples reported in this report. The laboratory evaluates system performance based on the LCS and LCSD control limits.

ALS Group USA, Corp.  
dba ALS Environmental

QA/QC Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25  
**Date Received:** 05/08/25  
**Date Analyzed:** 05/19/25

**Duplicate Matrix Spike Summary**  
**Inorganic Parameters**

**Sample Name:** Sperry-81 Drinking Fountain Music Wing RIGHT  
**Lab Code:** R2505176-030  
**Analysis Method:** 200.8

**Units:** ug/L  
**Basis:** NA

| Analyte Name | Sample Result | Matrix Spike<br>R2505176-030MS |              |       |        | Duplicate Matrix Spike<br>R2505176-030DMS |       |              |     | RPD Limit |
|--------------|---------------|--------------------------------|--------------|-------|--------|-------------------------------------------|-------|--------------|-----|-----------|
|              |               | Result                         | Spike Amount | % Rec | Result | Spike Amount                              | % Rec | % Rec Limits | RPD |           |
|              |               |                                |              |       |        |                                           |       |              |     |           |
| Lead, Total  | 12.2          | 33.7                           | 20.0         | 107   | 32.8   | 20.0                                      | 103   | 70-130       | 3   | 20        |

Results flagged with an asterisk (\*) indicate values outside control criteria.

Results flagged with a pound (#) indicate the control criteria is not applicable.

Percent recoveries and relative percent differences (RPD) are determined by the software using values in the calculation which have not been rounded.

Matrix Spike and Matrix Spike Duplicate Data is presented for information purposes only. The matrix may or may not be relevant to samples reported in this report. The laboratory evaluates system performance based on the LCS and LCSD control limits.

ALS Group USA, Corp.  
dba ALS Environmental

QA/QC Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Collected:** 05/08/25  
**Date Received:** 05/08/25  
**Date Analyzed:** 05/20/25

**Duplicate Matrix Spike Summary**  
**Inorganic Parameters**

**Sample Name:** Sperry-85 Kleeman Stadium Concession Bottle  
Filler Green Mon  
**Lab Code:** R2505176-034  
**Analysis Method:** 200.8

**Units:** ug/L  
**Basis:** NA

| Analyte Name | Sample Result | Result | Matrix Spike |       | Result | Duplicate Matrix Spike |       | % Rec Limits | RPD | RPD Limit |
|--------------|---------------|--------|--------------|-------|--------|------------------------|-------|--------------|-----|-----------|
|              |               |        | Spike Amount | % Rec |        | Spike Amount           | % Rec |              |     |           |
| Lead, Total  | 1.0 U         | 19.7   | 20.0         | 99    | 19.4   | 20.0                   | 97    | 70-130       | 2   | 20        |

Results flagged with an asterisk (\*) indicate values outside control criteria.

Results flagged with a pound (#) indicate the control criteria is not applicable.

Percent recoveries and relative percent differences (RPD) are determined by the software using values in the calculation which have not been rounded.

Matrix Spike and Matrix Spike Duplicate Data is presented for information purposes only. The matrix may or may not be relevant to samples reported in this report. The laboratory evaluates system performance based on the LCS and LCSD control limits.

ALS Group USA, Corp.  
dba ALS Environmental

QA/QC Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Analyzed:** 05/19/25

**Lab Control Sample Summary**  
**Inorganic Parameters**

**Units:**ug/L  
**Basis:**NA

**Lab Control Sample**  
R2505176-LCS1

| Analyte Name | Analytical Method | Result | Spike Amount | % Rec | % Rec Limits |
|--------------|-------------------|--------|--------------|-------|--------------|
| Lead, Total  | 200.8             | 19.3   | 20.0         | 96    | 85-115       |

ALS Group USA, Corp.  
dba ALS Environmental

QA/QC Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Analyzed:** 05/19/25

**Lab Control Sample Summary**  
**Inorganic Parameters**

**Units:**ug/L  
**Basis:**NA

**Lab Control Sample**  
R2505176-LCS2

| Analyte Name | Analytical Method | Result | Spike Amount | % Rec | % Rec Limits |
|--------------|-------------------|--------|--------------|-------|--------------|
| Lead, Total  | 200.8             | 19.6   | 20.0         | 98    | 85-115       |

ALS Group USA, Corp.  
dba ALS Environmental

QA/QC Report

**Client:** Rush-Henrietta Central School District  
**Project:** Sperry Senior High School  
**Sample Matrix:** Drinking Water

**Service Request:** R2505176  
**Date Analyzed:** 05/20/25

**Lab Control Sample Summary**  
**Inorganic Parameters**

**Units:**ug/L  
**Basis:**NA

**Lab Control Sample**  
R2505176-LCS3

| Analyte Name | Analytical Method | Result | Spike Amount | % Rec | % Rec Limits |
|--------------|-------------------|--------|--------------|-------|--------------|
| Lead, Total  | 200.8             | 20.7   | 20.0         | 104   | 85-115       |