



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

June 08, 2026

John Harding
Eagle Point School District #9
PO Box 548
Eagle Point, OR 97524
TEL: (541) 830-1240
FAX (541) 830-6375

RE: Lead Retests

Order No.: 26051041

Dear John Harding:

Neilson Research Corporation received 3 sample(s) on 5/27/2026 for the analyses presented in the following report.

The results relate only to the parameters tested or to the sample as received by the laboratory. This report shall not be reproduced except in full, without the written approval of Neilson Research Corporation. Testing was completed in accordance with the laboratory's established quality system and standard operating procedures. For the current scope of NELAP-accredited analytes and certifications, please refer to the Certifications section at www.nrclabs.com. If you have any questions regarding these test results, please feel free to call.

Sincerely,
Neilson Research Corporation

Tamra Schmedemann
Director of Project Management
245 S Grape St
Medford, OR 97501





Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Case Narrative

WO#: 26051041
Date: 6/8/2026

CLIENT: Eagle Point School District #9
Project: Lead Retests

The analyses were performed according to the guidelines in the Neilson Research Corporation Quality Assurance Program. This report contains analytical results for the sample(s) as received by the laboratory.

Neilson Research Corporation certifies that this report is in compliance with the requirements of NELAP. No unusual difficulties were experienced during analysis of this batch except as noted below or qualified with data flags on the reports.

Original



Neilson Research Corporation
 245 S Grape St
 Medford, OR 97501
 TEL: (541) 770-5678 FAX: (541) 770-2901
 Website: www.nrclabs.com

Analytical Report

WO#: 26051041
 Date Reported: 6/8/2026

Eagle Point School District #9
 PO Box 548
 Eagle Point, OR 97524

Lab Order: 26051041
Received Date: 5/27/2026 9:09:00 AM
Reported Date: 6/8/2026 10:14:31 AM

Sample Information:

Lab ID: 26051041-01 Client Sample ID: 063CF
 Collection Date: 5/26/2026 7:18:00 AM Collected By: Loni O'Connor
 Matrix: Drinking Water Sample Location: 063CF

Trace Metals by EPA 200.8 ICP-MS								Analyst; PR
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	1.77		0.500	ppb	1	6/2/2026	15.0	A

Lab ID: 26051041-02 Client Sample ID: 007CF
 Collection Date: 5/26/2026 1:02:00 PM Collected By: Loni O'Connor
 Matrix: Drinking Water Sample Location: 007CF

Trace Metals by EPA 200.8 ICP-MS								Analyst; PR
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	1.51		0.500	ppb	1	6/2/2026	15.0	A

Lab ID: 26051041-03 Client Sample ID: 077BF
 Collection Date: 5/26/2026 1:05:00 PM Collected By: Loni O'Connor
 Matrix: Drinking Water Sample Location: 077BF

Trace Metals by EPA 200.8 ICP-MS								Analyst; PR
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	ND		0.500	ppb	1	6/2/2026	15.0	A

QUALIFIERS	* Value exceeds Maximum or Minimum Contaminant Level.	C1 Sample container temperature is out of limit as specified at testcod
	DF Dilution Factor	E Value above quantitation range
	H Holding times for preparation or analysis exceeded	J Analyte detected below quantitation limits
	MI Recovery outside control limits due to Matrix Interference	ND Not Detected at the Reporting Limit
	R RPD outside accepted recovery limits	RL Reporting Limit

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original



Neilson Research Corporation
 245 S Grape St
 Medford, OR 97501
 TEL: (541) 770-5678 FAX: (541) 770-2901
 Website: www.nrclabs.com

QC SUMMARY REPORT

WO#: 26051041
 08-Jun-26

Client: Eagle Point School District #9
Project: Lead Retests

TestCode: LEAD_SCHOOLS

Sample ID: MB-37018	SampType: MBLK	TestCode: LEAD_SCHO	Units: ppb	Prep Date: 6/2/2026	RunNo: 70119						
Client ID: PBW	Batch ID: 37018	TestNo: E200.8	E200.8	Analysis Date: 6/2/2026	SeqNo: 1156619						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Lead ND 0.500

Sample ID: LCS-37018	SampType: LCS	TestCode: LEAD_SCHO	Units: ppb	Prep Date: 6/2/2026	RunNo: 70119						
Client ID: LCSW	Batch ID: 37018	TestNo: E200.8	E200.8	Analysis Date: 6/2/2026	SeqNo: 1156620						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Lead 104 0.500 100 0 104 85 115

Sample ID: 26051172-01AMS	SampType: MS	TestCode: LEAD_SCHO	Units: ppb	Prep Date: 6/2/2026	RunNo: 70119						
Client ID: BatchQC	Batch ID: 37018	TestNo: E200.8	E200.8	Analysis Date: 6/2/2026	SeqNo: 1156622						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Lead 138 0.500 100 35.9 102 70 130

Sample ID: 26051172-01AMSD	SampType: MSD	TestCode: LEAD_SCHO	Units: ppb	Prep Date: 6/2/2026	RunNo: 70119						
Client ID: BatchQC	Batch ID: 37018	TestNo: E200.8	E200.8	Analysis Date: 6/2/2026	SeqNo: 1156623						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Lead 136 0.500 100 35.9 101 70 130 138 0.947 20

Qualifiers: CI Sample container temperature is out of limit as specified at testcode DF Dilution Factor H Holding times for preparation or analysis exceed
 MI Recovery outside control limits due to Matrix Interference ND Not Detected at the Reporting Limit RL Reporting Limit

Original

Chain of Custody Record

This Chain of Custody is a LEGAL DOCUMENT and must be filled out accurately.

Section A Required Client Information		Section B Required Project Information		Section C Invoice Information		Section D Rush Status (Subject to Scheduling)	
Company: <u>JCS P9</u>		Project Name: <u>Re-tests</u>		Attention: <u>John Harding</u>		<input type="checkbox"/> Standard: 10 Business Days <input type="checkbox"/> Priority: 5 Business Days (List x 1.50) <input type="checkbox"/> Express: 3 Business Days (List x 1.75) <input type="checkbox"/> Rush: 2 Business Days (List x 2.00) <input type="checkbox"/> Rush: 1 Business Day (List x 2.50) <input type="checkbox"/> Rush: Same Day (List x 3.00)	
Address: <u>Eagle Pt School Dist.</u>		Project Number:		Company Name:		<input type="checkbox"/> Authorized <input type="checkbox"/> Yes <input type="checkbox"/> No	
Email:		Report To:		Address:			
Phone: _____ Fax: _____		Copy To:		P.O. #			
Collected By (Print): <u>Loni Oconnor</u>							
Collected By (Sign): <u>[Signature]</u>							
Email Report <input type="checkbox"/> Mail Report <input type="checkbox"/> Fax Report <input type="checkbox"/>							

Section E Sample Information					Analysis Requested												NRC Workorder # (Lab Use Only)		
Sample ID	Comp/Grab	Matrix*	Date Collected	Time Collected	No. of Containers												Remarks / Field Data	NRC Sample # (Use Only)	(Lab)
<u>063CF</u>			<u>5/26/26</u>	<u>7:18</u>	<u>1</u>	Lead 											<u>01</u>		
<u>007 CF</u>				<u>1:02</u>	<u>1</u>												<u>02</u>		
<u>077 BF</u>				<u>1:05</u>	<u>1</u>												<u>03</u>		

*Matrix: DW - Drinking Water WW - Wastewater W - Water S - Soil/Solid SL - Sludge O - Oil WP - Wipe OT - Other

Section F Relinquish/Receive		Section G Lab Use Only	
Relinquished By: <u>[Signature]</u>	Received By: <u>Loni Oconnor</u>	Temp: <u>Amb</u> IR Therm ID: <u>N/A</u>	<input type="checkbox"/> ≤6°C: Yes <input type="checkbox"/> No <input type="checkbox"/> Received on Ice: Yes <input type="checkbox"/> No Number of Bottles Received: _____ pH Checked: _____ COC Seals Intact: <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> NA Field Blank Included: <input type="checkbox"/> Yes <input type="checkbox"/> No
Relinquished By:	Received By:	Date: <u>5/26/26</u> Time: <u>9:09</u>	Received Via: <input type="checkbox"/> UPS <input type="checkbox"/> FedEX <input type="checkbox"/> Other <input checked="" type="checkbox"/> Hand
Relinquished By:	Received By:	Date: <u>5/27/26</u> Time: <u>9:09</u>	Payment: <input type="checkbox"/> Invoice <input type="checkbox"/> Cash <input type="checkbox"/> VISA, M/C <input type="checkbox"/> Check # _____ Amount _____
Relinquished By:	Received By Laboratory: <u>[Signature]</u>	Date: <u>5/27/26</u> Time: <u>9:09</u>	