



Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

June 17, 2026

John Harding
Eagle Point School District #9
PO Box 548
Eagle Point, OR 97524
TEL: (541) 830-1240
FAX (541) 830-6375

RE: Maint Shop Lead Testing

Order No.: 26060263

Dear John Harding:

Neilson Research Corporation received 5 sample(s) on 6/4/2026 for the analyses presented in the following report.

The results relate only to the parameters tested or to the sample as received by the laboratory. This report shall not be reproduced except in full, without the written approval of Neilson Research Corporation. Testing was completed in accordance with the laboratory's established quality system and standard operating procedures. For the current scope of NELAP-accredited analytes and certifications, please refer to the Certifications section at www.nrclabs.com. If you have any questions regarding these test results, please feel free to call.

Sincerely,
Neilson Research Corporation

Tamra Schmedemann
Director of Project Management
245 S Grape St
Medford, OR 97501





Neilson Research Corporation
245 S Grape St
Medford, OR 97501
TEL: (541) 770-5678 FAX: (541) 770-2901
Website: www.nrclabs.com

Case Narrative

WO#: 26060263
Date: 6/17/2026

CLIENT: Eagle Point School District #9
Project: Maint Shop Lead Testing

The analyses were performed according to the guidelines in the Neilson Research Corporation Quality Assurance Program. This report contains analytical results for the sample(s) as received by the laboratory.

Neilson Research Corporation certifies that this report is in compliance with the requirements of NELAP. No unusual difficulties were experienced during analysis of this batch except as noted below or qualified with data flags on the reports.

Original



Neilson Research Corporation
 245 S Grape St
 Medford, OR 97501
 TEL: (541) 770-5678 FAX: (541) 770-2901
 Website: www.nrclabs.com

Analytical Report

WO#: 26060263
 Date Reported: 6/17/2026

Eagle Point School District #9
 PO Box 548
 Eagle Point, OR 97524

Lab Order: 26060263
Received Date: 6/4/2026 1:36:00 PM
Reported Date: 6/17/2026 9:00:09 AM

Sample Information:

Lab ID: 26060263-01 Client Sample ID: 001BF
 Collection Date: 6/3/2026 11:33:00 AM Collected By: Loni O'Connor
 Matrix: Drinking Water Sample Location: 001BF

| Trace Metals by EPA 200.8 ICP-MS | | | | | | | | Analyst; PR |
|----------------------------------|--------|------|-------|-------|----|---------------|------|--------------|
| Analyses | Result | Qual | MRL | Units | DF | Date Analyzed | MCL | NELAP Status |
| Lead | ND | | 0.500 | ppb | 1 | 6/12/2026 | 15.0 | A |

Lab ID: 26060263-02 Client Sample ID: 002KF
 Collection Date: 6/3/2026 11:36:00 AM Collected By: Loni O'Connor
 Matrix: Drinking Water Sample Location: 002KF

| Trace Metals by EPA 200.8 ICP-MS | | | | | | | | Analyst; PR |
|----------------------------------|--------|------|-------|-------|----|---------------|------|--------------|
| Analyses | Result | Qual | MRL | Units | DF | Date Analyzed | MCL | NELAP Status |
| Lead | ND | | 0.500 | ppb | 1 | 6/12/2026 | 15.0 | A |

Lab ID: 26060263-03 Client Sample ID: 003BF
 Collection Date: 6/3/2026 11:37:00 AM Collected By: Loni O'Connor
 Matrix: Drinking Water Sample Location: 003BF

| Trace Metals by EPA 200.8 ICP-MS | | | | | | | | Analyst; PR |
|----------------------------------|--------|------|-------|-------|----|---------------|------|--------------|
| Analyses | Result | Qual | MRL | Units | DF | Date Analyzed | MCL | NELAP Status |
| Lead | ND | | 0.500 | ppb | 1 | 6/12/2026 | 15.0 | A |

| | | |
|-------------------|---|--|
| QUALIFIERS | * Value exceeds Maximum or Minimum Contaminant Level. | C1 Sample container temperature is out of limit as specified at testcode |
| | DF Dilution Factor | E Value above quantitation range |
| | H Holding times for preparation or analysis exceeded | J Analyte detected below quantitation limits |
| | MI Recovery outside control limits due to Matrix Interference | ND Not Detected at the Reporting Limit |
| | R RPD outside accepted recovery limits | RL Reporting Limit |

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original



Neilson Research Corporation
 245 S Grape St
 Medford, OR 97501
 TEL: (541) 770-5678 FAX: (541) 770-2901
 Website: www.nrclabs.com

Analytical Report

WO#: 26060263
 Date Reported: 6/17/2026

Eagle Point School District #9
 PO Box 548
 Eagle Point, OR 97524

Lab Order: 26060263
Received Date: 6/4/2026 1:36:00 PM
Reported Date: 6/17/2026 9:00:09 AM

Sample Information:

Lab ID: 26060263-04 Client Sample ID: 004BF
 Collection Date: 6/3/2026 11:43:00 AM Collected By: Loni O'Connor
 Matrix: Drinking Water Sample Location: 004BF

| Trace Metals by EPA 200.8 ICP-MS | | | | | | | | Analyst; PR |
|----------------------------------|--------|------|-------|-------|----|---------------|------|--------------|
| Analyses | Result | Qual | MRL | Units | DF | Date Analyzed | MCL | NELAP Status |
| Lead | ND | | 0.500 | ppb | 1 | 6/12/2026 | 15.0 | A |

Lab ID: 26060263-05 Client Sample ID: 005BF
 Collection Date: 6/3/2026 11:47:00 AM Collected By: Loni O'Connor
 Matrix: Drinking Water Sample Location: 005BF

| Trace Metals by EPA 200.8 ICP-MS | | | | | | | | Analyst; PR |
|----------------------------------|--------|------|-------|-------|----|---------------|------|--------------|
| Analyses | Result | Qual | MRL | Units | DF | Date Analyzed | MCL | NELAP Status |
| Lead | ND | | 0.500 | ppb | 1 | 6/12/2026 | 15.0 | A |

| | | |
|-------------------|---|--|
| QUALIFIERS | * Value exceeds Maximum or Minimum Contaminant Level. | C1 Sample container temperature is out of limit as specified at testcode |
| | DF Dilution Factor | E Value above quantitation range |
| | H Holding times for preparation or analysis exceeded | J Analyte detected below quantitation limits |
| | MI Recovery outside control limits due to Matrix Interference | ND Not Detected at the Reporting Limit |
| | R RPD outside accepted recovery limits | RL Reporting Limit |

NELAP A Accredited in accordance with NELAP ORELAP 100016, OR-028

Original



Neilson Research Corporation
 245 S Grape St
 Medford, OR 97501
 TEL: (541) 770-5678 FAX: (541) 770-2901
 Website: www.nrclabs.com

QC SUMMARY REPORT

WO#: 26060263

17-Jun-26

Client: Eagle Point School District #9
Project: Maint Shop Lead Testing

TestCode: LEAD_SCHOOLS

| | | | | | | | | | | | |
|----------------------------|------------------------|----------------------------|-------------------|---------------------------------|-----------------------|----------|-----------|-------------|------|----------|------|
| Sample ID: MB-37161 | SampType: MBLK | TestCode: LEAD_SCHO | Units: ppb | Prep Date: 6/12/2026 | RunNo: 70433 | | | | | | |
| Client ID: PBW | Batch ID: 37161 | TestNo: E200.8 | E200.8 | Analysis Date: 6/12/2026 | SeqNo: 1163496 | | | | | | |
| Analyte | Result | PQL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | RPD Ref Val | %RPD | RPDLimit | Qual |
| Lead | ND | 0.500 | | | | | | | | | |

| | | | | | | | | | | | |
|-----------------------------|------------------------|----------------------------|-------------------|---------------------------------|-----------------------|----------|-----------|-------------|------|----------|------|
| Sample ID: LCS-37161 | SampType: LCS | TestCode: LEAD_SCHO | Units: ppb | Prep Date: 6/12/2026 | RunNo: 70433 | | | | | | |
| Client ID: LCSW | Batch ID: 37161 | TestNo: E200.8 | E200.8 | Analysis Date: 6/12/2026 | SeqNo: 1163497 | | | | | | |
| Analyte | Result | PQL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | RPD Ref Val | %RPD | RPDLimit | Qual |
| Lead | 97.6 | 0.500 | 100 | 0 | 97.6 | 85 | 115 | | | | |

| | | | | | | | | | | | |
|----------------------------------|------------------------|----------------------------|-------------------|---------------------------------|-----------------------|----------|-----------|-------------|------|----------|------|
| Sample ID: 26060232-01AMS | SampType: MS | TestCode: LEAD_SCHO | Units: ppb | Prep Date: 6/12/2026 | RunNo: 70433 | | | | | | |
| Client ID: BatchQC | Batch ID: 37161 | TestNo: E200.8 | E200.8 | Analysis Date: 6/12/2026 | SeqNo: 1163501 | | | | | | |
| Analyte | Result | PQL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | RPD Ref Val | %RPD | RPDLimit | Qual |
| Lead | 93.6 | 0.500 | 100 | 0.715 | 92.9 | 70 | 130 | | | | |

| | | | | | | | | | | | |
|-----------------------------------|------------------------|----------------------------|-------------------|---------------------------------|-----------------------|----------|-----------|-------------|-------|----------|------|
| Sample ID: 26060232-01AMSD | SampType: MSD | TestCode: LEAD_SCHO | Units: ppb | Prep Date: 6/12/2026 | RunNo: 70433 | | | | | | |
| Client ID: BatchQC | Batch ID: 37161 | TestNo: E200.8 | E200.8 | Analysis Date: 6/12/2026 | SeqNo: 1163502 | | | | | | |
| Analyte | Result | PQL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | RPD Ref Val | %RPD | RPDLimit | Qual |
| Lead | 92.7 | 0.500 | 100 | 0.715 | 92.0 | 70 | 130 | 93.6 | 0.971 | 20 | |

Qualifiers: CI Sample container temperature is out of limit as specified at testcode DF Dilution Factor H Holding times for preparation or analysis exceeded
 MI Recovery outside control limits due to Matrix Interference ND Not Detected at the Reporting Limit RL Reporting Limit

Original



Sample Log-In Check List

Client Name: EAGLEPTSCHOOL Work Order Number: 26060263 RcptNo: 1

| | | | |
|---------------|--------------------|-----------------------|--|
| Logged by: | Ashley Spiegelberg | 6/4/2026 1:36:00 PM | |
| Completed By: | Kyle Woolsey | 6/10/2026 10:39:53 AM | |
| Reviewed By: | Dorie Evans | 6/17/2026 8:20:59 AM | |

Chain of Custody

1. Is Chain of Custody complete? Yes No Not Present
2. How was the sample delivered? Client

Log In

3. Coolers are present? Yes No NA
4. Shipping container/cooler in good condition? Yes No
- Custody seals intact on shipping container/cooler? Yes No Not Present NA
- No. Seal Date: Signed By:
5. Was an attempt made to cool the samples? Yes No NA
6. Were all samples received at a temperature of >0° C to 6.0°C Yes No NA
7. Sample(s) in proper container(s)? Yes No
8. Sufficient sample volume for indicated test(s)? Yes No
9. Are samples (except VOA and ONG) properly preserved? Yes No
10. Was preservative added to bottles? Yes No NA
11. Is the headspace in the VOA vials less than 1/4 inch or 6 mm? Yes No No VOA Vials
12. Were any sample containers received broken? Yes No
13. Does paperwork match bottle labels? Yes No
- (Note discrepancies on chain of custody)
14. Are matrices correctly identified on Chain of Custody? Yes No
15. Is it clear what analyses were requested? Yes No
16. Were all holding times able to be met? Yes No
- (If no, notify customer for authorization.)

Special Handling (if applicable)

17. Was client notified of all discrepancies with this order? Yes No NA

| | | | |
|----------------------|----------------------|-------|---|
| Person Notified: | <input type="text"/> | Date: | <input type="text"/> |
| By Whom: | <input type="text"/> | Via: | <input type="checkbox"/> eMail <input type="checkbox"/> Phone <input type="checkbox"/> Fax <input type="checkbox"/> In Person |
| Regarding: | <input type="text"/> | | |
| Client Instructions: | <input type="text"/> | | |

18. Additional remarks:

Cooler Information

| Cooler No | Temp °C | Condition | Seal Intact | Seal No | Seal Date | Signed By |
|-----------|---------|-----------|-------------|---------|-----------|-----------|
|-----------|---------|-----------|-------------|---------|-----------|-----------|

Chain of Custody Record

This Chain of Custody is a LEGAL DOCUMENT and must be filled out accurately.

Section A

Required Client Information

Company: JCS DG
 Address: _____
 Email: _____
 Phone: _____ Fax: _____
 Collected By (Print): Loni O'Connor
 Collected By (Sign): [Signature]
 Email Report ___ Mail Report ___ Fax Report ___

Section B

Required Project Information

Project Name: Maint. Shop
 Project Number: _____
 Report To: _____
 Copy To: _____

Section C

Invoice Information

Attention: John Harding
 Company Name: _____
 Address: _____
 P.O. # _____

Section D

Rush Status (Subject to Scheduling)

Standard: 10 Business Days
 Priority: 5 Business Days (List x 1.50)
 Express: 3 Business Days (List x 1.75)
 Rush: 2 Business Days (List x 2.00)
 Rush: 1 Business Day (List x 2.50)
 Rush: Same Day (List x 3.00)
 Authorized ___ Yes ___ No

Section E

Sample Information

| Sample ID | Comp/Grab | Matrix* | Date Collected | Time Collected | No. of Containers | Analysis Requested | NRC Workorder # (Lab Use Only) | Remarks / Field Data | NRC Sample # Use Only | (Lab) |
|-----------|-----------|---------|----------------|----------------|-------------------|--------------------|-----------------------------------|-------------------------|--------------------------|-------|
| 001 BF | 5080 | | 6/3/26 | 11:32 | | | 26060263 | | | |
| 002 KF | 5070 | | | 11:36 | | | | | 01 | |
| 003 BF | 5088 | | | 11:37 | | | | | 02 | |
| 004 BF | 5057 | | | 11:43 | | | | | 03 | |
| 005 BF | 5082 | | | 11:43 | | | | | 04 | |
| | | | | | | | | | 05 | |

*Matrix: DW - Drinking Water WW - Wastewater W - Water S - Soil/Solid SL - Sludge O - Oil WP - Wipe OT - Other

Section F

Relinquish/Receive

| Relinquished By: | Print | Date | Time |
|-------------------------|-------------------------|---------------|--------------|
| <u>[Signature]</u> | <u>Loni O'Connor</u> | <u>6/3/26</u> | <u>11:32</u> |
| Received By: | | | |
| Relinquished By: | | | |
| Received By: | | | |
| Relinquished By: | | | |
| Received By Laboratory: | <u>Ashley Spigelman</u> | <u>6/9/26</u> | <u>13:36</u> |

Section G

Lab Use Only

Temp: IR Therm ID: NA
 5°C: ___ Yes ___ No
 Received on Ice: ___ Yes ___ No
 Number of Bottles Received: _____
 pH Checked: _____
 COC Seals Intact: ___ Yes ___ No NA
 Field Blank Included: ___ Yes ___ No
 Received Via ___ UPS ___ FedEx ___ Other Hand
 Payment: ___ Invoice ___ Cash ___ VISA, M/C ___ Check # _____ Amount _____